

**Force Feedback for Patient Side Manipulator of Da Vinci  
Research Kit**

by

Anna Novoseltseva

A Thesis submitted to the Faculty of the

WORCESTER POLYTECHNIC INSTITUTE

in partial fulfillment of the requirements for the

Degree of Master of Science in Biomedical Engineering

Worcester, Massachusetts

May, 2018

© Anna Novoseltseva 2018

All rights reserved

Approved by:

---

Prof. Gregory S. Fischer, Advisor  
Worcester Polytechnic Institute

---

Prof. Karen Troy, Committee Member  
Worcester Polytechnic Institute

---

Prof. Loris Fichera, Committee Member  
Worcester Polytechnic Institute

# Abstract

statement of problem In order to get force feedback for da Vinci surgical robot was

procedure designed 2-DOF force-sensing instrument for patient side manipulator. In a future it can be used for haptic feedback.

results Designed instrument showed that accuracy and sensitivity, which complies to mostly all requirements.

conclusions Many things to be changed (add Z-direction, biocompatibility)

# Acknowledgments

I would like to express my gratitude to everybody in the world.

# Dedication

This dissertation is dedicated to everybody in the world.

# Contents

<b>Abstract</b>	<b>iii</b>
<b>Acknowledgments</b>	<b>iv</b>
<b>List of Tables</b>	<b>ix</b>
<b>List of Figures</b>	<b>x</b>
<b>1 Introduction</b>	<b>1</b>
<b>2 Background</b>	<b>3</b>
2.1 Teleoperated Surgical Robots . . . . .	3
2.2 Importance of Haptic Feedback . . . . .	7
2.3 Current Approaches . . . . .	8
2.4 Force Sensors . . . . .	12
2.5 Contributions . . . . .	14
<b>3 Methods and Results</b>	<b>15</b>
3.1 Force Measurement . . . . .	15

3.2	Sensor Placement Optimization . . . . .	16
3.2.1	Elasticity Modulus Measurements . . . . .	16
3.2.2	Density Measurements . . . . .	18
3.3	Requirements for the Device . . . . .	18
3.4	Mechanical Design . . . . .	18
3.4.1	Strain Gauge . . . . .	18
3.4.2	Installation of Strain Gauges . . . . .	19
3.4.3	Fabrication of the Sleeves . . . . .	20
3.5	Electrical and Software Design . . . . .	21
3.5.1	Circuit design . . . . .	21
3.5.2	Noise Analysis . . . . .	21
3.5.3	Microcontroller Software . . . . .	22
3.5.4	ROS Architecture . . . . .	23
3.6	Calibration . . . . .	23
3.6.1	Calibration Setup . . . . .	23
3.6.2	Calibration of the Load Cell . . . . .	23
3.6.3	Calibration Results . . . . .	23
3.7	Experiments . . . . .	25
3.7.1	Distance from the cannula to the tip dependence from readings	25
3.7.2	Temperature Dependence . . . . .	25
3.7.3	Hysteresis . . . . .	25

4 Discussion and Conclusion	26
References	28



# List of Tables

3.1	Table Title . . . . .	17
-----	-----------------------	----

# List of Figures

3.1	Block diagram . . . . .	16
3.2	Setup to measure elasticity modulus . . . . .	17
3.3	Z-direction force feedback sensor . . . . .	20
3.4	Z-direction force feedback sensor - section view . . . . .	21
3.5	Noise analysis of the output signal . . . . .	22
3.6	X-component . . . . .	24
3.7	Y-component . . . . .	24
3.8	Calibration results . . . . .	24

Disclaimer: certain materials are included under the fair use exemption of the U.S. Copyright Law and have been prepared according to the fair use guidelines and are restricted from further use.

# Acronyms

**PSM** Patient Side Manipulator

**DOF** Degrees of Freedom

**PCB** Printed Circuit Board

**ROS** Robot Operating System

# Chapter 1

## Introduction

Teleoperated da Vinci surgical system is a robot-assisted surgical system that enhances surgeons performance in minimally invasive surgeries by allowing highly precise translation of surgeon's hand movements to the instrument's movements.

The currently available da Vinci surgery system has a laparoscopic camera, providing visual feedback to guide doctors during surgery. However, the system does not have any kinesthetic or cutaneous feedback, known as haptics. [1]

During open surgeries, doctors usually get haptic feedback directly or through the surgical tools. In minimally invasive surgeries interaction with patients via long shafts leads to the loss of some force and tactile sense. In robotic surgery systems, surgeons have to manipulate robots indirectly, which leads to an elimination of any haptic feedback. [2]

It is believed that the addition of haptic feedback in the da Vinci surgery robot will

help to reduce the amount of surgical errors and intra-operative injuries, which will lead to faster post-surgery recovery time and decreased rate of unsuccessful surgeries. [2-4]

There are many technical challenges to overcome in order to implement the haptic feedback in da Vinci robot. One of them is getting accurate force readings from the patient side manipulator (PSM). To address this issue, we are trying to create force-feedback device, that can be easily added to the existing surgery system.

# Chapter 2

## Background

In this section we talk about all background knowledge necessary to

### 2.1 Teleoperated Surgical Robots

Recently robots started to be extensively used for surgical procedures. The use of robots allows doctors to perform surgical procedures with high accuracy, repeatability and reliability. Which in turn results in reducing operation time, errors and post-operation injuries. Minimally invasive surgeries are beneficial for accurate procedures with minimal access to operated organs, e.g. neurosurgery, eye surgery, cardiac surgery, intravascular surgeries and etc. Use of robots in minimally invasive procedures improves precision and reliability of operations. [?]

There are two types of devices used for surgeries, supporting and augmenting.

Supporting devices perform secondary functions to support the surgeon. Some of them used for positioning and stabilization purposes of cameras, endoscopic tools, ultrasound probes and etc. Others to increase device dexterity or autonomy (dexterous and autonomous endoscopes).

Augmenting devices are used to extend surgeon's ability in performing an operation. They can be divided in four categories. Hand-held tools are augmenting instruments that used for hand tremors reduction, for dexterity and navigation capability increase. Another type of augmenting devices are cooperatively-controlled tools, where the surgeon and the robot cooperatively manipulate the surgical device (e.g. ROBODOC system, Steady- Hand robot, LARS, the Neurobot, and the ACROBOT system). Teleoperated robots are type of augmenting tools, where surgeon (master) controls the movements of a surgical robot (slave) via a surgeon's console (e.g. the da Vinci and the Zeus systems). And autonomous tools, which can perform some tasks (suturing and knot tying) autonomously. [?]

Use of teleoperated robots in surgeries can solve many of the conventional surgery problems in terms of more precise manipulation capability, ergonomics, dexterity, and haptic feedback capability for the surgeon. They enhance dexterity by increase of instrument degrees of freedom, hand tremor compensation, and movements scaling that allows transformation of the control grips large movements into small motions inside the patient. 3-D view with depth perception gives surgeons ability to directly control a stable visual field with increased magnification and maneuverability. All of these

enhances the surgeon's operation performance. However, robot-assisted surgeries are high cost, need large operational room space, do not have established efficacy, and need for tableside assistants. For these reasons ability of hospitals to use surgical robots is low, making their use for routine surgeries improbable. [?]

Today, many surgical robotic systems have been commercially developed and approved by the FDA, such as the Da Vinci surgical system (Intuitive Surgical, Inc., Sunnyvale, CA) , Sensei X robotic catheter system (Hansen Medical Inc., Mountain View, CA), FreeHand v1.2 (FreeHand 2010 Ltd., Cardiff, UK), Invendoscopy E200 system (Invendo Medical GmbH, Germany), Flex robotic system (Medrobotics Corp., Raynham, MA), Senhance (TransEnterix, Morrisville, NC), Auris robotic endoscopy system (ARES; Auris Surgical Robotics, Silicon Valley, CA, USA), The NeoGuide Endoscopy System (NeoGuide Endoscopy System Inc, Los Gatos, CA). [?, ?]

There is also number of NON-FDA-approved platforms that currently under development or going through clinical trials. For example, MiroSurge (RMC, DLR, German Aerospace Center, Oberpfaffenhofen-Weling), The ViaCath system (BIOTRONIK, Berlin, Germany), SPORT surgical system (Titan Medical Inc., Toronto, Ontario), The SurgiBot (TransEnterix, Morrisville, NC), The Versius Robotic System (Cambridge Medical Robotics Ltd., Cambridge, UK), MASTER (Nanyang Technological University and National University Health System), Verb Surgical (Verb Surgical Inc., J & J/Alphabet, Mountain View, CA, USA), Miniature in vivo robot (MIVR) (MIVR, Virtual Incision, CAST, University of Nebraska Medical Center, Omaha,



Nebraska, USA), the Einstein surgical robot (Medtronic, Minneapolis, MN). [?]

The da Vinci Surgical System is one of the most commonly used robotic surgical systems. In 2015, over 3400 systems were in use around the world. More than 3 million surgeries were performed worldwide using da Vinci system [1]. The system has been approved for various types of surgeries such as cardiac, colorectal, thoracic, urological and gynecologic. However, new systems are emerging on the market, providing features that are absent currently in the da Vinci System. For example, in 2017 FDA approved Senhance robotic platform that provides actual haptic force feedback, allowing the surgeon to feel forces generated at the instruments end. In addition, the system uses eye-tracking technology to move the camera at the point the surgeon is looking at, while the da Vinci uses a footswitch panel to control the camera movement. Another example is Flex Robotic System, which consists of flexible endoscope for laparoendoscopic surgeries. This system is able to define a non-linear path to surgical target by advancing a flexible telescopic inner-outer mechanism with instruments inside it, whereas instruments in the da Vinci system can follow only non-flexible straight path. [?]

Effectiveness of da Vinci system in comparison to opened surgery and other systems [5]

## 2.2 Importance of Haptic Feedback

Write about studies with and without haptic feedback.

It has been shown that incorporating force feedback into teleoperated systems can reduce the magnitude of contact forces and therefore the energy consumption, the task completion time and the number of errors. In several studies [122, 147, 15], addition of force feedback is reported to achieve some or all of the following: reduction of the RMS force by 30

In [106], a scenario is proposed to incorporate force feedback into the Zeus surgical system by integrating a PHANToM haptic input device into the system. In [85], a dextrous slave combined with a modified PHANToM haptic master which is capable of haptic feedback in four DOFs is presented. A slave system which uses a modified Impulse Engine as the haptic master device is described in [30]. In [107], a telesurgery master-slave system that is capable of reflecting forces in three degrees of freedom (DOFs) is discussed. A master-slave system composed of a 6-DOF parallel slave micromanipulator and a 6-DOF parallel haptic master manipulator is described in [150]. Other examples of haptic surgical teleoperation include [93] and [11]. The haptics technology can also be used for surgical training and simulation purposes. For example, a 7-DOF haptic device that can be applied to surgical training is developed in [56]. A 5-DOF haptic mechanism that is used as part of a training simulator for urological operations is discussed in [146].

## 2.3 Current Approaches

1.3 Haptics for Robotic Surgery and Therapy Incorporating haptic sensation to robotic systems for surgery or therapy especially for minimally invasive surgery, which involves limited instrument maneuverability and 2-D camera vision, is a logical next step in the development of these systems. To do so, in addition to instrumentation of surgical tools, appropriate haptics-enabled user interfaces must be developed.

1.3.1 Haptic user interface technology In the following, examples of the currently available haptic devices are described. For a more complete survey of haptic devices, see [55].

1.3.1.1 PHANToM The PHANToM from Sensable Technologies Inc. ([www.sensable.com](http://www.sensable.com)) is one of the most commonly used haptic devices and comes in a number of models with different features. PHANToM 1.5A provides six DOFs input control. Of the six DOFs of the arm, depending on the model, some or all are force-reflective. In Figure 1.2a, a PHANToM 1.5/6DOF with force feedback capability in all of the six DOFs is shown.

1.3.1.2 Freedom-6S The Freedom-6S shown in Figure 1.2b is a 6-DOF device from MPB Technologies Inc. ([www.mpb-technologies.ca](http://www.mpb-technologies.ca)) that provides force feedback in all of the six degrees of freedom. The position stage is direct driven while the orientation stage is driven remotely by tendons. The Freedom-6S features static and dynamic balancing in all axes (see [54] for further design details).

1.3.1.3 Laparoscopic Impulse Engine and Surgical Workstation Originally as part of a laparoscopic

surgical simulator, the Laparoscopic Impulse Engine was designed by Immersion Corp. ([www.immersion.com](http://www.immersion.com)). The device can track the position of the instrument tip in five DOFs with high resolution and speed while providing force feedback in three DOFs. More recently, Immersion has developed the Laparoscopic Surgical Workstation (Figure 1.2c), which is capable of providing force feedback in five DOFs. An application example is the Virtual Endoscopic Surgery Trainer (VEST) from Select-IT VEST Systems AG ([www.select-it.de](http://www.select-it.de)). The VEST system uses the Laparoscopic Impulse Engine as its force-feedback input interface for simulating laparoscopic surgery interventions.

1.3.1.4 Xitact IHP The Xitact IHPTM from Xitact Medical Simulation ([www.xitact.com](http://www.xitact.com)) is a 4-DOF force feedback manipulator based on a spherical remote-center-of-motion mechanical structure and was originally designed for virtual reality based minimally invasive surgery simulation [42]. It features high output force capability, low friction, zero backlash and a large, singularity-free workspace. A picture of the Xitact IHP is shown in Figure 1.2d.

Placing force sensors on the surgical instrument [6]

They suggest to the measure of pulling and grasp forces at the tip of surgical instrument. For the design of the compliant forceps, the required compliance characteristics are first defined using a simple spring model with one linear and one torsional springs. This model may be directly realized as the compliant forceps. However, for the compact realization of the mechanism, we synthesize the spring model with two

torsional springs that has equivalent compliance characteristics to the linear-torsional spring model. Then, each of the synthesized torsional springs is realized physically by means of a flexure hinge. From this design approach, direct measurement of the pulling and grasp forces is possible at the forceps, and measuring sensitivity can be adjusted in the synthesis process. The validity of the design is evaluated by finite element analysis. Further, from the measured values of bending strains of two flexure hinges, a method to compute the decoupled pulling and grasp forces is presented via the theory of screws. Finally, force- sensing performance of the proposed compliant forceps is verified from the experiments of the prototype using some weights and load cells. 10.1109/TRO.2012.2194889

Making new surgical instrument design [7]

Method In this paper a force-feedback enabled surgical robotic system is described in which the developed force-sensing surgical tool is discussed in detail. The developed surgical tool makes use of a proximally located force/torque sensor, which, in contrast to a distally located sensor, requires no miniaturization or sterilizability. Results Experimental results are presented, and indicate high force sensing accuracies with errors  $\leq 0.09$  N. Conclusions It is shown that developing a force-sensing surgical tool utilizing a proximally located force/torque sensor is feasible, where a tool outer diameter of 12 mm can be achieved. For future work it is desired to decrease the current tool outer diameter to 10 mm.

Sensorless estimation methods

Vision based solution [8]

They proposed to use vision based solution with supervised learning to estimate the applied force and provide the surgeon with a suitable representation of it. The proposed solution starts with extracting the geometry of motion of the heart's surface by minimizing an energy functional to recover its 3D deformable structure. A deep network, based on a LSTM-RNN architecture, is then used to learn the relationship between the extracted visual-geometric information and the applied force, and to find accurate mapping between the two. Our proposed force estimation solution avoids the drawbacks usually associated with force sensing devices, such as biocompatibility and integration issues. We evaluate our approach on phantom and realistic tissues in which we report an average root-mean square error of 0.02 N.

SLiding pertrubation observer

This paper suggests a bilateral controller applying sliding perturbation observer based force estimation method. In the suggested bilateral controller, the master control uses impedance control and the slave control uses a sliding mode control (SMC). A torque and force sensorless teleoperation system can be implemented using the suggested bilateral control structure through an experimental evaluation. This paper presents a method of estimating the reaction force of the surgical robot instrument without sensors and attempts to use state observer of control algorithm. Sliding mode control with sliding perturbation observer (SMCSPO) is used to drive the instrument, where the sliding perturbation observer (SPO) computes the amount of perturbation

defined as the combination of the uncertainties and nonlinear terms where the major uncertainties arise from the reaction force. Based on this idea, this paper proposes a method to estimate the reaction force on the end-effector tip of the surgical robot instruments using only SPO and encoder without any additional sensors. To evaluate the validity of this paper, experiment was performed and the results showed that the estimated force computed from SPO is similar to the actual force.

Measuring the proximal guide wire force [9]

They measure the proximal guide wire force and the force between the surgeon's hand and the handle used on the Phantom, the force feedback closed-loop control can effectively eliminate the loss of mechanical impedance of force feedback information. The accuracy control of force feedback is greatly enhanced in the aspect of security and the operation efficiency.

Write about all disadvantages of previous methods.

Tools -  $\dot{\iota}$  limited lifetime - find citation

Vision -  $\dot{\iota}$  huge time delays, accuracy

Wire force -  $\dot{\iota}$  repeatability issue

## 2.4 Force Sensors

Two general principles dominate in force measurement: piezoelectric and strain-gauge sensors. [10]

Piezoelectric sensors consist of two crystal disks with an electrode foil in between. When force is applied, an electric charge, proportional to the applied force, is obtained and can be measured. Piezoelectric sensors show small deformation when force is applied, this results in a high resonance frequency. Also, piezoelectric sensors due to their principle of operation have significant linearity error and drift. [11]

In the strain gauge based force transducers the force causes deformation and subsequent linear change in resistance. Strain gauges are usually connected to a Wheatstone bridge circuit, where the output voltage is proportional to the applied force. Strain gauge based transducers provide small individual errors (200 ppm), show no drift, and are therefore appropriate for long-term monitoring tasks. However, they are relatively big, temperature dependent, and have lower resonance frequency in comparison to piezoelectric sensors. [10,11]

On the basis of the above mentioned, piezoelectric sensors are preferable for dynamic measurements of small forces while strain gauge sensors are better when large forces are measured. In this study, strain gauges were used since they show better accuracy and long-term stability. [10,11]

We can use QTC-pills, but it has no-linearity and hysteresis issues. We can test in the future.



## 2.5 Contributions

Force sensing devices for measuring forces in X-Y direction and one for Z-direction measurement were created. They allow to get accurate force readings from the da Vinci tools of the PSM. These devices can be easily added to the existing da Vinci system. Since we have to add created device on each robot arm only, it is cheaper than placement of sensors on each separate surgical tool. Moreover, created devices allow to get force data faster than through visual data processing method.

# Chapter 3

## Methods and Results

This chapter introduces methods and results of this work.

### 3.1 Force Measurement

Block diagram of the created system for 3-DOF force measurement is shown on figure 3.1. Forces that applied on the end of surgical tool are measured using strain gauges, which change their resistance with force. Using created printed circuit boards (PCBs), this resistance changes are measured and published within ROS. At the same time we measure current joint position of the tool, which is needed for the force calibration. On the PC position data and data from PCBs are used to find values of the force in X,Y,Z directions.




Figure 3.1: Block diagram

## 3.2 Sensor Placement Optimization

A finite element analysis was done in Solidworks to assess better placement of the strain gauges for the created device, shaft, and cannula. First we found elasticity modulus of the shaft and cannula, and used them for further analysis. The shaft and cannula material are unknown, so we studied their elasticity modulus, .. experimentally. At first we were trying to put sensors on the instrument shaft itself, on the cannula, on the sleeve to see where we get the most accurate results.

### 3.2.1 Elasticity Modulus Measurements

Elasticity Modulus of the shaft and the cannula were found experimentally (3.2). One end of the observing sample (shaft or cannula) was fixed and the force was applied on the other end. We used weights: 250g for the shaft and 555g for the cannula. The deformation was detected with dial indicator. Experiment was done 5 times, average displacement value was used.




Figure 3.2: Setup to measure elasticity modulus

Table 3.1: Elasticity Modulus Measurement Data

Component	$d_o$ , mm	$d_i$ , mm	$I$ , mm <sup>4</sup>	$m$ , g	$F$ , N	$L$ , mm	$L_{tot}$ , mm
Shaft	8.4	6	$1.808 \cdot 10^{-10}$	250	3.25	276.2	366.8
Cannula	10.54	8.75	$3.181 \cdot 10^{-10}$	555	6.011	95.5	105.55
Component	$\delta \pm SD$ , mm		$E \pm SD$ , GPa				
Shaft	$2.856 \pm 0.123$		$44.31 \pm 1.86$				
Cannula	$0.086 \pm 0.004$		$63.92 \pm 2.97$				

Elasticity Modulus was found using following equation:

$$E = \frac{FL^3}{3\delta I} \quad (3.1)$$

where  $F$  - force,  $L$  - length from the fixed point to indicator,  $I$  - area moment of inertia,  $\delta$  - displacement.

Area moment of Inertia:

$$I = \frac{\pi(d_o^4 - d_i^4)}{64} \quad (3.2)$$

where  $d_o$  - cylinder outside diameter,  $d_i$ - cylinder inside diameter.

Force acting on indicator:

$$F = \frac{L_{tot}}{L}mg \quad (3.3)$$

where  $L_{tot}$  - total length of the object,  $m$  - mass of the weight,  $g$  - gravitational constant.

### 3.2.2 Density Measurements

## 3.3 Requirements for the Device

From the literature review, following requirements for the device were developed:

Biocompatibility

Linearity

Range of forces Highest gripping force in Da Vinci tool was seen in the Hemolok[R] clip applier (39.92 [+ or -] 0.89 N). 0.3 - 11 N. Tensile strength and failure load of sutures for robotic surgery Others

## 3.4 Mechanical Design

### 3.4.1 Strain Gauge

According to the manual for strain gauge selection provided by Vishay Micro-Measurements. The strain gauge should have following parameters: Gauge length is 0.8 mm. Ideally gauge length should be 10 % of the shaft radius; Single grid; Isoelastic (D alloy) that has higher gauge factor with E backing; Encapsulated with

pre attached leads; Resistance 120 Ohms; Gauge factor - 2; STC number (self-temperature-compensation): DY dynamic Since the diameter of the shaft is only 8.5 mm, one of the most important parameters, in this case, is the size of the strain gauge. Another important parameter is gauges resistance, sensitivity of the strain gauge drastically depends on this parameter. Strain gauges with different resistivity were used (350 and 120 Ohms and sizes of 6 x 2.5 mm).

### **3.4.2 Installation of Strain Gauges**

The shaft was assumed to be made of Tecamax, but for gauge installation purposes materials for plastic was chosen as it is comparable in preparation. [12].

First the working surface (glass) and tweezers were cleaned with Neutralizer 5. After that shaft surface preparation was started, using solvent degreaser GC-6 Isopropyl Alcohol. A gauge layout was then applied with a 4H drafting pencil. The surface was then conditioned with Conditioner A and the extra liquid was wiped with gauze. Finally, the surface was then neutralized with M-Prep Neutralizer 5A. [12]

The strain gauges were first placed on the glass and then transported using mylar tape onto the instrument surface. A thin layer of catalyst was applied on the strain gauge and given one minute to dry. Then adhesive M-BOND 200 was applied on the shaft, pressure was applied on the tape for one minute, then two more minutes to let it dry before the tape was removed. Then leads soldering was done by application of pats, and soldering them with thin wires. [13]




Figure 3.3: Z-direction force feedback sensor

The methodology of the strain gauge application more specifically described in [12].

In compliance with the literature [12] for application of the strain gauge on metals, the same materials and technique can be used. Therefore, the same method to apply strain gauges on the cannula was used.

On the shaft On the cannula On the aluminum sleeve On the nylon sleeve

### 3.4.3 Fabrication of the Sleeves

Do not use 3D printed sleeves - causes non-linearity issue Use 350 Ohm strain-gauges and full-bridge circuit. So far the best design is Sleeve 1 - it is simple with linear output.




Figure 3.4: Z-direction force feedback sensor - section view

## 3.5 Electrical and Software Design

Using Altium Designer 15.1 PCB was developed and manufactured elsewhere.

Write about how to calibrate the PCB itself and how it works.

### 3.5.1 Circuit design

Write about how circuit works.

### 3.5.2 Noise Analysis

Noise analysis was performed using oscilloscope ... Make FFT analysis






Figure 3.5: Noise analysis of the output signal

### 3.5.3 Microcontroller Software

We use arduino for .. Filtering was implemented on microcontroller by finding average of recent 5 readings Filtering - ideally small time delay. Total teleoperation cycle delay - less than 100ms (visually noticeable delay). Kalman filter? Use same parameters for both signals to get identical time delay. good article about changing baud rate.

### **3.5.4 ROS Architecture**

## **3.6 Calibration**

First we used set of weights applied in two directions, but since the system was too sensitive for direction of force, we had to change our approach of determining the force direction. It was decided to use optical tracking system for that purpose.

### **3.6.1 Calibration Setup**

Tell how calibration with optical tracking system is working

### **3.6.2 Calibration of the Load Cell**

Was performed using following setup cite figure .. Calibration equation we got ..

### **3.6.3 Calibration Results**

Find mean square root error




Figure 3.6: X-component




Figure 3.7: Y-component

Figure 3.8: Calibration results

## **3.7 Experiments**

### **3.7.1 Distance from the cannula to the tip dependence from readings**

Distance from the cannula to the tip - measure - 3 inch. Maximum distance is 9 inches.

### **3.7.2 Temperature Dependence**

Try in 36.6 Celsius and room temperature Effect of the surrounding temperature on the device was ..

Do all experiments at least three times and calculate all statistic things you can (SD)

### **3.7.3 Hysteresis**

Hysteresis was checked .. We written separate program to check hysteresis

# Chapter 4

## Discussion and Conclusion

Compare shaft, cannula, sleeve - accuracy, hysteresis.

This is our conclusion)

For future work .. However, disadvantages would be addition of the cost to already expensive system and possible biocompatibility and sterilization issues. Also addition of the weight to the arm could alter robot performance, however, since the device will be placed close to center of rotation of the robot arm, it will have minimal affect on the moment of inertia in comparison to sensors added to the grippers.

Bio-compatibility and coatings

Sterilization techniques

Z-direction. We tried to measure force in Z-direction, but unfortunately results shown that developed system was not accurate and sensitive enough. To improve system we can suggest to change strain gauges to more sensitive ones. Since we have

small room for deformation - around 0.3 mm, we can not afford more deformation by using thinner or longer plates. Therefore, we decided to use motor current readings for z-directional reading of the force.

Increase speed from 588 Hz to something higher by change ADC from SPI communication to parallel communication. Right now maximum frequency is 250ksps, which is  $250k/32 = 7.812$  KBps.

Change communication channel between microcontroller and PC to faster one. Right now we use serial port, with max speed 7.1 KBps

Change microcontroller to the faster one.

We can use QTC-pills in future, promising approach, higher sensitivity

# References

- [1] Intuitive surgical, inc. - da vinci surgical system. [Online]. Available: <https://www.intuitivesurgical.com/>
- [2] A. M. Okamura, “Haptic feedback in robot-assisted minimally invasive surgery,” vol. 19, no. 1, pp. 102–107.
- [3] C. E. Reiley, T. Akinbiyi, D. Burschka, D. C. Chang, A. M. Okamura, and D. D. Yuh, “Effects of visual force feedback on robot-assisted surgical task performance,” vol. 135, no. 1, pp. 196–202.
- [4] O. A. J. van der Meijden and M. P. Schijven, “The value of haptic feedback in conventional and robot-assisted minimal invasive surgery and virtual reality training: a current review,” vol. 23, no. 6, pp. 1180–1190.
- [5] J. Yu, Y. Wang, Y. Li, X. Li, C. Li, and J. Shen, “The safety and effectiveness of da vinci surgical system compared with open surgery and laparoscopic surgery: a rapid assessment,” vol. 7, no. 2, pp. 121–134.
- [6] M. B. Hong and Y. H. Jo, “Design and evaluation of 2-DOF compliant forceps with force-sensing capability for minimally invasive robot surgery,” vol. 28, no. 4, pp. 932–941.
- [7] W. Schwalb, B. Shirinzadeh, and J. Smith, “A forcesensing surgical tool with a proximally located force/torque sensor,” vol. 13, no. 1.
- [8] A. I. Aviles, S. M. Alsaleh, J. K. Hahn, and A. Casals, “Towards retrieving force feedback in robotic-assisted surgery: A supervised neuro-recurrent-vision approach,” vol. 10, no. 3, pp. 431–443.

- [9] S. M. Yoon, W. J. Kim, and M. C. Lee, “Design of bilateral control for force feedback in surgical robot,” vol. 13, no. 4, pp. 916–925.
- [10] HBM. (2017) Piezoelectric or Strain Gauge Based Force Transducers? [Online]. Available: <https://www.hbm.com/en/3719/piezoelectric-or-strain-gauge-based-force-transducers/>
- [11] A. sensors. (2017) Piezoelectric Sensors and Strain Gauge-based Force Transducers: Principles in Force Measurement. [Online]. Available: <http://www.azosensors.com/article.aspx?ArticleID=281>
- [12] MICRO-MEASUREMENTS. (2014) Surface Preparation for Strain Gage Bonding. [Online]. Available: <http://www.vishaypg.com/docs/11129/11129B129.pdf>
- [13] V. P. Group. Cea Strain Gage Installation with M-Bond 200 Adhesive (Training Video) - Micro-Measurements. Youtube. [Online]. Available: <https://www.youtube.com/watch?v=SjXpF61HRys>