

Tasks Measurement Intake Sheet

Steps N°	Time .Group	Device	Task	m's EMG y/n	Scores & comments (time : see EMG data)
1	0	-	Init. installation w EMG & adjustmts of: Baseline: DynAREach on WC Device: DynAREach on TBO	1: Biceps brachii 2: Triceps brachii 3: Anterior deltoid 4: Posterior deltoid 5:N/A 6: N/A	
2	1	Baseline	Installation		
3	1.1	Baseline	Rest (45°)		
4	1.1	Baseline	Shoulder flexion (45° to 90°)		
5	1.1	Baseline	Elbow flexion (45° to 120°)		~1 sec maintaining effort at the end
6	1.2	Baseline	Pointing time 5x5 pictogram board (for 5 same & random picto as shown)		
7	1.2	Baseline	HFT Large Light Objects		Final rest slightly cut?
8	1.2	Baseline	Switch Joystick & Grid game: center start, time taken to reach 4 walls of grid		Each movement by a series of clicks (like in session 1)
9	1.3	Baseline	Palmar Grip EMG & score		Kg: 53
10	1.3	Baseline	HFT Simulating Feeding		Small hesitation at the beginning?
11	1.3	Baseline	BBT		Blocks: 69
12	2	Device	Installation		
13	2.1	Device	Rest (30°)		
14	2.1	Device	Shoulder flexion (30° to 90°)		

15	2.1	Device	Elbow flexion (30° to 120°):		
16	2	Baseline	Installation		
17	2.2	Baseline	Pointing time 5x5 pictogram board (for 5 same & random picto as shown)		
18	2.2	Baseline	HFT Large Light Objects		
19	2.2	Baseline	Switch Joystick & Grid game: center start, time taken to reach 4 walls of grid		SW-Joystick + HitchUSB ArrowKeys
20	2.3	Baseline	Palmar Grip EMG & score		Kg: 55
21	2.3	Baseline	HFT Simulating Feeding		
22	2.3	Baseline	BBT		Blocks: 66
					15 min pause and battery charging
23	3	Device	Installation		
24	3.1	Device	Rest (30°)		
25	3.1	Device	Shoulder flexion (30° to 90°)		
26	3.1	Device	Elbow flexion (30° to 120°):		
27	3.2	Device	Pointing time 5x5 pictogram board (for 5 same & random picto as shown)		
28	3.2	Device	HFT Large Light Objects		
29	3.2	Device	Switch Joystick & Grid game: center start, time taken to reach 4 walls of grid		SW-Joystick + HitchUSB ArrowKeys
30	3	Baseline	Installation		
31	3.3	Baseline	Palmar Grip EMG & score		Kg: 52
32	3.3	Baseline	HFT Simulating Feeding		First one slightly difficult to take
33	3.3	Baseline	BBT		Blocks: 71
34	4	Device	Installation		
35	4.1	Device	Rest (30°)		
36	4.1	Device	Shoulder flexion (30° to 90°)		
37	4.1	Device	Elbow flexion (30° to 120°):		
38	4.2	Device	Pointing time 5x5 pictogram board		~5 sec without motion at the beginning

			(for 5 same & random picto as shown)		
39	4.2	Device	HFT Large Light Objects		
40	4.2	Device	Switch Joystick & Grid game: center start, time taken to reach 4 walls of grid		SW-Joystick + HitchUSB ArrowKeys
41	4.3	Device	Palmar Grip EMG & score		Kg: 54
42	4.3	Device	HFT Simulating Feeding		4 th one: fallen/taken twice
43	4.3	Device	BBT		Blocks: 74. Channel 3 replaced by 5
44			Uninstall sensors & device(s)		Channel 3 replaced by 5

NOTES : N/A