

TEMPERATURE SENSOR USER MANUAL



Temperature sensor user manual

Copyright © 2024 · XYMA ANALYTICS PVT LTD

Version Feb 2024

TEMPERATURE SENSOR USER MANUAL

XYMA ANALYTICS PVT LTD

TABLE OF CONTENTS

- WARRANTY AND LIABILITY 5
- CONGRATULATIONS WITH YOUR NEW PRODUCT! 6
 - Support, ordering, and contact information 6
- OVERVIEW 8
- GETTING STARTED 9
 - Unpacking a new Sensor 9
 - Connect the temperature Sensor 9
 - Calibration 9
- MEASUREMENTS 11
 - Microsensors 11
 - Macrosensors 11
 - Electrical noise 11
- STORAGE AND MAINTENANCE 12
 - Storage 12
 - Cleaning the Sensor 12
- REFERENCES 13
- TROUBLE SHOOTING 14
- SPECIFICATIONS 15

WARRANTY AND LIABILITY

Notice to Purchaser

This product is for research use only . Not for use in human diagnostic or therapeutic procedures .

Warning

Microsensors have very pointed tips and must be handled with care to avoid personal injury and only by trained personnel .

XYMA recommends users to attend instruction courses to ensure proper use of the products .

Warranty and liability

The Temperature sensor is covered by a 90 days (glass sensor) or a one year (Temp Amp + TP2000) limited warranty . Microsensors are a consumable . XYMA Analytics will only replace dysfunctional sensors if they have been tested according with the instructions in the manual within 14 days of receipt of the sensor(s) . In no event will XYMA Analytics be liable for any direct, indirect, consequential or incidental damages, including lost profits, or for any claim by any third party, arising out of the use, the results of use, or the inability to use this product .

XYMA mechanical and electronic laboratory instruments must only be used under normal laboratory conditions in a dry and clean environment . XYMA Analytics assumes no liability for damages on laboratory instruments due to unintended field use or exposure to dust, humidity or corrosive environments .

Repair or Adjustment

Sensors and electrodes cannot be repaired . Equipment that is not covered by the warranty will, if possible, be repaired by XYMA Analytics with appropriate charges paid by the customer . In case of return of equipment please contact us for return authorization .

For further information please see the document General Terms of Sale and Delivery of XYMA Analytics as well as the manuals for the respective products .

CONGRATULATIONS WITH YOUR NEW PRODUCT!

Support, ordering, and contact information

If you wish to order additional products or if you encounter any problems and need scientific/technical assistance, please do not hesitate to contact our sales and support team . We will respond to your inquiry within one working day .

E-mail: info@xyma.in

XYMA Analytics
B4-04, 4th Floor, B Block,
IITM Research Park, Kanagam,
Tharamani, Chennai,
Tamil Nadu 600113.
Tel: [+91-9442949347](tel:+91-9442949347)

Further documentation and support is available at our website www.xyma.in.

REPLACEMENT OF SENSORS

Unisense will replace sensors that have been damaged during shipment provided that:

- *The sensors were tested within 2 weeks after receipt in accordance with the delivery note and the manual*
- *The seal is still intact (TP-type sensors, Table 1).*
- *The sensors are returned to Unisense for inspection within two weeks.*
- *The sensors are correctly packed for return to Unisense, in accordance with the note included in the sensor box.*

A sensor with sign of physical damage and with a broken seal, indicating that it has been removed from the protection tube, will not be replaced (TP-type sensors, Table 1)

SenSor	Amplifier	Calibration	Principle
Temp-UniAmp	UniAmp	Pre-calibrated	PT1000
Opto-Temp Field	FOM-UniAmp	Pre-calibrated	PT1000
Op-Temp	MOM	Pre-calibrated	PT1000
Op-Temp Field	FOM	Pre-calibrated	PT1000
TP-200	UniAmp, MMM,FMM, In situ	Calibration by user	Thermocouple
TP-500	UniAmp, MMM,FMM, In situ	Calibration by user	Thermocouple
TP-MR	UniAmp, MMM,FMM, In situ	Calibration by user	Thermocouple
TP-N	UniAmp, MMM,FMM, In situ	Calibration by user	Thermocouple
TP-NP	UniAmp, MMM,FMM, In situ	Calibration by user	Thermocouple
TP-2000	MMM, FMM	Calibration by user	Thermocouple

Table 1: Sensor types

Amplifier types:

UniAmp = UniAmp laboratory amplifiers

MMM = Microsensor Multimeter and Microsensor Monometer

FMM = Field Multimeter

In Situ = In Situ UniAmp and In Situ Amplifier

FOM-UniAmp = Field Opto UniAmp

MOM = MicroOptode Meter

FOM = Field Optode Meter

See <https://www.xyma.in/temperature> for detailed specifications and customizations .

OVERVIEW

This manual covers all XYMA Analytics temperature sensors as listed in Table 1 .

All sensors based on a PT1000 element (see Table 1) are pre-calibrated and the temperature will be shown in °C in the SensorTrace software .

Sensors based on thermocouple elements will give mV signals in the SensorTrace software and will need to be calibrated to show the temperature in °C .

The TP2000 and the PT1000 based temperatures sensors are all robust sensors with a tip diameter of 2 mm (see Table 2) . They are general purpose sensors and can be used in many applications where temperature measurements are needed for monitoring or temperature compensation of other sensor signals .

The XYMA glass temperature microsensor consists of a thermocouple inside a tapered glass capillary . This sensor can be used to determine temperature micro gradients in different environments e .g . hot spring biofilms, compost piles, and sediments with steep temperature gradients due to volcanic activity . With tip diameters down to 200 μm , the XYMA temperature microsensor is ideal for temperature measurements with a very high spatial resolution .

WARNING

XYMA sensors are neither intended nor approved for use on humans



Temp-UniAmp

2000

TP-



TP-200

GETTING STARTED

Unpacking a new sensor

The sensors must be tested within 2 weeks after receipt to verify that they have survived the transport in order to maintain the warranty (please see 1.3 Warranty and Liability) .

The PT1000 based sensors (Temp-UniAmp, Opto-Temp Field, Op-Temp, Op-Temp Field) can be used right out of the box without calibration .

When receiving a new microsensor remove the shock-absorbing grey plastic net . **Do not remove the seal and protective tube before the sensor has been tested found fully functional.**

WARNING

Do not remove seal and protective plastic tube before the following steps are successfully completed.

Connect The Temperature sensor

The temperature sensors must be connected to the temperature connector on the amplifier:

- Laboratory UniAmp, Multimeter, and Monometer: Connector is labelled T .
- Field Microsensor Multimeter: Channel 8 .
- Field Opto UniAmp and Field Optodometer: Connector for temperature sensor (only shallow version, deep version has built in temperature sensor) .
- **Calibration**

The thermocouple-based temperature sensors respond linearly to changes in temperature and only a two-point calibration is required .

Prepare two solutions with known temperature, one below and one above the temperature range where the measurements will be done . Measure the temperature of these solutions with a high accuracy thermometer and perform a two-point calibration in the SensorTrace software . Please see the SensorTrace Suite User Manual which can be downloaded from: <https://www.unisense.com/manuals/> .

MEASUREMENTS

Microsensors

The glass microsensors with a tip diameter of 200 or 500 μm (TP-200, TP-500) are fragile and should be positioned with a Micromanipulator . We recommend that measurements are performed in a stabilized set-up fixed on a sturdy table, free of moving or vibrating devices . We recommend the XYMA Lab Stand (LS) and the XYMA micromanipulator MM-33 or MM33-2 for laboratory use . For in situ use we recommend our In Situ Stand (IS19) and a micromanipulator .

The needle type sensors are glass sensors mounted in a needle and should be treated with care . It is important not to bend the needle as this may cause the glass sensor inside to break .

The TP-MR is designed for use with the XYMA Microrespiration System which will safely guide the sensor into the respiration chambers .

Macrosensors

The macro temperature sensors (Temp-UniAmp, Opto-Temp Field, Op-Temp, Op-Temp Field, TP-2000) may be used as a normal temperature sensor . **Electrical noise**

The electrical current generated by the high-impedance temperature microsensor is very small and may be affected by electrical noise . The XYMA temperature microsensor is shielded against electrical noise, however, strong electrical fields may interfere with the sensor signal . Minimize this by switching off unnecessary electrical/mechanical equipment and avoid touching sensor or cables during operation . On suspicion of sensor damage, repeat calibration and consult 8 Troubleshooting .

WARNING

Always introduce and retract the temperature microsensor axially using a micromanipulator and a stable stand when measuring in solid or semisolid substrate like sediment, tissue, biofilms, microbial mats etc.

STORAGE AND MAINTENANCE

Storage

Store the sensor the same way it was shipped . Mechanical shock should be avoided for the glass sensors . **Cleaning The sensor**

The glass sensors can be cleaned with different solutions . The standard method is to rinse it with ethanol, followed by 0 .01 M HCl and finally water . This will remove most substances . Alternatively, it is possible to rinse with 0 .1M NaOH, isopropanol or a detergent .

The macro sensors may be cleaned in a similar way as the glass sensor but the exposure to acidic or alkaline solutions should be brief, followed by rinsing with water .

REFERENCES

- Revsbech, N . P ., and B . B . Jørgensen . 1986 .
Microsensors: Their Use in Microbial Ecology, p
. 293-352 . In K . C . Marshall (ed .), *Advances in
Microbial Ecology*, vol . 9 . Plenum, New York .

TROUBLE SHOOTING

Problem	Sensor signal drifts
Possible cause	Sensor tip is broken
Solution	Replace the sensor
Problem	Noisy signal
Possible cause	Electrical interference
Solution 1	Turn off other equipment in the vicinity that could possibly emit electrical noise
Solution 2	Move the setup to a room without such equipment e.g. an office. If the noise disappears, it is likely that there is equipment in the original location that emits noise.

If you encounter other problems and need scientific/technical assistance, please contact sales@unisense.com for online support (we will answer you within one workday)

SPECIFICATIONS

	tp2000	temp-UniAmp*	GIASS
Outside tip diameter	2 mm	2 mm	TP-200 (180-220µm) TP-500 (400-600µm) TP-N (1.1 x 40 mm) TP-NP (1.6 x 40 mm) TP-MR (400-600 µm)
Diameter 20 mm from tip	2 mm	2 mm	<2 mm
Diameter 50 mm from tip	-	2 mm	8 mm
Total length	40 mm	50 mm	150-250 mm
Shaft diameter	8 mm	6 mm	8 mm
Cabling etc.			
Cable	Habia	4-wire PTFE	Standard thermocouple cable (type T)
Cable length	1.5-2 m	2 m	1.5-2 m

Connector	Lemo	Lemo	Standard thermocouple connector (type T)
Characteristics			
Type	Mineral insulated thermocouple	PT1000	Thermocouple, copper/constantan (type T)
Waterproof	Yes	Yes	Yes
Temperature range	-40°C-80°C	-40°C-80°C	-10°C-70°C
Temperature range tip	-40°C-100°C	-40°C-80°C	-20°C-100°C
Guaranteed lifetime	1 year	1 year	3 months
Expected lifetime	Years	Years	>1 year
Spatial resolution	N/A	N/A	About 2 times outside tip diameter
Output	40 µV per °C	40 µV per °C	40 µV per °C
Resolution	0.1°C	<0.1°C	0.1°C
Response time (90%)	< 3 s	< 3 s	< 3 s

Table 2: Temperature sensor specifications

* Temp-UniAmp, Opto-Temp Field, Op-Temp, Op-Temp Field have same specifications but different connectors .



XYMA ANALYTICS PVT LTD, CHENNAI

www.xyma.in · info@xyma.in