

IS 3055 (Part 1):1994 - Clinical Thermometers â Solid stem Type

SCOPE: Requirements for solid stem clinical thermometers.

CLAUSE 3.2 - ACCURACY

Maximum permissible error: $\pm 0.2^{\circ}\text{C}$ in range 35°C to 42°C .

TABLE 3 - THERMOMETER TOLERANCES

Temperature Range | Permissible Error

35°C - 42°C | $\pm 0.2^{\circ}\text{C}$

Below 35°C | $\pm 0.3^{\circ}\text{C}$

MARKING: Each thermometer shall bear IS 3055 (Part 1):1994.