



VOLTAGRAM TESTING & CALIBRATION

Plot No.-C 88&89, Sector:- B3, TDS City, Ghaziabad, UP, Pin 201 103
MOBILE:- 09717318883, E-MAIL: voltagramindia@gmail.com



CALIBRATION CERTIFICATE

Certificate No. VTC/OEG/ 2021060704

Calibrated for (Name & Address)

M/s. Onkar Engine & Generator Pvt. Ltd.

D - 255, Sector - 63,
Noida, UP - 201 301

Received on : 07/06/2021
Calibrated on : 07/06/2021
Calibration Due on : 06/06/2022
Issued on : 07/06/2021
Page(s) : 1 of 1

Description of Equipment / Instrument

Instrument Name : Dial Pressure Gauge
Make / Model : Wika
Range : 0 - 300 Psi
Least Count / Resolution : 5 Psi
Location : Lab

Serial No. : ---
Accuracy : ---
Type : ---
Condition of Equip. : Normal
Equip. / Party ID : ---

Description of Standard /Reference Equipment(s) / Instrument(s) used for calibration

S. No.	Nomenclature	Make / Model	Certificate No.	Calibrated By	Cal. Due on
1	Pressure Gauge	Keller	CC220921000000329F	UTC, Karnal	10/03/2022

Reference Standard/Specification	Calibration Performed at	Calibration Procedure
Direct Method	Lab	VTC/QP/ 42

Parameter Pressure

S. No.	UUC Value(S) Psi	Avg. Standard Value (M) Psi				(±)MU** %
		Ascending	Error (S-M)	Descending	Error (S-M)	
1	25	24.869	0.131	24.675	0.325	1.2
2	75	75.024	-0.024	75.021	-0.021	
3	150	150.207	-0.207	150.212	-0.212	
4	200	200.711	-0.711	200.714	-0.714	
5	300	300.015	-0.015	300.017	-0.017	



C.P. Saxena
07/06/2021
C.P. SAXENA
APPROVED BY (QM/TM)

Note:- 1) Calibration temperature:- 25±2°C; Relative Humidity:- 45 to 75%; 2) The calibration results reported in this calibration certificate are valid at the time of & under stated condition; 3) This certificate cannot be reproduced except in full without our prior permission in writing; 4) This certificate refers only to the particular item(s) submitted for calibration; 5) DUC/UUC – Device Under Calibration / Unit Under Calibration; 6) Measurements are traceable to National standards through unbroken chain of calibration from NABL accredited/Competent laboratory; 7) **Measurement of Uncertainty at approx 95.45% confidence level and coverage factor k=2. {Revision-02, Revision Date-01/04/2021, VTC/QF/19}