



Multiple-Function Analysis

ARKA Health, Inc. — Pre-Submission · Document 05 · Version 1.0 · June 9, 2026

This analysis follows FDA's guidance *Multiple Function Device Products: Policy and Considerations*. ARKA is a single platform with distinct software functions; each function is presented with its regulatory basis so the function under review can be assessed on its own merits.

1. Function map

Function	What it does	Regulatory basis presented	Status requested
ARKA-CLIN	Imaging-appropriateness recommendations from structured data; ranked, evidence-cited options with reviewable basis	§520(o)(1)(E) — Non-Device CDS (four criteria)	Concurrence (Question 1)
ARKA-INS	Benefit-eligibility determination, claims-based utilization/cost analysis, prior-authorization documentation, cost transparency, scheduling	§520(o)(1)(A) — administrative support of a health care facility	Concurrence (Question 2)
Reference viewer	Displays previously acquired DICOM studies as non-diagnostic thumbnails for study identification	Separate display function; contributes nothing to any recommendation	Out of scope — not under review

2. Why the functions do not contaminate one another

ARKA-CLIN never consumes images or signals. Input validation accepts only structured FHIR resources, and continuous-integration guards (`scripts/regulatory-checks.ts` and the import-lint in `scripts/lint-scope-boundary.ts`, executed on every push and pull request by `.github/workflows/go-live.yml`) fail the build if image- or signal-processing code enters in-scope CDS paths. The reference viewer is import-fenced behind a documented scope boundary (`docs/SCOPE_BOUNDARY.md`); no CDS card content is derived from its code path.

ARKA-INS outputs are administrative, not clinical. Coverage, denial-risk, cost, and scheduling outputs concern benefits, finance, and workflow. They do not alter ARKA-CLIN's guideline-anchored clinical logic, and they are visually and semantically distinct card types.

No function changes the analysis of another. Under the multiple-function framework, ARKA-CLIN — the function for which a CDS determination is requested — is assessed on its own. The other functions introduce no image or signal processing, no clinical directives, and no non-reviewable outputs into ARKA-CLIN's path.



3. Requested concurrence

ARKA requests FDA's concurrence with this mapping: ARKA-CLIN assessed as Non-Device CDS under §520(o)(1)(E); ARKA-INS as administrative support under §520(o)(1)(A); and the reference viewer treated as a separate display function not part of this review.

4. Evidence attached with this document

- Scope boundary documentation (docs/SCOPE_BOUNDARY.md)
- CI enforcement configuration (.github/workflows/go-live.yml ; scripts/regulatory-checks.ts ; scripts/lint-scope-boundary.ts)