

Agilent OpenLab CDS Chromatography Report

Equipment Class: Chromatography

System ID: AGILENT_1290_UV01

Manufacturer: Agilent Technologies

Model: 1290 Infinity II LC System

Detector Type: UV-VIS

Wavelength: 254.0 nm

Flow Rate: 1.2 mL/min

Sample ID: BATCH_AGILENT_PDF_01

Sample Name: Aspirin Tablets 325mg

Analyte Name: Acetylsalicylic Acid

Ret. Time (min): 3.85

Area (mAU*min): 1850.40

Height (mAU): 340.20

USP Tailing: 1.15

N (theoretical plates): 9200

CV (%): 0.32