



INTERNATIONAL JOURNAL OF CREATIVE RESEARCH THOUGHTS (IJCRT)

An International Open Access, Peer-reviewed, Refereed Journal

ENHANCED RESOLUTION TECHNIQUES FOR MEDICAL IMAGE PROCESSING

¹**Author: Shams Tabrez, Research Scholar at ECE department at Sri Satya Sai University of Technology & Medical Sciences, Sehare-MP**

²**Author: Dr. Vijay Prakash Singh, Professor at ECE department at Sri Satya Sai University of Technology & Medical Sciences, Sehare-MP**

Abstract: Medical imaging, video surveillance, and astronomy all benefit from high-resolution pictures, which may also be used for other purposes. Images are obtained in medical imaging for a variety of reasons. for the goal of medical investigation and for the purpose of giving information concerns the structure, physiology, and metabolism of the human body below-the-skin volume Medical imaging is crucial. a diagnostic tool for determining the existence of a specific disease illnesses. As a result, boosting the picture resolution should be considered. considerably enhance the capacity to diagnose remedial action treatment. Furthermore, a higher resolution may have a significant impact. Various ways for generating and forming algorithms that may be utilised to expand on this idea of Super resolution have been developed over the years. This study discusses a few different forms of medical imagination, as well as the numerous ways for super resolution and the current developments. This research shows how to use a unique interpolation approach to recreate three-dimensional medical pictures. Interpolation of images is frequently necessary. in the reconstruction of three dimensions This is a paper about examines and compares the limitations of established methods. Grey-based interpolation is an interpolation algorithm. interpolation based on shape In order to overcome the Drawbacks of standard algorithms, a revolutionary algorithm has been developed. There is a suggested interpolation algorithm that combines the advantages of both traditional and non-traditional interpolationalgorithms. Extensive research has revealed that the It is possible to improve the efficiency of three-dimensional reconstruction. By utilising the unique method, we were able to significantly increase our results.

Keywords: high-resolution medical pictures; Grey-based interpolation, computed tomography (CT).

1. INTRODUCTION

Digital image processing is the adjusting of advanced pictures with the help of an advanced PC. A computerized photo is one that has been changed over from an actual arrangement to an advanced one and afterward handled. We need to upgrade our picture or concentrate valuable information from a specific area of interest. Picture handling is a significant report region in designing and software engineering. To work on the nature of computerized pictures, advanced handling strategies are utilized to address surrenders in crude information from imaging sensors on different stages. Such photos must be subjected to long-term processing in order to overcome anomalies and aid in the recovery of the original data. All sorts of visual data go through three stages. Pre-Processing, Enhancing, and Information Extraction are steps to take when employing digital picture enhancement techniques. It is desirable to achieve an in practically every

application. a photo with an extremely high resolution In medical imaging, a high-resolution image may improve image classification or help pinpoint the exact location of a tumour with more accuracy, but it can also increase a patient's cancer risk. consumer satisfaction with high-definition video and online photographs on high-definition TVs The size of a picture has a direct correlation to the resolution of that image. The picture acquisition device's resolution As a result, As the resolution of the picture produced by a sensor improves, a cost-effective solution As a result, the sensor's price rises, making it an unaffordable option. As a result, the most crucial thing to know is whether or not there is a technique to increase the image's resolution. Super resolution is characterized in this work, as well as the way things are utilized in clinical picture handling.

2. SUPER RESOLUTION AS A CONCEPT

Recuperation of a high-resolution picture from a progression of low-goal photographs is the objective of the "Super Resolution" and "Super Resolution Reconstruction" techniques. Additionally, Super Resolution uses other information, for example, an information base that learns the connection among low and high-goal pictures (SR). Target identification, recognition, and tracking are just a few of the uses of superresolution in image processing and computer vision.

Picture memory is used in modern television receivers for upsampling and decoding. High-quality images may be retrieved from a sequence of low-quality shots using this method. Noise, blur, and downscaling have all taken place in these LR images. This method's crucial thought is to blend some low-goal or loud shots gotten some margin to create a solitary, high-goal picture or series of pictures. A top notch picture is resampled into an inferior quality one to make another picture. The last advance is to recover the high-goal picture. Whenever this high-goal picture is resampled in light of the info photos and the imaging model, the low-goal noticed pictures are made. The incredible majority of the superior quality pictures. It is classified "three-layered remaking of clinical imaging" (TDRMI) when a 3D model of the analyzed designs is made from a progression of 2D pictures. CT, MR, and SPECT [2] cuts are utilized to develop 2D photos of these surfaces, which assist specialists with figuring out the intricate designs in the cuts and help in the determination and treatment. Clinicians might profit from utilizing three-layered recreation of clinical pictures to dissect pathology in gatherings [3]. This technology may significantly boost medical diagnostic accuracy.

Figure 1 depicts the entire procedure. The original pictures were created in accordance with the DICOM 3.0 standard [4]. Filtering and interpolation of pictures are examples of image preprocessing. After preprocessing, picture segmentation and reconstruction are carried out. The insertion strategy is utilized to address the surmised guess of sectional pictures or voxels, which is the main piece of picture preprocessing. There are two methods for completing introduction: To complete voxels interpolation while resampling data and to construct a new picture from two adjacent images according to a set of defined criteria, respectively. This research employs the second approach.




Fig 1: The three-dimensional reconstruction flow chart

Because of a number of factors, it's hard to collect data at the correct discrete level every time: The first problem is that we may not be aware at the time of collection of what the needed discrete level will be. Second, the imaging device's capabilities limited the level that could be achieved. The third reason is that data collection takes less time, and the fourth reason is the massive quantity of storage that would otherwise be necessary.

Finally, X-ray technology such as CT scanners may have a limit on the amount of radiation they can deliver to a person [5]. As a consequence, interpolation technology is necessary for the reconstruction of medical images in three dimensions.

2.1 Sectional image interpolation methods

The following three conditions must be met when interpolating between sectional images: Sectional pictures must be comparable to each other; the distance between them determines how closely the interpolation image resembles its source images; and the interpolation image sequence must demonstrate a progressive shift from one source to another.

An interpolation technique's quality is critical, yet there are presently no uniform standards. There are two types of common criteria for measuring picture quality, based on the characteristics of digital images and human vision: level headed and emotional models for assessment In this venture, we utilize abstract assessment rules. Interjection strategies incorporate shape-and dark based addition. In the dark based introduction approach, addition focuses are utilized to describe the local dim qualities between the source pictures, and afterward the dim upsides of the insertion point are processed utilizing this data. Dim based addition approaches incorporate closest neighbor introduction, straight interjection, Lagrange insertion, spline addition, and different techniques. Shape-based insertion fabricates the form of the addition picture straightforwardly from the state of existing area pictures. Shape-based insertion incorporates distance change based and morphology-based interjection. Both depend on changes. This article will cover Lagrange introduction as well as distance change based insertion.

2.2 The interpolation method of Lagrange

Using numerical computation, the Lagrange interpolation polynomial may be utilised to describe many points using polynomials. Consequently, the Lagrange interpolation polynomial is a helpful computing tool in approximation theory.

The polynomial of Lagrange interpolation can be written as:

$$L_m(x) = \sum_{j=0}^m y_j l_j(x) = \sum_{j=0}^m y_j \prod_{\substack{i=0 \\ i \neq j}}^m \frac{x - x_i}{x_j - x_i} \quad (1)$$

There are many ways to write this interpolation's foundational function in Lagrange notation $(0, 1, \dots, j, \dots, m)$. This polynomial has $m+1$ things of m -request premise capacities, and the mix coefficients are the upsides of the hubs work.

Whenever $m = 0, 1$, the Lagrange technique addresses the closest neighborhood introduction and direct addition calculations. The premise capacity of two introduction hubs ($m = 1$) might be addressed as follows:

$$l_i(x) = \frac{x - x_{i+1}}{x_i - x_{i+1}} \quad l_{i+1}(x) = \frac{x - x_i}{x_{i+1} - x_i} \quad (2)$$

The Lagrange interpolation polynomial can thus be written as:

$$\begin{aligned} L_h(x) &= y_i l_i(x) + y_{i+1} l_{i+1}(x) \\ &= y_i \frac{x - x_{i+1}}{x_i - x_{i+1}} + y_{i+1} \frac{x - x_i}{x_{i+1} - x_i} \end{aligned} \quad (3)$$

$[X_0, X_m]$ is separated into many interpolation sub-ranges when m is big. The span off not set in stone at the hubs, and an insertion polynomial is produced for each sub-range utilizing those focuses.

The distance transform-based interpolation technique

Three-dimensional tooth reconstruction by Bors [7] makes use of Binary morphological shape-based interpolation, an approach that draws on morphological information for the first time. Simple and fast shape-related interpolation may be achieved using the distance transform method. The two sectional pictures are labelled S_i and S_{i+1} , respectively, with the corresponding borders labelled C_i and C_{i+1} . The distance between any point (P_i, P_{i+1}) and their associated limits is marked as D_i and D_{i+1} in the original sectional pictures.

It follows that: (1) if point P is on the border C , then $D > 0$; (2) if point P is outside the border C , then $D = 0$; (3) when point P is on the border C , then D equals 0.

Finding the closest points P_t on the interpolated picture to the original images I_S and I_{1S} along the Z -axis, this method seeks the shortest distance from P_t on I_S and I_{1S} to P_i and P_{i+1} on the original images, then uses equations 4 to calculate P 's distance from his corresponding boundary.

$$D_t = (x - x') \times D_k + x' \times D_{k+1} \quad (4)$$

2.3 Improvements to the interpolation algorithm

The grey-based interpolation technique, as is generally known, is simple and fast, but it performs poorly. The interpolated image's contour will be blurry and pixels regional fracture phenomena will occur if the graydistributing of medical sectional images is difficult. To fix these faults in interpolation, a shape-based approach would require a long time. This paper proposes a better interjection approach that consolidates the extraordinary matchability of dim based insertion with the great edge impact shape-based addition to resolve these issues. The method also incorporates the notion of multi-volume interpolation. We label the two original pictures with S_i and S_{i+1} in the image sequences, and the image to be interpolated with S_t . The following are the particular measures to take:

1st step: The contour of S_i and S_{i+1} is determined using the enhanced interpolation procedure. Tissues of varying densities have varied grey values.

Soft tissue, fat, and air all have lower grey values than bone. Uses structural components with certain forms to measure and dilate images, and then derives the proper boundaries of C_i and C_{i+1} combined grey value distribution from these images.

Step 2: Using the distance transform-based interpolation approach, the best interpolation algorithm finds the border marked C_t of the interpolated image. On the interpolated picture t S , we pick one point P_t . In order to determine how far apart the two matching spots are from each other, the computer compares the two original images. Using Equation 4, the new method calculates the distance between point P_t and its border and establishes the relationship between P_t and C_t . The method derives the boundary C_t by traversing all locations on the interpolated picture.

Step 3: Interpolated image locations are sought for by the new algorithm. Several previous original photographs of the image are marked.

$$L_h(x) = \begin{cases} L_h^{(0)}(x) & x \in [x_0, x_1] \\ L_h^{(1)}(x) & x \in [x_1, x_2] \\ \dots & \dots \\ L_h^{(i)}(x) & x \in [x_i, x_{i+1}] \\ \dots & \dots \\ L_h^{(m-1)}(x) & x \in [x_{m-1}, x_m] \end{cases} \quad (5)$$

Using the revised approach, we can retrieve the grey value of P_t , after which we visit every point on the interpolated picture in the same manner, yielding the final interpolated image S_t . The following equation is used to determine the value of i .

The outcomes of the experiment

$$L_h^{(i)}(t) = y_i \frac{t - x_{i+1}}{x_i - x_{i+1}} + y_{i+1} \frac{t - x_i}{x_{i+1} - x_i} \quad t \in [x_i, x_{i+1}] \quad (6)$$

The value of i is determined using the following formula.

$$i = \begin{cases} 0 & t \leq x_1 \\ i & x_i < t \leq x_{i+1} \quad (1 \leq i \leq m-1) \\ m-1 & t \geq x_{m-1} \end{cases} \quad (7)$$






Fig 2: Original CT scans in four sections

Four sequential, sectional, and filtered CT images were used in this investigation, as shown in Fig. 2. All of the photos have a 256x256 pixel resolution. When contrasted with the Lagrange introduction procedure and distance change based addition calculation, the better interjection approach runs tests between the initial two bits of sectional photos and the last three (c). To evaluate the exploratory outcomes on an additional individual level, we developed the middle part of every insertion picture.


Figure 3 depicts the experimental outcomes. The Lagrange interpolation technique produced the picture (a). For medical photos with complicated structures, the interpolated image contour is obviously ambiguous. Although its contour is more defined, image (b) has a covering shadow peculiarity that can be seen in specific spots of the halfway broadening picture. This is the aftereffect of a distance change based insertion approach. The improved introduction strategy created another photo (c).



(a) The Lagrange interpolation algorithm



(b) The distance transform-based interpolation



(c) Algorithms for better interpolation

Fig 3: Algorithms used for interpolation

Combining the advantages of two previous interpolation techniques, the new method uses a large number of photographs to interpolate, resulting in a crisp interpolated image with rich grey value information.

CONCLUSION

Targeting some of the classic interpolation's limitations This study presents a more efficient interpolation approach. This is determined by the algorithm. The interpolated picture is contoured, then interpolated. by combining the grey values of four different filtered CT images This creates an interpolation picture. algorithm combines the benefits of two classic methods. algorithms. The interpolated picture resembles the original. initial CT scans and in accordance with the progressive The picture sequences are processed in this way. The details and the edges The image's metadata is likewise nicely preserved. VC++ is used to implement all algorithms. The results of the experiments show that the improved The approach outperforms classical interpolation. In terms of subjective evaluation criteria, algorithms are used.

REFERENCES

- [1] P.V.Suhocki, "Say it with multimedia: new presentation tools for the radiologist", The American Journal of Roentgenology, 1998, vol. 171, no. 2, pp. 313-319.
- [2] DushyantV.Sahani, SanjeevaP.Kalva, "Tumor perfusion and treatment response in rectal cancer with multisession CT", Initial Observations Radiology' 2005, vol.23:785-792.
- [3] Ren Jingying, Chen Shuyue, "Research on interpolation methods for cross-sections slice image", Computer Measurement and Control, 2005, vol. 13, no. 7, pp. 729-733.
- [4] Jing Jin, Qiang Wang, Yi Shen, "An improved marching cubes method for surface reconstruction of volume data", Cambridge University Press, 2006, vol. 6, no.7, pp. 34-67.
- [5] Grevera G J, Udupa J K, "Shape-Based Interpolation of Multidimensional Grey-Level Images", IEEE Transactions On Medical Imaging, 1996, vol. 15, no. 6, pp. 881-892.

- [6] Chen M J, Huang C H, Lee W L, "A fast edge-oriented algorithm for image interpolation", Image and Vision Computing, 2005, vol. 23, no. 9, pp. 791-798.
- [7] Rors A G, Kechagias L, Pitas I, "Binary morphological shape-based interpolation applied to 3-D tooth reconstruction", IEEE Trans. Med. Imaging, 2002, vol. 21, no. 2, pp. 100-108
- [8] Keh-Shih Chuang, Chun-Yuan Chen, Liq-Ji Yuan and Ching-Kai Yeh, "Shape-based grey-level image interpolation", Phys.Med.Biol,1999, vol.44, no.6,pp. 1565-1577
- [9] S. Chaudhari, Department of Electrical Engineering, Indian Institute of Technology – Bombay, Mumbai, India 400 076, "Super Resolution Imaging". New York, Boston, Dordrecht, London, Moscow, Kluwer Academic Publishers.
- [10] D. Trinh, M. Luong, F. Dibos, J. M. Rocchisani, C. D. Pham and T. Q. Nguyen, "Novel Example-Based Method for Super-Resolution and Denoising of Medical Images", IEEE Transactions on Image processing, Vol. 23, No. 4, April 2014.
- [11] H. Okuhata, R. Imai, M. Ise, R. Y. Omaki1, H. Nakamura, S. Hara and I. Shirakawa, "Implementation of Dynamic-Range Enhancement and Super-Resolution Algorithms for Medical Image Processing" in 2014 IEEE International Conference on Consumer Electronics (ICCE).
- [12] F. Rousseau, D. Gounot and C. Studholme, "On High-Resolution Image Estimation using Low-Resolution Brain MRI", 35th Annual International Conference of the IEEE EMBS, Osaka, Japan, 3 - 7 July, 2013.
- [13] W. E. Hakimi and S. Wesarg, "Accurate Super-Resolution Reconstruction for CT and MR Images".
- [14] R. Islam, A. J. Lambert and M. R. Pickering, "Super Resolution of 3D MRI Images Using A Gaussian Scale Mixture Model Constraint", 978- 1-4673-0046-9/12/\$26.00 ©2012 IEEE.
- [15] D. Glasner, S. Bagon and M. Irani, "Super-Resolution from a Single Image", IEEE 12th International Conference on Computer Vision (ICCV), 2009.
- [16] A. Ozcelikkale, G. B. Akar, H. M. Ozaktas, "Accurate Super-Resolution Reconstruction for CT and MR Images", IEEE 17th International Conference on Image Processing, September 26-29, 2010, Hong Kong.
- [17] A. Souza and R. Senn, "Model-Based Super-Resolution for MRI", 30th Annual International IEEE EMBS Conference, Vancouver, British Columbia, Canada, August 20-24, 2008.