



Eidgenössische Technische Hochschule Zürich
Swiss Federal Institute of Technology Zurich

D MAVT



Sensory-Motor Systems Lab

Armando Daniel Amoros Lozano

Evaluation of Master Devices for TAVI/TAVR Teleoperated Robot

Semester Thesis

Sensory-Motor Systems Lab
Swiss Federal Institute of Technology (ETH) Zurich

Supervision

Prof. Dr. Robert Riener
PD Dr. med. Andre Plass Leitender

Thesis Advisor

Prof. Dr. Robert Riener

October 2019

Abstract

This thesis work intends to find the most suitable master device type for a TAVI/TAVR teleoperated robot, capable of controlling the 2 DOF (translation and rotation) of the different catheters and guide wires used during the intervention.

After a state-of-the-art research in catheter handling teleoperated robots, four different master devices were tested. Each device controls each one of the 2 DOF independently: The first device with completely digital inputs (Keyboard type), the second device hybrid with 1 digital input (translation) and 1 analogical input (rotation) (Remote Controller type), and the remaining two devices with completely analogical inputs (Joystick type and CatheterLike type).

The experiments were performed by 15 candidates from which 1 was an expert TAVI surgeon. Each device was tested under three different experiments and by the appreciation of the users, the first two experiments were designed as a follow the target task assessing the precision and response of each DOF independently. The third experiment was designed as a navigation task, using both DOF, measuring the time and smoothness of the movements and path followed until reaching the goal.

The results of the experiments and the user's poll responses indicate that the Joystick type device has a better overall performance for controlling the 2 DOF of a regular catheter used in TAVI-/TAVR surgery.

Acknowledgements

I want to thank all the people, family and friends, that supported me one way or another before and during the course of my masters. Special thanks to Joaquin Campos, who supported my academic way even before this journey started, to my brothers and parents that have been there to support me even from far away whenever I needed it. Finally, special thanks to Tanya, Claudia and Roland Hansalik, for being the family I needed in Switzerland.

Declaration of Originality

I hereby declare that the written work I have submitted entitled

Evaluation of Master Devices for TAVI/TAVR Teleoperated Robot

is original work which I alone have authored and which is written in my own words.¹

Author(s)

Armando Daniel Amoros Lozano

Supervision

Prof. Dr. Robert Riener
PD Dr. med. Andre Plass Leitender

With the signature I declare that I have been informed regarding normal academic citation rules and that I have read and understood the information on 'Citation etiquette' (http://www.ethz.ch/students/exams/plagiarism_s_en.pdf). The citation conventions usual to the discipline in question here have been respected.

The above written work may be tested electronically for plagiarism.

Place and date

Signature

¹Co-authored work: The signatures of all authors are required. Each signature attests to the originality of the entire piece of written work in its final form.

Symbols

Acronyms and Abbreviations

TAVI/TAVR	Transcatheter aortic valve implantation/replacement
AS	Aortic Stenosis
FDA	Food and Drug Administration
DOF	Degrees of freedom
RMSE	Root Mean Squared Error
DSJ	Dimensionless Squared Jerk
ETH	Swiss Federal Institute of Technology
SMS	Sensory-Motor Systems Lab
PVL	Paravalvular Leakage

Contents

Abstract	i
Acknowledgements	iii
Declaration of Originality	v
Symbols	vii
1 Introduction	1
1.1 TAVI/TAVR	1
1.2 Motivation	3
1.3 Teleoperation	5
1.4 Objective	5
2 State of the art	7
3 Chosen Master Devices and Characteristics	9
3.1 Keyboard	10
3.2 Remote	10
3.3 Joystick	11
3.4 CatheterLike	12
3.5 Mapping Types, Advantages and Disadvantages	13
4 Experiments and Results	17
4.1 Experiment Setup	17
4.2 Graphic Environment	17
4.3 MATLAB/Arduno Interface	18
4.4 Arduino/Device Interface	19
4.5 Physical Setup	19
4.6 Participants Statistics	19
4.7 Instructions	20
4.8 Practice Round	20
4.9 Experiment 1st and 2nd DOF	20
4.9.1 Overview	20
4.9.2 Results	21
4.10 Experiment Maze	28
4.10.1 Overview	28
4.10.2 Results	28
4.11 Poll Results	35
5 Conclusion	39
5.1 Conclusions and Discussions	39
5.2 Future work	41

A Experiments Instructions	43
B Participants Survey	47

List of Figures

1.1	Aortic valve with AS (image taken from [2])	1
1.2	Balloon catheter retracting the calcified valve (image taken from [2])	2
1.3	Artificial valve deployment (image taken from [2])	2
1.4	TAVI surgery room with fluoroscopy on the left (image taken from [11])	3
1.5	Lead suit commonly used for TAVI procedure (image taken from [6])	4
1.6	Teleoperation diagram	5
2.1	Comparison between the state of the art devices in catheter teleoperation	8
2.2	State of the art devices, (1) The Amigo Remote System, (2) Sensei X, (3) Monarch, (4) Stereoaxis and (5) CorPath GRX	8
3.1	TAVI catheter degrees of freedom	9
3.2	Keypad model V7-KP1019-USB-4EB with USB connection	10
3.3	Remote device used for the experiments	11
3.4	CAD design of the remote device (push buttons C&K 8532T1ZBE2) (Rotatory sensor CTS 285CCDFSAAB4C1)	11
3.5	Joystick used for the experiments	12
3.6	CatheterLike device used for the experiment	13
3.7	CAD design of the catheter like design (Slide Pot PTA3043-2210CIB103) (Rotatory sensor CTS 285CCDFSAAB4C1)	13
3.8	Table of advantages and disadvantages of each mapping type	15
4.1	Experiment setup communication diagram	17
4.2	Default experiment graphic environment	18
4.3	Physical setup environment used for the experiments	19
4.4	Initial state of the 1st DOF experiment	21
4.5	Initial state of the 2nd DOF experiment	21
4.6	1st DOF RMSE per device across all worlds	22
4.7	1st DOF RMSE per device at each world	23
4.8	1st DOF RMSE per attempt, participant with no experience	23
4.9	1st DOF RMSE per attempt, participant with experience	24
4.10	2nd DOF Activation error while performing 1st DOF experiment	24
4.11	2nd DOF RMSE per device across all worlds	25
4.12	2nd DOF RMSE per device at each world	25
4.13	2nd DOF RMSE per attempt, participant with no experience	26
4.14	2nd DOF RMSE per attempt, participant with experience	26
4.15	1st DOF Activation error while performing 2nd DOF experiment	27
4.16	Initial state of the experiment maze	29
4.17	Collision during maze experiment	30
4.18	End of maze experiment when colliding with upper wall	30
4.19	Maze experiment average time per world	31
4.20	Maze experiment average time per device across all worlds	31
4.21	Maze experiment time training curve of all participants	32

4.22	Maze experiment average number of collisions per world	32
4.23	Maze experiment average number of collisions per device across all worlds	33
4.24	Maze experiment number of collisions training curve of all participants	33
4.25	Maze experiment DSJ per world	34
4.26	Maze experiment DSJ per device across all worlds	34
4.27	Participants percentage of preference per device on the 1st DOF	35
4.28	Participants percentage of preference per device on the 2nd DOF	36
4.29	Participants percentage of preference per device on the overall	36
4.30	Number of participants with experience with each master device	37
5.1	Devices qualitative table from users experience	40
5.2	Attempts of the expert TAVI surgeon in the 1st DOF experiment	40

Chapter 1

Introduction

1.1 TAVI/TAVR

TAVI/TAVR stands for Transcatheter Aortic Valve Implantation/Replacement (in the following only named TAVI), which is a minimal invasive surgery meant to treat AS (Aortic Stenosis), a condition caused by the calcification of the aortic valve (Figure 1.1), making it harder and thicker [4]. This condition in the valve disturbs the blood flow going into the aortic artery, making the heart work harder.



Figure 1.1: Aortic valve with AS (image taken from [2])

TAVI surgery is performed through an incision in the groin, where a sheath is placed in order to give access to a variety of guide wires and catheters to the aortic artery. Different guide wires and catheters are used to first gain access to the aortic arch, then go through it and finally gain access to the left heart ventricle crossing the calcified aortic valve. Each one of these steps requires a combination of axial and rotational movements performed by the surgeon to avoid damages in the artery and valve.

Once access to the left ventricle of the heart is granted, a balloon-expandable mounted on a catheter is inserted and positioned in the calcified valve. When in position, the balloon is expanded in order to retract the leaflets of the diseased valve (Figure 1.2).

Last, the new aortic valve is inserted mounted on a catheter and placed over the retracted old




Figure 1.2: Balloon catheter retracting the calcified valve (image taken from [2])

calcified valve. After the new valve is in position and fully expanded (Figure 1.3), the leaflets start working regulating the blood flow. If the placement is successful, the catheter and guide wire are retracted and the groin incision is closed.




Figure 1.3: Artificial valve deployment (image taken from [2])

During the whole surgery, the physicians have visual guidance help provided by fluoroscopy image (2D image), in order to track the position of the catheter and wires at any time, as well as the positioning of the balloon-expandable and the new aortic valve when being deployed.

1.2 Motivation

TAVI procedure has become more popular since the FDA approved the procedure for intermediate risk patients that present sever AS in 2016 [15], and estimates predict that numbers in North America and Europe will raise more than double (270 000 patients annually) if it is approved for low risk patients [5].

Each one of these surgeries presents a risk not only for the patients, but for every interventionalist present in the room, given that per each intervention interventionalists are exposed to a median of 5.5 mRad produced by the fluoroscopy imaging [1], with cardiologists, the medical professionals, exposed to the highest amounts of radiation [18]. Recent studies have shown that:

- 85% of the brain tumors found in interventionalists are located on the left side of the brain, consistent with the closest side of the brain exposed to the radiation source in the surgery room(Figure 1.4) [18].
- 50% of the interventionalists have significant posterior subcapsular lens changes, causing propensity to cataracts [22].



Figure 1.4: TAVI surgery room with fluoroscopy on the left (image taken from [11])

In order to mitigate the risks inherent to radiation exposure, a common practice is wearing a leaded suit as protective equipment. These suits have to be worn at any time while the fluoroscopy imaging system is turned on. Such suits are commonly made of lead and may weight up to 7 Kgs (Figure 1.5). This additional weight may cause orthopedic issues for interventionists as proven in recent studies:

- 60% have suffered spine issues after 21 years in practice [3].
- 33% miss work due to orthopedic issues [8].

This is why we believe TAVI procedures should be performed assisted by a teleoperated robot, which will set the interventionalists away from the radiation source. Although many teleoperated



Figure 1.5: Lead suit commonly used for TAVI procedure (image taken from [6])

robots already exist in the market for surgeries involving the use of catheters, none has been developed specifically for TAVI and its specific needs.

1.3 Teleoperation

A teleoperated surgical robot allows the surgeon to be located away from the operation table, and thus, for TAVI procedure away from the radiation source. Teleoperation has proven in similar surgeries to reduce the radiation in patients by 17% [20] and to reduce as well the exposure to the primary surgeon in 95% [23], beside the orthopedic benefits implied by not wearing the lead suits at all times during the surgery.

As depicted in Figure 1.6 the working station for the surgeon can be located away from the operation table, together with the fluoroscopy imaging screen and the master device, which may be actuated for haptic feedback, giving the surgeon another dimension given that the visual cues are poorly displayed in a black and white 2D image.




Figure 1.6: Teleoperation diagram

1.4 Objective

This thesis work intends to find the most suitable master device type for a TAVI/TAVR teleoperated robot, capable not only of controlling the 2 DOF (translation and rotation) of the different catheters and guide wires used during the intervention, but also capable of allowing the surgeons to succeed in the most complex maneuvers involving both DOF movements at the same time when crossing the aortic valve (Long maneuver times at this state increases the risk of cerebral embolism) [9] [13] and to perform accurate movements to position the artificial valve before deployment, mitigating this way paravalvular leak complications present in 8% of the TAVI procedures [7] [12].

Chapter 2

State of the art

Teleoperated robots for similar types of surgeries as TAVI have been developed. Figure 2.1 presents the robot's characteristics, such as the type of catheter they handle, the kind of intervention they were created for and the kind of master slave they operate with. On the other side figure 2.2 shows the pictures of the master devices.

Devices like Niobe [21] and Monarch [10] are highly costly and complex (much more than needed for a TAVI intervention, not mentioning that TAVI catheters could not be operated by Niobe magnetic fields, since they are plastic); on the other hand The Amigo system was designed to overcome these points having a simpler and cheaper designs [19]. Nevertheless, all these systems are designed to operate steerable catheters with 3 or more DOF, which make them an overkill for TAVI surgery, however, the concept behind their robotic devices could be simplified and adapted to only manage the 2 DOF necessary for TAVI.

Moreover, the CorPath GRX [17] is the system with the most similarities to what is needed for TAVI, handling 2 DOF catheters, guide wires and a stent balloon. However, TAVI surgery requires more than one catheter and guide wire to gain access to the left ventricle of the heart, as well as managing the new aortic valve deployer catheter.

These differences in characteristics and needs are the reason the devices in section 3 where selected, taking into account the inherent research and effort the devices of the state of the art suppose, some of them were just simplified enough to cover TAVI surgery needs.

Company – Robot name	Catheter Type	Surgery type	Master device type
Catheter Robotics Inc. – The Amigo remote control system	At least 3 DOF, steerable catheter	Electrophysiology (EP), and radiofrequency catheter ablation of arrhythmias	Remote Controller
Stereoaxis – Niobe	At least 3 DOF, steerable catheter	Endocardial ablation, gastrointestinal endoscopy, and others.	Regular 2D or 3D mouse
Hansen Medical – Sensei X	At least 3 DOF, steerable catheter	Atrial Fibrillation (AF), and other arrhythmia procedures.	Combination of keyboard, and haptic device in delta robot configuration.
Auris – Monarch	At least 3 DOF, steerable catheter	Bronchoscopic visualization and access to patient airways for diagnostic and therapeutic procedures.	Videogame console type remote controller with buttons and joysticks.
Commercial Robots – CorPath GRX	2 DOF catheters and guide wires.	Percutaneous coronary intervention (PCI) and Pulmonary vein isolation (PVI)	Joysticks

Figure 2.1: Comparison between the state of the art devices in catheter teleoperation



Figure 2.2: State of the art devices, (1) The Amigo Remote System, (2) Sensei X, (3) Monarch, (4) Stereoaxis and (5) CorPath GRX

Chapter 3

Chosen Master Devices and Characteristics

Every device was chosen to control independently each one of the catheter's DOF as shown in figure 3.1. In this section the mechanics, electronics and functioning of every input device is explained. After, a comparison between all device's characteristics, advantages and disadvantages is made.




Figure 3.1: TAVI catheter degrees of freedom

3.1 Keyboard

The keyboard device intends to represent any other device only controlled by on-off buttons, working in a digital configuration. For this experiment a numerical keypad was used (figure 3.2).

The control of the catheter was performed with the arrow key, with the UP and Down arrow key controlling the 1st DOF of the catheter. The 2nd DOF was controlled with the RIGHT and LEFT arrow keys.

Both types of movements are mapped to the simulated catheter with a PressedTime-Velocity mapping. For information about the mapping refer to section 3.5.



Figure 3.2: Keypad model V7-KP1019-USB-4EB with USB connection

3.2 Remote

The Remote device is a combination of on-off buttons and a semi-analogical sensor (figure 3.4 and 3.3). The shell of the device was 3D printed in ABS. It is also equipped with two push buttons, and a continuous 360 degrees rolling disk.

The 1st DOF of the catheter is controlled using the two push buttons in a digital configuration. Movements in this DOF are mapped to the simulated catheter with a PressedTime-Velocity mapping.

The 2nd DOF is controlled by a non-contacting rotatory position sensor attached to the rolling disk, which gives a semi-analogical reading with a 12 BIT resolution per 360 degrees. It is important to notice that the experiment setup as can be seen in section 4.4 uses an Arduino with a 10BIT ADC channel, which trims the initial sensor's 12BIT resolution. Movements in this DOF are mapped using a Velocity-Velocity mapping.

For information about the mapping refer to section 3.5.




Figure 3.3: Remote device used for the experiments




Figure 3.4: CAD design of the remote device (push buttons C&K 8532T1ZBE2) (Rotatory sensor CTS 285CCDFSAAB4C1)

3.3 Joystick

The Joystick device is implemented with fully analogical sensors (figure 3.5). Originally the Joystick had 3 DOF available, plus a push button on the top. The roll DOF and the button are not connected at all. Thus, performing any movement on them will not make any difference in the outputs.

The 1st DOF of the catheter is controlled by a regular 5Kohm potentiometer, mapped on the pitch movement of the joystick. The movements in this DOF are translated to the simulation catheter with a Distance-Velocity mapping.

The 2nd DOF is also controlled by regular 5Kohm potentiometer mapped to the yaw movement of the joystick.

The two DOF used, pitch and yaw, are limited on their angle and brought back automatically to

initial position by springs once any force is applied. The movements in both DOF are translated into the simulated catheter with a Position-Velocity mapping. For information about the mapping refer to section 3.5.

The potentiometers have essentially infinite resolution, only trimmed by the Arduino's ADC channel (See section 4.4 for experiment setup) used to read out the sensors.



Figure 3.5: Joystick used for the experiments

3.4 CatheterLike

The CatheterLike device is a combination of an analogic sensor and a semi-analogic sensor, figure 3.6 and 3.7. The shell of the device was 3D printed in ABS. The device has a handle on the outside that can be pushed and pulled axially coming back automatically to resting position (activated by springs), and can also be rotated over its own axis continuously 360 degrees.

The 1st DOF of the catheter is controlled by a 10Kohm and 30mm slide potentiometer. Activated by the push and pull movement on the device, it is mapped to the simulated catheter with a Position-velocity mapping.

The 2nd DOF is activated by the rotation of the device, mapped Velocity-Velocity to the simulated catheter and controlled by a non-contacting rotatory position sensor, which gives a semi-analogical

reading with a 12 BIT resolution per 360 degrees. It is important to notice that the experiment setup as can be seen in section 4.4 uses an Arduino with a 10BIT ADC channel, which trims the initial sensor's 12BIT resolution.

For information about the mapping refer to section 3.5.




Figure 3.6: CatheterLike device used for the experiment




Figure 3.7: CAD design of the catheter like design (Slide Pot PTA3043-2210CIB103) (Rotatory sensor CTS 285CCDFSAAB4C1)

3.5 Mapping Types, Advantages and Disadvantages

Every device with its different type of sensor is better adapted for a specific mapping. This means, how specific movements in each one of the devices, in each one of its degrees of freedom, translates to movements on the simulation catheter.

The **PressedTime-Velocity** mapping was created to mitigate the clear disadvantage devices with on-off buttons have against analogical/semi-analogical ones. This disadvantage is due to the fact that on-off devices have always the same speed applied over the catheter. If that speed is too high, it is not possible to apply small movement. On the other hand, if it is too low, it would take too much time to perform long movements. On the contrary, analogical sensor devices have a wide range resolution from where low and high speeds can be applied in the same configuration.

This mapping consists in incrementing the output velocity linearly with a defined slope in relation with the time the button was pressed, starting from an initial offset, eq 3.1.

$$outputVel = constant + (constant * pressedTime) \quad (3.1)$$

The **Position-Velocity** mapping is used in devices with analogical sensors, mapping linearly with a defined slope the position of the sensors to the velocity of the output catheters, eq 3.2.

$$outputVel = sign(position) * max(0, abs(position - threshold)) * constant \quad (3.2)$$

The threshold variable is meant to create a dead zone around the resting point of the device, given that the mechanism that automatically returns the device to the resting point is never perfect and it is also important to avoid accidental activation movements on the simulated catheter.

The **Velocity-Velocity** mapping is used for the semi-analogical rotatory sensors, which are able to rotate continuously 360 degrees (multi turn). This method maps the velocity of the device to the velocity of the simulated catheter, multiplied by a constant, eq 3.3.

$$outputVel = \frac{deviceCount(k) - deviceCount(k-1)}{Ts} * constant \quad (3.3)$$

Using these type of sensors simplifies the mechanics of the device but adds an additional challenge when getting the information from the Arduino to Matlab (See Experiment Environment section 4.3 for setup). This happens because the sample rate the Arduino board has to sample the device is much faster than the sample rate from Matlab to the Arduino, which makes the simulation take noisy data and the behavior of the simulation catheter look highly erratic. In order to overcome this and make the user experience smoother, the Arduino sample rate has to match with the Matlab simulation sample rate (Being T_s on eq 3.3 now the Matlab sample rate), making Arduino to implicitly take the average of all the movements between samples.

All the advantages and disadvantages are presented in figure 3.8.

Sensor Type	Mapping	Device	Advantages	Disadvantages
Analogic	Position - Velocity	<ul style="list-style-type: none"> • Joystick (Axial & Rotation) • Catheter (Axial) 	<ul style="list-style-type: none"> • High Resolution with high control 	<ul style="list-style-type: none"> • No instant stop, prone to overshoot • Additional mechanics
	Velocity - Velocity	<ul style="list-style-type: none"> • Remote (Rotation) • Catheter (Rotation) 	<ul style="list-style-type: none"> • Simplified mechanics • Instant stop 	<ul style="list-style-type: none"> • Resolution highly dependent on sampling rate (Arduino) in combination with communication rate (Matlab)
Digital	Pressed Time* - Velocity	<ul style="list-style-type: none"> • Keyboard (Axial & Rotation) • Remote (Axial) 	<ul style="list-style-type: none"> • Simpler for first time users • Instant stop when releasing the 	<ul style="list-style-type: none"> • Once released, it is necessary to build up velocity again • No control in velocity

Figure 3.8: Table of advantages and disadvantages of each mapping type

Chapter 4

Experiments and Results

4.1 Experiment Setup

The hardware of the experiment setup consists of a computer running the Matlab graphic simulation, the four master devices (Keyboard, Remote, Joystick and CatheterLike) and an Arduino board, used to read the sensors of the devices (the keyboard was connected directly to the computer) and communicated with Matlab through USB serial. The diagram is shown in figure 4.1.




Figure 4.1: Experiment setup communication diagram

4.2 Graphic Environment

The graphic environment is fully developed on Matlab, it consist of a representation of a catheter (With a green square following the tip) in a simulated fluoroscopy image, as surgeons would see it in the operation room, figure 4.2. This means the image showed on screen is a 2D black and white plain representation. The catheter can be moved in its 2 DOF, being the 1st DOF the axial movement (defined in simulation by the Y coordinate), the catheter would move up and down only. The 2nd DOF is the rotation (defined in simulation as radians starting from the left most possible

position of the catheter's tip), which is translated to a left and right movement on the screen.

Since the 2nd DOF is defined by the rotation (in radians) of the catheter, some dynamics are implicitly stated in the simulation during the left to right movements. As the image shown is intended to be a 2D representation of an actual 3D catheter in real life, when the catheter is rotating over its own axis a sinusoidal movement can be observed, moving slower when the tip of the catheter is on the most left/right side of the screen and moving fast when the tip is near the center of the catheter.




Figure 4.2: Default experiment graphic environment

4.3 MATLAB/Arduino Interface

The Arduino board and Matlab are connected through USB serial communication. The Arduino board is responsible to sense the devices' sensors. When the Matlab simulation starts, Matlab sends the necessary commands to Arduino in order to set it up for the specific device that is going to be in use, after this the Arduino board starts to capture the states of the sensors instantly.

Matlab simulation runs in a loop after setting up all the initial conditions. First it requests the Arduino board for the current states of the device. After it updates the graphics accordingly with the movements reported by Arduino. Lastly it stores the states in a log file, for later data analysis.

The Arduino board is programed with serial communication interruptions, which means it is constantly getting the current states of the device and storing them into a buffer. Once the Matlab code request the information, the Arduino board code will be interrupted in order to send the last saved states from the device; once this is done the Arduino board will go back to its state reading task.

The only device that works in a different fashion is the keyboard, which is connected directly to the computer through USB. This device interrupts Matlab directly when one of the keys is pressed and stores the device state directly in the simulation environment. In order to avoid adding/subtracting external latency, even though the Keyboard does not interact directly with the Arduino board, the simulation in Matlab runs the communication with the Arduino board, which returns by default the constant from equation 3.1. The constant form equation 3.1 is substituted by the `keyboardConst` defined in equation 4.1, with $\{-1,0,1\}$ the possible states of `keyboardMatlabState`. Then the specific equation for the keyboard changes from 3.1 to 4.2.

$$outputVel = keyboardConst = constant * keyboardMatlabState \quad (4.1)$$

$$outputVel = keyboardConst + keyboardConst * pressedTime \quad (4.2)$$

4.4 Arduino/Device Interface

The Arduino board has analogical and digital input in which the devices are directly connected to, according each device sensor types. Once the Arduino board is initialized by Matlab, readings to the corresponding inputs are performed and processed according to the programmed mapping type (see section 3.5). The processed information is stored in a buffer and the loop keeps running in the same way.

4.5 Physical Setup

All the experiments were run in the same environment under the same conditions. An external monitor was setup to run the graphic simulation over a desk with a chair for the experiment candidate to sit. Every device was presented in front the candidate to be taken when necessary according the running experiment, as shown in figure 4.3.

All the simulations start only once the candidate had performed the first movement. In that moment the simulation starts recording data, and once the simulation time or the goal is reached, it closes itself and immediately launches the next simulation, waiting for the candidate to perform the first movement again, until all the repetitions are finished.




Figure 4.3: Physical setup environment used for the experiments

4.6 Participants Statistics

The experiment was conducted by 15 participants, from the 15 participants:

- Only 1 of the participants was an expert surgeon in TAVI procedure
- The average age of the participants was: 30.23

- 46.15 were Male
- 53.85 were Female

4.7 Instructions

Before initiating the experiments, a sheet with the instructions set as the one shown in appendix A was handed to each one of the participants.

4.8 Practice Round

Before any experiment was performed, each participant had time to try and play with every device and how it interacts with the graphic simulation on each one of the DOF. An empty world simulation was displayed as shown in figure 4.2. The devices were randomly selected and handed to the participant, who could manipulate the simulated catheter until he/she would decide to stop or 2 minutes maximum had passed.

4.9 Experiment 1st and 2nd DOF

4.9.1 Overview

This experiment was designed to test the performance of the 1st DOF and 2nd DOF of every device independently. The objective is to try to collocate the tip of the catheter (green square) inside a target (red square) that is only moving up and down for the 1st DOF experiment (figure 4.4) and left to right for the 2nd DOF experiment (figure 4.5). The target is always initially collocated above the initial tip of the catheter position for the 1st DOF and to the right of the initial tip of the catheter position for the 2nd DOF experiment. The target starts moving automatically once the simulation detects the first movement of the user, following a predefined path for 10 seconds.

For each DOF three different predefined paths (worlds) for the target were created, trying to cover two main purposes, the first one was to exploit the capabilities of every device, reaction time, resolution, change of direction, acceleration and deceleration. The second purpose was to simulate movements surgeons would face in a normal TAVI procedure: 1) for the 1st DOF having to pull back rapidly after moving forward and small precise movements when placing the new valve (preventing PVA [7] [12]) or trying to cross the aortic valve, 2) for the 2nd DOF having to rotate to avoid a blockage in the artery or contacting with the wall of the aortic artery, and small precise rotation movements surgeons need to perform as technique for crossing the aortic valve [9] [13].

During the 10 seconds of simulation the RMSE between the tip of the catheter and the target is recorded in X and Y coordinates independently. This measurement was selected as it encloses information and gives insight of the experiment purpose previously mentioned.

For each one of the DOF every participant had to complete five times each one of the three target's predefined paths, which means 15 randomly ordered repetitions per device. The repetitions were all executed, one device at a time; however, the devices were ordered randomly at the beginning of the experiment. Also, the order on which every DOF experiment was performed, was randomized for every participant.




Figure 4.4: Initial state of the 1st DOF experiment




Figure 4.5: Initial state of the 2nd DOF experiment

4.9.2 Results

All the data from the 15 participants was gathered and the RMSE in both coordinates X and Y for each one of the two DOF computed.

As can be observed in figure 4.7, the 1st DOF experiment shows that the Joystick device has a better performance than the rest of the devices at every different world, followed by the keyboard. However, in figure 4.6 when showing all the samples from the 3 worlds together, two important things are observed. First, the joystick device seems to still show a better performance, but this cannot be concluded when comparing with the keyboard, due to interaction P-Value of 0.006. Secondly, the Joystick device shows a high number of atypical values far away from the box plot, which at first seemed to be failed experiments. However, if those atypical values are traced, the majority of them belong to the first trials of different participants. Thus, if each number of trials is plotted separately as in figure 4.8 (a participant without previous training using Joysticks), a pronounced training curve can be observed. On the other hand, if we plot the same information from a participant that reported experience with the type of the devices used for the experiment as in figure 4.9, the superior performance of the Joystick can be observed again followed by the keyboard at every attempt.

Another important thing to look at, is the accidental activation error, which means movements in the DOF that was not being tested at each experiment. In order to quantify these involuntary movements, since some of them were too small to be quantified, a percentage of trials where an activation error was present was calculated, as can be observed in figure 4.10. It is intuitive that the Keyboard has 0% of errors due to its digital characteristics, something that would be expected as well from the Remote device, however, that was not the case, since 2.7% of the cases had an activation error, which can be attributed to experiment errors or issues with the ergonomics of the device. Finally, it can be observed that the catheter had the higher activation error, followed by the joystick.

On the 2nd DOF the figure 4.12 shows that the keyboard had a better performance at every world than any other device, without a clear differentiation between the other three. Also, in the box-plots of figure 4.11 it can be observed how the Joystick, Remote and Catheter devices have similar distributions, and the keyboard has a better performance with a P-Value of 4.5e-10.

The same training behavior as in the 1st DOF experiment can be observed in the 2nd DOF for some of the devices, as shown in figure 4.13 from a candidate without experience manipulating devices. Also, in figure 4.15 the accidental activation error, of the 1st DOF in this case, shows a 0% in the Keyboard and Remote (devices involving digital inputs in at least one of the two DOF), followed by the Joystick with 3.5% of test cases and the Catheter with the worst performance with 8%. Both devices with 0% proven to be better than the Catheter with P-Value of 0.001, but failed to prove superiority over the Joystick due to an interaction P-Value of 0.03.




Figure 4.6: 1st DOF RMSE per device across all worlds




Figure 4.7: 1st DOF RMSE per device at each world




Figure 4.8: 1st DOF RMSE per attempt, participant with no experience




Figure 4.9: 1st DOF RMSE per attempt, participant with experience




Figure 4.10: 2nd DOF Activation error while performing 1st DOF experiment




Figure 4.11: 2nd DOF RMSE per device across all worlds




Figure 4.12: 2nd DOF RMSE per device at each world




Figure 4.13: 2nd DOF RMSE per attempt, participant with no experience




Figure 4.14: 2nd DOF RMSE per attempt, participant with experience




Figure 4.15: 1st DOF Activation error while performing 2nd DOF experiment

4.10 Experiment Maze

4.10.1 Overview

This experiment was designed to test the joint performance of the two DOF of every device. The objective is to navigate the tip of the simulated catheter (green square) through a one-way maze as shown in figure 4.16, until reaching the top part of the maze. The users were instructed (as shown in appendix A) to navigate the maze keeping the maximum distance to the walls as possible, avoiding collisions and reaching the top wall of the maze in the minor time possible.

The simulation starts recording the data when the first movement of the user is detected. In case of collision with a wall, the tip of the catheter stops moving and changes color from green to red (as shown in figure 4.17), indicating a collisions status. In order to clear the collision the user has to back off, any other movement that would keep the collision is not permitted and would not move the simulated catheter at all. Once the tip of the simulated catheter touches the upper wall (as shown in figure 4.18) the simulation stops.

In order to exploit all the capabilities of each device, two different mazes (worlds) were used for this experiment. The first maze (world1) has longer stretches in the 1st DOF, occasionally stopped by a short rotation movement, in order to simulate the first fast strokes the surgeons have to perform in order to reach the aortic arch, if a blockage is found, a small rotation is made and then the surgeon keeps moving forward. The second one (world2) with higher alternation between left/right and up/down movements, simulating scenarios where surgeons have to maneuver both DOF at the same time, such as when they cross the aortic valve [9] [13].

For this experiment three main metrics were measured, Time to complete the experiment, Number of collisions and Dimensionless squared Jerk. These metrics were selected since they enclose relevant information on the capabilities of the devices and how suitable to perform the tasks they are. Time to complete the tasks is of high importance, since one of the objectives of a Teleoperated robots is to reduce the radiation exposure both, to patients and practitioners. Number of collisions is important to give an insight of the maneuverability of every device (acceleration rate, deceleration rate, precision, among others) which is highly important for patient safety. The dimensionless squared jerk has been previously used in similar studies to asses performance of catheter trajectories [16] [14], it is useful to describe the smoothness of the path as well as the accelerations and decelerations, while being independent of time and longitude of the path. The formula to calculate this is described in equation 4.3 where TP is the experiment time (from t_i to t_e) and PL the path length of the trajectory followed by the user.

$$Jd = \left(\int_{t_i}^{t_e} (\ddot{x}(t))^2 + (\ddot{y}(t))^2 dt \right) * \frac{Tp^5}{PL^2} \quad (4.3)$$

For each one of the mazes every participant had to complete 5 repetitions, which means 10 randomly ordered repetitions per device. The repetitions were executed, one device at a time; however, the devices were ordered randomly at the beginning of the experiment.

4.10.2 Results

The Time result of the experiment were not statistically conclusive, since interaction with P-Value < 0.05 was always present in the comparison between all the devices. As can be seen in figure 4.19, the Keyboard and the Joystick had a close average time in world1, however, in world2 (with the long stretches) Keyboard dominates. This can be related to the easy control of the keyboard without any training given the immediate stop advantage, as mentioned in figure 3.8, this allows the user to advance confidently with high velocity and without expecting a collision. Overall it can be observed in figure 4.20 that the lowest times were achieved with the Keyboard and Joystick; however,




Figure 4.16: Initial state of the experiment maze

the distribution in the upper quartile in the joystick device as with the remote is widely spread, which as in the experiment for the 1st DOF and 2nd DOF may be attributed to the learning curve.

The average number of collisions (per repetition) shown in figure 4.22 states that the Keyboard and Remote are significantly better than the Joystick and Catheter with a P-Value < 0.01. This result gives again intuition about the importance of the advantages and disadvantages shown in figure 3.8, if taking into account the information from figure 4.23, where analogical sensor devices have a significantly higher amount of collisions in world2 (world with longer stretches) than in world1, combined with the proneness of analogical sensor devices to overshoot shows how important participants training is, and coming back to the Time results, it confirms how these collisions are related. This can be directly corroborated with the Joystick when putting together figure 4.24 showing the collision training curve and figure 4.21 showing the time training curve, where it can be appreciated how time per attempt goes down as collisions per attempt go down.

The joystick got a lower average Dimensionless Squared Jerk number in both worlds as shown in figure 4.25, although not significant due to an interaction P-Value lower than 0.05, immediately followed by the keyboard device (figure 4.26). It is important to remark that part of the good result of the keyboard in the DSJ is due to the linear increment of the velocity over time of the PressedTime-Velocity mapping; thus, if the function is changed to something non linear, this will have a high impact in this metric.

Also, it is important to remark how even though the Joystick had a worse performance than the keyboard in the collision average, and this collisions caused higher DSJ values, the performance overall in the DSJ was better. This may be due to the fact that the Keyboard can stop instantly, giving really high jerking values, which in a real world is also translated to high stress in the actuators.




Figure 4.17: Collision during maze experiment




Figure 4.18: End of maze experiment when colliding with upper wall




Figure 4.19: Maze experiment average time per world




Figure 4.20: Maze experiment average time per device across all worlds




Figure 4.21: Maze experiment time training curve of all participants




Figure 4.22: Maze experiment average number of collisions per world




Figure 4.23: Maze experiment average number of collisions per device across all worlds




Figure 4.24: Maze experiment number of collisions training curve of all participants




Figure 4.25: Maze experiment DSJ per world




Figure 4.26: Maze experiment DSJ per device across all worlds

4.11 Poll Results

After completing the experiment, each one of the participants were requested to answer questions about the performance of the devices and give any additional comments about the experiment/devices. The exact questions as they were handed to the participants can be found in appendix B.

Since the poll requested the participants to enumerate the preference for the devices from 1-4, the figure 4.27, figure 4.28 and figure 4.29, were calculated as a percentage by adding all the points each device got and dividing by the total amount of points. It can be seen for the 1st DOF device preference, the Joystick has the higher punctuation, followed by the Keyboard and then the Remote, leaving the CatheterLike last. Moreover, the 2nd DOF has in first place the Joystick again, in this case followed by the remote, then the Keyboard and last the CatheterLike device. The overall poll had the same results as in the 1st DOF, but note that the difference between keyboard and remote was significantly reduced.

Another one of the questions asked to the participants was the experience with the devices, shown in figure 4.30, which shows how many participants out of the 15 declared experience in any of the devices, with the Keyboard being the most used, followed by the Joystick.

The most relevant comments made by the participants are as follow:

- Remote appears to be significantly slower than keyboard in the 1st DOF
- Takes some time to operate the joystick, but with time it feels the easiest to use
- Rotation in joystick has too much dead zone and no low velocity
- (From Surgeon) The final user interface could use the Joystick as main device and the Remote as detachable device for being able to operate near the patient if necessary.




Figure 4.27: Participants percentage of preference per device on the 1st DOF




Figure 4.28: Participants percentage of preference per device on the 2nd DOF




Figure 4.29: Participants percentage of preference per device on the overall




Figure 4.30: Number of participants with experience with each master device

Chapter 5

Conclusion

5.1 Conclusions and Discussions

The discussion raised along the experiments were mostly focused on if the quality of the in house-made devices (Remote and CatheterLike) were too low compared with the produced devices (Keyboard and Joystick). Figure 5.1 shows the qualitative measurements of basic parameters both reported by the participants of the experiments (reported through the survey in appendix B) and noted by myself during the sessions. The CatheterLike device had the most reported issues; however, as explained in 3.4, this device tries to replicate the usual handling of a catheter in TAVI and as can be seen in figure 5.2 (the figure shows the attempts in the 1st DOF experiment of the expert TAVI surgeon), the CatheterLike device had the best performance. This shows how with proper training the device has the mechanic capabilities to perform better than the rest of the devices, explaining the lower performance of the CatheterLike device mostly due to the lack of training of the remaining participants.

Considering how the experiments developed and the data gathered from:

- The 1st DOF Experiment results (figure 4.6 and 4.7).
- The maze experiment DSJ result (figure 4.25 and 4.26).
- The training curves of the 1st DOF errors, Maze-Times and Maze-collisions (figure 4.8, 4.21 and 4.24).
- The poll results and comments (figure 4.29).

I consider the Joystick to be the most suitable option as a master device for a TAVI teleoperated robot.

Device	Delay/Latency	Friction	Reliability	Ergonomics
Keyboard	LOW	LOW	HIGH	MEDIUM The UP and DOWN arrows in the keypad are not spaced correctly for an appropriate one hand operation.
Remote	MEDIUM The used buttons have a pre-travel of 1.27, only activating at the end of the buttons total-travel. Giving the user the impression of non immediate action or delay, (actually reported in the polls by users).	LOW	HIGH	HIGH
Joystick	LOW	LOW	HIGH	HIGH
CatheterLike	LOW	HIGH Both, rotation and translation movements present friction, making the movements unstable at occasions.	MEDIUM The spring system intended to return automatically the catheter inside the dead zone on the 1 st DOF is not working properly, making the catheter to keep moving even in resting position.	HIGH

Figure 5.1: Devices qualitative table from users experience




Figure 5.2: Attempts of the expert TAVI surgeon in the 1st DOF experiment

5.2 Future work

Explore different functions to increment the velocity over time in the PressedTime-Velocity mapping; for this thesis work a linear function was used.

Explore different mapping functions for the analogical sensor devices in the Position-velocity mapping; for this thesis work a linear function was used.

Explore feedback strategies for the possibility of making the devices haptic.

Explore the possibility to have two master devices, e.g. adding a detachable secondary master device.

Appendix A

Experiments Instructions

Instruction set given to the participants at the beginning of the experiment.

Welcome to the “**Choosing a master device for TAVI teleoperated surgery**” experiment. The experiment is divided in 4 sections, Training, Experiment 1, Experiment 2 and Experiment 3. Experiment 1 to 3 are not necessarily executed in that order; they will be randomized for means of the experiment. You will get informed which experiment is to be executed next.

Every device control 2 degrees of freedom (is able to make 2 kind of movements), the first one gets reflected up and down on the screen, and the second one gets reflected left and right on the screen.

All the experiments are a simulation of the control of a catheter (shown in white), displayed on the screen as a 2D projection (as it would be seen in fluoroscopy image).

The 4 sections are described as follows:

Training:


In this section, you will have the opportunity to try and get familiar with every device moving the catheter freely around the screen. (3 mins max. each).

Experiment 1:

In this experiment there is a floating red square (target) and a green square attached to the tip of the catheter (in white).

This experiment starts moving the target as soon as any movement in the device is performed. The task is to follow the target as close as possible with the green square commanded by the device.

In this experiment the target is only moving up and down.




Experiment 2:

In this experiment there is a floating red square (target) and a green square attached to the tip of the catheter (in white).

This experiment starts moving the target as soon as any movement in the device is performed. The task is to follow the target as close as possible with the green square commanded by the device.

In this experiment the target is only moving left and right.


**Experiment 3:**

In this experiment there is a maze with only one way possible and a green square attached to the tip of the catheter (in white).

The task is to navigate through the maze and arrive to the upper part of the maze, the experiment will finish once the upper wall is touched with the green square commanded by the device.

Objective:

- Avoid collisions of the green square with the walls of the maze. If a collision occurs, the square will turn red and it's necessary to move back in order to keep moving, once the square turns green again it is possible to move in any direction again.
- Keep the green square as far from the walls as possible, meaning, try to move as centered in the maze as possible
- Try to reach the upper wall as fast as possible.
- Try to keep the same strategy consistently through all the devices.



Appendix B

Participants Survey

Survey handed to the participants after completing the experiment.

The following information is only for statistical purposes and is not going to be linked to your name in any way. If you do not feel comfortable answering any of them, leave them blank.

Which device did you find simpler to use in the 1st DOF (up and down)? Enumerate from 1 to 4, being 1 the simplest one

Keyboard	<input type="checkbox"/>	Joystick	<input type="checkbox"/>	Remote	<input type="checkbox"/>	Catheter	<input type="checkbox"/>
----------	--------------------------	----------	--------------------------	--------	--------------------------	----------	--------------------------

Which device did you find simpler to use in the 2nd DOF (left and right)? Enumerate from 1 to 4, being 1 the simplest one

Keyboard	<input type="checkbox"/>	Joystick	<input type="checkbox"/>	Remote	<input type="checkbox"/>	Catheter	<input type="checkbox"/>
----------	--------------------------	----------	--------------------------	--------	--------------------------	----------	--------------------------

Which device did you find simpler to use overall? Enumerate from 1 to 4, being 1 the simplest one

Keyboard	<input type="checkbox"/>	Joystick	<input type="checkbox"/>	Remote	<input type="checkbox"/>	Catheter	<input type="checkbox"/>
----------	--------------------------	----------	--------------------------	--------	--------------------------	----------	--------------------------

Do you think you had any advantage using any of the devices because of some previous experience or work (gaming, operation of robots, office work)? Please specify.

Do you have any comment or suggestion about the experiment?

Age:

Gender:

Bibliography

- [1] Alejandro Aquino, Ali J. Khiabani, Matthew C. Henn, Alan Zajarias, Spencer J. Melby, Marc Sintek, John Lasala, Puja Kachroo, Eric Novak, and Hersh S. Maniar. Radiation exposure during transcatheter valve replacement: What cardiac surgeons need to know. *The Annals of Thoracic Surgery*, 2019.
- [2] Maimonides Medical Center. Transcatheter aortic valve replacement (tavr) program. video on a webpage. <https://www.maimonidesmed.org/heart-vascular-institute/centers/structural-heart-center/tavr>.
- [3] Gregory J. Dehmer and Members of the Joint Inter-Society Task Force on Occupational Hazards in the Catheterization Laboratory. Occupational hazards for interventional cardiologists. *Catheterization and Cardiovascular Interventions*, 68(6):974–976, 2006.
- [4] J.G. Diez. Transcatheter aortic valve implantation (tavi): the hype and the hope. *Texas Heart Institute journal*, 40(3):298–301, 2013.
- [5] Andras P Durko, Ruben L Osnabrugge, Nicolas M Van Mieghem, Milan Milojevic, Darren Mylotte, Vuyisile T Nkomo, and A Pieter Kappetein. Annual number of candidates for transcatheter aortic valve implantation per country: current estimates and future projections. *European Heart Journal*, 39(28):2635–2642, 2018.
- [6] EIZO. Revolution full overlap lumbar vest & skirt - 703. webpage. <https://www.infabcorp.com/product/revolution-full-overlap-lumbar-vest-skirt-703/>.
- [7] Tarantini G, Gasparetto V, Napodano M, Fraccaro C, Gerosa G, and Isabella G. Valvular leak after transcatheter aortic valve implantation: a clinician update on epidemiology, pathophysiology and clinical implications. *Am J Cardiovasc Dis*, 1(3):312–320, 2011.
- [8] James A. Goldstein, Stephen Balter, Michael Cowley, John Hodgson, and Lloyd W. Klein. Occupational hazards of interventional cardiologists: Prevalence of orthopedic health problems in contemporary practice. *Catheterization and Cardiovascular Interventions*, 63(4):407–411, 2004.
- [9] J. Kevin Harrison, Charles J. Davidson, Harry R. Phillips, Michael B. Harding, Katherine B. Kisslo, and Thomas M. Bashore. A rapid, effective technique for retrograde crossing of valvular aortic stenosis using standard coronary catheters. *Catheterization and Cardiovascular Interventions*, 21(1):51–54, 1990.
- [10] Auris Health. Transforming medical intervention. webpage. <https://www.aurishealth.com/monarch-platform>.
- [11] Infab. Shimane prefectoral central hospital. webpage. <https://www.eizoglobal.com/solutions/casestudies/spch/index.html>.
- [12] Wilczek K, Bujak K, Regula R, and Chodor Pand Osadnik T. Risk factors for paravalvular leak after transcatheter aortic valve implantation. *Kardiochir Torakochirurgia Pol*, 12(2):89–94, 2015.

- [13] Albert M. Kasel, Anupama Shivaraju, Wolfgang von Scheidt, Adnan Kastrati, and Christian Thilo. Anatomic guided crossing of a stenotic aortic valve under fluoroscopy: Right cusp rule, part iii. *JACC: Cardiovascular Interventions*, 8(1):119–120, 2015.
- [14] Jae-Man Kwak, Erica Kholinne, Maulik Gandhi, Arnold Adikrishna, Hanpyo Hong, Yucheng Sun, Kyoung-Hwan Koh, and In-Ho Jeon. Improvement of arthroscopic surgical performance using a new wide-angle arthroscope in the surgical training. *PLoS ONE*, 14(3), 2019.
- [15] Young M Mahmaljy H. Transcatheter aortic valve replacement (tavr/tavi, percutaneous replacement). StatPearls [Internet], Jan 2019. Available from: <https://www.ncbi.nlm.nih.gov/books/NBK431075/>.
- [16] E.B. Mazomenos, PL. Chang, and R.A. et al Rippel. Catheter manipulation analysis for objective performance and technical skills assessment in transcatheter aortic valve implantation. *International Journal of Computer Assisted Radiology and Surgery*, 11(6):1121–1131, 2016.
- [17] Corindus Vascular Robotics. Corindus vascular robotics. webpage. <https://www.corindus.com/>.
- [18] Ariel Roguin, Jacob Goldstein, and Olivier Bar. Brain tumours among interventional cardiologists: a cause for alarm? *EuroIntervention*, 7(9):1081–1086, 2012.
- [19] Zohaib A. Shaikh, Michael F. Eilenberg, and Todd J. Cohen. The amigo remote catheter system: From concept to bedside. *The Journal of Innovations in Cardiac Rhythm Management*, pages 2795–2802, 2017.
- [20] Nathaniel R. Smilowit, Jeffrey W. Moses, Fernando A. Sosa, Benjamin Lerman, Yasir Qureshi, Kate E. Dalton, Lauren T. Privitera, Diane Canone-Weber, Varinder Singh, Martin B. Leon, and Giora Weisz. Robotic-enhanced pci compared to the traditional manual approach. *J Invasive Cardiol.*, 6(7):318–321, 2014.
- [21] Stereotaxis. Stereotaxis products. webpage. <http://www.stereotaxis.com/products/#!/niobe>.
- [22] Eliseo Vano, Norman J. Kleiman, Ariel Duran, Mariana Romano-Miller, and Madan M. Rehani. Radiation-associated lens opacities in catheterization personnel: Results of a survey and direct assessments. *Journal of Vascular and Interventional Radiology*, 24(2):197–204, 2013.
- [23] Giora Weisz, D. Christopher Metzger, Ronald P. Caputo, Juan A. Delgado, J. Jeffrey Marshall, George W. Vetrovec, Mark Reisman, Ron Waksman, Juan F. Granada, Victor Novack, Jeffrey W. Moses, and Joseph P. Carrozza. Safety and feasibility of robotic percutaneous coronary intervention: Precise (percutaneous robotically-enhanced coronary intervention) study. *Journal of the American College of Cardiology*, 61(15):1596–1600, 2013.