

Multi-isotope generator calculation

Artem Lebedev^{1*}

^{1*}Department of Chemistry and Chemical Biology, McMaster
University, Hamilton, ON, Canada.

Corresponding author(s). E-mail(s): artem@indiechemistry.com;

Abstract

A brief explanation of the math behind solving arbitrarily complex isotope generator.

1 Introduction

Recent clinical developments [1] with the therapeutic agents based on Ac-225 prompted a vigorous search for practical methods to produce this isotope at scale. The clinical value of this isotope is estimated so high, that current market price of 1 patient dose of the isotope alone currently goes as high as 10,000 USD, higher than any other isotope ever tried in humans. This high value started a frenzy of development and over a dozen of approaches were suggested employing many types of accelerators, nuclear reactors and isotopes generators. However, practical considerations of production at scale, with the right purity, in a distributed manner and robust enough to satisfy pharmaceutical manufacturing constraints eliminate most of the suggested methodologies. 3 methods currently supply most of the market: Th-229/Ac-225 generator, Ra-226(p,2n)Ac-225 reaction enabled by low energy cyclotron and Ra-226(γ ,2n)Ac-225 process enabled by high energy eLINACs.

The last methodology is rapidly gaining popularity. The reason for this popularity is not the great advantages of the process, but rather lack of significant problems typically associated with the competitors. A common (p,2n) approach produces product contaminated with Ac-227; additionally potential failure of the radium target in this process can irreversibly contaminate the entire accelerator. Th-229/Ac-225 generator potentially is a very attractive technology, but the starting material is available commercially to only one company in the world, making this approach of no interest for any other developer. Photonuclear route also comes with significant problems. First,

radium starting material is scantily available and hard to deal with in routine production. Second, eLINACs of sufficient power are unique instruments with less than a dozen of suitable units built in the entire world ever. Third, low conversion efficiency requires chemical processing of curies of Ra-226 to obtain meaningful quantities of Ac-225, necessitating large and complex GMP hot cell facilities. None of these problems, however are of fundamental nature, they all can be solved with the right investment and a capable team. As the old Jewish saying goes, "if a problem can be solved with money, it isn't a problem, it's an expense".

This review aims to provide a chemists perspective to the photonuclear production route.

This review focuses on the accelerator-driven experimental setups. While using photonuclear reaction to produce Ac-225 using a nuclear reactor [2] is an intriguing idea, the practical considerations of that approach are so vastly different from the accelerator-driven methods, that it is hard to draw meaningful comparison.

2 Two solutions to the same equations

Let's consider the most important isotope generator in nuclear medicine: Tc generator, that is still a staple of the clinical practice, used for millions upon millions of SPECT scans every year, all around the world. Mo-99 decays two ways: ~87.6% to Tc-99m (branching ratio b) and the remaining ~12.4% directly to (virtually) stable Tc-99, which does not factor into the generator considerations. From here, a conventional chemist's thinking can easily lead us to the following system of equations:

$$\begin{cases} \frac{d[{}^{99}\text{Mo}]}{dt} = -\lambda_{\text{Mo}}[{}^{99}\text{Mo}] \\ \frac{d[{}^{99\text{m}}\text{Tc}]}{dt} = b\lambda_{\text{Mo}}[{}^{99}\text{Mo}] - \lambda_{\text{Tc}}[{}^{99\text{m}}\text{Tc}] \end{cases} \quad (1)$$

where $\lambda_{\text{Mo}} = \ln 2/65.94h$, $\lambda_{\text{Tc}} = \ln 2/6.01h$ and branching ratio $b = 0.876$

What this is saying is that the rate of decay of ${}^{99}\text{Mo}$ is proportional to its concentration, a definition of the radioactive decay. For the ${}^{99\text{m}}\text{Tc}$ the picture is more complex, but is still intuitive: its rate of decay is *negatively* proportional to its own concentration $[{}^{99\text{m}}\text{Tc}]$, but also *positively* proportional to the parent's concentration $[{}^{99}\text{Mo}]$. Also, let's not forget that the $[{}^{99}\text{Mo}]$ is scaled to the branching ratio here.

****Matrix form****

$$\frac{d}{dt} \begin{pmatrix} [{}^{99}\text{Mo}] \\ [{}^{99\text{m}}\text{Tc}] \end{pmatrix} = \begin{pmatrix} -\lambda_{\text{Mo}} & 0 \\ b\lambda_{\text{Mo}} & -\lambda_{\text{Tc}} \end{pmatrix} \begin{pmatrix} [{}^{99}\text{Mo}] \\ [{}^{99\text{m}}\text{Tc}] \end{pmatrix}$$

Only the subdiagonal entry picks up the factor b . Diagonalizing exactly as before (eigenvalues still $-\lambda_{\text{Mo}}$, $-\lambda_{\text{Tc}}$ since those live on the diagonal regardless of b):

$$\begin{pmatrix} [{}^{99}\text{Mo}] \\ [{}^{99\text{m}}\text{Tc}] \end{pmatrix} = \begin{pmatrix} e^{-\lambda_{\text{Mo}} t} & 0 \\ \frac{b\lambda_{\text{Mo}}}{\lambda_{\text{Tc}} - \lambda_{\text{Mo}}} (e^{-\lambda_{\text{Mo}} t} - e^{-\lambda_{\text{Tc}} t}) & e^{-\lambda_{\text{Tc}} t} \end{pmatrix} \begin{pmatrix} [{}^{99}\text{Mo}]_0 \\ [{}^{99\text{m}}\text{Tc}]_0 \end{pmatrix}$$

So Bateman becomes:

$$N_2(t) = N_1^0 \frac{\lambda_1}{\lambda_2 - \lambda_1} \left(e^{-(\lambda_1 - \lambda_2)t} - e^{-\lambda_2 t} \right)$$

— literally just the un-branched formula times b . This makes sense: branching ratio only ever multiplies the *production* term, never the decay-loss term, so it factors out linearly and leaves the transient/secular equilibrium math (the classic 6-hour Mo-99/Tc-99m generator “milking” curve) otherwise unchanged.

****Where it stops being this simple****

If you extend the chain further (e.g. tracking Tc-99 ground state buildup too, or handling isotopes with 3+ branches like your Ac-225 chain has at Bi-213 $\rightarrow$ Tl-209/Po-213), you get multiple daughters fed from one parent. That’s not a single chain anymore — A becomes a genuine tree structure, and you’d want one column feeding multiple rows (each weighted by its own b_i , with $\sum b_i = 1$) rather than a simple bidiagonal matrix. The matrix exponential machinery still works fine, it’s just no longer tridiagonal/bidiagonal, so ‘expm’ is basically mandatory rather than a convenience at that point.

Acknowledgements. Author thanks Ian Horwath for a stimulating discussion and Andrew Knight for the feedback on the web app version of these computations.

Declarations

- Funding: The work was entirely self-funded
- Conflict of interest: The author is an employee of BWXT Medical, Ltd.
- Data availability: The article and the app code are available on GitHub

References

- [1] Jang, A., Kendi, A.T., Johnson, G.B., Halfdanarson, T.R., Sartor, O.: Targeted alpha-particle therapy: A review of current trials **24**(14) <https://doi.org/10.3390/ijms241411626>
- [2] Heibel, M.D., Lopresti, B.J., Congedo, T.V.: Producing Ac-225 Using Gamma Radiation.
WO2023086762A2. <https://patents.google.com/patent/WO2023086762A2/en?q=WO2023086762> Accessed 2026-06-19