



BHARATBENZ

## Daimler India Commercial Vehicles Pvt. Ltd -Central Lab

SIPCOT Industrial Estate, Oragadam Village, Singaperumal Koil Road,  
Mathur(PO), Sriperumbadur Taluk, Kanchipuram - 602105, Tamil Nadu, India.

Phone: +91 44 4969 4456,

Website: www.daimler-indiacv.com / www.bharatbenz.com



CR20220001 A29

# CERTIFICATE OF CALIBRATION

## I. Descriptions of DUC

**Certificate No** : DICV/CL/22/00000007 **Report Issues Date** : 30-08-2022  
**Date Of Calibration** : 27-07-2022 **Next Calibration Due** : Jan-2023  
**Name Of The Custom** : Gopi Kannayiram **Date of Receipt** : 27-07-2022  
**Department** : TRANS-PRD **Description of Equipment** : Depth Measurement gauge

Range	Least Count	Serial No	Make
5.50/4.50mm	NA	PAG305032907	PARI

**Identification Number** : TtPDMG-3853

**Condition of DUC** : Good

**Calibration Procedure /  
Reference Standard :** : IS3455

## II. Master & Equipment Used For Calibration & Traceability

Sl.	Master Equipment	Id No	Make	Traceability	Certificate No	Validity
-----	------------------	-------	------	--------------	----------------	----------

The master Equipment used for calibration are calibrated and traceable to National / International Standard

**Environmental Condition** :  $20 \pm 1.0^{\circ}\text{C}$  /  $50 \pm 10\%$  RH

**Uncertainty Of Measurement** : NA

The stated Expanded uncertainty is estimated at a confidence level approximately 95% with a coverage factor  $k = 2$

### Calibrated By

Karthikeyan V  
Engineer

### Authorized Signatory

Gurushev P  
Quality Manager

- The Calibration results are valid at the time of & under the stated conditions of measurement.
- The report should not be reproduced either in full or in part without out permission in writing.
- Next calibration due date given in this cetificate as per customer request irrespective of the status or condition of the IMTE.

DAIMLER TRUCK

## Daimler India Commercial Vehicles Pvt. Ltd -Central Lab

SIPCOT Industrial Estate, Oragadam Village, Singaperumal Koil Road,  
Mathur(PO), Sriperumbadur Taluk, Kanchipuram - 602105, Tamil Nadu, India.



BHARATBENZ

Phone: +91 44 4969 4456,  
Website: www.daimler-indiacv.com / www.bharatbenz.com



### III. Mechanical Calibration Results

Certificate No DICV/CL/22/00000007

Report Issue Date : 30-08-2022

All Values are in mm

Nominal	Trial-1	Trial-2	Trial-3	Average
5.50000	5.49000	5.49000	5.49000	5.49000
4.50000	4.52000	4.52000	4.52000	4.52000

### IV. Result Of Calibration

: Calibrated parameters found OK

### V. Remarks

: OK

Statement of confirmity is given based on decision rule agreed by requestor.

### Calibrated By

Karthikeyan V  
Engineer

### Authorized Signatory

Gurushev P  
Quality Manager

\*\*\*End of Certificate\*\*\*

1. The Calibration results are valid at the time of & under the stated conditions of measurement.
2. The report should not be reproduced either in full or in part without out permission in writing.
3. Next calibration due date given in this cetificate as per customer request irrespective of the status or condition of the IMTE.

DAIMLER TRUCK