

**Daimler India Commercial Vehicles Pvt. Ltd -Central Lab**SIPCOT Industrial Estate, Oragadam Village, Singaperumal Koil Road,  
Mathur(PO), Sriperumbadur Taluk, Kanchipuram - 602105, Tamil Nadu, India.

Phone: +91 44 4969 4456,

Website: www.daimler-indiacv.com / www.bharatbenz.com



CR20220001 A16

**CERTIFICATE OF CALIBRATION****I. Descriptions of DUC**

**Certificate No** : DICV/CL/22/00000001 **Report Issues Date** : 25-07-2022

**ULR No** : CC202122000000001F

**Date Of Calibration** : 21-07-2022 **Next Calibration Due** : Jul-2023

**Name Of The Custom** : Gopi Kannayiram **Date of Receipt** : 16-07-2021

**Department** : LAB-QA **Description of Equipment** : Master Ring Gauge

Range	Least Count	Serial No	Make
37.4974mm	NA	GF0950	BAKER

**Identification Number** : LQMRG-004**Condition of DUC** : Good**Calibration Procedure / Reference Standard** : IS 3485:1983**II. Master & Equipment Used For Calibration & Traceability**

Sl.	Master Equipment	Id No	Make	Traceability	Certificate No	Validity
1	ULM	CM-01	Trimos	NABL Accredited Lab No:CC-2656	CGIL/2021/0309/H-001	08-03-2023
2	Master Setting Ring 40.00	CM-22	Trimos	NABL Accredited Lab No:CC-2428	JRPM-CCTR-DIM- 2020-0254	07-02-2022

The master Equipment used for calibration are calibrated and traceable to National / International Standard

**Environmental Condition** :  $20 \pm 1.0^{\circ}\text{C}$  /  $50 \pm 10\%$  RH**Uncertainty Of Measurement** :The stated Expanded uncertainty is estimated at a confidence level approximately 95% with a coverage factor  $k = 2$ **Calibrated By**

Karthikeyan V

Engineer

**Authorized Signatory**

Gurushev P

Quality Manager

- The Calibration results are valid at the time of & under the stated conditions of measurement.
- The report should not be reproduced either in full or in part without out permission in writing.
- Next calibration due date given in this cetificate as per customer request irrespective of the status or condition of the IMTE.

DAIMLER TRUCK

**Daimler India Commercial Vehicles Pvt. Ltd -Central Lab**

SIPCOT Industrial Estate, Oragadam Village, Singaperumal Koil Road,  
Mathur(PO), Sriperumbadur Taluk, Kanchipuram - 602105, Tamil Nadu, India.

Phone: +91 44 4969 4456,

Website: www.daimler-indiacv.com / www.bharatbenz.com

**III. Mechanical Calibration Results**

Certificate No DICV/CL/22/00000001

Report Issue Date : 25-07-2022

All Values are in mm

Nominal	Trial-1	Trial-2	Trial-3	Average
37.50000	37.50000	37.50000	37.50000	37.50000

**IV. Result Of Calibration**

: Testing

**V. Remarks**

: Testing

Statement of conformity is given based on decision rule Exclude MU agreed by requestor.

**Calibrated By**

Karthikeyan V

Engineer

**Authorized Signatory**

Gurushev P

Quality Manager

\*\*\*End of Certificate\*\*\*

1. The Calibration results are valid at the time of & under the stated conditions of measurement.
2. The report should not be reproduced either in full or in part without out permission in writing.
3. Next calibration due date given in this cetificate as per customer request irrespective of the status or condition of the IMTE.

DAIMLER TRUCK