



### **CATALOG**

#### **PORTABLE TEST TUBE WARMER**



# **Specification:**

- Stable heating environment.
- Panel easily removable for decontamination
- 14 chambers for falcon test tubes 12ml / 5ml.
- Temp accuracy ± 0.2 degree Celsius.
- LED display.
- Rectangular warming blocks to hold tubes.





### **VERTICAL LAMINAR AIR FLOW FOR ANDROLOGY LAB.**



# **Specification:**

IUI workstation Made of Heavy thick Galvanized Steel sheet duly powder coated paint and work table made of 304 Grade stainless steel 2 mm thick,

# Other specification:-

- a) Adjustable Luminosity.
- b) Full operated with Microprocessor control Panel
- c) Have the following functions:-
- d) Control Panel LCD Touch panel full operated microprocessor base Independent On/Off for UV , FL, Sockets, Air flow speed, with usb data output.
- e) Electrical Power Sockets ( Outlets )
- f) Fixed Support Stand (Height 80-85 Cm) Aprox
- g) 1 Nos HEPA Filter.
- h) 1 Nos Pre Filter 10 Micron
- i) Fluorescent light to provide 500 -600 Lux & UV Light
- j) Fan Blower "EBM NADI" 1 Nos.
- k) Overall Dimensions: 1000x800x2000mm (WxDxH)



## **ANTI-VIBRATION TABLE**



#### **Specification:**

- Suitable for placing the inverted microscope with micromanipulator
- Should be designed for assurance of stable operation, 3 vibration isolators with rolling diaphragm to give excellent isolation from surface borne vibrations in all directions,
- Thread for adjusting height. 3 vibration isolators with rolling diaphragm to give excellent isolation from surface borne vibrations in all directions.
- The inner table shall be heavily weighted pre-fabricated.
- Should be mounted on special rubber mounts to dampen the minutest vibration.
- Rust proof construction.
- The top SS 304 grade and Pipe with powder coated paint



#### VERTICAL LAMINAR AIR FLOW FOR IVF LAB



## **Specification:**

IVF workstation Made of Heavy thick Galvanized Steel sheet duly powder coated paint and work table made of food grade stainless steel 2-3 mm thick,

#### Other specification:-

- a. 6x2 feet IVF workstation with 2 stereo Microscopes, one Left and one right side .
- b. 80-100 mm heating glass stage to minimize temp. fluctuation, With independent Temp. controllers. Noise Level between 48-52 dBA.
- c. 6" LCD touch Screen have the following functions.

Auto Start and shut down function. Independent on/off for Lights , blower speed , total time, & inbuilt Temp. Controllers for table.

- d. 2 Nos HEPA Filters Class (H14).
- e. 2 Nos Pre Filters 10 Micron.
- f. 2 Nos Dry Bath with warming blocks.
- g. 21 "LCD display LG



# JD Medical Systems

**Quality Products For Quality Hospitals** 

- h. Unit have ant vibration adjustable legs.
- i. Fluorescent light & UV Ligh
- j. Fan Blower "EBM NADI" 2 Nos.
- k. Light source Stand.
- l. Extra Power sockets inside the workstation. ( 9 Nos. )
- m. No Side glass window and No motorized glass front Door Provide Cable Manager provide both side

