EUDAMED Enumerations

UDI/Devices

Production v 2.18.1
2025

UDI/Device - Enumerations - 2.18.1 (PROD), 3.18 (PG)

1 - Introduction

This "Enumerations" document contains the value lists for drop down elements and lists where a limited set of values can be selected.

2 - Purpose

This purpose of this document is to provide an overview of the possible values fields can contain to be valid information for EUDAMED.

3 - Scope

We opted to provide enumerations and their detailed descriptions by module. This document refers to UDI/Device module Enumerations only.

4 - Changelog

No changes

4 - Enumerations

<table>
<tr><th>Identifier</th><th>Status</th><th>Description</th><th>Previously used</th></tr>
<tr><td>BR-CRF-004 : Certificate type - ENUM_CRF_CertificateType</td><td>RESOLVED</td><td>For MDR certificates (see) the certificate type can be:

<table>
<tr><th>Label</th><th>Value</th><th>Sort order</th></tr>
<tr><td>(MDR) EU Quality Management System certificate (Annex IX Chapter I)</td><td>MDR_QUALITY_MANAGEMENT_SYSTEM</td><td>1</td></tr>
<tr><td>(MDR) EU Technical Documentation certificate (Annex IX Chapter II)</td><td>MDR_TECHNICAL_DOCUMENTATION</td><td>2</td></tr>
<tr><td>(MDR) EU Type Examination certificate (Annex X)</td><td>MDR_TYPE_EXAMINATION</td><td>3</td></tr>
<tr><td>(MDR) EU Quality Assurance certificate (Annex XI Part A)</td><td>MDR_QUALITY_ASSURANCE</td><td>4</td></tr>
<tr><td>(MDR) EU Product verification certificate (Annex XI Part B)</td><td>MDR_PRODUCT_VERIFICATION</td><td>5</td></tr>
</table>

For IVDR certificates (see) the certificate type can be:

<table>
<tr><th>Label</th><th>Value</th><th>Sort order</th></tr>
<tr><td>(IVDR) EU Quality Management System certificate (Annex IX Chapter I)</td><td>IVDR_QUALITY_MANAGEMENT_SYSTEM</td><td>1</td></tr>
<tr><td>(IVDR) EU Technical Documentation certificate (Annex IX Chapter II)</td><td>IVDR_TECHNICAL_DOCUMENTATION</td><td>2</td></tr>
<tr><td>(IVDR) EU Type Examination certificate (Annex X)</td><td>IVDR_TYPE_EXAMINATION</td><td>3</td></tr>
<tr><td>(IVDR) EU Production Quality Assurance certificate (Annex XI)</td><td>IVDR_PRODUCTION_QUALITY_ASSURANCE</td><td>4</td></tr>
</table>
</td><td></td></tr>
</table>

BR-UDID-646 : List of Applicable Legislations for Devices and System or Procedure Pack

RESOLVED

Applicable Legislation valid for a Regulation Device are :

Label	Value	Notes
REGULATION (EU) 2017/745 on medical devices	MDR	
REGULATION (EU) 2017/746 on in vitro diagnostic medical devices	IVDR	

Applicable Legislation valid for a Legacy Device are:

Label	Value	Notes
Council Directive 93/42/EEC on Medical devices	MDD	
Council Directive 90/385/EEC - Approximation of the laws of the Member States relating to active implantable medical devices	AIMDD	
Directive 98/79/EC on in vitro diagnostic medical devices	IVDD	

Applicable Legislation valid for a System or Procedure Pack :

Label	Value	Notes
REGULATION (EU) 2017/745 on medical devices	MDR	

BR-UDID-800 : Clinical Size Measure Unit - ENUM_UDID_ClinicalSizeMeasureUnit

RESOLVED

Name	Code
%	MU01
/litre (/L)	MU02
/millilitre (/mL)	MU03
/millimole (/mmol)	MU04
Active	MU138
Adult	MU168
ampere (A)	MU107
Angled	MU140
atomic mass unit: u	MU87
atomic mass unit: u per 12h	MU89
atomic mass unit: u per 1h	MU88
atomic mass unit: u per 2h	MU90
atomic mass unit: u per d	MU91
atomic mass unit: u per g	MU92
atomic mass unit: u per kg	MU93
atomic mass unit: u per mL	MU94
bar (bar)	MU126
becquerel (Bq)	MU122
bel (B)	MU133
Belt	MU154
candela (cd)	MU109
centilitre (cL)	MU09
centimetre (cm)	MU10
centimetre per second (cm/s)	MU11
Child	MU167
Concave	MU158
Contour	MU151
Convex	MU156
Convex light	MU157
coulomb (C)	MU113
cubic centimetre (cu cm or cm^3)	MU12
cubic millimetre (cu mm or mm^3)	MU13
cup	MU14
Curved	MU175
dalton (Da)	MU131
day (d)	MU15
decibel (dB)	MU134
decilitre (dL)	MU16
decimetre (dm)	MU17
degree (°)	MU105
degree Celsius (°C)	MU18
dioptre (D)	MU104
electronvolt (eV)	MU127
Extra large (XL)	MU164

Extra small (XS)	MU160
farad (F)	MU117
femtolitre (fL)	MU21
femtometre (fm)	MU20
femtomole/litre (fmol/L)	MU22
Flat	MU159
foot (ft)	MU23
French (F)	MU170
gram (g)	MU24
gram per litre (g/L)	MU25
gray (Gy)	MU123
hectare (ha)	MU129
Heel	MU150
henry (H)	MU120
hertz (Hz)	MU27
hour (h)	MU26
inch (in)	MU28
Infant	MU166
international unit [IU]	MU05
international unit per day [IU]/day ([IU]/d)	MU06
international unit per litre [IU]/litre [IU]/L	MU07
international unit per millilitre [IU]/millilitre [IU]/mL	MU08
joule (J)	MU112
J-tip	MU141
katal (kat)	MU125
kelvin (K)	MU108
kilogram (kg)	MU29
kilolitre (KL)	MU30
kilometre (km)	MU31
kilopascal (kPa)	MU32
kiloU per litre (kU/L)	MU33
Large (L)	MU163
left	MU172
litre (L)	MU34
lumen (lm)	MU106
lux (lx)	MU121
Medium (M)	MU162
metre (m)	MU36
micro[iU] per millilitre (µ[iU]/mL or µ[iU]/ml)	MU95
microgram (µg)	MU40
microgram per day (µg/d)	MU96
microgram per decilitre µg/dL	MU135
microgram per litre (µg/L)	MU97
microgram per millilitre (µg/mL)	MU98
microgram per minute (µg/min)	MU41
microlitre (µL or µl)	MU42
micrometre (µm)	MU43
micrometre per second (µm/s)	MU99
micromole (µmol)	MU44
micromole per gram (µmol/g)	MU100
micromole per litre (µmol/L)	MU45
micromoles per minute (=1 mol/min)	MU19
milli international unit per litre (m[iU]/L)	MU35
millibar (mbar)	MU46
milliequivalent (mEq)	MU47
milligram (mg)	MU37
milligram per decilitre (mg/dL)	MU136
milligram per litre (mg/L)	MU38
milligram per millilitre (mg/mL)	MU39
millilitre (mL)	MU48
millilitre per day (mL/d)	MU56
millilitre per minute (mL or ml/min)	MU57
millilitre per second (mL/s)	MU49
millimetre (mm)	MU50
millimetre of mercury (mm Hg)	MU51
millimetre per hour (mm/h)	MU58

millimole (mmol)	MU52
millimole per gram (mmol/g)	MU59
millimole per kilogram (mmol/kg)	MU60
millimole per kilogram[H2O] (mmol/kg[H2O])	MU61
millimoles per litre (mmol/L)	MU53
millisecond (ms)	MU54
Mini	MU147
minute (min)	MU55
mole (mol)	MU62
month	MU63
Multishape	MU149
nanogram (ng)	MU64
nanogram per litre (ng/L)	MU67
nanogram per millilitre (ng/mL)	MU68
nanolitre (nL)	MU65
nanometre (nm)	MU66
nanomole per day (nmol/d)	MU69
nanomole per gram (nmol/g)	MU70
nanomole per hour per millilitre (nmol/h/mL or nmol/h/ml)	MU71
nanomole per litre (nmol/L)	MU72
Neonate	MU165
neper (Np)	MU132
newton (N)	MU110
ohm ()	MU115
pascal (Pa)	MU111
Passive	MU137
pH (pH)	MU169
picogram (pg)	MU73
picogram per millilitre (pg/mL)	MU74
picolitre (pL)	MU75
picometre (pm)	MU76
picomole per gram (pmol/g)	MU77
picomole per hour per milligram (pmol/h/mg)	MU78
picomole per hour per millilitre (pmol/h/mL or pmol/h/ml)	MU79
picomole per litre (pmol/L)	MU80
Rectangular	MU153
right	MU173
Sacral	MU148
second (s)	MU81
siemens (S)	MU116
sievert (Sv)	MU124
Small (S)	MU161
Soft angled	MU144
Soft straight	MU142
Square	MU152
square centimetre (sq cm or cm^2)	MU82
square foot (sq ft)	MU83
square inch (sq in)	MU84
square metre (sq m or m^2)	MU85
square millimetre (sq mm or mm^2)	MU86
Standard	MU176
Stiff angled	MU145
Stiff J-tip	MU146
Stiff straight	MU143
Straight	MU139
tesla (T)	MU119
tonne (metric ton)	MU130
unified atomic mass unit (u)	MU128
volt (V)	MU114
watt (W)	MU103
weber (Wb)	MU118
week	MU101
year	MU102
OTHER	MU999

BR-UDID-803 : CMR Type - ENUM_UDID_CMRType

RESOLVED

Label	Value	Notes	Sort Order
1A	CMR_1A		1
1B	CMR_1B		2

BR-UDID-804 : Contained Item Status - ENUM_UDID_ContainedItemStatus

RESOLVED

Label	Value	Notes	Sort Order
On the EU market	ON_THE_MARKET		1
No longer placed on the EU market	NO_LONGER_PLACED_ON_THE_MARKET		2
Not intended for the EU market	NOT_INTENDED_FOR_EU_MARKET		3

BR-UDID-805 : Critical Warnings - ENUM_UDID_CriticalWarnings (Part II)

RESOLVED

Label	Value	Notes
P307 – IF exposed: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 301	
P308 – IF exposed or concerned: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 302	
P309 – IF exposed or if you feel unwell: P312 – Call a POISON CENTER or doctor /physician if you feel unwell.	CW 303	
P337 – If eye irritation persists: P313 – Get medical advice/attention.	CW 304	
P332 – If skin irritation occurs: P313 – Get medical advice/attention.	CW 305	
P333 – If skin irritation or rash occurs: P313 – Get medical advice/attention.	CW 306	
P301 – IF SWALLOWED: P313 – Get medical advice/attention.	CW 307	
P302 – IF ON SKIN: P313 – Get medical advice/attention.	CW 308	
P303 – IF ON SKIN (or hair): P313 – Get medical advice/attention.	CW 309	
P304 – IF INHALED: P313 – Get medical advice/attention.	CW 310	
P305 – IF IN EYES: P313 – Get medical advice/attention.	CW 311	
P307 – IF exposed: P313 – Get medical advice/attention.	CW 312	
P308 – IF exposed or concerned: P313 – Get medical advice/attention.	CW 313	
P309 – IF exposed or if you feel unwell: P313 – Get medical advice/attention.	CW 314	
P337 – If eye irritation persists: P314 – Get medical advice/attention if you feel unwell.	CW 315	
P332 – If skin irritation occurs: P314 – Get medical advice/attention if you feel unwell.	CW 316	
P333 – If skin irritation or rash occurs: P314 – Get medical advice/attention if you feel unwell.	CW 317	
P301 – IF SWALLOWED: P314 – Get medical advice/attention if you feel unwell.	CW 318	
P302 – IF ON SKIN: P314 – Get medical advice/attention if you feel unwell.	CW 319	
P303 – IF ON SKIN (or hair): P314 – Get medical advice/attention if you feel unwell.	CW 320	
P304 – IF INHALED: P314 – Get medical advice/attention if you feel unwell.	CW 321	
P305 – IF IN EYES: P314 – Get medical advice/attention if you feel unwell.	CW 322	
P306 – IF ON CLOTHING: P314 – Get medical advice/attention if you feel unwell.	CW 323	
P307 – IF exposed: P314 – Get medical advice/attention if you feel unwell.	CW 324	
P308 – IF exposed or concerned: P314 – Get medical advice/attention if you feel unwell.	CW 325	
P309 – IF exposed or if you feel unwell: P314 – Get medical advice/attention if you feel unwell.	CW 326	
P337 – If eye irritation persists: P315 – Get immediate medical advice/attention.	CW 327	
P332 – If skin irritation occurs: P315 – Get immediate medical advice/attention.	CW 328	
P333 – If skin irritation or rash occurs: P315 – Get immediate medical advice/attention.	CW 329	
P301 – IF SWALLOWED: P315 – Get immediate medical advice/attention.	CW 330	
P302 – IF ON SKIN: P315 – Get immediate medical advice/attention.	CW 331	
P303 – IF ON SKIN (or hair): P315 – Get immediate medical advice/attention.	CW 332	

P304 – IF INHALED: P315 – Get immediate medical advice/attention.	CW 333	
P305 – IF IN EYES: P315 – Get immediate medical advice/attention.	CW 334	
P307 – IF exposed: P315 – Get immediate medical advice/attention.	CW 335	
P308 – IF exposed or concerned: P315 – Get immediate medical advice/attention.	CW 336	
P309 – IF exposed or if you feel unwell: P315 – Get immediate medical advice/attention.	CW 337	
P337 – If eye irritation persists: P320 – Specific treatment is urgent (see on this label).	CW 338	
P332 – If skin irritation occurs: P320 – Specific treatment is urgent (see on this label).	CW 339	
P333 – If skin irritation or rash occurs: P320 – Specific treatment is urgent (see on this label).	CW 340	
P301 – IF SWALLOWED: P320 – Specific treatment is urgent (see on this label).	CW 341	
P302 – IF ON SKIN: P320 – Specific treatment is urgent (see on this label).	CW 342	
P303 – IF ON SKIN (or hair): P320 – Specific treatment is urgent (see on this label).	CW 343	
P304 – IF INHALED: P320 – Specific treatment is urgent (see on this label).	CW 344	
P305 – IF IN EYES: P320 – Specific treatment is urgent (see on this label).	CW 345	
P306 – IF ON CLOTHING: P320 – Specific treatment is urgent (see on this label).	CW 346	
P307 – IF exposed: P320 – Specific treatment is urgent (see on this label).	CW 347	
P308 – IF exposed or concerned: P320 – Specific treatment is urgent (see on this label).	CW 348	
P309 – IF exposed or if you feel unwell: P320 – Specific treatment is urgent (see on this label).	CW 349	
P337 – If eye irritation persists: P321 – Specific treatment (see on this label).	CW 350	
P332 – If skin irritation occurs: P321 – Specific treatment (see on this label).	CW 351	
P333 – If skin irritation or rash occurs: P321 – Specific treatment (see on this label).	CW 352	
P301 – IF SWALLOWED: P321 – Specific treatment (see on this label).	CW 353	
P302 – IF ON SKIN: P321 – Specific treatment (see on this label).	CW 354	
P303 – IF ON SKIN (or hair): P321 – Specific treatment (see on this label).	CW 355	
P304 – IF INHALED: P321 – Specific treatment (see on this label).	CW 356	
P305 – IF IN EYES: P321 – Specific treatment (see on this label).	CW 357	
P306 – IF ON CLOTHING: P321 – Specific treatment (see on this label).	CW 358	
P307 – IF exposed: P321 – Specific treatment (see on this label).	CW 359	
P308 – IF exposed or concerned: P321 – Specific treatment (see on this label).	CW 360	
P309 – IF exposed or if you feel unwell: P321 – Specific treatment (see on this label).	CW 361	
P337 – If eye irritation persists: P322 – Specific measures (see on this label).	CW 362	
P301 – IF SWALLOWED: P322 – Specific measures (see on this label).	CW 363	
P302 – IF ON SKIN: P322 – Specific measures (see on this label).	CW 364	
P303 – IF ON SKIN (or hair): P322 – Specific measures (see on this label).	CW 365	
P304 – IF INHALED: P322 – Specific measures (see on this label).	CW 366	
P305 – IF IN EYES: P322 – Specific measures (see on this label).	CW 367	
P306 – IF ON CLOTHING: P322 – Specific measures (see on this label).	CW 368	
P307 – IF exposed: P322 – Specific measures (see on this label).	CW 369	
P308 – IF exposed or concerned: P322 – Specific measures (see on this label).	CW 370	
P309 – IF exposed or if you feel unwell: P322 – Specific measures (see on this label).	CW 371	
P307 – IF exposed: P330 – Rinse mouth.	CW 372	
P308 – IF exposed or concerned: P330 – Rinse mouth.	CW 373	
P309 – IF exposed or if you feel unwell: P330 – Rinse mouth.	CW 374	
P307 – IF exposed: P331 – Do NOT induce vomiting.	CW 375	

P308 – IF exposed or concerned: P331 – Do NOT induce vomiting.	CW 376	
P309 – IF exposed or if you feel unwell: P331 – Do NOT induce vomiting.	CW 377	
P333 – If skin irritation occurs or rash occurs: P334 – Immerse in cool water/wrap in wet bandages.	CW 378	
P302 – IF ON SKIN: P334 – Immerse in cool water/wrap in wet bandages.	CW 379	
P303 – IF ON SKIN (or hair): P334 – Immerse in cool water/wrap in wet bandages.	CW 380	
P337 – If eye irritation persists: P334 – Immerse in cool water/wrap in wet bandages.	CW 381	
P305 – IF IN EYES: P334 – Immerse in cool water/wrap in wet bandages.	CW 382	
P333 – If skin irritation occurs or rash occurs: P335 – Brush off loose particles from skin.	CW 383	
P302 – IF ON SKIN: P335 – Brush off loose particles from skin.	CW 384	
P303 – IF ON SKIN (or hair): P335 – Brush off loose particles from skin.	CW 385	
P306 – IF ON CLOTHING: P335 – Brush off loose particles from skin.	CW 386	
P307 – IF exposed: P335 – Brush off loose particles from skin.	CW 387	
P308 – IF exposed or concerned: P335 – Brush off loose particles from skin.	CW 388	
P309 – IF exposed or if you feel unwell: P335 – Brush off loose particles from skin.	CW 389	
P333 – If skin irritation occurs or rash occurs: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 390	
P302 – IF ON SKIN: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 391	
P303 – IF ON SKIN (or hair): P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 392	
P305 – IF IN EYES: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 393	
P307 – IF exposed: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 394	
P308 – IF exposed or concerned: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 395	
P309 – IF exposed or if you feel unwell: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW 396	
P305 – IF IN EYES: P338 – Remove contact lenses, if present and easy to do. Continue rinsing.	CW 397	
P307 – IF exposed: P338 – Remove contact lenses, if present and easy to do. Continue rinsing.	CW 398	
P308 – IF exposed or concerned: P338 – Remove contact lenses, if present and easy to do. Continue rinsing.	CW 399	
P309 – IF exposed or if you feel unwell: P338 – Remove contact lenses, if present and easy to do. Continue rinsing.	CW 400	
P304 – IF INHALED: P340 – Remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW 401	
P307 – IF exposed: P340 – Remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW 402	
P308 – IF exposed or concerned: P340 – Remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW 403	
P309 – IF exposed or if you feel unwell: P340 – Remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW 404	
P332 – If skin irritation occurs: P350 – Gently wash with plenty of soap and water.	CW 405	
P302 – IF ON SKIN: P350 – Gently wash with plenty of soap and water.	CW 406	
P303 – IF ON SKIN (or hair): P350 – Gently wash with plenty of soap and water.	CW 407	
P306 – IF ON CLOTHING: P350 – Gently wash with plenty of soap and water.	CW 408	
P307 – IF exposed: P350 – Gently wash with plenty of soap and water.	CW 409	
P308 – IF exposed or concerned: P350 – Gently wash with plenty of soap and water.	CW 410	
P309 – IF exposed or if you feel unwell: P350 – Gently wash with plenty of soap and water.	CW 411	
P333 – If skin irritation or rash occurs: P351 – Rinse cautiously with water for several minutes.	CW 412	
P332 – If skin irritation occurs: P351 – Rinse cautiously with water for several minutes.	CW 413	
P302 – IF ON SKIN: P351 – Rinse cautiously with water for several minutes.	CW 414	
P303 – IF ON SKIN (or hair): P351 – Rinse cautiously with water for several minutes.	CW 415	
P305 – IF IN EYES: P351 – Rinse cautiously with water for several minutes.	CW 416	
P306 – IF ON CLOTHING: P351 – Rinse cautiously with water for several minutes.	CW 417	
P307 – IF exposed: P351 – Rinse cautiously with water for several minutes.	CW 418	

P308 – IF exposed or concerned: P351 – Rinse cautiously with water for several minutes.	CW 419	
P309 – IF exposed or if you feel unwell: P351 – Rinse cautiously with water for several minutes.	CW 420	
P332 – If skin irritation occurs: P352 – Wash with plenty of soap and water.	CW 421	
P302 – IF ON SKIN: P352 – Wash with plenty of soap and water.	CW 422	
P303 – IF ON SKIN (or hair): P352 – Wash with plenty of soap and water.	CW 423	
P306 – IF ON CLOTHING: P352 – Wash with plenty of soap and water.	CW 424	
P307 – IF exposed: P352 – Wash with plenty of soap and water.	CW 425	
P308 – IF exposed or concerned: P352 – Wash with plenty of soap and water.	CW 426	
P309 – IF exposed or if you feel unwell: P352 – Wash with plenty of soap and water.	CW 427	
P332 – If skin irritation occurs: P353 – Rinse skin with water/shower.	CW 428	
P302 – IF ON SKIN: P353 – Rinse skin with water/shower.	CW 429	
P303 – IF ON SKIN (or hair): P353 – Rinse skin with water/shower.	CW 430	
P306 – IF ON CLOTHING: P353 – Rinse skin with water/shower.	CW 431	
P307 – IF exposed: P353 – Rinse skin with water/shower.	CW 432	
P308 – IF exposed or concerned: P353 – Rinse skin with water/shower.	CW 433	
P309 – IF exposed or if you feel unwell: P353 – Rinse skin with water/shower.	CW 434	
P332 – If skin irritation occurs: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 435	
P302 – IF ON SKIN: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 436	
P303 – IF ON SKIN (or hair): P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 437	
P306 – IF ON CLOTHING: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 438	
P307 – IF exposed: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 439	
P308 – IF exposed or concerned: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 440	
P309 – IF exposed or if you feel unwell: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW 441	
P332 – If skin irritation occurs: P361 – Remove/Take off immediately all contaminated clothing.	CW 442	
P302 – IF ON SKIN: P361 – Remove/Take off immediately all contaminated clothing.	CW 443	
P303 – IF ON SKIN (or hair): P361 – Remove/Take off immediately all contaminated clothing.	CW 444	
P306 – IF ON CLOTHING: P361 – Remove/Take off immediately all contaminated clothing.	CW 445	
P307 – IF exposed: P361 – Remove/Take off immediately all contaminated clothing.	CW 446	
P308 – IF exposed or concerned: P361 – Remove/Take off immediately all contaminated clothing.	CW 447	
P309 – IF exposed or if you feel unwell: P361 – Remove/Take off immediately all contaminated clothing.	CW 448	
P332 – If skin irritation occurs: P362 – Take off contaminated clothing and wash before reuse.	CW 449	
P302 – IF ON SKIN: P362 – Take off contaminated clothing and wash before reuse.	CW 450	
P303 – IF ON SKIN (or hair): P362 – Take off contaminated clothing and wash before reuse.	CW 451	
P306 – IF ON CLOTHING: P362 – Take off contaminated clothing and wash before reuse.	CW 452	
P307 – IF exposed: P362 – Take off contaminated clothing and wash before reuse.	CW 453	
P308 – IF exposed or concerned: P362 – Take off contaminated clothing and wash before reuse.	CW 454	
P309 – IF exposed or if you feel unwell: P362 – Take off contaminated clothing and wash before reuse.	CW 455	
P332 – If skin irritation occurs: P363 – Wash contaminated clothing before reuse.	CW 456	
P302 – IF ON SKIN: P363 – Wash contaminated clothing before reuse.	CW 457	
P303 – IF ON SKIN (or hair): P363 – Wash contaminated clothing before reuse.	CW 458	
P306 – IF ON CLOTHING: P363 – Wash contaminated clothing before reuse.	CW 459	
P307 – IF exposed: P363 – Wash contaminated clothing before reuse.	CW 460	
P308 – IF exposed or concerned: P363 – Wash contaminated clothing before reuse.	CW 461	

		P309 – IF exposed or if you feel unwell: P363 – Wash contaminated clothing before reuse.	CW 462	
		P371 – In case of major fire and large quantities: P374 – Fight fire with normal precautions from a reasonable distance.	CW 463	
		P371 – In case of major fire and large quantities: P375 – Fight fire remotely due to the risk of explosion.	CW 464	
		P371 – In case of major fire and large quantities: P380 – Evacuate area.	CW 465	
		P371 – In case of major fire and large quantities: P381 – Eliminate all ignition sources if safe to do so.	CW 466	
		OTHER **	CW 999	
		**: requires to provide additional free text with translations		
BR-UDID-805 : Critical Warnings - ENUM_UDID_CriticalWarnings (Part I)	RESOLVED			

Label	Value	Notes
Do not resterilize	CW 001	
Defibrillation-proof type B applied part	CW 006	
Do not use if package is damaged	CW 007	
Biological risks	CW 008	
Do not re-use	CW 009	
Consult instructions for use	CW 010	
Caution	CW 011	
Caution: Electrostatic sensitive, do not touch directly.	CW 012	
Radiation of laser apparatus	CW 013	
Type BF applied part	CW 014	
Defibrillation-proof type BF applied part	CW 015	
Type B applied part	CW 016	
Defibrillation-proof type CF applied part	CW 017	
Refer to instruction manual/booklet	CW 018	
Do not sterilize	CW 020	
Do not cut	CW 022	
Consider hazards from magnetic fields	CW 023	
Do not store near magnets or magnetic devices	CW 024	
MR SAFE	CW 025	
MR CONDITIONAL	CW 026	
MR UNSAFE	CW 027	
Exposure to radiation when using the device	CW 028	
Caution, sharp edges	CW 029	

Consult healthcare professional before use	CW030	
Implant timeframe limitations	CW031	
Do not use past indicated date	CW032	
Restriction in use of des infectors,	CW035	
Restrictions in use for combination of other devices	CW036	
GHS01 – Danger - Unstable, Explosive	CW037	
GHS02 - Danger or Warning - Flammable	CW038	
GHS03- Danger or Warning -Oxidising	CW039	
GHS04- Warning -Compressed gas	CW040	
GHS05 -Danger or Warning -Corrosive cat. 1	CW041	
GHS06 – Danger -Toxic cat. 1 - 3	CW042	
GHS07 - Warning -Toxic cat. 4, Irritant cat. 2 or 3, Lower systemic health hazards	CW043	
GHS08 - Danger or Warning - Systemic health hazards	CW044	
GHS09 - Warning (for cat. 1)- (for cat. 2 no signal word) Environment	CW045	
H200 – Unstable explosives.	CW046	
H201 – Explosive; mass explosion hazard.	CW047	
H202 – Explosive, severe projection hazard.	CW048	
H203 – Explosive; fire, blast or projection hazard.	CW049	
H204 – Fire or projection hazard.	CW050	
H205 – May mass explode in fire.	CW051	
H220 – Extremely flammable gas.	CW052	
H221 – Flammable gas.	CW053	
H222 – Extremely flammable aerosol.	CW054	
H223 – Flammable aerosol.	CW055	
H224 – Extremely flammable liquid and vapour.	CW056	
H225 – Highly flammable liquid and vapour.	CW057	
H226 – Flammable liquid and vapour.	CW058	
H228 – Flammable solid.	CW059	
H240 – Heating may cause an explosion.	CW060	
H241 – Heating may cause a fire or explosion.	CW061	
H242 – Heating may cause a fire.	CW062	
H250 – Catches fire spontaneously if exposed to air.	CW063	

H251 – Self-heating: may catch fire.	CW064	
H252 – Self-heating in large quantities; may catch fire.	CW065	
H260 – In contact with water releases flammable gases which may ignite spontaneously.	CW066	
H261 – In contact with water releases flammable gases.	CW067	
H270 – May cause or intensify fire; oxidiser.	CW068	
H271 – May cause fire or explosion; strong oxidiser.	CW069	
H272 – May intensify fire; oxidiser.	CW070	
H280 – Contains gas under pressure; may explode if heated.	CW071	
H281 – Contains refrigerated gas; may cause cryogenic burns or injury.	CW072	
H290 – May be corrosive to metals.	CW073	
H300 – Fatal if swallowed.	CW074	
H301 – Toxic if swallowed.	CW075	
H302 – Harmful if swallowed.	CW076	
H304 – May be fatal if swallowed and enters airways.	CW077	
H310 – Fatal in contact with skin.	CW078	
H311 – Toxic in contact with skin.	CW079	
H312 – Harmful in contact with skin.	CW080	
H314 – Causes severe skin burns and eye damage.	CW081	
H315 – Causes skin irritation.	CW082	
H317 – May cause an allergic skin reaction.	CW083	
H318 – Causes serious eye damage.	CW084	
H319 – Causes serious eye irritation.	CW085	
H330 – Fatal if inhaled.	CW086	
H331 – Toxic if inhaled.	CW087	
H332 – Harmful if inhaled.	CW088	
H334 – May cause allergy or asthma symptoms or breathing difficulties if inhaled.	CW089	
H335 – May cause respiratory irritation.	CW090	
H336 – May cause drowsiness or dizziness.	CW091	
H340 – May cause genetic defects exposure cause the hazard>.	CW092	
H341 – Suspected of causing genetic defects.	CW093	
H350 – May cause cancer.	CW094	
H351 – Suspected of causing cancer.	CW095	

H360 – May damage fertility or the unborn child.	CW096	
H361 – Suspected of damaging fertility or the unborn child.	CW097	
H362 – May cause harm to breast-fed children.	CW098	
H370 – Causes damage to organs.	CW099	
H371 – May cause damage to organs.	CW100	
H372 – Causes damage to organs through prolonged or repeated exposure exposure cause the hazard>.	CW101	
H373 – May cause damage to organs through prolonged or repeated exposure.	CW102	
H400 – Very toxic to aquatic life.	CW103	
H410 – Very toxic to aquatic life with long lasting effects.	CW104	
H411 – Toxic to aquatic life with long lasting effects.	CW105	
H412 – Harmful to aquatic life with long lasting effects.	CW106	
H413 – May cause long lasting harmful effects to aquatic life.	CW107	
EUH 001 – Explosive when dry.	CW108	
EUH 006 – Explosive with or without contact with air.	CW109	
EUH 014 – Reacts violently with water.	CW110	
EUH 018 – In use, may form flammable/explosive vapour-air mixture.	CW111	
EUH 019 – May form explosive peroxides.	CW112	
EUH 044 – Risk of explosion if heated under confinement.	CW113	
EUH 029 – Contact with water liberates toxic gas.	CW114	
EUH 031 – Contact with acids liberates toxic gas.	CW115	
EUH 032 – Contact with acids liberates very toxic gas.	CW116	
EUH 066 – Repeated exposure may cause skin dryness or cracking.	CW117	
EUH 070 – Toxic by eye contact.	CW118	
EUH 071 – Corrosive to the respiratory tract.	CW119	
EUH 059 – Hazardous to the ozone layer.	CW120	
EUH 201 – Contains lead. Should not be used on surfaces liable to be chewed or sucked by children.	CW121	
EUH 201A – Warning! Contains lead.	CW122	
EUH 202 – Cyanoacrylate. Danger. Bonds skin and eyes in seconds. Keep out of the reach of children.	CW123	
EUH 203 – Contains chromium (VI). May produce an allergic reaction.	CW124	
EUH 204 – Contains isocyanates. May produce an allergic reaction.	CW125	
EUH 205 – Contains epoxy constituents. May produce an allergic reaction.	CW126	
EUH 206 – Warning! Do not use together with other products. May release dangerous gases (chlorine)	CW127	

EUH 207 – Warning! Contains cadmium. Dangerous fumes are formed during use. See information supplied by the manufacturer. Comply with the safety instructions.	CW128	
EUH 208 – Contains (name of sensitising substance). May produce an allergic reaction.	CW129	
EUH 209 – Can become highly flammable in use.	CW130	
EUH 209A – Can become flammable in use.	CW131	
EUH 210 – Safety data sheet available on request.	CW132	
EUH 401 – To avoid risks to human health and the environment, comply with the instructions for use.	CW133	
P101 – If medical advice is needed, have product container or label at hand.	CW134	
P102 – Keep out of reach of children.	CW135	
P103 – Read label before use.	CW136	
P201 – Obtain special instructions before use.	CW137	
P202 – Do not handle until all safety precautions have been read and understood.	CW138	
P210 – Keep away from heat/sparks/open flames/hot surfaces. — No smoking.	CW139	
P211 – Do not spray on an open flame or other ignition source.	CW140	
P220 – Keep/Store away from clothing/…/combustible materials.	CW141	
P221 – Take any precaution to avoid mixing with combustibles…	CW142	
P222 – Do not allow contact with air.	CW143	
P223 – Keep away from any possible contact with water, because of violent reaction and possible flash fire.	CW144	
P231 – Handle under inert gas.	CW146	
P232 – Protect from moisture.	CW147	
P233 – Keep container tightly closed.	CW148	
P234 – Keep only in original container.	CW149	
P235 – Keep cool.	CW150	
P240 – Ground/bond container and receiving equipment.	CW151	
P241 – Use explosion-proof electrical/ventilating/lighting/…/equipment.	CW152	
P242 – Use only non-sparking tools.	CW153	
P243 – Take precautionary measures against static discharge.	CW154	
P244 – Keep reduction valves free from grease and oil.	CW155	
P250 – Do not subject to grinding/shock/…/friction.	CW156	
P251 – Pressurized container: Do not pierce or burn, even after use.	CW157	
P260 – Do not breathe dust/fume/gas/mist/vapours/spray.	CW158	
P261 – Avoid breathing dust/fume/gas/mist/vapours/spray.	CW159	
P262 – Do not get in eyes, on skin, or on clothing.	CW160	

P263 – Avoid contact during pregnancy/while nursing.	CW161	
P264 – Wash thoroughly after handling.	CW162	
P270 – Do not eat, drink or smoke when using this product.	CW163	
P271 – Use only outdoors or in a well-ventilated area.	CW164	
P272 – Contaminated work clothing should not be allowed out of the workplace.	CW165	
P273 – Avoid release to the environment.	CW166	
P280 – Wear protective gloves/protective clothing/eye protection/face protection.	CW167	
P281 – Use personal protective equipment as required.	CW168	
P282 – Wear cold insulating gloves/face shield/eye protection.	CW169	
P283 – Wear fire/flame resistant/retardant clothing.	CW170	
P284 – Wear respiratory protection.	CW171	
P285 – In case of inadequate ventilation wear respiratory protection.	CW172	
P231 + P232 – Handle under inert gas. Protect from moisture.	CW173	
P301 – IF SWALLOWED: P310 – Immediately call a POISON CENTER or doctor /physician.	CW175	
P302 – IF ON SKIN: P310 – Immediately call a POISON CENTER or doctor/physician.	CW176	
P303 – IF ON SKIN (or hair): P310 – Immediately call a POISON CENTER or doctor /physician.	CW177	
P304 – IF INHALED: P310 – Immediately call a POISON CENTER or doctor/physician.	CW178	
P305 – IF IN EYES: P310 – Immediately call a POISON CENTER or doctor/physician.	CW179	
P306 – IF ON CLOTHING: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW180	
P307 – IF exposed: P310 – Immediately call a POISON CENTER or doctor/physician.	CW181	
P308 – IF exposed or concerned: P310 – Immediately call a POISON CENTER or doctor /physician.	CW182	
P309 – IF exposed or if you feel unwell: P310 – Immediately call a POISON CENTER or doctor/physician.	CW183	
P342 – If experiencing respiratory symptoms: P310 – Immediately call a POISON CENTER or doctor/physician.	CW184	
P342 – If experiencing respiratory symptoms: P311 – Call a POISON CENTER or doctor /physician.	CW185	
P342 – If experiencing respiratory symptoms: P312 – Call a POISON CENTER or doctor /physician if you feel unwell.	CW186	
P342 – If experiencing respiratory symptoms: P313 – Get medical advice/attention.	CW187	
P342 – If experiencing respiratory symptoms: P314 – Get medical advice/attention if you feel unwell.	CW188	
P342 – If experiencing respiratory symptoms: P315 – Get immediate medical advice /attention.	CW189	
P342 – If experiencing respiratory symptoms: P320 – Specific treatment is urgent (see on this label).	CW190	
P342 – If experiencing respiratory symptoms: P321 – Specific treatment (see on this label).	CW191	
P342 – If experiencing respiratory symptoms: P322 – Specific measures (see on this label).	CW192	
P301 – IF SWALLOWED: P330 – Rinse mouth.	CW193	

P301 – IF SWALLOWED: P331 – Do NOT induce vomiting.	CW194	
P332 – If skin irritation occurs: P322 – Specific measures (see on this label).	CW195	
P333 – If skin irritation or rash occurs: P322 – Specific measures (see on this label).	CW196	
P332 – If skin irritation occurs: P334 – Immerse in cool water/wrap in wet bandages.	CW197	
P332 – If skin irritation occurs: P335 – Brush off loose particles from skin.	CW198	
P332 – If skin irritation occurs: P336 – Thaw frosted parts with lukewarm water. Do no rub affected area.	CW199	
P337 – If eye irritation persists: P338 – Remove contact lenses, if present and easy to do. Continue rinsing.	CW201	
P342 – If experiencing respiratory symptoms: P340 – Remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW202	
P341 – If breathing is difficult, remove victim to fresh air and keep at rest in a position comfortable for breathing.	CW203	
P333 – If skin irritation or rash occurs: P350 – Gently wash with plenty of soap and water.	CW206	
P337 – If eye irritation persists: P351 – Rinse cautiously with water for several minutes.	CW207	
P333 – If skin irritation or rash occurs: P352 – Wash with plenty of soap and water.	CW208	
P333 – If skin irritation or rash occurs: P353 – Rinse skin with water/shower.	CW209	
P333 – If skin irritation or rash occurs: P360 – Rinse immediately contaminated clothing and skin with plenty of water before removing clothes.	CW210	
P333 – If skin irritation or rash occurs: P361 – Remove/Take off immediately all contaminated clothing.	CW211	
P333 – If skin irritation or rash occurs: P362 – Take off contaminated clothing and wash before reuse.	CW212	
P333 – If skin irritation or rash occurs: P363 – Wash contaminated clothing before reuse.	CW213	
P372 – Explosion risk in case of fire.	CW216	
P373 – DO NOT fight fire when fire reaches explosives.	CW217	
P370 – In case of fire: P374 – Fight fire with normal precautions from a reasonable distance.	CW218	
P370 – In case of fire: P375 – Fight fire remotely due to the risk of explosion.	CW219	
P376 – Stop leak if safe to do so.	CW220	
P377 – Leaking gas fire: Do not extinguish, unless leak can be stopped safely.	CW221	
P370 – In case of fire: P380 – Evacuate area.	CW223	
P370 – In case of fire: P381 – Eliminate all ignition sources if safe to do so.	CW224	
P390 – Absorb spillage to prevent material damage.	CW225	
P391 – Collect spillage.	CW226	
Do not Use Tap Water	CW254	
Do Not use saliva	CW255	
Warning; Crushing of hands	CW257	
Warning; Hot surface	CW258	
Warning; Electricity	CW259	

Warning; Counterrotating rollers	CW260	
Warning; Optical radiation	CW261	
Single patient multiple use	CW262	
Contains human blood or plasma derivatives	CW263	
Contains a medicinal substance	CW264	
Contains hazardous substances	CW265	
Contains biological material of human origin	CW266	
Contains biological material of animal origin	CW267	
H229 Pressurized container: may burst if heated	CW268	
H230 May react explosively even in the absence of air	CW269	
H231 May react explosively even in the absence of air at elevated pressure and/or temperature	CW270	
H420 Harms public health and the environment by destroying ozone in the upper atmosphere	CW271	
P364 And wash it before reuse.	CW272	
Caution, risk of electric shock	CW273	
Non-ionizing electromagnetic radiation	CW274	
Ask your pharmacist or doctor for advice before using or taking the product (e.g. if pregnant or breastfeeding)	CW275	
Potential Interactions with other products (e.g. medicines, food)	CW276	
If any symptoms are troublesome or last more than a few days or if you notice any side effects not mentioned, please inform your doctor or pharmacist	CW277	
Avoid contact with non-sterile water	CW278	
P337 – If eye irritation persists: P310 – Immediately call a POISON CENTER or doctor /physician.	CW279	
P332 – If skin irritation occurs: P310 – Immediately call a POISON CENTER or doctor /physician.	CW280	
P333 – If skin irritation or rash occurs: P310 – Immediately call a POISON CENTER or doctor/physician.	CW281	
P337 – If eye irritation persists: P311 – Call a POISON CENTER or doctor/physician.	CW282	
P332 – If skin irritation occurs: P311 – Call a POISON CENTER or doctor/physician.	CW283	
P333 – If skin irritation or rash occurs: P311 – Call a POISON CENTER or doctor /physician.	CW284	
P301 – IF SWALLOWED: P311 – Call a POISON CENTER or doctor/physician.	CW285	
P302 – IF ON SKIN: P311 – Call a POISON CENTER or doctor/physician.	CW286	
P303 – IF ON SKIN (or hair): P311 – Call a POISON CENTER or doctor/physician.	CW287	
P304 – IF INHALED: P311 – Call a POISON CENTER or doctor/physician.	CW288	
P305 – IF IN EYES: P311 – Call a POISON CENTER or doctor/physician.	CW289	
P307 – IF exposed: P311 – Call a POISON CENTER or doctor/physician.	CW290	
P308 – IF exposed or concerned: P311 – Call a POISON CENTER or doctor/physician.	CW291	

P309 – IF exposed or if you feel unwell: P311 – Call a POISON CENTER or doctor /physician.	CW 292			
P337 – If eye irritation persists: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 293			
P332 – If skin irritation occurs: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 294			
P333 – If skin irritation or rash occurs: P312 – Call a POISON CENTER or doctor /physician if you feel unwell.	CW 295			
P301 – IF SWALLOWED: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 296			
P302 – IF ON SKIN: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 297			
P303 – IF ON SKIN (or hair): P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 298			
P304 – IF INHALED: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 299			
P305 – IF IN EYES: P312 – Call a POISON CENTER or doctor/physician if you feel unwell.	CW 300			

BR-UDID-806 : Device Recall Scope - ENUM_UDID_DeviceRecallScope

RESOLVED

Label	Value	Notes	Sort Order
Lot Number	LOT_NUMBER		1
Serial Number	SERIAL_NUMBER		2
Software Version	SOFTWARE_VERSION		3
UDI-DI	UDI-DI		4

BR-UDID-807 : Device Status - ENUM_UDID_DeviceStatus

RESOLVED

Label	Value	Applicable for the OCM*	Sort Order
On the EU market	ON_THE_MARKET	Yes	1
No longer placed on the EU market	NO_LONGER_PLACED_ON_THE_MARKET	Yes	2
Not intended for the EU market	NOT_INTENDED_FOR_EU_MARKET	No	3

- OCM – old/custom-made devices aka NRD

BR-UDID-809 : Intruction for Use Type - ENUM_UDID_IFUType

RESOLVED

Label	Value	Notes	Sort Order
Critical Warnings or Contra Indications	CW		1
Storage and Handling Conditions	SHC		2

BR-UDID-812 : Non Medical Device Types - ENUM_UDID_NonMedicalDeviceType

RESOLVED

Label	Value	Notes	Sort Order
Brain Electrostimulation	BRAIN_ELECTROSTIMULATION		1
Contact Lenses	CONTACT_LENSES		2
EMR	EMR		3
EQUIPMENT FOR ADIPOSE TISSUE	EQUIPMENT_FOR_ADIPOSE_TISSUE		4
FILLING BY INJECTION	FILLING_BY_INJECTION		5
PRODUCT IN BODY	PRODUCT_IN_BODY		6

BR-UDID-813 : Production Identifier - ENUM_UDID_ProductionIdentifier

RESOLVED

Label	Value	Notes	Sort Order
BATCH NUMBER	BATCH_NUMBER		1
SERIAL NUMBER	SERIALISATION_NUMBER		2
MANUFACTURING DATE	MANUFACTURING_DATE		3
EXPIRATION DATE	EXPIRATION_DATE		4
SOFTWARE IDENTIFICATION	SOFTWARE_IDENTIFICATION		5

BR-UDID-814 : Public Filter Device Type - ENUM_UDID_PublicFilter_DeviceType

RESOLVED

Label	Sort order
Active device	1
CMR substances	2
Cells or substances of microbial origin	3
Companion diagnostics	4
Containing latex	5
Contains substances considered medicinal product derived from human blood or plasma	6
Contains substances considered to be a medicinal product	7
Endocrine disrupting substances	8
IIb Implantable well established technology	9
Implantable	10
Instrument	11
Intended for near-patient-testing	12
Intended for self-testing	13
Intended to administer/remove a medicinal substance	14
KIT	15
Made to order soft contact lenses	16
Measuring function	17
Needs sterilisation	18
New Device	19
Orthopedic	20
Procedure Pack which is a device in itself	21
Professional testing	22
Reagent	23
Reprocessed	24
Reusable surgical element	25
Rigid gas permeable (RGP) Contact Lenses	26
Software	27
Spectacle frames	28
Spectacle lenses	29
Sterile	30
System which is a device in itself	31
Tissue or cells from animal origin	32
Tissue or cells from human origin	33

BR-UDID-815 : Public Filter Legislation - ENUM_UDID_PublicFilter_Leg

RESOLVED

Label	Value	Notes
MDR	MDR	
IVDR	IVDR	

BR-UDID-816 : Public Filter System or Procedure Pack - ENUM_UDID_PublicFilter_SystemProcedurePack

RESOLVED

Label	Value	Sort order	Notes
Device	Device	1	
System	System	2	
Procedure Pack	Procedure_Pack	3	

BR-UDID-818 : Special Device Type - ENUM_UDID_SpecialDevice

RESOLVED

Label	Value	Order Number	Applicable legislation
Software	IVDR_SOFTWARE	1	Applicable for the following Applicable Legislations : IVDR
	IVDD_SOFTWARE	1	Applicable for the following Applicable Legislations : IVDD
	MDR_SOFTWARE	1	Applicable for the following Applicable Legislations : MDR
	MDD_SOFTWARE	1	Applicable for the following Applicable Legislations : MDD
	AIMDD_SOFTWARE	1	Applicable for the following Applicable Legislations : AIMDD
Standard soft contact lenses	MDR_STANDARD_SOFT_CONTACT_LENSES	2	Applicable for the following Applicable Legislations : MDR
	MDD_STANDARD_SOFT_CONTACT_LENSE	2	Applicable for the following Applicable Legislations : MDD
Rigid Gas Permeable (RGP) Contact Lenses	MDR_RIGID_GAS_PERMEABLE	3	Applicable for the following Applicable Legislations : MDR
	MDD_RIGID_GAS_PERMEABLE	3	Applicable for the following Applicable Legislations : MDD
Made to order soft contact lenses	MDR_MADE_TO_ORDER	4	Applicable for the following Applicable Legislations : MDR
	MDD_MADE_TO_ORDER	4	Applicable for the following Applicable Legislations : MDD
Made to order Rigid Gas Permeable (RGP) contact lenses	MDR_MADE_TO_ORDER_RIGID_GAS_PERMEABLE	5	Applicable for the following Applicable Legislations : MDR
Spectacle frames	MDR_SPECTACLES_FRAMES	6	Applicable for the following Applicable Legislations : MDR
	MDD_SPECTACLES_FRAMES	6	Applicable for the following Applicable Legislations : MDD
Spectacle lenses	MDR_SPECTACLES_LENSES	7	Applicable for the following Applicable Legislations : MDR
	MDD_SPECTACLES_LENSES	7	Applicable for the following Applicable Legislations : MDD
Ready-made reading spectacles	MDR_READY_MADE_SPECTACLES	8	Applicable for the following Applicable Legislations : MDR
	MDD_READY_MADE_SPECTACLES	8	Applicable for the following Applicable Legislations : MDD
Orthopaedic	MDR_ORTHOPEDIC	9	Applicable for the following Applicable Legislations : MDR CHANGE: ~~EUDAMEDMDR-39564~~
	MDD_ORTHOPEDIC	9	Applicable for the following Applicable Legislations : MDD,
	AIMDD_ORTHOPEDIC	9	Applicable for the following Applicable Legislations : AIMDD

BR-UDID-821 : Substance types - ENUM_UDID_SubstType

RESOLVED

Label	Value	Notes	Sort Order
MEDICINAL PRODUCT	MEDICINAL_PRODUCT_SUBSTANCE		1
HUMAN PRODUCT	HUMAN_PRODUCT_SUBSTANCE		2

BR-UDID-823 : Workflow State - ENUM_UDID_WorkflowState

RESOLVED

Label	Value	Notes	Sort Order
Draft	DRAFT		1
Submitted	SUBMITTED		2
Registered	REGISTERED		3
Discarded	DISCARDED		4

BR-UDID-823b : Multi-component Device Type - ENUM_UDID_MultiComponentType

RESOLVED

Sort Order	Label	Value	Applicable for Vigilance OCM
1	System which is a device in itself	SYSTEM	Yes
2	Procedure Pack which is a device in itself	PROCEDURE_PACK	Yes
3	SPP System	SPP_SYSTEM	No
4	SPP Procedure Pack	SPP_PROCEDURE_PACK	No

BR-UDID-824 : Certificate Types used for Device Certificate Information - Enum_UDI_Device_Certificate_Info_CertTypes

RESOLVED

The Certificate types that can be returned by EUDAMED in the Device Certificate information section (actual Certificates linked to Devices in EUDAMED) ca be the following - based on the Applicable legislation:

For MDR Applicable Legislation, the types of the Certificate can be:

Label	Value	Notes
EU Quality Management System certificate (Annex IX Chapter I)	MDR_QUALITY_MANAGEMENT_SYSTEM	
EU Technical Documentation certificate (Annex IX Chapter II)	MDR_TECHNICAL_DOCUMENTATION	
EU Type Examination certificate (Annex X)	MDR_TYPE_EXAMINATION	
EU Quality Assurance certificate (Annex XI Part A)	MDR_QUALITY_ASSURANCE	
EU Product verification certificate (Annex XI Part B)	MDR_PRODUCT_VERIFICATION	

For IVDR Applicable Legislation, the types of the Certificate can be:

Label	Value	Notes
EU Quality Management System certificate (Annex IX Chapter I)	IVDR_QUALITY_MANAGEMENT_SYSTEM	
EU Technical Documentation certificate (Annex IX Chapter II)	IVDR_TECHNICAL_DOCUMENTATION	
EU Type Examination certificate (Annex X)	IVDR_TYPE_EXAMINATION	
EU Production Quality Assurance certificate (Annex XI)	IVDR_PRODUCTION_QUALITY_ASSURANCE	

For MDD Applicable Legislation, the types of the Certificate can be:

Label	Value	Notes
Directive 93/42/EEC Annex II excluding section 4	MDD_II_EX_4	
Directive 93/42/EEC Annex II section 4	MDD_II_4	
Directive 93/42/EEC Annex III	MDD_III	
Directive 93/42/EEC Annex IV	MDD_IV	
Directive 93/42/EEC Annex V	MDD_V	
Directive 93/42/EEC Annex VI	MDD_VI	

For AIMDD Applicable Legislation, the types of the Certificate can be:

Label	Value	Notes
Directive 90/385/EEC Annex 2 excluding section 4	AIMDD_II_EX_4	
Directive 90/385/EEC Annex 2 section 4	AIMDD_II_4	
Directive 90/385/EEC Annex 3	AIMDD_III	
Directive 90/385/EEC Annex 4	AIMDD_IV	
Directive 90/385/EEC Annex 5	AIMDD_V	

For IVDD Applicable Legislation, the types of the Certificate can be:

Label	Value	Notes
Directive 98/79/EC Annex III section 6	IVDD_III_6	
Directive 98/79/EC Annex IV excl. section 4 and 6	IVDD_IV_EX_4_6	
Directive 98/79/EC Annex IV section 4	IVDD_IV_4	
Directive 98/79/EC Annex IV section 6	IVDD_IV_6	
Directive 98/79/EC Annex V	IVDD_V	
Directive 98/79/EC Annex VI	IVDD_VI	
Directive 98/79/EC Annex VII excluding section 5	IVDD_VII_EX_5	
Directive 98/79/EC Annex VII section 5	IVDD_VII_5	

BR-UDID-830 : Device link relationship - ENUM_Device_Link_Relationship_Type	RESOLVED	<table><tr><th>Label</th><th>Value</th><th>Notes</th><th>Sort Order</th></tr><tr><td>Legacy-Regulation Device Link</td><td>LEGACY_REGULATION_DEVICE_LINK</td><td></td><td>1</td></tr></table>	
DeviceSubstatusEnum	PUBLISHED	1. RECALLED 2. FIELD_SAFETY_CORRECTIVE_ACTION_INITIATED	BR-UDID-808: ENUM_UDID_DeviceSubStatus, DeviceSubstatusEnum
DeviceSystemProcedurePackTypeEnum	PUBLISHED	1. DEVICE 2. PROCEDURE_PACK 3. SYSTEM	BR-UDID-811: ENUM_UDID_DeviceSystemProcedurePackType , DeviceSystemProcedurePackTypeEnum
StorageHandlingConditionEnum	PUBLISHED	1. SHC001 2. SHC002 3. SHC003 4. SHC004 5. SHC005 6. SHC006 7. SHC007 8. SHC008 9. SHC009 10. SHC010 11. SHC012 12. SHC013 13. SHC014 14. SHC015 15. SHC016 16. SHC017 17. SHC018 18. SHC019 19. SHC021 20. SHC022 21. SHC023 22. SHC024 23. SHC025 24. SHC033 25. SHC035 26. SHC036 27. SHC037 28. SHC038 29. SHC042 30. SHC044 31. SHC045 32. SHC046 33. SHC099	BR-UDID-820: ENUM_UDID_StorageHandlingCondition, StorageHandlingConditionEnum
SystemProcedurePackTypeEnum	PUBLISHED	1. PROCEDURE_PACK 2. SYSTEM	BR-UDID-822: ENUM_UDID_SystemProcedurePackType, SystemProcedurePackTypeEnum

clinical-size-precisionEnum	PUBLISHED	1. range 2. text 3. value	BR-UDID-801: ENUM_UDID_ClinicalSizePrecision, clinical-size-precision
clinical-size-typeEnum	PUBLISHED	1. CST33 2. CST66 3. CST67 4. CST999 5. CST1 6. CST2 7. CST3 8. CST4 9. CST5 10. CST6 11. CST7 12. CST8 13. CST9 14. CST10 15. CST11 16. CST12 17. CST13 18. CST14 19. CST15 20. CST16 21. CST17 22. CST18 23. CST19 24. CST20 25. CST21 26. CST22 27. CST23 28. CST24 29. CST25 30. CST26 31. CST27 32. CST28 33. CST29 34. CST30 35. CST31 36. CST32 37. CST34 38. CST35 39. CST36 40. CST37 41. CST38 42. CST39 43. CST40 44. CST41 45. CST42 46. CST43 47. CST44 48. CST45 49. CST46 50. CST47 51. CST48 52. CST49 53. CST50 54. CST51 55. CST52 56. CST53 57. CST54 58. CST55 59. CST56 60. CST57 61. CST58 62. CST59 63. CST60 64. CST61 65. CST62 66. CST63 67. CST65	BR-UDID-802: ENUM_UDID_ClinicalSizeType, clinical-size-type
risk-classEnum	PUBLISHED	1. aimdd 2. class-a 3. class-b 4. class-c 5. class-d 6. class-i 7. ivd-annex-ii-list-a 8. class-iia 9. ivd-annex-ii-list-b 10. class-iib 11. ivd-devices-self-testing 12. class-iii 13. ivd-general	BR-UDID-817: ENUM_UDID_RiskClass, risk-class

European
Commission