

Table of contents

\\USER

2019_MBC_PROTOCOLS_VE11C

SB.2002.60 MR DEVELOPMENT - RS

SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104

localizer	*
gre_field_mapping	*
t1_fl2d_sag	*
t2_tse_tra_p2	*
t1_mp2rage	*
t2s_megre	*
ep2d_se_ap	*
ep2d_se_pa	*
ep2d_fid_basic_bold_p2_task	*
ep2d_se_ap	*
ep2d_se_pa	*
ep2d_fid_basic_bold_p2_rest	*
cmrr_ep2d_se_ap	*
cmrr_ep2d_se_apinvrope	*
cmrr_mbep2d_diff	*
pasl_3d_tra	*
csi_slaser	*

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\localizer *

TA: 0:10 PM: REF Voxel size: 0.5×0.5×7.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Contrast - Dynamic

Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
Base resolution	256
Phase resolution	91 %
Phase partial Fourier	Off
Interpolation	On

Routine

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
TE	3.69 ms
Averages	2
Concatenations	3
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	250 mm
FoV phase	100.0 %
Slice thickness	7.0 mm
TR	7.5 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	3

Contrast - Common

TR	7.5 ms
TE	3.69 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	2
Averaging mode	Short term

Geometry - AutoAlign

Slice group	1
Position	L0.0 A20.0 H0.0 mm

Geometry - AutoAlign

Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High

System - Tx/Rx

Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.5 ms
Concatenations	3
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	250 mm
FoV phase	100.0 %
Phase resolution	91 %

Physio - PACE

Resp. control	Off
Concatenations	3

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	20 deg
Measurements	1
Contrasts	1
TR	7.5 ms
TE	3.69 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No

Sequence - Part 1

Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\gre_field_mapping *

TA: 0:26 PM: REF Voxel size: 3.0×3.0×6.0 mmRel. SNR: 1.00 : fm_r

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	18
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	200.0 ms
TE 1	4.92 ms
TE 2	7.38 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HE1-4

Contrast - Common

TR	200.0 ms
TE 1	4.92 ms
TE 2	7.38 ms
MTC	Off
Flip angle	60 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	18
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	200.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	192 mm
A >> P	192 mm
F >> H	134 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Flow comp.	Yes
Multi-slice mode	Interleaved
Bandwidth	290 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\t1_fl2d_sag *

TA: 0:17 PM: REF Voxel size: 1.4×1.4×6.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	12
Dist. factor	30 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Phase oversampling	30 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	100.0 ms
TE	2.46 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Contrast - Common

TR	100.0 ms
TE	2.46 ms
MTC	Off
Magn. preparation	None
Flip angle	70 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
Base resolution	160
Phase resolution	80 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	12
Dist. factor	30 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	100.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	Head > Basis
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	100.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	100.0 %
Phase resolution	80 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off

Inline - MIP

Save original images	On
----------------------	----

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	70 deg
Measurements	1
Contrasts	1
TR	100.0 ms
TE	2.46 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Performance
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	TR
Max. TR	550.0 ms
Allowed delay	60 s

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\t2_tse_tra_p2 *

TA: 0:14 PM: REF Voxel size: 0.9×0.9×6.0 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	38
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Routine

Slice group	1
Slices	12
Dist. factor	30 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	30 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	2000.0 ms
TE	26.0 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	12
Dist. factor	30 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
TR	2000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	2000.0 ms
TE	26.0 ms
MTC	Off
Magn. preparation	None
Flip angle	150 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	6.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	On

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T

System - Miscellaneous

Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2000.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	100.0 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	8.82 ms
Bandwidth	219 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	6
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Normal
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	17

Sequence - Assistant

Mode	TR
Max. TR	7500.0 ms
Allowed delay	60 s

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\t1_mp2rage *

TA: 0:24 PM: REF Voxel size: 1.1×1.1×5.0 mmPAT: 1.0 Rel. SNR: 1.00 : WIP_cmp

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	40
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	5.00 mm
TR	2600.0 ms
TE	2.82 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	2600.0 ms
TE	2.82 ms
Magn. preparation	Non-sel. IR
TI 1	600 ms
TI 2	2000 ms
Flip angle 1	5 deg
Flip angle 2	5 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	93.8 %
Slice thickness	5.00 mm
Base resolution	224
Phase resolution	100 %
Slice resolution	100 %

Resolution - Common

Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	40
FoV read	256 mm
FoV phase	93.8 %
Slice thickness	5.00 mm
TR	2600.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P

System - Miscellaneous

Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Performance
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	256 mm
F >> H	200 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2600.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI 1	600 ms
TI 2	2000 ms
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	93.8 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	T1 map
Flip angle 1	5 deg
Flip angle 2	5 deg
Measurements	1
Contrasts	1
TR	2600.0 ms
TE	2.82 ms

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	6.6 ms
Bandwidth	240 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	40

Sequence - Special

Sparse Sampling	On
US	4.6 x
Samples/TR	175
Density	0.50
Jitter Radius	1.2
Reference Scan	External
No. Ref-Lines	32
Centric	Off
Virtual Coils	Off
No. Iterations	20
CSM RO Resolution	0
Regularisation INV1	0.00100
Regularisation INV2	0.00040
Uniform	On
Denoised UNI	On
FLAWS	Off
FLAWS-hc	Off
FLAWS-hc inv.	Off
Division image	Off
Synthetic TI 0	410 ms
Synthetic TI 1	1100 ms
Denoise Lambda	1
Scaling	0 10^
Echo Averaging	Off

Sequence - Special

FID Monitoring	Off
----------------	-----

Sequence - Assistant

Mode	Off
------	-----

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\t2s_megre *

TA: 0:24 PM: FIX Voxel size: 3.6×3.6×5.0 mmPAT: 3 Rel. SNR: 1.00 : fl_r

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	F >> H
AutoAlign	---
Phase oversampling	10 %
Slice oversampling	0.0 %
Slices per slab	30
FoV read	230 mm
FoV phase	90.6 %
Slice thickness	5.00 mm
TR	25.0 ms
TE 1	10.00 ms
TE 2	15.00 ms
TE 3	20.00 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Contrast - Common

TR	25.0 ms
TE 1	10.00 ms
TE 2	15.00 ms
TE 3	20.00 ms
MTC	Off
Magn. preparation	None
Flip angle	15 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	230 mm
FoV phase	90.6 %

Resolution - Common

Slice thickness	5.00 mm
Base resolution	64
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	F >> H
Slice oversampling	0.0 %
Slices per slab	30
FoV read	230 mm
FoV phase	90.6 %
Slice thickness	5.00 mm
TR	25.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	F >> H
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Coronal

Geometry - Saturation

Saturation mode	Standard
-----------------	----------

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Coronal
Rotation	90.00 deg
F >> H	209 mm
R >> L	230 mm
A >> P	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	25.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	230 mm
FoV phase	90.6 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Noise threshold	15
Save original images	On
MapIt	R2* map
Flip angle	15 deg
Measurements	1
Contrasts	3
TR	25.0 ms
TE 1	10.00 ms
TE 2	15.00 ms
TE 3	20.00 ms

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	3
Flow comp. 1	Yes
Readout mode	Monopolar
Multi-slice mode	Interleaved
Bandwidth 1	310 Hz/Px
Bandwidth 2	310 Hz/Px
Bandwidth 3	310 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_se_ap *

TA: 0:12 PM: REF Voxel size: 1.5×1.5×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
TE	60.0 ms
Averages	1
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	3000 ms
TE	60.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

Accel. mode	GRAPPA
-------------	--------

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-0.01 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-0.01 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_se_pa *

TA: 0:12 PM: REF Voxel size: 1.5×1.5×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
TE	60.0 ms
Averages	1
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	3000 ms
TE	60.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

Accel. mode	GRAPPA
-------------	--------

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_fid_basic_bold_p2_task *

TA: 0:20 PM: REF Voxel size: 3.0×3.0×4.0 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	2800 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HE1-4

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	2800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	2800 ms
TE	30.0 ms
MTC	Off
Flip angle	90 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	4
Delay in TR	1000 ms
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2800 ms
Concatenations	1

Perf

GBP	Off
PBP	Off
TTP	Off
relCBV	Off
relCBF	Off
relMTT	Off
relCBVCorr	Off
Measurements	4

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_se_ap *

TA: 0:12 PM: REF Voxel size: 1.5×1.5×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
TE	60.0 ms
Averages	1
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HE1-4

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	3000 ms
TE	60.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	-0.01 deg
Initial Orientation	Transversal

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	On

Geometry - Navigator

System - Miscellaneous

Resolution - iPAT

Accel. mode	GRAPPA
-------------	--------

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-0.01 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_se_pa *

TA: 0:12 PM: REF Voxel size: 1.5×1.5×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
TE	60.0 ms
Averages	1
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HE1-4

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	180.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	3000 ms
TE	60.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

Accel. mode	GRAPPA
-------------	--------

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Slice
Diff. directions	1
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\ep2d_fid_basic_bold_p2_rest *

TA: 0:31 PM: REF Voxel size: 3.0×3.0×4.0 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	24
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	2800 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HE1-4

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	2800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Contrast - Common

TR	2800 ms
TE	30.0 ms
MTC	Off
Flip angle	90 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	8
Delay in TR	1000 ms
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

Accel. mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2800 ms
Concatenations	1

Perf

GBP	Off
PBP	Off
TTP	Off
relCBV	Off
relCBF	Off
relMTT	Off
relCBVCorr	Off
Measurements	8

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - pTX Pulses

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\cmrr_ep2d_se_ap *

TA: 0:12 PM: REF Voxel size: 3.0×3.0×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3000 ms
TE	60.00 ms
Multi-band accel. factor	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	3000 ms
TE	60.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
----------	--------

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Multi-band accel. factor	1

Physio - PACE

Resp. control	Off
Multi-band accel. factor	1

Diff - Neuro

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\cmrr_ep2d_se_apinvrope *
TA: 0:12 PM: REF Voxel size: 3.0×3.0×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Routine

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3000 ms
TE	60.00 ms
Multi-band accel. factor	1
Filter	Prescan Normalize
Coil elements	HE1-4

Geometry - Common

Slice group	1
Slices	25
Dist. factor	25 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Contrast - Common

TR	3000 ms
TE	60.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled
Special sat.	None

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Geometry - Navigator**System - Miscellaneous****Resolution - iPAT**

PAT mode	GRAPPA
----------	--------

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	124 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3000 ms
Multi-band accel. factor	1

Physio - PACE

Resp. control	Off
Multi-band accel. factor	1

Diff - Neuro

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm ²

Diff - Body

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	64
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	On
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\cmrr_mbep2d_diff *

TA: 0:38 PM: REF Voxel size: 1.1×1.1×4.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	27
Dist. factor	30 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3800 ms
TE	81.00 ms
Multi-band accel. factor	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	3800 ms
TE	81.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
Base resolution	192
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
----------	--------

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	40
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	27
Dist. factor	30 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.00 mm
TR	3800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled
Special sat.	None

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off

System - Miscellaneous

AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	140 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3800 ms
Multi-band accel. factor	1

Physio - PACE

Resp. control	Off
Multi-band accel. factor	1

Diff - Neuro

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Bipolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
FA maps	On
Mosaic	On
Tensor	On
Noise level	40

Diff - Body

Diffusion mode	MDDW
Diff. directions	6
Diffusion Scheme	Bipolar

Diff - Body

Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	On
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.78 ms
Bandwidth	1446 Hz/Px

Sequence - Part 2

EPI factor	192
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	Off
Force Maxwell corr.	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\pasl_3d_tra *

TA: 0:20 PM: ISO Voxel size: 1.9×1.9×4.5 mmRel. SNR: 1.00 : tgse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	16
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	4.50 mm
TR	2000 ms
TE	20.26 ms
Averages	1
Concatenations	1
Filter	Raw filter, Distortion Corr.(2D), Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	2000 ms
TE	20.26 ms
Flip angle	180 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	2
Delay in TR	0 ms
Multiple series	Off

Contrast - ASL

Perfusion mode	FAIR QII
Suppression Mode	GRAY-WHITE
Bolus Duration	800 ms
Inversion Time	1000 ms
Averaging mode	CONSTANT
Inversion Array Size	1

Resolution - Common

FoV read	240 mm
FoV phase	100.0 %
Slice thickness	4.50 mm
Base resolution	64
Phase resolution	97 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	On

Resolution - Filter Image

Distortion Corr.	On
Mode	2D
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
Slice oversampling	0.0 %
Slices per slab	16
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	4.50 mm
TR	2000 ms
Multi-slice mode	Interleaved
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	ISO
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L

System - Miscellaneous

Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-180.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	72 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.237844 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2000 ms
Concatenations	1
Segments	2

Sequence - Part 1

Introduction	Off
Dimension	3D
Reordering	Centric
Multi-slice mode	Interleaved
Echo spacing	0.49 ms
Bandwidth	2442 Hz/Px

Sequence - Part 2

EPI factor	31
Segments	2
RF pulse type	Normal
Gradient mode	Performance
Turbo factor	16

\\USER\2019_MBC_PROTOCOLS_VE11C\SB.2002.60 MR DEVELOPMENT - RS\SB.2002.60 MR DEVELOPMENT - PROTOCOLQC - 104\csi_slaser *

TA: 0:37 PM: REF Voxel size: 20.0×20.0×20.0 mmRel. SNR: 1.00 : csi_slaser

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Position	Isocenter
Orientation	Transversal
Rotation	0 deg
Slices	1
Vol A >> P	100 mm
Vol R >> L	80 mm
FoV A >> P	160 mm
FoV R >> L	160 mm
Thickness F >> H	20 mm
TR	1200 ms
TE	50 ms
Averages	1
Filter	Prescan Normalize, Hamming
Coil elements	HE1-4

Contrast

TR	1200 ms
TE	50 ms
Averages	1
Averaging mode	Long term
Flip angle	65 deg
Water suppr.	Weak water suppr.
Water suppr. BW	50 Hz
Measurements	1

Resolution - Common

FoV R >> L	160 mm
FoV A >> P	160 mm
Thickness F >> H	20 mm
Scan res. R >> L	8
Scan res. A >> P	8
Interpol. res. R >> L	8
Interpol. res. A >> P	8
Hamming	On
Width	50
Prescan Normalize	On
Vector size	256

Geometry - Common

Position	Isocenter
Orientation	Transversal
Rotation	0 deg
FoV R >> L	160 mm
FoV A >> P	160 mm

Geometry - Common

Thickness F >> H	20 mm
Vol R >> L	80 mm
Vol A >> P	100 mm
Sat. region	1
Thickness	40 mm
Position	L61.0 P0.0 H0.0 mm
Orientation	Sagittal
Sat. delta frequ.	-3.40 ppm
Sat. region	2
Thickness	40 mm
Position	R61.0 P0.0 H0.0 mm
Orientation	Sagittal
Sat. delta frequ.	-3.40 ppm
Sat. region	3
Thickness	40 mm
Position	L0.0 P61.0 H0.0 mm
Orientation	Coronal
Sat. delta frequ.	-3.40 ppm
Sat. region	4
Thickness	40 mm
Position	L0.0 A61.0 H0.0 mm
Orientation	Coronal
Sat. delta frequ.	-3.40 ppm

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Adj. water suppr.	On
Adjust with body coil	Off
Confirm freq. adjustment	On
Only after freq. change	On
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	80 mm
A >> P	100 mm
F >> H	20 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.237844 MHz
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Sequence - Common

Preparation scans	2
Dimension	2D
Delta frequency	-2.70 ppm
Phase encoding	Weighted
Bandwidth	2000 Hz
Acquisition duration	128 ms
Remove oversampling	Off