

No. Laporan: KALX62509425099R

Sertifikat Kalibrasi Eksternal

Identitas Pemilik : Laboratorium Telkom Test House (TTH)
Jl. Gegerkalong Hilir No. 47 Bandung 40152

Deskripsi Alat yang Dikalibrasi

Nama Alat : ~2.2 km LENGTH OF SINGLE MODE OPTICAL FIBRE
Merek : CORNING
Tipe/Model : ~2.2 km
Nomor Seri : 100537530456D
Identifikasi Alat : 009/KAL

Informasi Kalibrasi

Tanggal Diterima CSTTH : 07 November 2025
Tanggal Diterima Lab : 19 November 2025
Penyedia Kalibrasi : Lab Kalibrasi Telkom Jakarta (LK-016-IDN)
Tempat Kalibrasi : Lab Kalibrasi Telkom Jakarta
Tanggal Pengiriman : 20 November 2025
Tanggal Pengambilan : 20 November 2025
Nomor Sertifikat : JCL/25/3209/DY
Tanggal Kalibrasi : 20 November 2025
Tanggal Kalibrasi Ulang : 20 November 2026

Bandung, 24 Desember 2025

Diperiksa Oleh:



ADI PERMADI

Mgr. Standardization & Suitability Assurance

Disetujui Oleh:



GUNADI DWI HANTORO

SM Infrastructure Service Research & Assurance



CERTIFICATE OF CALIBRATION

P T T E L K O M I N D O N E S I A T b k
KAN ACCREDITED CALIBRATION LABORATORY (LK-016-IDN)

Jakarta : Jl. Percetakan Negara No.17-19 Jakarta Pusat 10570 Phone 021-21480341
Medan : Jl. Gaharu No. 1 Medan 20235 Phone 061-4531311
Surabaya : Jl. Gayungan PTT No. 17-19 Surabaya 60235 Phone 031-8283678
Makassar : Jl. AP Pettarani No. 4 Makassar 90221 Phone 0411-861153

CS Support
☎ 081-16-4-17025

Certificate Number

JCL/25/3209/DY

Approved Signatory


M. Ilyas

Date of Issue : 21-Nov-2025
Submitted by : PT. TELKOM INDONESIA, Tbk.
Laboratorium Telkom Test House (TTH)
Attention : Mgr. Standardization & Suitability Assurance
Description : ~2.2 km Length of Single Mode Optical Fibre
Manufacturer/Model : Corning/ ~2.2 km
Serial Number : 100537530456D / (009/KAL)
Environment : 23.4 °C ± 0.2 °C 56.2 % ± 0.1 % RH
Calibration Location : In-House
Date Calibrated : 20-Nov-2025
Recommended Due Date : 19-Nov-2026
Date Received : 20-Nov-2025

PT. Telkom Indonesia, Tbk. - Telkom Certifies that this instrument has been calibrated under the environmental condition stated above using telkom standard (s) which is (are) traceable to the international system of Units (SI) through MIKROKOM, Czech Republic.

Method of Calibration :

The calibration was carried out based on the established Standard Telkom Calibration Procedure, TCL-WI-0009 and by comparison against the following standard(s) and instrument(s) which is (are) calibrated on schedule to maintain traceability and required accuracy :

Description	S/N	Cal Date	Due Date
1. OTDR	08795	04-Sep-2024	03-Sep-2027

CERTIFICATE OF CALIBRATION

Checked By : 

Certificate Number : JCL/25/3209/DY



1. Length Measurements λ 1310 nm

TEST No.	EXPECTED VALUE OF UUT	MEASURED VALUE	CORRECTION	UNCERTAINTY
1 above	2230.72 m	2230.48 m	-0.24 m	1.7 m

2. Length Measurements λ 1550 nm

TEST No.	EXPECTED VALUE OF UUT	MEASURED VALUE	CORRECTION	UNCERTAINTY
1.	2231.64 m	2231.44 m	-0.20 m	1.7 m