

MDF: Magnetic Particle Imaging Data Format

T. Knopp^{1,2}, T. Viereck³, G. Bringout⁴, M. Ahlborg⁴, J. Rahmer⁵, M. Hofmann^{1,2}

¹Section for Biomedical Imaging, University Medical Center Hamburg-Eppendorf, Germany

²Institute for Biomedical Imaging, Hamburg University of Technology, Germany

³Institute of Electrical Measurement and Fundamental Electrical Engineering, TU Braunschweig, Germany

⁴Institute of Medical Engineering, University of Lübeck, Germany

⁵Philips GmbH Innovative Technologies, Research Laboratories, Röntgenstraße 24-26, 22315 Hamburg, Germany

February 9, 2016

Version **1.0.1**

Abstract

Magnetic particle imaging (MPI) is a tomographic method to determine the spatial distribution of magnetic nanoparticles. In this document a file format for the standardized storage of MPI data is introduced. The aim of the Magnetic Particle Imaging Data Format (MDF) is to provide a coherent way of exchanging MPI data acquired with different MPI scanners worldwide. The focus of the file format is on sequence parameters, raw measurement data, calibration data, and reconstruction data. The format is based on the hierarchical document format (HDF) in version 5 (HDF5).

1 Introduction

The purpose of this document is to introduce a file format for exchanging Magnetic Particle Imaging (MPI) data. The Magnetic Particle Imaging Data Format (MDF) is based on the hierarchical document format (HDF) in version 5 (HDF5). HDF5 allows to store multiple datasets within a single file and is thus very flexible to use. To

allow the exchange of MPI data, one has to specify a naming scheme within HDF5 files which is the purpose of this document. In order to create and access HDF5 data, an Open Source C library is available. For most programming languages bindings to this library exist. Matlab supports HDF5 by the functions `h5read` and `h5write`. For

Python the `h5py` package exists. The Julia programming language provides access to HDF5 files via the `HDF5` package. For languages based on the .NET framework the `HDF5DotNet` library is available.

In this initial version of the file format the focus is on sequence parameters, raw measurement data, calibration data, and reconstruction data. The format can store three different dataset types

1. Measurement data
2. System calibration data
3. Reconstruction data

It is possible to combine measurement data and reconstruction data into a single file. However, calibration data has to be stored in an independent HDF5 file.

1.1 Datatypes

For most parameters a fixed datatype is used, i.e. the drive-field amplitudes are stored as `Float64` values. In case of the measurement data and the calibration data the datatype is not restricted such that maximum flexibility is given. In case of no restriction we refer to the datatype as `Any`.

Since storing complex data in HDF5 is not standardized, we extend the dimensionality of an existing array and store the real and imaginary part in the last dimension with size 2 (index 0 = real part, index 1 = imaginary part). In this way the real and imaginary part of a complex datum is stored sequentially on disk. When loading the data it is possible to cast it to a complex array in most programming languages.

1.2 Units

Physical quantities are given in SI units with one exception. The field strength is reported in $T\mu_0^{-1} = 4\pi\text{ Am}^{-1}\mu_0^{-1}$. This convention has been proposed in the first MPI publication and since that time consistently used in most MPI related publications. The aim of this convention is to report the numbers on a Tesla scale, which most readers with a background in MRI are familiar with, but, on the other hand still use the correct unit for the magnetic field strength.

1.3 Sanity Check

In order to check if a generated MDF file is valid, we provide a sanity check script that can be found in the gitub repository:

<https://github.com/MagneticParticleImaging/MDF>

The code is written in the Julia programming language, which has to be downloaded from:

<http://julialang.org>.

More detailed instructions can be found in the README of the repository.

1.4 Contact

If you find mistakes in this document or the specified file format or if you want to discuss extensions to this specification, please open an issue on github:

<https://github.com/MagneticParticleImaging/MDF>

As the file format is versionized it will be possible to extend it for future needs of MPI. The current version discussed in this document is version 1.0.1.

2 Data (group: /)

Remarks: Within the root group metadata about the file itself is stored. Within several subgroups, metadata about the experimental setting, the MPI tracer, and the MPI scanner are stored. The actual data is stored in dedicated groups on measurement data, calibration data, and/or reconstruction data.

Parameter	Type	Dim	Unit/Format	Optional	Description
version	String	1	0.1	no	Version of the file format
uuid	String	1	f81d4fae-7dec-11d0-a765-00a0c91e6bf6	no	Universally Unique Identifier (RFC 4122)
date	String	1	yyyy-mm-ddThh:mm:ss.ms	no	UTC creation time of MDF data set

2.1 Study Description (group: /study/)

Remarks: The study description group describes the experimental setting under which the MPI data was recorded. The study field may be used as a name tag for several experiments, which are related. The dataset at hand may then be described by a number and a short description. Additionally, the name of the imaged subject can be provided and the starting time of the MPI measurement can be provided. The reference field may be used to indicate, if the background signal of the scanner was recorded.

Parameter	Type	Dim	Unit/Format	Optional	Description
name	String	1		yes	Name of the study
experiment	String	1		yes	Experiment number within study
description	String	1		yes	Short description of the experiment
subject	String	1		yes	Name of the subject that was imaged
reference	Int64	1		yes	Flag indicating if field of view was empty during the measurement
simulation	Int64	1		yes	Flag indicating if the data in this file is simulated rather than measured

2.2 Tracer Parameters (group: /tracer/)

Remarks: The tracer parameter group contains information about the MPI tracer used during the experiment such as the tracer name, its vendor, the tracer concentration, and the total volume applied.

Note that the injection clock recording the injection time should be synchronized with the clock, which provides the starting time of the measurement.

Parameter	Type	Dim	Unit/Format	Optional	Description
name	String	1		yes	Name of tracer used in experiment
batch	String	1		yes	Batch of tracer
vendor	String	1		yes	Name of tracer supplier
volume	Float64	1	L	yes	Total volume of applied tracer
concentration	Float64	1	mol(Fe)/L	yes	Concentration of tracer
time	String	1	yyyy-mm-ddThh:mm:ss.ms	yes	UTC time at which tracer injection started

2.3 Scanner Parameters (group: /scanner/)

Remarks: The scanner parameter group contains information about the MPI scanner used such as the manufacturer, the model, and the facility where the scanner is installed.

Parameter	Type	Dim	Unit/Format	Optional	Description
facility	String	1		no	Facility where the MPI scanner is installed
operator	String	1		no	User who operates the MPI scanner
manufacturer	String	1		no	Scanner manufacturer
model	String	1		no	Scanner model
topology	String	1		no	Scanner topology (e.g. FFP or FFL)

2.4 Acquisition Parameters (group: /acquisition/)

Remarks: The acquisition parameter group can describe different imaging protocols and trajectory settings. The corresponding data is organized into general information, a subgroup containing data on the particle excitation and a subgroup containing data on the receive channels.

In general each MPI dataset consists of the measurement data of L frames. A frame groups all data together that will be used to

reconstruct an image/volume. On certain MPI scanners the drive-field field-of-view (FOV) can be shifted either by magnetic fields or by mechanical movement. Therefore, a frame may consist of J sub-measurements. For instance a Cartesian 2D trajectory with 100 lines would be realized by setting `numPatches` = 100.

Parameter	Type	Dim	Unit/Format	Optional	Description
<code>numFrames</code>	Int64	1	1	no	Number of available frames, denoted by L
<code>framePeriod</code>	Float64	1	s	no	Complete time to acquire a full frame
<code>numPatches</code>	Int64	1	1	no	Number of patches within a frame denoted by J
<code>gradient</code>	Float64	3 or $J \times 3$	$\text{Tm}^{-1}\mu_0^{-1}$	no	Gradient strength of the selection field in x , y , and z directions
<code>time</code>	String	1	yyyy-mm-ddThh:mm:ss.ms	no	UTC start time of MPI measurement

2.4.1 Drive Field (group: `/acquisition/drivefield/`)

Remarks: The drive field subgroup describes the details on the imaging protocol and trajectory settings. On the lowest level each MPI scanner contains D channels for particle excitation.

These excitation signals are usually sinusoidal and can be described by D amplitudes (drive field strengths), a base frequency, and D dividers. Depending on the base frequency and the divider a

periodic excitation signal is generated defining the sampling trajectory of the field-free point or field-free line. The trajectory may cover a 1D, 2D, or 3D area.

Certain parameters such as `strength`, `fieldOfViewCenter`, and `fieldOfView` can either be defined globally for the entire multi-patch sequence or individually for each patch of the sequence.

Parameter	Type	Dim	Unit/Format	Optional	Description
numChannels	Int64	1	1	no	Number of drive field channels, denoted by D
strength	Float64	D or $J \times D$	$T\mu_0^{-1}$	no	Applied drive field strength
baseFrequency	Float64	1	Hz	no	Base frequency to derive drive field frequencies
divider	Int64	D	1	no	Divider for drive fields frequencies (baseFrequency / divider)
period	Float64	1	s	no	Drive field trajectory period
averages	Int64	1		no	Number of internal averages (applied in hardware/software)
repetitionTime	Float64	1	s	no	Time to complete averaged DF trajectory (averages * period)
fieldOfView	Float64	3 or $J \times 3$	m	no	Approximate size of the area/volume captured by the trajectory
fieldOfViewCenter	Float64	3 or $J \times 3$	m	no	Center of the drive field trajectory (relative to origin/center)

2.4.2 Receiver (group: /acquisition/receiver/)

Remarks: The receiver subgroup describes details on the MPI receiver. For a multi-patch sequence it is assumed, that signal acquisition only takes place during particle excitation. During each drive-field cycle, C receive channels record the superposition of the change of the particle magnetization at Z equidistant time points.

Parameter	Type	Dim	Unit/Format	Optional	Description
numChannels	Int64	1		no	Number of receive channels C
bandwidth	Float64	1	Hz	no	Bandwidth of the receiver unit
numSamplingPoints	Int64	1		no	Number of sampling point within one drive-field period denoted by Z
frequencies	Float64	K	Hz	no	Vector containing recorded frequencies
transferFunction	Float64	$C \times K \times 2$		yes	Transfer function of the receive channel

2.5 Measurements (group: /measurement/)

Remarks: Measured data can be stored in Fourier domain (FD) representation as well as time domain (TD) representation. Usually only one of the representation is stored and one has to calculate the missing representation if it is needed but not available.

One measurement consists of the voltages recorded in all receive channels for all patches. The number of measurements in time domain is thus given by $\tilde{M} = JCZ$. In frequency domain one measure-

ment has $M = KCZ$ data points. On disc the temporal/frequency index is the fastest, while the index over the patches is the slowest. If several measurements are acquired (indicated by *numFrames*), the measurements are concatenated. The number of measurements is denoted by L .

Parameter	Type	Dim	Unit/Format	Optional	Description
dataFD	Any	$L \times C \times K \times 2$ or $L \times J \times C \times K \times 2$		yes	Measurement data stored in Fourier domain representation. The last dimension is used for storing the complex data.
dataTD	Any	$L \times C \times Z$ or $L \times J \times C \times Z$		yes	Measurement data stored in time domain representation

2.6 Calibration (group: /calibration/)

Remarks: Calibration data is usually acquired by shifting a delta sample through the FOV of the scanner. The resulting system matrix can be stored in time or frequency domain. The number of spatial points is indicated by N . For practical reasons the system matrix is stored in a transposed way compared to the measurement data. Hence the index over the positions is the fastest while the index over all data points of a particular measurement is the slowest. Depending on the programming language used (column-major or row-major

order) the system matrix may appear in a transposed representation when loaded into main memory.

If a regular grid is used for sampling, by default the fastest index is in x direction, the second fastest index is in y direction, while the slowest index is in z direction. If a different ordering is used this can be documented using the optional parameter **order**. For non-regular sampling points there is the possibility to explicitly store all positions.

Parameter	Type	Dim	Unit/Format	Optional	Description
dataFD	Any	$C \times K \times N \times 2$ or $J \times C \times K \times N \times 2$		yes	System matrix stored in its Fourier space representation. The last dimension is used for storing the complex data.
snrFD	Float64	$C \times K$		yes	Signal-to-noise estimate for recorded frequency components
dataTD	Float64	$C \times Z \times N$ or $J \times C \times Z \times N$		yes	System matrix stored in its time domain representation
fieldOfView	Float64	3	m	yes	Field of view of system matrix
fieldOfViewCenter	Float64	3	m	yes	Center of the system matrix (relative to origin/-center)
size	Int64	3		yes	Number of voxels in each dimension
order	String	1		yes	Ordering of the dimensions, default is <i>xyz</i>
positions	Float64	$N \times 3$	m	yes	Position of each of the grid points, stored as (x, y, z) tripels
deltaSampleSize	Float64	3	m	yes	Size of delta Sample used for calibration scan
method	String	1		yes	Method used to obtain calibration data. Can for instance be robot, hybrid, or simulation

2.7 Reconstruction Results (group: /reconstruction/)

Reconstruction results are stored in the parameter **data**. Since the grid of the reconstruction data can be different than the system matrix grid, the grid parameter are mirrored in the reconstruction parameter group.

Usually the data is stored in a real data format but it is also possible to store complex data if the reconstruction output is complex.

Parameter	Type	Dim	Unit/Format	Optional	Description
data	Any	$L \times N$		yes	Reconstructed data
fieldOfView	Float64	3	m	yes	Field of view of reconstructed data
fieldOfViewCenter	Float64	3	m	yes	Center of the reconstructed data (relative to origin/center)
size	Int64	3		yes	Number of voxels in each dimension
order	String	1		yes	Ordering of the dimensions, default is <i>xyz</i>
positions	Float64	$N \times 3$	m	yes	Position of each of the grid points, stored as (x, y, z) tripels