

GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION
NATIONAL LEGAL METROLOGY TESTING CENTRE

TYPE EVALUATION TEST REPORT

Non-Automatic Weighing Instrument (NAWI) as per OIML R 76-1:2006 (E)

Report Number: REP-2026-OIML-001	Date: 15-Sep-2026
Session ID: TS-2026-001-MW3000	Report Status: FINALIZED (v1.0)
Overall Result: PASS	Standard: OIML R 76-1 Edition 2006 (E)

1. Applicant & Manufacturer Information

Applicant Name:	Demo WeighTech Pvt. Ltd.
Applicant Address:	Plot 45, Okhla Industrial Area Phase III, New Delhi 110020, India
Manufacturer:	Demo WeighTech Pvt. Ltd.
Manufacturer Address:	Plot 45, Okhla Industrial Area Phase III, New Delhi 110020, India
Country of Manufacture:	India

2. Technical & Metrological Characteristics

Instrument Model:	MW-3000	Accuracy Class:	Class III
Max Capacity (Max):	3000.0 kg	Min Capacity (Min):	20.0 kg
Scale Interval (e):	1.0 kg	Division (d):	1.0 kg
Interval Count (n):	3000	Serial Number:	MW3000-2026-IND-001
Load Receptor:	Platform (4 pts)	Electronic:	Yes
Power Supply:	230 V AC, 50 Hz	Temperature Range:	-10.0°C to 40.0°C

3. Laboratory Environmental Conditions

Ambient Temperature:	22.4 °C	Relative Humidity:	52.0 %
Atmospheric Pressure:	1012.8 hPa	Mains Voltage:	230.2 V AC

4. OIML R-76 Test Applicability & Coverage Matrix

Code	Clause	Test Procedure	Applicability	Execution	Status
A.1	Clause A.1	Administrative Examination	APPLICABLE	NOT TESTED	PENDING
A.2	Clause A.2	Compare Construction with Documentation	APPLICABLE	NOT TESTED	PENDING
A.3	Clause A.3	Initial Examination	APPLICABLE	NOT TESTED	PENDING
A.4.2.1	Clause 4.5.1 / A.4.2.1	Range of Zero-Setting	APPLICABLE	NOT TESTED	PENDING
A.4.2.3	Clause A.4.2.3	Accuracy of Zero-Setting	APPLICABLE	EXECUTED	PASS

A.4.4.1	Clause A.4.4.1 & A.4.4.3	Weighing Performance Test	APPLICABLE	EXECUTED	PASS
A.4.7	Clause A.4.7	Eccentricity Test	APPLICABLE	EXECUTED	PASS
A.4.8	Clause A.4.8	Discrimination Test	APPLICABLE	NOT TESTED	PENDING
A.4.10	Clause A.4.10	Repeatability Test	APPLICABLE	EXECUTED	PASS
A.4.11.1	Clause A.4.11.1	Creep Test	APPLICABLE	NOT TESTED	PENDING
A.4.11.2	Clause A.4.11.2	Zero Return Test	APPLICABLE	NOT TESTED	PENDING
A.5.1	Clause A.5.1	Tilting Test	APPLICABLE	NOT TESTED	PENDING
A.5.3.1	Clause A.5.3.1	Static Temperatures Test	APPLICABLE	NOT TESTED	PENDING
A.5.3.2	Clause A.5.3.2	Temperature Effect on No-Load Indication	APPLICABLE	NOT TESTED	PENDING
A.5.4	Clause A.5.4	Voltage Variations Test	APPLICABLE	NOT TESTED	PENDING
A.6	Clause A.6	Endurance Test	APPLICABLE	NOT TESTED	PENDING
B.2	Clause B.2	Damp Heat, Steady State	APPLICABLE	NOT TESTED	PENDING
B.3	Clause B.3	Performance Tests for Disturbances (EMC / ESD / Dips)	APPLICABLE	NOT TESTED	PENDING
B.4	Clause B.4	Span Stability Test	APPLICABLE	NOT TESTED	PENDING
Annex G	Annex G (Cl. G.1 - G.4)	Software-Controlled Instrument Examination	APPLICABLE	NOT TESTED	PENDING

5. Summary of Executed Test Results & Metrological Compliance

Clause	Test Description	Observed Result	Permissible Limit	Result
Clause A.4.2.3	Accuracy of Zero-Setting	E0 = +0.050 kg	Limit = ±0.250 kg (0.25e)	PASS
Clause A.4.4.1 & A.4.4.3	Weighing Performance Test	5 points evaluated (Max Ec = 0.200 kg)	Tiered MPE as per Table 6 for Class III	PASS
Clause A.4.7	Eccentricity Test	Max eccentricity error = 0.090 kg	Limit = ±1.000 kg	PASS
Clause A.4.10 & 3.6.1	Repeatability Test	ΔE (Max-Min) = 0.030 kg	Limit = 1.000 kg	PASS

6. Review & Authorization

Evaluated By: _____ Test Engineer Date: 15-Sep-2026	Reviewed By: _____ Technical Reviewer Status: FINALIZED	Approved By: _____ Director / Lab In-Charge Status: FINALIZED / SEALED Date: 13-Sep-2026
---	---	---

LEGAL DISCLAIMER / NOTICE: This document is a prototype evaluation test report generated by METRIREPORT for SIH 2026. It does not constitute an official Legal Metrology Model Approval Certificate under Section 22 of the Legal Metrology Act, 2009 without statutory Gazette notification.