

GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION
NATIONAL LEGAL METROLOGY TESTING CENTRE

TYPE EVALUATION TEST REPORT

Non-Automatic Weighing Instrument (NAWI) as per OIML R 76-1:2006 (E)

Report Number: REP-2026-OIML-CF407F	Date: 14-Sep-2026
Session ID: TS-20260914-4171B7	Report Status: READY_FOR_REVIEW (v1.0)
Overall Result: PASS	Standard: OIML R 76-1 Edition 2006 (E)

1. Applicant & Manufacturer Information

Applicant Name:	WeighTech Metrology India Ltd.
Applicant Address:	Plot 42, Electronics City, Bangalore, India
Manufacturer:	WeighTech Metrology India Ltd.
Manufacturer Address:	Plot 42, Electronics City, Bangalore, India
Country of Manufacture:	India

2. Technical & Metrological Characteristics

Instrument Model:	MW-3000	Accuracy Class:	Class III
Max Capacity (Max):	3000.0 kg	Min Capacity (Min):	20.0 kg
Scale Interval (e):	1.0 kg	Division (d):	1.0 kg
Interval Count (n):	3000	Serial Number:	SN-E2E-TEST-F795AF20
Load Receptor:	Platform (4 pts)	Electronic:	Yes
Power Supply:	230 V AC, 50 Hz	Temperature Range:	-10.0°C to 40.0°C

3. Laboratory Environmental Conditions

Ambient Temperature:	22.5 °C	Relative Humidity:	55.0 %
Atmospheric Pressure:	1013.25 hPa	Mains Voltage:	230.0 V AC

4. Summary of Test Results & Compliance (OIML R 76-1:2006)

Clause	Test Description	Observed Result	Permissible Limit	Result
Clause A.4.2.3	Accuracy of Zero-Setting	E0 = +0.050 kg	Limit = ± 0.250 kg (0.25e)	PASS
Clause A.4.4.1 & A.4.4.3	Weighing Performance Test	5 points evaluated (Max Ec = 0.200 kg)	Tiered MPE as per Table 6 for Class III	PASS
Clause A.4.7	Eccentricity Test	Max eccentricity error = 0.090 kg	Limit = ± 1.000 kg	PASS
Clause A.4.10 & 3.6.1	Repeatability Test	ΔE (Max-Min) = 0.030 kg	Limit = 1.000 kg	PASS

5. Review & Authorization

Evaluated By: Test Engineer Date: 14-Sep-2026	Reviewed By: Technical Reviewer Status: APPROVED	Approved By: Director / Lab In-Charge Seal & Signature Placeholder
--	---	---

LEGAL DISCLAIMER / NOTICE: This document is a prototype evaluation test report generated by METRIREPORT for SIH 2026. It does not constitute an official Legal Metrology Model Approval Certificate under Section 22 of the Legal Metrology Act, 2009 without statutory Gazette notification.