

# GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION  
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION  
NATIONAL LEGAL METROLOGY TESTING CENTRE

## TYPE EVALUATION TEST REPORT

Non-Automatic Weighing Instrument (NAWI) as per OIML R 76-1:2006 (E)

|                                     |                                               |
|-------------------------------------|-----------------------------------------------|
| Report Number: REP-2026-OIML-D0A3BD | Date: 14-Sep-2026                             |
| Session ID: TS-20260914-F53C1D      | Report Status: <b>READY_FOR_REVIEW (v1.0)</b> |
| Overall Result: <b>PASS</b>         | Standard: OIML R 76-1 Edition 2006 (E)        |

### 1. Applicant & Manufacturer Information

|                         |                                                      |
|-------------------------|------------------------------------------------------|
| Applicant Name:         | Retail Scales Co.                                    |
| Applicant Address:      | Guindy Industrial Estate, Chennai 600032, Tamil Nadu |
| Manufacturer:           | Retail Scales Co.                                    |
| Manufacturer Address:   | Guindy Industrial Estate, Chennai 600032, Tamil Nadu |
| Country of Manufacture: | India                                                |

### 2. Technical & Metrological Characteristics

|                     |                  |                     |                   |
|---------------------|------------------|---------------------|-------------------|
| Instrument Model:   | RL-30            | Accuracy Class:     | Class III         |
| Max Capacity (Max): | 30.0 kg          | Min Capacity (Min): | 0.1 kg            |
| Scale Interval (e): | 0.005 kg         | Division (d):       | 0.005 kg          |
| Interval Count (n): | 6000             | Serial Number:      | RL30-2026-IND-551 |
| Load Receptor:      | Platform (4 pts) | Electronic:         | Yes               |
| Power Supply:       | 230 V AC, 50 Hz  | Temperature Range:  | -10.0°C to 40.0°C |

### 3. Laboratory Environmental Conditions

|                       |             |                    |            |
|-----------------------|-------------|--------------------|------------|
| Ambient Temperature:  | 22.5 °C     | Relative Humidity: | 54.0 %     |
| Atmospheric Pressure: | 1013.25 hPa | Mains Voltage:     | 230.0 V AC |

### 4. Summary of Test Results & Compliance (OIML R 76-1:2006)

| Clause                   | Test Description          | Observed Result                          | Permissible Limit                       | Result      |
|--------------------------|---------------------------|------------------------------------------|-----------------------------------------|-------------|
| Clause A.4.2.3           | Accuracy of Zero-Setting  | E0 = +0.000 kg                           | Limit = $\pm 0.001$ kg (0.25e)          | <b>PASS</b> |
| Clause A.4.4.1 & A.4.4.3 | Weighing Performance Test | 4 points evaluated (Max  Ec  = 0.001 kg) | Tiered MPE as per Table 6 for Class III | <b>PASS</b> |
| Clause A.4.7             | Eccentricity Test         | Max eccentricity error = 0.000 kg        | Limit = $\pm 0.005$ kg                  | <b>PASS</b> |
| Clause A.4.10 & 3.6.1    | Repeatability Test        | $\Delta E$ (Max-Min) = 0.000 kg          | Limit = 0.007 kg                        | <b>PASS</b> |

5. Review & Authorization

|                                                                                |                                                                                   |                                                                                                         |
|--------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------|
| <div>Evaluated By:</div> <div>Test Engineer</div> <div>Date: 14-Sep-2026</div> | <div>Reviewed By:</div> <div>Technical Reviewer</div> <div>Status: APPROVED</div> | <div>Approved By:</div> <div>Director / Lab In-Charge</div> <div>Seal &amp; Signature Placeholder</div> |
|--------------------------------------------------------------------------------|-----------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------|

**LEGAL DISCLAIMER / NOTICE:** This document is a prototype evaluation test report generated by METRIREPORT for SIH 2026. It does not constitute an official Legal Metrology Model Approval Certificate under Section 22 of the Legal Metrology Act, 2009 without statutory Gazette notification.