

# GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION  
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION  
NATIONAL LEGAL METROLOGY TESTING CENTRE

## TYPE EVALUATION TEST REPORT

Non-Automatic Weighing Instrument (NAWI) as per OIML R 76-1:2006 (E)

|                                  |                                        |
|----------------------------------|----------------------------------------|
| Report Number: REP-2026-OIML-001 | Date: 14-Sep-2026                      |
| Session ID: TS-2026-001-MW3000   | Report Status: <b>FINALIZED (v1.0)</b> |
| Overall Result: <b>PASS</b>      | Standard: OIML R 76-1 Edition 2006 (E) |

### 1. Applicant & Manufacturer Information

|                         |                                                                   |
|-------------------------|-------------------------------------------------------------------|
| Applicant Name:         | Demo WeighTech Pvt. Ltd.                                          |
| Applicant Address:      | Plot 45, Okhla Industrial Area Phase III, New Delhi 110020, India |
| Manufacturer:           | Demo WeighTech Pvt. Ltd.                                          |
| Manufacturer Address:   | Plot 45, Okhla Industrial Area Phase III, New Delhi 110020, India |
| Country of Manufacture: | India                                                             |

### 2. Technical & Metrological Characteristics

|                     |                  |                     |                     |
|---------------------|------------------|---------------------|---------------------|
| Instrument Model:   | MW-3000          | Accuracy Class:     | Class III           |
| Max Capacity (Max): | 3000.0 kg        | Min Capacity (Min): | 20.0 kg             |
| Scale Interval (e): | 1.0 kg           | Division (d):       | 1.0 kg              |
| Interval Count (n): | 3000             | Serial Number:      | MW3000-2026-IND-001 |
| Load Receptor:      | Platform (4 pts) | Electronic:         | Yes                 |
| Power Supply:       | 230 V AC, 50 Hz  | Temperature Range:  | -10.0°C to 40.0°C   |

### 3. Laboratory Environmental Conditions

|                       |            |                    |            |
|-----------------------|------------|--------------------|------------|
| Ambient Temperature:  | 22.4 °C    | Relative Humidity: | 52.0 %     |
| Atmospheric Pressure: | 1012.8 hPa | Mains Voltage:     | 230.2 V AC |

### 4. Summary of Test Results & Compliance (OIML R 76-1:2006)

| Clause                   | Test Description          | Observed Result                          | Permissible Limit                       | Result      |
|--------------------------|---------------------------|------------------------------------------|-----------------------------------------|-------------|
| Clause A.4.2.3           | Accuracy of Zero-Setting  | E0 = +0.050 kg                           | Limit = $\pm 0.250$ kg (0.25e)          | <b>PASS</b> |
| Clause A.4.4.1 & A.4.4.3 | Weighing Performance Test | 5 points evaluated (Max  Ec  = 0.200 kg) | Tiered MPE as per Table 6 for Class III | <b>PASS</b> |
| Clause A.4.7             | Eccentricity Test         | Max eccentricity error = 0.090 kg        | Limit = $\pm 1.000$ kg                  | <b>PASS</b> |
| Clause A.4.10 & 3.6.1    | Repeatability Test        | $\Delta E$ (Max-Min) = 0.030 kg          | Limit = 1.000 kg                        | <b>PASS</b> |

5. Review & Authorization

|                                                                                          |                                                                                             |                                                                                                                   |
|------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------|
| <div>Evaluated By:</div> <div><div>Test Engineer</div><div>Date: 14-Sep-2026</div></div> | <div>Reviewed By:</div> <div><div>Technical Reviewer</div><div>Status: APPROVED</div></div> | <div>Approved By:</div> <div><div>Director / Lab In-Charge</div><div>Seal &amp; Signature Placeholder</div></div> |
|------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------|

**LEGAL DISCLAIMER / NOTICE:** This document is a prototype evaluation test report generated by METRIREPORT for SIH 2026. It does not constitute an official Legal Metrology Model Approval Certificate under Section 22 of the Legal Metrology Act, 2009 without statutory Gazette notification.