

# GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION  
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION  
NATIONAL LEGAL METROLOGY TESTING CENTRE

## TYPE EVALUATION TEST REPORT

Non-Automatic Weighing Instrument (NAWI) as per OIML R 76-1:2006 (E)

|                                                   |                                               |
|---------------------------------------------------|-----------------------------------------------|
| Report Number: REP-2026-OIML-E52F61               | Date: 15-Sep-2026                             |
| Session ID: TS-20260915-8ADCD7                    | Report Status: <b>READY_FOR_REVIEW (v1.0)</b> |
| Overall Result: <b>MANUAL REVIEW / INCOMPLETE</b> | Standard: OIML R 76-1 Edition 2006 (E)        |

### 1. Applicant & Manufacturer Information

|                         |                                             |
|-------------------------|---------------------------------------------|
| Applicant Name:         | WeighTech Metrology India Ltd.              |
| Applicant Address:      | Plot 42, Electronics City, Bangalore, India |
| Manufacturer:           | WeighTech Metrology India Ltd.              |
| Manufacturer Address:   | Plot 42, Electronics City, Bangalore, India |
| Country of Manufacture: | India                                       |

### 2. Technical & Metrological Characteristics

|                     |                  |                     |                      |
|---------------------|------------------|---------------------|----------------------|
| Instrument Model:   | MW-3000          | Accuracy Class:     | Class III            |
| Max Capacity (Max): | 3000.0 kg        | Min Capacity (Min): | 20.0 kg              |
| Scale Interval (e): | 1.0 kg           | Division (d):       | 1.0 kg               |
| Interval Count (n): | 3000             | Serial Number:      | SN-E2E-TEST-1237C3B3 |
| Load Receptor:      | Platform (4 pts) | Electronic:         | Yes                  |
| Power Supply:       | 230 V AC, 50 Hz  | Temperature Range:  | -10.0°C to 40.0°C    |

### 3. Laboratory Environmental Conditions

|                       |             |                    |            |
|-----------------------|-------------|--------------------|------------|
| Ambient Temperature:  | 22.5 °C     | Relative Humidity: | 55.0 %     |
| Atmospheric Pressure: | 1013.25 hPa | Mains Voltage:     | 230.0 V AC |

### 4. OIML R-76 Test Applicability & Coverage Matrix

| Code    | Clause                 | Test Procedure                          | Applicability | Execution  | Status  |
|---------|------------------------|-----------------------------------------|---------------|------------|---------|
| A.1     | Clause A.1             | Administrative Examination              | APPLICABLE    | NOT TESTED | PENDING |
| A.2     | Clause A.2             | Compare Construction with Documentation | APPLICABLE    | NOT TESTED | PENDING |
| A.3     | Clause A.3             | Initial Examination                     | APPLICABLE    | NOT TESTED | PENDING |
| A.4.2.1 | Clause 4.5.1 / A.4.2.1 | Range of Zero-Setting                   | APPLICABLE    | NOT TESTED | PENDING |
| A.4.2.3 | Clause A.4.2.3         | Accuracy of Zero-Setting                | APPLICABLE    | EXECUTED   | PASS    |

|          |                          |                                                       |            |            |         |
|----------|--------------------------|-------------------------------------------------------|------------|------------|---------|
| A.4.4.1  | Clause A.4.4.1 & A.4.4.3 | Weighing Performance Test                             | APPLICABLE | EXECUTED   | PASS    |
| A.4.7    | Clause A.4.7             | Eccentricity Test                                     | APPLICABLE | EXECUTED   | PASS    |
| A.4.8    | Clause A.4.8             | Discrimination Test                                   | APPLICABLE | NOT TESTED | PENDING |
| A.4.10   | Clause A.4.10            | Repeatability Test                                    | APPLICABLE | EXECUTED   | PASS    |
| A.4.11.1 | Clause A.4.11.1          | Creep Test                                            | APPLICABLE | NOT TESTED | PENDING |
| A.4.11.2 | Clause A.4.11.2          | Zero Return Test                                      | APPLICABLE | NOT TESTED | PENDING |
| A.5.1    | Clause A.5.1             | Tilting Test                                          | APPLICABLE | NOT TESTED | PENDING |
| A.5.3.1  | Clause A.5.3.1           | Static Temperatures Test                              | APPLICABLE | NOT TESTED | PENDING |
| A.5.3.2  | Clause A.5.3.2           | Temperature Effect on No-Load Indication              | APPLICABLE | NOT TESTED | PENDING |
| A.5.4    | Clause A.5.4             | Voltage Variations Test                               | APPLICABLE | NOT TESTED | PENDING |
| A.6      | Clause A.6               | Endurance Test                                        | APPLICABLE | NOT TESTED | PENDING |
| B.2      | Clause B.2               | Damp Heat, Steady State                               | APPLICABLE | NOT TESTED | PENDING |
| B.3      | Clause B.3               | Performance Tests for Disturbances (EMC / ESD / Dips) | APPLICABLE | NOT TESTED | PENDING |
| B.4      | Clause B.4               | Span Stability Test                                   | APPLICABLE | NOT TESTED | PENDING |
| Annex G  | Annex G (Cl. G.1 - G.4)  | Software-Controlled Instrument Examination            | APPLICABLE | NOT TESTED | PENDING |

5. Summary of Executed Test Results & Metrological Compliance

| Clause                   | Test Description          | Observed Result                          | Permissible Limit                       | Result |
|--------------------------|---------------------------|------------------------------------------|-----------------------------------------|--------|
| Clause A.4.2.3           | Accuracy of Zero-Setting  | E0 = +0.050 kg                           | Limit = ±0.250 kg (0.25e)               | PASS   |
| Clause A.4.4.1 & A.4.4.3 | Weighing Performance Test | 5 points evaluated (Max  Ec  = 0.200 kg) | Tiered MPE as per Table 6 for Class III | PASS   |
| Clause A.4.7             | Eccentricity Test         | Max eccentricity error = 0.090 kg        | Limit = ±1.000 kg                       | PASS   |
| Clause A.4.10 & 3.6.1    | Repeatability Test        | ΔE (Max-Min) = 0.030 kg                  | Limit = 1.000 kg                        | PASS   |

6. Review & Authorization

|                                                                                                 |                                                                                                                                               |                                                                                                                        |
|-------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------|
| <div>Evaluated By:</div> <div>_____</div> <div>Test Engineer</div> <div>Date: 15-Sep-2026</div> | <div>Reviewed By:</div> <div>_____</div> <div>Technical Reviewer</div> <div>Status: <b>READY FOR REVIEW</b></div> <div>(Pending Review)</div> | <div>Approved By:</div> <div>_____</div> <div>Director / Lab In-Charge</div> <div>Status: Pending Final Approval</div> |
|-------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------|

LEGAL DISCLAIMER / NOTICE: This document is a prototype evaluation test report generated by METRIREPORT for SIH 2026. It does not constitute an official Legal Metrology Model Approval Certificate under Section 22 of the Legal Metrology Act, 2009 without statutory Gazette notification.