Faisalabad Institute of Cardiology, Faisalabad. Flow meter wall type

	30	pressure 4.2 kg/cm. cm-60psi, Probe (DIN/BSS/JP Type) should be me make of SS material,	ا.
		less steel ball, other pieces in brass, polymide and NBR. Latex free. Litters/min from 0 to 15l/min. input	
		pressure and external tube in polycarbonate, stain	
		For instant flow measuring, un-breakable type,	
	Qty	Specifications	Sr. No.

Medical Superintendent
Faisalabad Institute of Cardiology,

Faisalabad Institute of Cardiology,

