

FAKULTÄT FÜR INFORMATIK


DER TECHNISCHEN UNIVERSITÄT MÜNCHEN

Clinical Project

Simulation of CT metal artefacts in C

Author: Alexander Winkler

Advisors: Dipl.-Inf. Philip Stefan,
Dipl.-Inf. Patrick Wucherer,
Pascal Fallavollita, Ph.D.



Abstract

Blablabla

Contents

Abstract	iii
1. Introduction	1
2. Causes of metal artefacts in CT	3
2.1. Examples of metal artefacts in CT	3
2.2. Fundamentals of X-Ray physics	3
2.2.1. Generation of X-Ray radiation[6]	4
2.2.2. Lambert-Beer's Law[3]	4
2.2.3. Beam hardening	5
2.3. Fundamentals of CT reconstruction	5
3. Simulation of CT	7
3.1. Forward Projection	7
3.1.1. Overview over the forward projection	7
3.1.2. Implementation of line integrals	7
3.1.3. Implementation of beam hardening	7
3.2. Back Projection	7
3.3. Parts of the simulator	7
3.3.1. Segmented CT slice	7
3.3.2. Simulation of X-Ray tube	7
3.3.3. Look up tables for attenuation values	7
3.4. Miscellaneous parts of the simulator	7
3.4.1. Logger	7
3.4.2. Image reading and writing	7
3.5. Future Work	7
3.5.1. More easily implementable artefacts	7
3.5.2. Unresolved bugs	7
4. Results	9
5. Conclusion	11
Appendix	15
A. Real metal artefacts images for reference	15
Bibliography	19

1. Introduction

Almost 100 years ago the Austrian mathematician Johann Radon described the mathematical theory computed tomography is based on: He showed that the knowledge of a function $f(x, y)$ is equivalent to the knowledge of all its line integrals. Thus the function can be reconstructed from an infinite set of its projections.

X-Ray is an imaging modality which produces projection images of a volume which show also the internals of the object. So in a perfect world by taking X-Ray images from all angles around an object orthogonal slices of the object can be reconstructed perfectly: In an idealized situation with high radiation dose, monochromatic X-Rays, infinite detector resolution, perfect detectors, no motion and no scatter CT images would be a perfect reflection of reality.[2]

In reality however these conditions are not met. Thus errors in the reconstruction occur resulting in image artefacts.

Usually especially for commercial CT scanners which are used in medical imaging the reduction of these artefacts is of special interest. But to better understand the causes of these artefacts as well as training of medical doctors to make reasonable diagnoses in the presence of these artefacts simulation is also of interest. In this clinical project we developed a simulator of a CT which produces CT images from segmented patient datasets in which the artefacts usually visible in medical imaging are also simulated in a realistic fashion. Especially the simulation of metal artefacts is of major importance as the resulting image artefacts are a major concern in medical imaging as these artefacts degrade image quality considerably.

2. Causes of metal artefacts in CT

2.1. Examples of metal artefacts in CT

To know why metal artefacts in CT should be simulated, we should first know what these metal artefacts are and how they look like.

The human body usually does not contain any metal objects. However many medical implants are composed of metal as well as many surgical tools. Thus metal can indeed be found quite often in patients who are scanned in CT scanners: Artificial hip bones, dental fillings, pacemakers, intramedullary rods, screws or in the case of interventional CT surgical tools like needles, trocars etc. are found in many patients. If these patients are scanned in CT artefacts in the tomographic reconstruction are usually encountered, how severe these artefacts are depends mainly on their size and the material they are composed of.

These artefacts usually consist of star like streaks emanating from the metal object going through the whole or large portions of the reconstruction as can be seen in figure 2.1. More examples can be found in appendix A.




Figure 2.1.: Bilateral hip replacements (with streaks obscuring perirectal lymphadenopathy from rectal cancer)[7]

2.2. Fundamentals of X-Ray physics

To understand the generation of CT metal artefacts we first need to understand how X-Ray generation and attenuation works in general.

2.2.1. Generation of X-Ray radiation[6]

To generate X-Ray radiation an X-Ray tube is used. It consists of a cathode and a rotating anode (usually made of tungsten) which is located within an evacuated tube housing. When electrons from the cathode hit the anode two physical effects take place which result in the emission of X-Ray radiation: Bremsstrahlung and characteristic X-Rays.

Bremsstrahlung results in a continuous spectrum of EM radiation which is produced by the abrupt deceleration of charged particles ("Bremsstrahlung" is German for "braking radiation"). This deceleration is caused by deflection of electrons in the Coulomb field of the nuclei of the anode material.

Characteristic rays are caused by the removal of inner shell electrons and subsequent filling of the holes with electrons from the outer shells. Shell energy differences determine the energies of the characteristic X-rays (this strongly depends on the anode material).

Both effects together result in the total emission spectrum of the X-Ray source. This spectrum depends on tube voltage and heating current and anode material and anode angle. This spectrum usually looks like as depicted in figure 2.2.




Figure 2.2.: Total X-Ray spectrum of a tungsten anode at an acceleration voltage of $U_a = 80\text{kV}$ [6]

2.2.2. Lambert-Beer's Law[3]

Usually all physical mechanisms that lead to the attenuation of radiation intensity behind a homogeneous object are summed up in a single attenuation coefficient μ . In this simple model the total attenuation of a monochromatic X-Ray beam after passing a distance of $\Delta\eta$ through an object can be calculated as

$$I(\eta + \Delta\eta) = I(\eta) - \mu(\eta)I(\eta)\Delta\eta$$

where I is the radiation intensity - which is proportional to the number of photons.

By reordering and taking the limit $\Delta\eta \rightarrow 0$ this leads to the differential quotient

$$\lim_{\Delta\eta \rightarrow 0} \frac{I(\eta + \Delta\eta) - I(\eta)}{\Delta\eta} = \frac{dI}{d\eta} = -\mu(\eta)I(\eta)$$

If the medium is assumed to be homogeneous the medium can be described by a single attenuation constant $\mu(\eta) = \mu$. By solving this ordinary linear and homogeneous, first-order differential equation with constant coefficients this gives the equation

$$I(\eta) = I_0 e^{-\mu\eta}$$

known as Beer's law of attenuation.

The full derivation of the formula with all assumptions and mathematical intermediate steps can be found in [3].

2.2.3. Beam hardening

non-linear relation between the attenuation values, μ , and the measured values of the projection due to the fact that different bands of the frequency spectrum are differently attenuated soft X-ray beams, are more strongly absorbed than the high-energy, hard X-ray beams. This is the reason why this effect is named hardening of the X-ray spectrum and the corresponding image error is named beam-hardening artefact.[3]

2.3. Fundamentals of CT reconstruction

3. Simulation of CT

par bla[4]

3.1. Forward Projection

3.1.1. Overview over the forward projection

3.1.2. Implementation of line integrals

3.1.3. Implementation of beam hardening

3.2. Back Projection

blabla[1]

3.3. Parts of the simulator

3.3.1. Segmented CT slice

3.3.2. Simulation of X-Ray tube

3.3.3. Look up tables for attenuation values

blabla[5]

3.4. Miscellaneous parts of the simulator

3.4.1. Logger

3.4.2. Image reading and writing

3.5. Future Work

3.5.1. More easily implementable artefacts

Ring artefacts

Poisson noise

Motion artefacts

3.5.2. Unresolved bugs

4. Results

5. Conclusion

Appendix

A. Real metal artefacts images for reference




Figure A.1.: Aneurysm clip (with streaks obscuring an acute hemorrhage)[7]




Figure A.2.: Bilateral hip replacements (with streaks obscuring a fluid collection adjacent to the joint)[7]



Figure A.3.: Aneurysm clip (with streaks obscuring an acute infarct)[7]



Figure A.4.: Cholecystectomy clips[7]




Figure A.5.: Dental fillings[7]




Figure A.6.: Hip replacement[7]




Figure A.7.: Bilateral hip replacements[7]



Figure A.8.: Splenectomy clips[7]



Figure A.9.: Embolization coils[7]




Figure A.10.: Embolization coils[7]

Bibliography

- [1] Adam Arnesen. Cuda ct backprojection reconstruction, February 2011. Available online at http://wiki.arnesenfamily.net/doku.php?id=byu:projects:cuda_ct visited on July 30th 2014.
- [2] F Edward Boas and Dominik Fleischmann. Ct artifacts: causes and reduction techniques. *Imaging in Medicine*, 4(2):229–240, 2012.
- [3] T.M. Buzug. *Computed Tomography: From Photon Statistics to Modern Cone-Beam CT*. Springer, 2008.
- [4] Bruno De Man. Iterative reconstruction for reduction of metal artifacts in computed tomography. *status: published*, 2001.
- [5] S. M. Seltzer J. H. Hubbell. Tables of x-ray mass attenuation coefficients and mass energy-absorption coefficients from 1 kev to 20 mev for elements $z = 1$ to 92 and 48 additional substances of dosimetric interest, May 1996. Available online at <http://www.nist.gov/pml/data/xraycoef/index.cfm> visited on July 30th 2014.
- [6] Markus Kowarschik. Interventional Imaging & Image Processing: Angiographic X-ray imaging, Basics of X-ray imaging (generation, absorption, and detection of X-rays). November 2012.
- [7] ReVision Radiology. Ct metal artifact reduction using the metal deletion technique (mdt)., February 2014. Available online at <http://www.revisionrads.com/index.html> visited on August 8th 2014.