



# Simulation of CT metal artefacts

Alexander Winkler

Advisors: Dipl.-Inf. Philipp Stefan,

Dipl.-Inf. PatrickWucherer, Pascal Fallavollita, Ph.D.


Clinical project final presentation





# **Examples of metal artefacts in CT**

- The human body usually does not contain any metal objects
  - many medical implants are composed of metal
    - · Artificial hip bones
    - · dental fillings
    - Pacemakers
    - intramedullary rods
    - Screws
    - Needles
    - Trocars
    - etc.




artefacts usually consist of dark stripes between metal objects with light,
 pinstriped lines covering the surrounding tissue





### **Generation of X-Ray radiation**

- to generate X-Ray radiation an X-Ray tube is used
- consists of a cathode and a rotating (tungsten) anode
- electrons from the cathode hit the anode
  - Bremsstrahlung
  - Characteristic emission lines
- Both effects together result in the total emission spectrum of the X-Ray source







### Attenuation of X-Ray photons: Lambert-Beer's Law

- Usually all physical mechanisms that lead to the attenuation of radiation intensity behind a homogeneous object are summed up in a single attenuation coefficient µ
- In this simple model the total attenuation of a monochromatic X-Ray beam after passing a distance of through an object can be calculated as

$$I(\eta + \Delta \eta) = I(\eta) - \mu(\eta)I(\eta)\Delta \eta$$

By reordering and taking the limit this leads to

$$\lim_{\eta \to 0} \frac{I(\eta + \Delta \eta) - I(\eta)}{\Delta \eta} = \frac{\Delta I}{\Delta \eta} = -\mu(\eta)I(\eta)$$

 Solving this ordinary linear and homogeneous, firstorder differential equation with constant coefficients this gives


$$I(\eta) = I_0 e^{-\mu\eta}$$





### Beam hardening


- Colloquially, X-ray is known to have the property of effectively penetrating material.
- But radiation attenuation is not only dependent on path length
  - but is a function of the specific, wavelengthdependent interaction between X-ray and the material concerned.
- Lambert-Beer's Law is a simplification
  - it does not take beam hardening into account








### Beam hardening



- Colloquially, X-ray is known to have the property of effectively penetrating material.
- But radiation attenuation is not only dependent on path length
  - but is a function of the specific, wavelengthdependent interaction between X-ray and the material concerned.
- Lambert-Beer's Law is a simplification
  - it does not take beam hardening into account







### Beam hardening

Taking energy dependence into account this leads to

$$I(s) = \int_{0}^{E_{\text{max}}} I_0(E)e^{-\int_{0}^{s} \mu(\eta, E)d\eta} dE$$

- If a beam is completely attenuated this is called photon starvation
- behind a thick metal object the system detects an infinitely high attenuation
- Filtered backprojection does not cope with these inconsistencies
  - lines through the object are encountered with extremely high numerical values, which spread across the entire image and are not compensated for by any other projection direction





#### **Forward Projection**

- Forward projection is essentially a Radon Transform
  - With some special conditionals to simulate physical effects not present in the mathematical model
- For this simulation parallel beam geometry was chosen
  - Equivalent to fan beam geometry
- Simulation of
  - Attenuation according to Lambert-Beer's Law modified to take energy dependence into account
  - X-Ray tube spectra at different voltages
  - Polychromaticy by several monochromatic runs
- Complete forward projection can be formulated as


$$\sum_{p\in P}\sum_{en\in E}\sum_{d\in D}\sum_{l\in L}I_0\cdot e^{-\sum_{\eta\in M}\mu(\eta,x,y,en)}$$
 with  $x=l\cdot\sin(p)+d\cdot\cos(p)$  ,  $y=-l\cdot\cos(p)+d\cdot\sin(p)$ 







## **Segmented CT slice**


















### **Lookup tables for X-Ray spectra**




Sampled from:
J. Anthony Seibert. X-ray imaging physics for nuclear medicine technologists. part 1:
Basic principles of x-ray production. J. Nucl. Med. Technol., 32 no. 3,:139–147, 2004.





### Lookup tables for attenuation coefficients



Tables by:
S. M. Seltzer J. H.
Hubbell. Tables of x-ray
mass attenuation
coefficients and mass
energy-absorption
coefficients from 1 kev to
20 mev for elements z =
1 to 92 and 48
additional substances of
dosimetric interest, May
1996.





### **Configuration file**

```
pathToSlice=slices/Segmentation2
pathToOutputReconstruction=outData.pgm
pathToOutputSinogram=simulatedSinogram.pqm
pathToXRaySpectra=Data/XRaySpectra
minEnergy=20
maxEnergy=140
energyLevels=8
numberOfProjectionAngles=64
numberOfThreads=4
tubeEnergy=140
detectorThreshold=100
attenuationMultiplicator=10
```





#### Miscellaneous parts of the simulator

- Out paradigm was not to rely on any third party libraries
  - So some more code had to be implemented...
- Own logger (inspired by loggers like log4j)
  - Different loglevels: ALL > TRACE > DEBUG > INFO > WARN > ERR > FATAL > OFF
  - Heavily relies on macros to guarantee no major impact on performance
  - Outputs loglevel, source code file, function, line and message. In "verbose mode"
  - Can also write output to file
- Image reading and writing
  - portable graymap format .pgm (Portable Anymap or Netpbm format)
  - One of the easiest image formats
    - Super short header followed by color values in ASCII separated by whitespaces






### Results


- Images take long to compute (5-60 minutes on my laptop)
- CPU speed and number of threads/cores is the limiting factor
- Computation speed is direct proportional to
  - Number of projections
  - Number of energy levels
- Computation speed is indirect proportional to the number of threads
- We think that simulation of the CT would be a good candidate for the calculation on the graphics card
  - All rays from all projections can be run in parallel, no dependencies on each other
  - Current simulator is written in C, a port to CUDA-C should be possible with reasonable amount of work





## **Results**











## **Results**









### **Future Work**

### More easily implementable CT artefacts

- Ring artefacts
  - Caused by miscalibrated detector
  - Multiply/Add intensity value by random numbers at the end of the ray
- Poisson Noise
  - Caused by discrete particle nature of light at low X-Ray doses
  - Modify relative photon count at the start of the ray
- Motion artefacts
  - Caused by the patient moving during the scan
  - Merge sinograms from two or more simulation runs



