



DEPARTMENT OF LABORATORY SERVICES

Patient Name	Mrs. SIKHA GORAI	Bill No	OPCA321/159429
Source	OPD	Sample Date	13/11/2021 5:14PM
UHID	496850	Receiving Date	13/11/2021 6:15PM
Age/Gender	52 Yrs/Female	Report Date	13/11/2021 9:53PM
Bed No/Ward	OPD	Report Status	Final
Referred By	Dr. SELF	Lab No	392115400
		ManualDept No.	SRF
		ID:	1970400404100
Investigation Name	Result	Units	Bio. Ref.Interval

MOLECULAR DIAGNOSTICS

SARS-CoV2 (COVID19) QUALITATIVE REAL TIME PCR (TRUNAT)

SPECIMEN	NASOPHARYNGEAL & OROPHARYNGEAL SWABS
RESULT	Negative

SARS COVID-19 QUALITATIVE REAL TIME PCR is a chip-based Real Time duplex Reverse Transcription Polymerase Chain Reaction (RT PCR) test for the detection of SARS CoV-2 RNA in human oropharyngeal and nasopharyngeal swab specimen in TRUNAT platform. The test detects the E and Orf1a genes of the virus. Optimal performance of this test requires appropriate specimen collection, handling, storage and transport to the test site. Though very rare, mutations within the highly conserved regions of the target genome where the assay primers and/or probe bind may result in the under-quantitation of or a failure to detect the presence of the concerned pathogen. A specimen for which the assay reports "Not Detected" cannot be concluded to be negative for the concerned pathogen. As with any diagnostic test, results from the assay should be interpreted in the context of other clinical and laboratory findings. Limit of detection of Orf1a and E gene was estimated to be 480 and 487 genome copies/ml respectively. Invalid samples have to be repeated with fresh specimen from the sample preparation stage.

ICMR Advisory on Interpretation of Ct Values

-Ct values differ from one kit to the other. Comparability of Ct values among different kits is a challenge as labs are using a mixed basket of kits now with different Ct cut-offs and different gene targets.

- Ct values also depend on how the sample has been collected, type of specimen, transport temperature, technical competence of the person performing the test, calibration of equipment and pipettes and analytical skills of the interpreters amongst others.

- Severity of COVID-19 disease largely depends on host factors besides the viral load. Some patients with low viral load may land up in very severe disease due to triggering of the immunological responses. Hence, again high Ct value may give a false sense of security.

- Moreover, the RT-PCR test presently being conducted is qualitative in nature. Ct values may give a rough estimate of viral load. However, more specialized standards are required for quantitative assays which are currently unavailable for SARS-CoV-2.

In view of the above, it is not recommended to rely on numerical Ct values for determining infectiousness of COVID-19 patients and deciding patient management protocols.

*ICMR Registration number for Covid-19 is TMHDMCWB.



Dr. PALASH DAS
DPH,MD(MICROBIOLOGY)
ASSOCIATE CONSULTANT

The Mission Hospital, Durgapur | 219(P) Immon Kalyan Sarani, Sector-2C, Bidhan Nagar, Durgapur - 713212
P: 0343 2535555 | M: 9233355555 | F: 0343 2532550 | Email: hospital@themissionhospital.in | www.themissionhospital.com

Call: 9800881600 for free Ambulance within City Limits of Durgapur
A Unit of Durgapur Medical Centre Pvt. Ltd. CIN: U85110WB1987PTC042580



DEPARTMENT OF LABORATORY SERVICES

Patient Name Mrs. SIKHA GORAI
Source OPD
UHID 496850
Age/Gender 52 Yrs/Female
Bed No/Ward OPD
Referred By Dr. SELF

Bill No OPCA321/159429
Sample Date 13/11/2021 5:14PM
Receiving Date 13/11/2021 6:15PM
Report Date 13/11/2021 9:53PM
Report Status Final
Lab No 392115400
ManualDept No. SRF
ID:1970400404100

-----**End Of Report**-----




Dr. PALASH DAS
DPH,MD(MICROBIOLOGY)
ASSOCIATE CONSULTANT

The Mission Hospital, Durgapur | 219(P) Immon Kalyan Sarani, Sector-2C, Bidhan Nagar, Durgapur - 713212
P: 0343 2535555 | M: 9233355555 | F: 0343 2532550 | Email: hospital@themissionhospital.in | www.themissionhospital.com

Call: 9800881600 for free Ambulance within City Limits of Durgapur

A Unit of Durgapur Medical Centre Pvt. Ltd. CIN: U85110WB1987PTC042580