mda-load library v1.4.0 manual

Dohn Alexander Arms dohnarms@anl.gov

July 20, 2016

This library is used to load MDA files created by the **saveData** program that is part of **EPICS**. It will load files with MDA versions of 1.2, 1.3, and 1.4, which are the only versions used currently.

1 Compiling

Compiling requires that you specify the **mda-load** library. This can be done directly by specifying the filename libmda-load.a and its path. If the library is in a standard search location for libraries, you can use the "-1 mda-load" option with gcc or ld.

If you are using Solaris, you will also need to specify the Networking Services Library (nsl). This can be done with the "-l nsl" option.

2 Code

An MDA file structure has a defined structure that is mimicked by the data structures this library uses. There first is a header describing the scan(s), followed by the highest dimensional scan, then all the lower dimensional scans, and finally by an optional section of extra PVs.

2.1 Functions

There are functions for loading the entire file or just parts of it. The function and structure definitions are in the include file mda-load.h.

In the following functions, fptr is a FILE pointer corresponding to an open MDA file, normally gotten from fopen() and removed by fclose(). It doesn't matter what position the pointer is at in the file when calling these functions.

```
struct mda_file *mda_load( FILE *fptr)
```

This function loads the entire MDA file at once, returning a pointer to an mda_file structure.

```
void mda_unload( struct mda_file *mda)
```

This will free the memory occupied by the mda_file to which mda points.

```
struct mda_header *mda_header_load( FILE *fptr)
```

This function loads the MDA file header, returning a pointer to a mda_header structure.

```
void mda_header_unload( struct mda_header *header)
```

This will free the memory occupied by the mda_header to which header points.

```
struct scan *mda_scan_load( FILE *fptr)
```

This function loads the entire scan structure (all dimensions) for an MDA file, returning a pointer to a mda_scan structure. It is the same as mda_subscan_load(), where depth is zero and recursive is one.

```
struct mda_scan *mda_subscan_load( FILE *fptr, int depth, int *indices, int recursive)
This function loads only a part of an MDA file's scans, returning a pointer to the resulting mda_scan
structure. The depth refers to how many dimensions down into the scan the subscan is located (it has to be
```

less than the number of dimensions in the file), and to the number of members in indices. The indices array contains the location of the subscan, starting with the highest dimensional index; if depth is zero, indices can be set to NULL. The recursive flag determines whether the lower dimensional scans will also be read (if any exist).

```
void mda_scan_unload( struct mda_scan *scan)
```

This will free the memory occupied by the mda_scan to which scan points.

```
struct mda_extra *mda_extra_load( FILE *fptr)
```

This function attempts to load the extra PVs for an MDA file. If they exist, it returns a pointer to an mda_extra structure, otherwise it returns NULL.

```
void mda_extra_unload( struct mda_extra *extra)
```

This will free the memory occupied by the mda_extra to which extra points. It is safe to call this function if extra is set to NULL.

```
struct mda_file *mda_update( FILE *fptr, struct mda_file *previous_mda)
```

This function is the same as mda_load(), except that it will attempt to use the data from a previous mda_load() or mda_update() to reduce the amount of data that needs to be read. This is intended for use when an MDA file is being read while the scan is in progress, allowing only the newly written part of the file to be read, reducing the CPU load. The previous_data pointer that is passed needs to be either point to a valid struct mda_struct or is NULL. If previous_data is not NULL, its value after the function call is invalid, even if the function returned a NULL due to an error; the data to which it pointed was either taken by the newly returned struct mda_struct or was released from memory, .

2.2 Example

The simplest way to use this library is to simply load the entire MDA file, using the template below. Then there are only two functions needed, mda_load() and mda_unload(), with the rest of the work coming from accessing the MDA data structure.

```
#include "mda-load.h"
...
FILE *fptr;
struct mda_file *mda;
...
if( (fptr = fopen(filename, "r")) == NULL)
        exit(1);
if( (mda = mda_load(fptr)) == NULL)
        exit(1);
fclose(fptr);
...
/* Access the mda structure here. */
...
mda_unload(mda);
...
```

3 Data

Once you have a pointer to the allocated mda_file, you can access its members directly. The definition of the various data structures are in mda-load.h. Assuming that the data pointer is called mda, here's how you access the data.

The code takes advantage of the C99 integer data types in stdint.h, which allows explicitly declaring the size of an integer. Any of the int8_t, int16_t, or int32_t types can simply be used as an int if it makes your code simpler (such as using %i with printf).

3.1 Header

This section contains the global data values for the MDA file. version signifies the MDA format version, normally 1.3. scan_number is the number assigned by saveData to the scan. data_rank show the number of dimensions to the scan (for a 3-D scan, this is 3). The dimensions array (with data_rank elements) contains the number of elements for each dimension of the scan, starting with the highest dimensional scan; for a 4-D scan, dimensions[1] is the number of elements to the 3-D scans. regular signifies whether the dimensions of any of the scans were changed while the overall scan was running. extra_pvs_offset gives, for the section of extra PV's, the offset in bytes from the beginning of the file; if that section does not exist, this will be 0.

3.2 Scans

```
(struct mda_scan *) mda->scan
      (int16_t)
                  scan->scan_rank
      (int32_t)
                  scan->requested_points
      (int32_t)
                  scan->last_point
      (int32_t *) scan->offsets
      (char *)
                  scan->name
      (char *)
                  scan->time
      (int16_t)
                  scan->number_positioners
      (int16_t)
                  scan->number_detectors
                  scan->number_triggers
      (int16_t)
      (struct mda_positioner *) scan->positioners[n] ,
                                [n] = [0] to [scan->number_positioners - 1]
            (int16_t) positioners[n]->number
            (char *) positioners[n]->name
                      positioners[n]->description
            (char *)
            (char *) positioners[n]->step_mode
            (char *) positioners[n]->unit
            (char *)
                      positioners[n]->readback_name
            (char *) positioners[n]->readback_description
            (char *) positioners[n]->readback_unit
      (struct mda_detector * scan->detectors[n] ,
                                [n] = [0] to [scan->number_detectors - 1]
```

```
(int16_t) detectors[n]->number
     (char *) detectors[n]->name
     (char *) detectors[n]->description
     (char *) detectors[n]->unit
(struct mda_trigger *) scan->triggers[n] ,
                          [n] = [0] to [scan->numbers_triggers - 1]
     (int16_t) triggers[n]->number
     (char *) triggers[n]->name
               triggers[n]->command
      (float)
(double *) scan->positioners_data[n] , [n] = [0] to [scan->number_positioners - 1]
      (double) scan->positioners_data[n][m] ,
                          [m] = [0] to [scan->requested_points - 1]
(float *) scan->detectors_data[n] , [n] = [0] to [scan->number_detectors - 1]
     (float) scan->detectors_data[n][m] ,
                          [m] = [0] to [scan->requested_points - 1]
(struct mda_scan **) scan->sub_scans
```

This section includes the scan data. It is also recursive in nature due to it being able to handle arbitrary dimensions.

3.2.1 Structure

The overall structure for multidimensional files is dictated by scan_rank and sub_scans. As long as scan_rank is greater than one, sub_scans will not be NULL and will contain an array of the next lower dimensional scans (it will be NULL if mda_subscan_load() was used to retrieve the scan when the recursive parameter set to zero). For a multidimensional scan, this takes the form of a tree, since each sub-scan can also have its own sub-scans. For a higher dimensional scan, the values for the positioners and detectors apply to all scan with its sub_scans.

Suppose a $5\times8\times20$ scan, where you want to access the (3,7,x) 1-D scan, you would access it as mda->scan->sub_scans[2]->sub_scans[6]. However, if the scan was aborted, mda->scan->sub_scans[2] or mda->scan->sub_scans[2]->sub_scans[6] might be NULL, depending on where the scan was aborted. Using last_point can let you know the last "officially valid" sub-scan is sub_scans[last_point - 1]. The reason that I say "officially valid" is that another scan might exist at sub_scans[last_point], as it was the scan in progress that was aborted; one can use this data, but should take care.

3.2.2 Variables

As described before, scan_rank is the dimensionality of this scan. requested_points is how many points were wanted, while last_point tells how many actually were finished. offsets is an array of requested_points members, showing the distance from the beginning of the MDA file to the subscans; if the value is zero, then that scan does not exist. name is the name of the scanner in EPICS, while time is when this particular scan was started.

number_positioners tells how many positioners are moved as part of this scan. The positioners array, holding number_positioners members, has a description of each positioner and its readback. number is the internal number the scanRecord uses to identify this positioner, while name is what its called, and description describes it. step_mode is how the scan determined what step to use: it can be linear, where the spacing between steps is equal; table, where the step positions are read from an array; or fly, where the step positions are read back during an on-the-fly scan. unit is the associated unit of the positioner. Similarly, for the readback, there is readback_name, readback_description, and readback_unit.

The detector information is very similar to the positioners, as there is a detectors array with number_detectors elements. For each detectors, there is also a number, name, description, and unit.

The trigger information is again similar to the positioners, with a triggers array with number_triggers elements. Each trigger has a number and name associated to it, as well as a command, which is a value sent to name to trigger.

The positioner data values are held in an two dimensional array named positioners_data. Since one can't allocate a two dimensional array directly, it's actually an array of pointers (corresponding to each detector), pointing to arrays of more pointers (corresponding the the data). To access the 8th data point of the 12th detector, one would type (scan->positioners_data[11])[7]; the parentheses are not optional.

The detector data values, in detectors_data, is accessed similarly to the positioner data values.

The sub_scans variable is used for accessing lower dimensional scans (if they exist). It's described in Sec. 3.2.1.

3.3 Extra PV's

This section, which doesn't always exist (signified by extra being NULL), contains extra PV's recorded during the scan. number_pvs is the number of PV's contained, with the PV's being held in an array pvs.

For each PV, there is the name string and description string. type lets you know what kind of data type it is, with the correspondence seen in Table 1. If type isn't EXTRA_STRING, count gives the number of elements to the array and unit string gives the unit for the values. The values themselves are held in an array values.

Table 1. Extra r v data type			
type name	type value	C type	Description
EXTRA_PV_STRING	0	(char *)	zero-terminated string
EXTRA_PV_INT8	32	(int8_t *)	8-bit integer array
EXTRA_PV_INT16	29	(int16_t *)	16-bit integer array
EXTRA_PV_INT32	33	(int32_t *)	32-bit integer array
EXTRA_PV_FLOAT	30	(float *)	floating-point array
EXTRA_PV_DOUBLE	34	(double *)	double-precision floating-point array

Table 1: Extra PV data type

Accessing the values is done by setting pointer values to the correct type, according to type and Table 1. Suppose the third extra PV was of type EXTRA_DOUBLE, and you wanted to access its fifth member, this could be done using ((double *) pvs[2]->values)[4].

4 Basic Information Routines

It is possible to load only the basic information of an MDA file, saving load time and memory. This information includes everything in the header, as well as the detector, positioner, and trigger descriptions for each scan dimension. As this information is taken from the first scan of each dimension, there is an assumption that every other scan is correctly represented by the first scan of its dimensionality. This also means that for a multidimensional file, not all of the file is checked for errors; this can be done with mda_test(), which will cost some additional time.

4.1 Functions

```
int mda_test( FILE *fptr)
```

This function will check a file's integrity by loading the entire file structure (but not keeping all the information in memory to keep memory usage minimal), using the same error checking as mda_load(). A returned value of 0 means that there was no error, while 1 means there was an error.

```
struct mda_fileinfo *mda_info_load( FILE *fptr)
```

This function loads a fileinfo of the MDA file, returning a pointer to an mda_fileinfo structure. It does not check the validity of the entire file, which can be done with mda_test().

```
void mda_info_unload( struct mda_fileinfo *fileinfo)
```

This will free the memory occupied by the mda_fileinfo to which fileinfo points.

4.2 Structure

```
(struct mda_fileinfo *) fileinfo
      (float)
              fileinfo->version
      (int32_t) fileinfo->scan_number
      (int16_t) fileinfo->data_rank
      (int32_t) fileinfo->dimensions[n] , [n] = [0] to [data_rank - 1]
      (int16_t) fileinfo->regular
      (int32_t) fileinfo->last_topdim_point
      (char *) fileinfo->time
      (struct mda_scaninfo *) fileinfo->scaninfos[n] , [n] = [0] to [data_rank - 1]
            (int16_t) scaninfos[n]->scan_rank
            (int32_t) scaninfos[n]->requested_points
            (char *) scaninfos[n]->name
            (int16_t) scaninfos[n]->number_positioners
            (int16_t) scaninfos[n]->number_detectors
            (int16_t) scaninfos[n]->number_triggers
            (struct mda_positioner *) scaninfos[n]->positioners[n] ,
                                      [n] = [0] to [scan->number_positioners - 1]
                  (int16_t) positioners[n]->number
                  (char *) positioners[n]->name
                  (char *) positioners[n]->description
                  (char *) positioners[n]->step_mode
                  (char *) positioners[n]->unit
                  (char *) positioners[n]->readback_name
                  (char *) positioners[n]->readback_description
                  (char *) positioners[n]->readback_unit
            (struct mda_detector * scaninfos[n]->detectors[n] ,
                                      [n] = [0] to [scan->number_detectors - 1]
                  (int16_t) detectors[n]->number
                  (char *) detectors[n]->name
                  (char *) detectors[n]->description
                  (char *) detectors[n]->unit
            (struct mda_trigger *) scaninfos[n]->triggers[n] ,
                                      [n] = [0] to [scan->numbers_triggers - 1]
                  (int16_t) triggers[n]->number
                  (char *) triggers[n]->name
                  (float)
                            triggers[n]->command
```

The structure of the mda_fileinfo structure is nonrecursive, and is thus easier to access. The variables version, scan_number, data_rank, dimensions, and regular are the same as described in Section 3.1. The variable last_topdim_point is the number of completed scans in the highest dimensional scan, while time is the data and time of the start of the overall measurement. The information for each dimension's scans are in the scaninfos array. The variables contained in each entry is a subset of those found in the earlier describes scan, and all are described in in Section 3.2.2.

5 Library Changes

Between version 0.3 and 1.0, the structure names were changed to be prefixed by mda_.

From 1.1 to 1.2, integers use types that denote how many bits they use: char (as an 8-bit integer) \rightarrow int8_t, short \rightarrow int16_t, and long \rightarrow int32_t. The most likely change needed in user code for this is with printf(), where one would use %d instead of %ld, etc. The Extra PV data type names were also renamed, to have them start with EXTRA_PV_ and have the integer length in the relevant names.