


1 The diagram shows a view of an eyepiece graticule being calibrated using a stage micrometer.



Each division on the stage micrometer scale is 0.1 mm.

Which row shows a correct calculation to calibrate each eyepiece graticule unit and shows the appropriate units?

	calculation	units
A	$\frac{0.1}{27} \div 1000$	mm
B	$\frac{0.1}{52} \times 1000$	μm
C	$\frac{0.2}{54} \times 1000$	μm
D	$\frac{0.2}{79} \times 1\,000\,000$	nm