

2 A student calibrated an eyepiece graticule using a stage micrometer.

- Each division of the stage micrometer was 0.01 mm.
- With a $\times 10$ magnification objective lens, 10 eyepiece graticule units matched 10 divisions on the stage micrometer.

The same microscope was used with a $\times 40$, instead of a $\times 10$, magnification objective lens to measure the diameter of an alveolus. The diameter of the alveolus was found to be 96 eyepiece graticule units.

The eyepiece lens was not changed.

What is the best estimate for the diameter of the alveolus?

- A** 0.960 mm **B** 3.84 mm **C** 240 μm **D** 384 μm