

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

June 12, 2026

Patrick Miller
Cadillac, City of
1121 Plett Rd.
Cadillac, MI 49601

RE: Trace Project 26F0686
Client Project NPDES - Av. CN- June 2026

Enclosed are your analytical results. The results of this report relate only to the samples listed in the body of this report.

All reports were examined through Trace's validation process to ensure that requirements for quality and completeness were satisfied. All reported analytical results were obtained in accordance with the methods referenced on the reports. Every practical effort was made to meet the reporting limit specifications for this work, however, some results may have raised reporting limits to correct for percent solids.

For clients that require NELAP Accreditation, Trace certifies that these test results meet all requirements of the NELAP Standard, except for those analytes with a "N" notation. These analytes have not been evaluated by NELAP at Trace's discretion and will not be reported unless requested by client.

If you have questions concerning this report, please contact me at 231.773.5998 or by email at tbrewer@trace-labs.com.

Sincerely,

A handwritten signature in black ink that reads "Timothy W. Brewer".

Tim Brewer
Project Manager
Enclosures



TNI EL V1:2016

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

SAMPLE SUMMARY

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID	Sample ID	Matrix	Collected By	Date Collected	Date Received
26F0686-01	Plant Influent	Aqueous	PM/DG	06/10/26 12:48	06/11/26 10:48
26F0686-02	Plant Effluent	Aqueous	PM/DG	06/10/26 12:42	06/11/26 10:48
26F0686-03	Industrial Manhole	Aqueous	PM/DG	06/10/26 11:51	06/11/26 10:48
26F0686-04	AAR 6th	Aqueous	PM/DG	06/10/26 11:19	06/11/26 10:48
26F0686-05	AAR FAC	Aqueous	PM/DG	06/10/26 10:14	06/11/26 10:48
26F0686-06	AAR RT	Aqueous	PM/DG	06/10/26 10:09	06/11/26 10:48
26F0686-07	Akwel 5th	Aqueous	PM/DG	06/10/26 11:28	06/11/26 10:48
26F0686-08	Akwel	Aqueous	PM/DG	06/10/26 11:31	06/11/26 10:48
26F0686-09	Akwel Cadillac	Aqueous	PM/DG	06/10/26 11:36	06/11/26 10:48
26F0686-10	Arvco	Aqueous	PM/DG	06/10/26 10:52	06/11/26 10:48
26F0686-11	Borg Warner	Aqueous	PM/DG	06/10/26 10:39	06/11/26 10:48
26F0686-12	CCI	Aqueous	PM/DG	06/10/26 11:41	06/11/26 10:48
26F0686-13	CRE	Aqueous	PM/DG	06/10/26 10:56	06/11/26 10:48
26F0686-14	Fiamm	Aqueous	PM/DG	06/10/26 11:10	06/11/26 10:48
26F0686-15	Hutchinson	Aqueous	PM/DG	06/10/26 11:23	06/11/26 10:48
26F0686-16	MI Rubber	Aqueous	PM/DG	06/10/26 11:13	06/11/26 10:48
26F0686-17	Pirahna Hose	Aqueous	PM/DG	06/10/26 11:04	06/11/26 10:48

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-01 Matrix: Aqueous Date Collected: 06/10/26 12:48
Sample ID: Plant Influent Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184245

Cyanide (Available)	*	65 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	---	---------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-02 Matrix: Aqueous Date Collected: 06/10/26 12:42
Sample ID: Plant Effluent Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184245

Cyanide (Available)	*	66 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	---	---------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-03 Matrix: Aqueous Date Collected: 06/10/26 11:51
Sample ID: Industrial Manhole Date Received: 06/11/26 10:48

PARAMETERS		RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
WET CHEMISTRY											
Analysis Method: EPA OIA 1677-09											
Batch: T184245											
Cyanide (Available)	*	160	ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-04 Matrix: Aqueous Date Collected: 06/10/26 11:19
Sample ID: AAR 6th Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs			5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-05 Matrix: Aqueous Date Collected: 06/10/26 10:14
Sample ID: AAR FAC Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	3.0 ug/L	2.0	1	06/11/26	Ijs	06/11/26	Ijs		5.9
---------------------	----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-06 Matrix: Aqueous Date Collected: 06/10/26 10:09
Sample ID: AAR RT Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-07 Matrix: Aqueous Date Collected: 06/10/26 11:28
Sample ID: Akwel 5th Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-08 Matrix: Aqueous Date Collected: 06/10/26 11:31
Sample ID: Akwel Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	4.5 ug/L	2.0	1	06/11/26	Ijs	06/11/26	Ijs	5.9
---------------------	----------	-----	---	----------	-----	----------	-----	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-09 Matrix: Aqueous Date Collected: 06/10/26 11:36
Sample ID: Akwel Cadillac Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	3.6 ug/L	2.0	1	06/11/26	Ijs	06/11/26	Ijs	5.9
---------------------	----------	-----	---	----------	-----	----------	-----	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-10 Matrix: Aqueous Date Collected: 06/10/26 10:52
Sample ID: Arvco Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs	5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-11 Matrix: Aqueous Date Collected: 06/10/26 10:39
Sample ID: Borg Warner Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-12 Matrix: Aqueous Date Collected: 06/10/26 11:41
Sample ID: CCI Date Received: 06/11/26 10:48

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	*	670 ug/L	50	25	06/11/26	Ijs	06/11/26	Ijs		5.9
---------------------	---	----------	----	----	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-13 Matrix: Aqueous Date Collected: 06/10/26 10:56
Sample ID: CRE Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184249

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-14 Matrix: Aqueous Date Collected: 06/10/26 11:10
Sample ID: Fiamm Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184250

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-15 Matrix: Aqueous Date Collected: 06/10/26 11:23
Sample ID: Hutchinson Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184250

Cyanide (Available)	3.7 ug/L	2.0	1	06/11/26	Ijs	06/11/26	Ijs		5.9
---------------------	----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-16 Matrix: Aqueous Date Collected: 06/10/26 11:13
Sample ID: MI Rubber Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184250

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26F0686
Client Project ID: NPDES - Av. CN- June 2026

Trace ID: 26F0686-17 Matrix: Aqueous Date Collected: 06/10/26 11:04
Sample ID: Pirahna Hose Date Received: 06/11/26 10:48

PARAMETERS	RESULTS UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T184250

Cyanide (Available)	<2.0 ug/L	2.0	1	06/11/26	ljs	06/11/26	ljs		5.9
---------------------	-----------	-----	---	----------	-----	----------	-----	--	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

QUALITY CONTROL RESULTS

Trace Project ID: 26F0686

Client Project ID: NPDES - Av. CN- June 2026

QC Batch: T184245

Analysis Description: Cyanide, Available

QC Batch Method: EPA OIA 1677-09

Analysis Method: EPA OIA 1677-09

METHOD BLANK: T184245-BLK1

Parameter	Units	Blank Result	Reporting Limit	Notes
Cyanide (Available)	ug/L	<2.0	2.0	

LABORATORY CONTROL SAMPLE: T184245-BS1

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limit	Notes
Cyanide (Available)	ug/L	100.0	100	100	86-118	

Trace Project ID: 26F0686

Client Project ID: NPDES - Av. CN- June 2026

QC Batch: T184249

Analysis Description: Cyanide, Available

QC Batch Method: EPA OIA 1677-09

Analysis Method: EPA OIA 1677-09

METHOD BLANK: T184249-BLK1

Parameter	Units	Blank Result	Reporting Limit	Notes
Cyanide (Available)	ug/L	<2.0	2.0	

LABORATORY CONTROL SAMPLE: T184249-BS1

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limit	Notes
Cyanide (Available)	ug/L	100.0	100	100	86-118	

MATRIX SPIKE / MATRIX SPIKE DUPLICATE: T184249-MSD1

Original: 26F0686-04

Parameter	Units	Original Result	Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limit	RPD	Max RPD	Notes
Cyanide (Available)	ug/L	0	100.0	97.0	97.3	97	97	82-130	0.3	11	

Trace Project ID: 26F0686

Client Project ID: NPDES - Av. CN- June 2026

QC Batch: T184250

Analysis Description: Cyanide, Available

QC Batch Method: EPA OIA 1677-09

Analysis Method: EPA OIA 1677-09

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

METHOD BLANK: T184250-BLK1

Parameter	Units	Blank Result	Reporting Limit	Notes
Cyanide (Available)	ug/L	<2.0	2.0	

LABORATORY CONTROL SAMPLE: T184250-BS1

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limit	Notes
Cyanide (Available)	ug/L	100.0	100	100	86-118	

MATRIX SPIKE / MATRIX SPIKE DUPLICATE: T184250-MSD1

Original: **26F0686-14**

Parameter	Units	Original Result	Spike Conc.	MS Result	MSD Result	MS % Rec	MSD % Rec	% Rec Limit	RPD	Max RPD	Notes
Cyanide (Available)	ug/L	0	100.0	97.7	97.5	98	98	82-130	0.2	11	

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

AN EXPLANATION OF TERMS AND SYMBOLS WHICH MAY OCCUR IN THIS REPORT

DEFINITIONS

LCS	Laboratory Control Sample
LCSD	Laboratory Control Sample Duplicate
MS	Matrix Spike
MSD	Matrix Spike Duplicate
RPD	Relative Percent Difference
DUP	Matrix Duplicate
RDL	Reporting Detection Limit
MCL	Maximum Contamination Limit
TIC	Tentatively Identified Compound
<, ND or U	Indicates the compound was analyzed for but not detected
*	Indicates a result that exceeds its associated MCL or Surrogate control limits
N	Indicates that the laboratory is not accredited by NELAP for this compound
NA	Indicates that the compound is not available.

NOTE: Samples for volatiles that have been extracted with a water miscible solvent were corrected for the total volume of the solvent/water mixture.
Solid matrices Method Blanks are at 100% solids as such results are the same wet or dry.

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com



Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673

Phone 231.773.5998
Fax 888.979.4469
www.trace-labs.com

CHAIN-OF-CUSTODY RECORD

Page 1 of 2

Report Results To:

Bill To:

Company Name: City of Cadillac Utilities WWTP	PO #:
Report To: Cindy Tomaszewski	Contact Name: Cindy Tomaszewski
Mailing Address: 1121 Platt Rd (for shipping only; UPS, FedEx)	Billing Address (if different): 200 N Lake St (for USPS only)
City, State, Zip Code: Cadillac MI 49601	City, State, Zip Code: Cadillac MI 49601
Office Phone: 231 775 2368	Cell Phone: 231 878 6230
Email Address: lab@cadillac.mi.net, ctomaszewski@cadillac.mi.net	Phone Number: 231 775 0181
	Billing Email Address: ap@cadillac.mi.net

Turnaround Requirements:

Matrix Key:

☒ Standard 5-10 Days
☐ 3 Day*
☐ 1 Day*
S = Soil / Solid
W = Water
SL = Sludge
OI = Oil
WI = Wipes
LW = Liquid Waste
A = Air
D = Drinking Water

*Results provided end of business day, requires prior approval.

Project Name: NPDES		Sampled By: PH / D4	
Trace No.	Date Collected	Time Collected	Client Sample ID
1	10-26-13	1243	Plant Influent
2	10-26-13	1242	Plant Effluent
3	11-51		Industrial Manhole
4	11-19		AAR 6th
5	10-14		AAR FAC
6	10-09		AAR RT
7	11-28		Akwel 5th
8	11-31		Akwel
9	11-31		Akwel Cadillac
10	10-26		Arvco
Released By: T. Jeyko		Received By: UPS GROUND	Date: 10-10-26 / 1500
3)		4)	

Metals Field Filtered (Y / N)	Matrix	Number of Containers	Preservation	Available CN	Analysis Requested	Remarks	Possible Health Hazards?
			Cool				
			HCl				
			HNO ₃				
			H ₂ SO ₄				
			NaOH				
			Other				

☐ Check this box if you would not like your samples analyzed if received outside of the conditions outlined in the Trace Sample Acceptance Policy at www.trace-labs.com/downloads.

Form 70-Z-1

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace ID No. 26F0686

Page 2 of 2

Trace Use:

Logged By:	CN
Checked By:	NE
Soil Volatiles Preserved (circle if applicable):	
MeOH	Low Level
	Lab
Sampling Time:	

$|O| = |O_I|$ D = Drinking Water

Project Name: NPDES			Sampled By: PH / DB		Available CN										Possible Health Hazard	
Trace No.	Date Collected	Time Collected	Client Sample ID	Metals Field Filtered (Y / N)	Matrix	Number of Containers	Preservation				Remarks	Grab				
							Cool	HCl	HNO ₃	H ₂ SO ₄						
							NaOH	Other								
11	6-10-26	1039	Borg Warner		W	1				X						
12		1141	CCI		W	1				X						
13		1058	CRE		W	1				X						
14		1110	Flamm		W	1				X						
15		1123	Hutchinson		W	1				X						
16		1113	IM Rubber		W	1				X						
17		1104	Piranha Hose		W	1				X						
					W	1				X						
					W	1				X						
					W	1				X						
Please Sign			Released By	Received By	Date	Time	Released By				Received By	Date	Time			
			T. Dogko	UPS Ground	6-10-26	1300	2) UPS				6/4/26	6/11/26	10:00			
			3)				4)									

In executing this Chain of Custody, the client acknowledges the terms as set forth at www.trace-labs.com/terms-of-agreement.

Page 24 of 25

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

Sample Log In Checklist

26F0686

Cadillac, City of

Project Manager: Tim Brewer

Date: 6/11/26	Original Observation	Corrected Temperature	IR-9 (CF: +0.4°C)	IR-12 (CF: +0.1°C)	IR-13 (CF: -1.0°C)	IR-14 (CF: -0.6°C)	SR1 (CF: -0.3°C)	SR2 (CF: -0.2°C)	Temp Blank	Client Sample
Time: 10:00										
Initials: CN										
Package Description: Cooler										
Package Temp °C	1.9	2.0	x							
Representative Sample Temp °C	6.1	6.2	x							x

Sample Receipt

Yes No

☒ ☐ Received on ice or other coolant

☒ ☐ Ice still present upon receipt

☐ ☒ Custody seals present

☐ Trace Courier ☐ Client Drop-off

☐ Yes ☐ No Custody seals intact (if applicable)

☒ UPS ☐ Fed Ex ☐ US Mail ☐ Other

Sample Condition

Yes No N/A

☒ ☐ All sample containers arrived unbroken and labeled

☒ ☐ Sufficient sample to run requested analyses

☒ ☐ Correct chemical preservative added to samples

☐ ☐ ☒ Samples preserved at Trace (if yes, fill in preservation information below)

Preservation - Reagent: _____ LIMS ID: _____ Date/Time: _____

☐ ☐ ☒ Chemical preservation verified, check EMD pH test strip used (if applicable)

☐ pH 0-2.5 (Lot: HC461264) ☐ pH 11.0-13.0 (Lot: HC022540) ☐ Other

☐ ☐ ☒ Air bubbles absent from VOAs (no bubble greater than 6mm) _____

Chain of Custody (COC)

Yes No

☒ ☐ All bottle labels agree with COC

☒ ☐ COC filled out properly

☒ ☐ COC signed by client

Notes:

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.