



Analytical Laboratory Report

Report ID: S67156.01(01)
Generated on 10/29/2024

Report to
Attention: Cindy Tomaszewski
City of Cadillac - WWTP
200 N Lake Street
Cadillac, MI 49601

Phone: 231-775-2368 FAX:
Email: lab@cadillac-mi.net

Report produced by
Merit Laboratories, Inc.
2680 East Lansing Drive
East Lansing, MI 48823

Phone: (517) 332-0167 FAX: (517) 332-6333

Contacts for report questions:
John Lavery (johnlavery@meritlabs.com)
Barbara Ball (bball@meritlabs.com)

Report Summary
Lab Sample ID(s): S67156.01
Project: Qtrly Biosolids
Collected Date(s): 10/08/2024
Submitted Date/Time: 10/09/2024 10:30
Sampled by: Taysha Laczko
P.O. #:

Table of Contents

- Cover Page (Page 1)
- General Report Notes (Page 2)
- Report Narrative (Page 2)
- Laboratory Accreditations (Page 3)
- Qualifier Descriptions (Page 3)
- Glossary of Abbreviations (Page 3)
- Method Summary (Page 4)
- Parameter Summary (Page 5)
- Sample Summary (Page 6)


Maya Murshak
Technical Director



Analytical Laboratory Report

General Report Notes

Analytical results relate only to the samples tested, in the condition received by the laboratory.

Methods may be modified for improved performance.

Results reported on a dry weight basis where applicable.

'Not detected' indicates that parameter was not found at a level equal to or greater than the reporting limit (RL).

When MDL results are provided, then 'Not detected' indicates that parameter was not found at a level equal to or greater than the MDL.

40 CFR Part 136 Table II Required Containers, Preservation Techniques and Holding Times for the Clean Water Act specify that samples for acrolein and acrylonitrile, and 2-chloroethylvinyl ether need to be preserved at a pH in the range of 4 to 5 or if not preserved, analyzed within 3 days of sampling.

QA/QC corresponding to this analytical report is a separate document with the same Merit ID reference and is available upon request.

Starred (*) analytes are not NY NELAP accredited.

Samples are held by the lab for 30 days from the final report date unless a written request to hold longer is provided by the client.

Report shall not be reproduced except in full, without the written approval of Merit Laboratories, Inc.

Limits for drinking water samples, are listed as the MCL Limits (Maximum Contaminant Level Concentrations)

PFAS requirement: Section 9.3.8 of U.S. EPA Method 537.1 states "If the method analyte(s) found in the Field Sample is present in the FRB at a concentration greater than 1/3 the MRL, then all samples collected with that FRB are invalid and must be recollected and reanalyzed."

Samples submitted without an accompanying FRB may not be acceptable for compliance purposes.

Wisconsin PFAs analysis: MDL = LOD; RL = LOQ. LOD and LOQ are adjusted for dilution.

All accreditations/certifications held by this laboratory are listed on page 3. Not all accreditations/certifications are applicable to this report.

For a specific list of accredited analytes, please feel free to contact the laboratory or visit <https://www.meritlabs.com/certifications>.

Report Narrative

There is no additional narrative for this analytical report

Laboratory Accreditations (For Reference Only)

Authority	Accreditation ID
Michigan DEQ	#9956
DOD ELAP & ISO/IEC 17025:2017	#69699 PJLA Testing
WBENC	#2005110032
Ohio VAP	#CL0002
Indiana DOH	#C-MI-07
New York NELAC	#11814
North Carolina DENR	#680
North Carolina DOH	#26702
Pennsylvania DEP	#68-05884
Wisconsin DNR	FID# 399147320

Qualifier Descriptions

Qualifier	Description
!	Result is outside of stated limit criteria
B	Compound also found in associated method blank
E	Concentration exceeds calibration range
F	Analysis run outside of holding time
G	Estimated result due to extraction run outside of holding time
H	Sample submitted and run outside of holding time
I	Matrix interference with internal standard
J	Estimated value less than reporting limit, but greater than MDL
L	Elevated reporting limit due to low sample amount
M	Result reported to MDL not RDL
O	Analysis performed by outside laboratory. See attached report.
R	Preliminary result
S	Surrogate recovery outside of control limits
T	No correction for total solids
X	Elevated reporting limit due to matrix interference
Y	Elevated reporting limit due to high target concentration
b	Value detected less than reporting limit, but greater than MDL
e	Reported value estimated due to interference
j	Analyte also found in associated method blank
o	Associated EIS outside of control limits
p	Benzo(b)Fluoranthene and Benzo(k)Fluoranthene integrated as one peak.
q	Qualifier ion ratio outside of control limits
x	Preserved from bulk sample

Glossary of Abbreviations

Abbreviation	Description
RL/RDL	Reporting Limit
MDL	Method Detection Limit
MS	Matrix Spike
MSD	Matrix Spike Duplicate
SW	EPA SW 846 (Soil and Wastewater) Methods
E	EPA Methods
SM	Standard Methods
LN	Linear
BR	Branched



Analytical Laboratory Report

Method Summary

Method	Version
ASTM D7968-17M	ASTM Method D7968 - 17 Modified (Isotopic Dilution)
SM2540B	Standard Method 2540 B 2020

Parameter Summary

Parameter	Synonym	Cas #
PFBA	Perfluorobutanoic Acid	375-22-4
PFPeA	Perfluoropentanoic Acid	2706-90-3
4:2 FTSA	4:2 Fluorotelomer Sulfonic Acid	757124-72-4
PFHxA	Perfluorohexanoic Acid	307-24-4
PFBS	Perfluorobutane sulfonic Acid	375-73-5
PFHpA	Perfluoroheptanoic Acid	375-85-9
PFPeS	Perfluoropentane Sulfonic Acid	2706-91-4
6:2 FTSA	6:2 Fluorotelomer Sulfonic Acid	27619-97-2
PFOA	Perfluorooctanoic Acid	335-67-1
PFHxS	Perfluorohexane Sulfonic Acid	355-46-4
PFHxS-LN	Perfluorohexane Sulfonic Acid - LN	355-46-4-LN
PFHxS-BR	Perfluorohexane Sulfonic Acid - BR	355-46-4-BR
PFNA	Perfluorononanoic Acid	375-95-1
8:2 FTSA	8:2 Fluorotelomer Sulfonic Acid	39108-34-4
PFHpS	Perfluoroheptane Sulfonic Acid	375-92-8
PFDA	Perfluorodecanoic Acid	335-76-2
N-MeFOSAA	N-methyl perfluorooctanesulfonamidoacetic acid	2355-31-9
EtFOSAA	N-Ethyl Perfluorooctane Sulfonamidoacetic Acid	2991-50-6
PFOS	Perfluorooctane Sulfonic Acid	1763-23-1
PFOS-LN	Perfluorooctane Sulfonic Acid - LN	1763-23-1-LN
PFOS-BR	Perfluorooctane Sulfonic Acid - BR	1763-23-1-BR
PFUnDA	Perfluoroundecanoic Acid	2058-94-8
PFNS	Perfluorononane Sulfonic Acid	68259-12-1
PFDoDA	Perfluorododecanoic Acid	307-55-1
PFDS	Perfluorodecane Sulfonic Acid	335-77-3
PFTTrDA	Perfluorotridecanoic Acid	72629-94-8
FOSA	Perfluorooctane Sulfonamide	754-91-6
PFTeDA	Perfluorotetradecanoic Acid	376-06-7
11Cl-PF3OUdS	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	763051-92-9
9Cl-PF3ONS	9-chlorohexadecafluoro-3-oxanone1-sulfonic acid	756426-58-1
ADONA	4,8-dioxa-3H-perfluorononanoic acid	919005-14-4
HFPO-DA	Hexafluoropropylene oxide dimer	13252-13-6
FHpPA (7:3 FTCA)	3-Perfluoroheptyl propanoic acid	812-70-4
FPePA (5:3 FTCA)	3-Perfluoropentyl propanoic acid	914637-49-3
FPrPA (3:3 FTCA)	3-Perfluoropropyl propanoic acid	356-02-5
PFBSA	Perfluorobutanesulfonamide	30334-69-1
PFECHS	Perfluoro-4-ethylcyclohexanesulfonate	67584-42-3
PFHxSA	Perfluorohexanesulfonamide	41997-13-1



Analytical Laboratory Report

Sample Summary (1 samples)

Sample ID	Sample Tag	Matrix	Collected Date/Time
S67156.01	BioSlids - DIG #3	Sludge	10/08/24 11:02



Analytical Laboratory Report

Lab Sample ID: S67156.01
Sample Tag: BioSlids - DIG #3
Collected Date/Time: 10/08/2024 11:02
Matrix: Sludge
COC Reference: 175370

Sample Containers

#	Type	Preservative(s)	Refrigerated?	Arrival Temp. (C)	Thermometer #
1	15mL Centrifuge Tube	None	Yes	6.0	IR
1	125mL Plastic	None	Yes	6.0	IR

Extraction / Prep.

Parameter	Result	Method	Run Date	Analyst	Flags
Initial wt. (g) / Final wt. (g) / Volume (ml)*	7.78/6.48/10	ASTM D7968-17M	10/14/24 11:30	CED	

Inorganics

Method: SM2540B, Run Date: 10/16/24 13:34, Analyst: MAM

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
Total Solids*	4.3	1		%	1		

Organics

34 PFAs, Method: ASTM D7968-17M, Run Date: 10/15/24 01:28, Analyst: CED

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
PFBA*	Not detected	3.6		ug/kg	179	375-22-4	
PFPeA*	Not detected	1.8		ug/kg	179	2706-90-3	
4:2 FTSA*	Not detected	1.8		ug/kg	179	757124-72-4	I
PFHxA*	5.4	1.8		ug/kg	179	307-24-4	
PFBS*	Not detected	1.8		ug/kg	179	375-73-5	
PFHpA*	Not detected	1.8		ug/kg	179	375-85-9	
PFPeS*	Not detected	1.8		ug/kg	179	2706-91-4	
6:2 FTSA*	Not detected	1.8		ug/kg	179	27619-97-2	I
PFOA*	Not detected	1.8		ug/kg	179	335-67-1	
PFHxS*	Not detected	1.8		ug/kg	179	355-46-4	
PFHxS-LN*	Not detected	1.8		ug/kg	179	355-46-4-LN	
PFHxS-BR*	Not detected	1.8		ug/kg	179	355-46-4-BR	
PFNA*	Not detected	1.8		ug/kg	179	375-95-1	
8:2 FTSA*	Not detected	1.8		ug/kg	179	39108-34-4	I
PFHpS*	Not detected	1.8		ug/kg	179	375-92-8	
PFDA*	2.1	1.8		ug/kg	179	335-76-2	
N-MeFOSAA*	29	1.8		ug/kg	179	2355-31-9	
EtFOSAA*	7	1.8		ug/kg	179	2991-50-6	
PFOS*	5.1	1.8		ug/kg	179	1763-23-1	
PFOS-LN*	4.2	1.8		ug/kg	179	1763-23-1-LN	
PFOS-BR*	Not detected	1.8		ug/kg	179	1763-23-1-BR	
PFUnDA*	Not detected	1.8		ug/kg	179	2058-94-8	I
PFNS*	Not detected	1.8		ug/kg	179	68259-12-1	
PFDODA*	Not detected	1.8		ug/kg	179	307-55-1	I
PFDS*	Not detected	1.8		ug/kg	179	335-77-3	
PFTDA*	Not detected	1.8		ug/kg	179	72629-94-8	I
FOSA*	2.3	1.8		ug/kg	179	754-91-6	
PFTeDA*	Not detected	1.8		ug/kg	179	376-06-7	
11Cl-PF3OUdS*	Not detected	1.8		ug/kg	179	763051-92-9	
9Cl-PF3ONS*	Not detected	1.8		ug/kg	179	756426-58-1	

I-Matrix interference with internal standard



Analytical Laboratory Report

Lab Sample ID: S67156.01 (continued)

Sample Tag: BioSlids - DIG #3

34 PFAs, Method: ASTM D7968-17M, Run Date: 10/15/24 01:28, Analyst: CED (continued)

Parameter	Result	RL	MDL	Units	Dilution	CAS#	Flags
ADONA*	Not detected	1.8		ug/kg	179	919005-14-4	
HFPO-DA*	Not detected	1.8		ug/kg	179	13252-13-6	
FHpPA (7:3 FTCA)*	82	1.8		ug/kg	179	812-70-4	
FPePA (5:3 FTCA)*	340	1.8		ug/kg	179	914637-49-3	
FPrPA (3:3 FTCA)*	Not detected	1.8		ug/kg	179	356-02-5	
PFBSA*	Not detected	1.8		ug/kg	179	30334-69-1	
PFECHS*	Not detected	1.8		ug/kg	179	67584-42-3	
PFHxSA*	Not detected	1.8		ug/kg	179	41997-13-1	

Merit Laboratories Login Checklist

Lab Set ID:S67156

Client:CADILLAC (City of Cadillac - WWTP)

Project: Qtrly Biosolids

Submitted: 10/09/2024 10:30 Login User: MMC

Attention: Cindy Tomaszewski

Address: City of Cadillac - WWTP
200 N Lake Street
Cadillac, MI 49601

Phone: 231-775-2368

FAX:

Email: lab@cadillac-mi.net

Selection	Description	Note
Sample Receiving		
01.	☒ Yes ☐ No ☐ N/A	Samples are received at 4C +/- 2C Thermometer # IR 6.0
02.	☒ Yes ☐ No ☐ N/A	Received on ice/ cooling process begun
03.	☒ Yes ☐ No ☐ N/A	Samples shipped UPS
04.	☐ Yes ☒ No ☐ N/A	Samples left in 24 hr. drop box
05.	☒ Yes ☐ No ☐ N/A	Are there custody seals/tape or is the drop box locked
Chain of Custody		
06.	☒ Yes ☐ No ☐ N/A	COC adequately filled out
07.	☒ Yes ☐ No ☐ N/A	COC signed and relinquished to the lab
08.	☒ Yes ☐ No ☐ N/A	Sample tag on bottles match COC
09.	☐ Yes ☒ No ☐ N/A	Subcontracting needed? Subcontacted to:
Preservation		
10.	☒ Yes ☐ No ☐ N/A	Do sample have correct chemical preservation
11.	☐ Yes ☐ No ☒ N/A	Completed pH checks on preserved samples? (no VOAs)
12.	☐ Yes ☒ No ☐ N/A	Did any samples need to be preserved in the lab?
Bottle Conditions		
13.	☒ Yes ☐ No ☐ N/A	All bottles intact
14.	☒ Yes ☐ No ☐ N/A	Appropriate analytical bottles are used
15.	☒ Yes ☐ No ☐ N/A	Merit bottles used
16.	☒ Yes ☐ No ☐ N/A	Sufficient sample volume received
17.	☐ Yes ☒ No ☐ N/A	Samples require laboratory filtration
18.	☒ Yes ☐ No ☐ N/A	Samples submitted within holding time
19.	☐ Yes ☐ No ☒ N/A	Do water VOC, TOX, DO or Alkalinity bottles contain

Corrective action for all exceptions is to call the client and to notify the project manager.

Client Review By: _____ Date: _____

