

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

July 02, 2026

Cindy Tomaszewski
Cadillac, City of
200 North Lake Street
Cadillac, MI 49601

RE: Trace Project 26G0072
Client Project NPDES Monthly Requirements- June 2026

Enclosed are your analytical results. The results of this report relate only to the samples listed in the body of this report.

All reports were examined through Trace's validation process to ensure that requirements for quality and completeness were satisfied. All reported analytical results were obtained in accordance with the methods referenced on the reports. Every practical effort was made to meet the reporting limit specifications for this work, however, some results may have raised reporting limits to correct for percent solids.

For clients that require NELAP Accreditation, Trace certifies that these test results meet all requirements of the NELAP Standard, except for those analytes with a "N" notation. These analytes have not been evaluated by NELAP at Trace's discretion and will not be reported unless requested by client.

If you have questions concerning this report, please contact me at 231.773.5998 or by email at tbrewer@trace-labs.com.

Sincerely,

A handwritten signature in black ink that reads "Timothy W. Brewer".

Tim Brewer
Project Manager
Enclosures



TNI EL V1:2016

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

SAMPLE SUMMARY

Trace Project ID: 26G0072
Client Project ID: NPDES Monthly Requirements- June 2026

Trace ID	Sample ID	Matrix	Collected By	Date Collected	Date Received
26G0072-01	CCI	Aqueous	DG/PM	06/30/26 11:25	07/01/26 12:56

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

ANALYTICAL RESULTS

Trace Project ID: 26G0072
Client Project ID: NPDES Monthly Requirements- June 2026

Trace ID: 26G0072-01 Matrix: Aqueous Date Collected: 06/30/26 11:25
Sample ID: CCI Date Received: 07/01/26 12:56

PARAMETERS	RESULTS	UNITS	RDL	DILUTION	PREPARED	BY	ANALYZED	BY	NOTES	MCL
------------	---------	-------	-----	----------	----------	----	----------	----	-------	-----

WET CHEMISTRY

Analysis Method: EPA OIA 1677-09

Batch: T185120

Cyanide (Available)	2.4 ug/L	2.0	1	07/01/26	Ijs	07/01/26	Ijs	5.9
---------------------	----------	-----	---	----------	-----	----------	-----	-----

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Trace Analytical Laboratories, Inc.
2241 Black Creek Road
Muskegon, MI 49444-2673



231-773-5998 Phone
888-979-4469 Fax
www.trace-labs.com

QUALITY CONTROL RESULTS

Trace Project ID: 26G0072

Client Project ID: NPDES Monthly Requirements- June 2026

QC Batch: T185120

Analysis Description: Cyanide, Available

QC Batch Method: EPA OIA 1677-09

Analysis Method: EPA OIA 1677-09

METHOD BLANK: T185120-BLK1

Parameter	Units	Blank Result	Reporting Limit	Notes
Cyanide (Available)	ug/L	<2.0	2.0	

LABORATORY CONTROL SAMPLE: T185120-BS1

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limit	Notes
Cyanide (Available)	ug/L	100.0	97.8	98	86-118	

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

AN EXPLANATION OF TERMS AND SYMBOLS WHICH MAY OCCUR IN THIS REPORT

DEFINITIONS

LCS	Laboratory Control Sample
LCSD	Laboratory Control Sample Duplicate
MS	Matrix Spike
MSD	Matrix Spike Duplicate
RPD	Relative Percent Difference
DUP	Matrix Duplicate
RDL	Reporting Detection Limit
MCL	Maximum Contamination Limit
TIC	Tentatively Identified Compound
<, ND or U	Indicates the compound was analyzed for but not detected
*	Indicates a result that exceeds its associated MCL or Surrogate control limits
N	Indicates that the laboratory is not accredited by NELAP for this compound
NA	Indicates that the compound is not available.

NOTE: Samples for volatiles that have been extracted with a water miscible solvent were corrected for the total volume of the solvent/water mixture.
Solid matrices Method Blanks are at 100% solids as such results are the same wet or dry.

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.

Phone 231.773.5998
FAX 888.979.4469
www.trace-labs.com

Page 1 of 1

Trace ID No.

260072

Trace Use:

Logged By:	CN
Checked By:	LJS
Soil Volatiles Preserved (circle if applicable):	
MeOH	Low Level
	Lab
Sampling Time:	

Matrix Key:

- ☒ Standard, 5-10 Days S = Soil / Solid WI = Wipes
- ☐ 3 Day^a W = Water LW = Liquid Waste
- ☐ 1 Day^a SL = Sludge A = Air
- OI = Oil D = Drinking Water
- ^aResults provided end of business day; requires prior approval.

*Results provided end of business day, requires prior approval

[illegible]

☐ Check this box if you would not like your samples analyzed if received outside of the conditions outlined in the Trace Sample Acceptance Policy at www.trace-labs.com/downloads

Form 70-Z.1

Page 6 of 7

26G0072

Cadillac, City of
Project Manager: Tim Brewer

Sample Log In Checklist

Date: 7/1/26	Original Observation	Corrected Temperature	IR-9 (CF: +0.4°C)	IR-12 (CF: +0.1°C)	IR-13 (CF: -1.0°C)	IR-14 (CF: -0.6°C)	SR1 (CF: -0.3°C)	SR2 (CF: -0.2°C)	Temp Blank	Client Sample
Time: 12:56										
Initials: CN										
Package Description: Cooler										
Package Temp °C	3.4	3.5	x							
Representative Sample Temp °C	6.1	6.2	x							x

Sample Receipt

Yes No

☒ ☐ Received on ice or other coolant

☐ ☒ Ice still present upon receipt

☐ ☒ Custody seals present

☐ Trace Courier ☐ Client Drop-off

☐ Yes ☐ No Custody seals intact (if applicable)

☒ UPS ☐ Fed Ex ☐ US Mail ☐ Other

Sample Condition

Yes No N/A

☒ ☐ ☐ All sample containers arrived unbroken and labeled

☒ ☐ ☐ Sufficient sample to run requested analyses

☒ ☐ ☐ Correct chemical preservative added to samples

☐ ☒ ☐ Samples preserved at Trace (if yes, fill in preservation information below)

Preservation - Reagent: _____ LIMS ID: _____ Date/Time: _____

☐ ☐ ☒ Chemical preservation verified, check EMD pH test strip used (if applicable)

☐ pH 0-2.5 (Lot: HC461264) ☐ pH 11.0-13.0 (Lot: HC022540) ☐ Other

☐ ☐ ☒ Air bubbles absent from VOAs (no bubble greater than 6mm) _____

Chain of Custody (COC)

Yes No

☒ ☐ All bottle labels agree with COC

☒ ☐ COC filled out properly

☒ ☐ COC signed by client

Notes:

CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Trace Analytical Laboratories, Inc.