

From: Cindy Tomaszewski <ctomaszewski@Cadillac-MI.net>
Sent: 2/14/2018 3:03:11 PM
To: "Brady, Donal (DEQ)" <BradyD6@michigan.gov>
Subject: RE: HACH NH3 ISE probe

Don –

Thank you.

When we switch from one “approved” method to another we run both methods side-by-side for 30 data points, covering all applicable matrices, to prove that the results are equivalent. We would apply the same scenario to this switch.

The ongoing QA/QC will be the same (as called for in SM 4500-NH3 D 2011 and our SOP): DOC; semi-annual PT (blind); weekly calibration, matrix spike/dup, blank; daily calibration checks.

If you would like to see something beyond that, just let me know. Thanks for your help,

Cindy A. Tomaszewski

Lab Supervisor

From: Brady, Donal (DEQ) [mailto:BradyD6@michigan.gov]
Sent: Wednesday, February 14, 2018 2:42 PM
To: Cindy Tomaszewski
Subject: RE: HACH NH3 ISE probe

Hi Cindy,

I have found out we have allowed/allow other “equivalent” methods. I have asked our permit staff in Lansing about this particular “equivalent” method and will let you know when I hear back from them. Do you have a plan for what QA/QC you would complete if you ran this method? Thanks.



Don Brady, P.E.
Michigan Department of Environmental Quality
Water Resources Division
Cadillac District Office
120 West Chapin Street
Cadillac, MI 49601
(231) 876-4478
BradyD6@michigan.gov

From: Cindy Tomaszewski [<mailto:ctomaszewski@Cadillac-MI.net>]
Sent: Monday, February 12, 2018 4:02 PM

To: Brady, Donal (DEQ) <BradyD6@michigan.gov>

Subject: HACH NH3 ISE probe

Don –

We are considering purchasing the HACH Ammonia ISE (ISENH3181) to use for analysis in conjunction with our NPDES reporting.

This probe and the associated published method (HACH Method 10001) are considered “USEPA method equivalent”. Since they are not designated specifically as “approved” or “accepted” for wastewater, we are required to check with our local regulator before reporting results for our NPDES.

Technical details regarding the HACH Ammonia ISE probe and method can be found at the following links:

https://support.hach.com/app/answers/answer_view/a_id/1000053

https://support.hach.com/app/answers/answer_view/a_id/1009762

<https://www.hach.com/hq440d-laboratory-ammonia-nh-ion-meter-package-with-isenh3181-ion-selective-electrode/product-details?id=7640592108&callback=qs>

Please review the details and let me know if we can use this probe/method for ammonia analysis and reporting in conjunction with our NPDES Permit. Thank you,

Cindy A. Tomaszewski

Lab Supervisor

City of Cadillac – Utilities Dept.

231-775-2368 (lab)

231-878-6230 (cell)