

From: Brady, Donal (DEQ)
Sent: 3/5/2018 4:10:38 PM
To: "Cindy Tomaszewski" <ctomaszewski@Cadillac-MI.net>
Subject: RE: HACH NH3 ISE probe

Hi Cindy,

Yes. They got back to me last Friday. DEQ-WRD NPDES permit staff and I agree you can proceed with Hach Method 10001 for monitoring ammonia in accordance with the Cadillac WWTP's NPDES permit (MI0020257). The DEQ-WRD understands Hach method 10001 is equivalent to a method referenced in 40 CFR 136 for ammonia analysis (SM 4500-NH3 D). Please proceed with the side-by-side evaluation and QA/QC as discussed in your 2/14/18 email below. Please let me know if you have any issues with the new method. Thanks!



Don Brady, P.E.
Michigan Department of Environmental Quality
Water Resources Division
Cadillac District Office
120 West Chapin Street
Cadillac, MI 49601
(231) 876-4478
BradyD6@michigan.gov

From: Cindy Tomaszewski [mailto:ctomaszewski@Cadillac-MI.net]
Sent: Monday, March 5, 2018 3:29 PM
To: Brady, Donal (DEQ) <BradyD6@michigan.gov>
Subject: RE: HACH NH3 ISE probe

Don –

Just following-up to see if you have rec'd any response from Lansing on this matter.

We would like to purchase a replacement probe/ISE soon, and need to know whether we can proceed with the HACH equipment.

Thanks for your help,

Cindy A. Tomaszewski
Lab Supervisor

From: Cindy Tomaszewski
Sent: Wednesday, February 14, 2018 3:03 PM
To: 'Brady, Donal (DEQ)'
Subject: RE: HACH NH3 ISE probe

Don –

Thank you.

When we switch from one “approved” method to another we run both methods side-by-side for 30 data points, covering all applicable matrices, to prove that the results are equivalent. We would apply the same scenario to this switch.

The ongoing QA/QC will be the same (as called for in SM 4500-NH3 D 2011 and our SOP): DOC; semi-annual PT (blind); weekly calibration, matrix spike/dup, blank; daily calibration checks.

If you would like to see something beyond that, just let me know. Thanks for your help,

Cindy A. Tomaszewski

Lab Supervisor

From: Brady, Donal (DEQ) [<mailto:BradyD6@michigan.gov>]

Sent: Wednesday, February 14, 2018 2:42 PM

To: Cindy Tomaszewski

Subject: RE: HACH NH3 ISE probe

Hi Cindy,

I have found out we have allowed/allow other “equivalent” methods. I have asked our permit staff in Lansing about this particular “equivalent” method and will let you know when I hear back from them. Do you have a plan for what QA/QC you would complete if you ran this method? Thanks.



Don Brady, P.E.

Michigan Department of Environmental Quality

Water Resources Division

Cadillac District Office

120 West Chapin Street

Cadillac, MI 49601

(231) 876-4478

BradyD6@michigan.gov

From: Cindy Tomaszewski [<mailto:ctomaszewski@Cadillac-MI.net>]

Sent: Monday, February 12, 2018 4:02 PM

To: Brady, Donal (DEQ) <BradyD6@michigan.gov>

Subject: HACH NH3 ISE probe

Don –

We are considering purchasing the HACH Ammonia ISE (ISENH3181) to use for analysis in conjunction with our NPDES reporting.

This probe and the associated published method (HACH Method 10001) are considered "USEPA method equivalent". Since they are not designated specifically as "approved" or "accepted" for wastewater, we are required to check with our local regulator before reporting results for our NPDES.

Technical details regarding the HACH Ammonia ISE probe and method can be found at the following links:

https://support.hach.com/app/answers/answer_view/a_id/1000053

https://support.hach.com/app/answers/answer_view/a_id/1009762

<https://www.hach.com/hq440d-laboratory-ammonia-nh-ion-meter-package-with-isenh3181-ion-selective-electrode/product-details?id=7640592108&callback=qs>

Please review the details and let me know if we can use this probe/method for ammonia analysis and reporting in conjunction with our NPDES Permit. Thank you,

Cindy A. Tomaszewski

Lab Supervisor

City of Cadillac – Utilities Dept.

231-775-2368 (lab)

231-878-6230 (cell)