

\\Nazeri Library\Spine\CSF Leak\CSF-leak-Arash\SAG T2 SPACE myelo LSP *

TA: 5:40 min Coil Selection: Auto Voxel Size: 0.7×0.7×0.7 mm³ Acc:: 7.5 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	L10.7 P39.1 F397.0 mm
Orientation	Sagittal
Phase Encoding Dir.	H >> F
Slices per Slab	96
Phase Oversampling	80 %
Slice Oversampling	8.3 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	0.70 mm
TR	1800.0 ms
TE	220.00 ms
Averages	2.0
Concatenations	1
AutoAlign	---

Contrast - Common

TR	1800.0 ms
TE	220.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	Constant
Flip Angle	100 deg
Fat-Water Contrast	SPAIR
Fat Saturation	Strong
Dark Blood	Off
Blood Suppression	Off
Wrap-up Magn.	Restore
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Contrast - Dynamic

Reordering	Linear
------------	--------

Resolution - Common

FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	0.70 mm
Base Resolution	384
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CS
Total Factor	7.5
Reference Scans	Integrated
Reference Lines PE	24
Reference Lines 3D	24
Phase Partial Fourier	Off
Slice Partial Fourier	Off
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	On
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	L10.7 P39.1 F397.0 mm
Orientation	Sagittal
Phase Encoding Dir.	H >> F
Slices per Slab	96
Phase Oversampling	80 %
Slice Oversampling	8.3 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	0.70 mm
TR	1800.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	L10.7 P39.1 F397.0 mm
Orientation	Sagittal
Phase Encoding Dir.	H >> F
AutoAlign	---
Initial Position	L10.7 P39.1 F397.0

Geometry - AutoAlign

L	10.7 mm
P	39.1 mm
F	397.0 mm
Initial Orientation	Sagittal
Initial Rotation	90.00 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	397 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combination	Adaptive Combine
Matrix Optimization	Performance
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L10.7 P39.1 F397.0 mm
Orientation	Sagittal
Rotation	90.00 deg
F >> H	280 mm
A >> P	280 mm
R >> L	68 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	123.245888 MHz
? Ref. Amplitude 1H	0.000 V

System - Tx/Rx

Reset	Off
Image Scaling	1.000
Gain	High

Physio - Signal

1st Signal/Mode	None
Trigger Delay	0 ms
TR	1800.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	SPAIR
Magn. Preparation	None
Dark Blood	Off
FOV Read	280 mm
FOV Phase	100.0 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	On
Number of Radial Views	19
Axis of Radial Views	H-F
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	spcR
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Flow Compensation	None
Reordering	Linear
Bandwidth	723 Hz/Px
Echo Spacing	4.16 ms

Sequence - Part 1

Turbo Factor	120
Echo Train Duration	508 ms

Sequence - Part 2

Introduction	On
--------------	----

Sequence - Assistant

SAR Assistant	TR > Flip Angle
Min Flip Angle	100 deg
Max. TR	2300.0 ms
Allowed Delay	30 s