Biospherical Instruments Inc.

Calibration Date: 04/14/26 **Job No.:** L-20498

Model Number: QCP2260
Serial Number: 20678
Operator: MW
Standard Lamp: V-037 (11/7/25)
Operating Voltage Range: 6 to 15VDC

To calculate irradiance, divide the net sensor reading in Volts by the appropriate Calibration Factor

Dry Calibration Factor:	7.22E-18	Volts / (quanta/(cm²·s))	4.35E+00	Volts / (µEinsteins/(cm²·s))
Wet Calibration Factor:	6.72E-18	Volts / (quanta/(cm²·s))	4.05E+00	Volts / (µEinsteins/(cm²·s))

Sensor Test Data and Results:

Sensor Supply Current (Dark): 3.4 mA
Sensor Supply Voltage: 6 Volts

Sensor Calibration Signal Voltage: 118.0 mV
Sensor Dark Voltage: 51.1 mV

Lamp Integrated PAR Irradiance: 9.27E+15 quanta/(cm^2·s) 0.01539 µEinsteins/(cm²·s)
Immersion Coefficient: 0.9310

Note:
1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).