Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

Field Case Study Report on Avare Biotech App Validation powered by MAKSA technology on Frozen-Thawed Bovine Semen (Punjab, Pakistan, 2025)

Date: January 2026
Location: Punjab, Pakistan

Conducted by: Dr.M. Ameer Hamza (Lahore, Pakistan)	Verified by: Avare Biotech Team (Dubai, UAE)
Theriogenologist | Embryologist | Andrologist	Maksim Mershiev, The CEO&Co-Founder of Avare Biotech
Signature:	Signature:

Signed by:
Maksim Mershiev
325996BF0B2B43A...

Signed by:
Maksim Mershiev
325996BF0B2B43A...

Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

Executive Summary / Conclusions

- The MAKSA technology showed convenient and effective performance when analyzing frozen–thawed Bovine semen samples.
- Total motility measurements obtained through MAKSA were quite robust.
- Progressive motility values recorded by MAKSA were generally higher, most likely due to the system's extra detectability of progressive motion of sperms in thawed samples.
- Accurate concentration analysis requires semen videos to be recorded exactly according to Avare's guidelines. The results are highly sensitive to recording quality, particularly magnification. Insufficient zoom and differences in cameras or microscopes are the primary reasons for variation in concentration values.
- The graphical demonstration of the results in a Pie chart in MAKSA report are a plus point.
- Video capture procedures followed AVARE guidelines, lead to most consistent and effective results

Profit

Using MAKSA technology, reproduction centers, insemination labs, breeding centers, and farmers can:

- **save around $26,000 per year** on CASA consumables (per 10,000 samples tested)
- **increase conception rates by ~30%** for their end clients
- **strengthen their reputation** and **attract more clients** through more consistent results.

Objective

To validate the smartphone-based MAKSA (Mobile-Assisted Key Semen Analysis) technology powering Avare Biotech App for F**rozen Bovine semen** evaluation against conventional expensive CASA systems

Materials and Methods

- Frozen semen samples from two bovine species were used in this study. The samples included buffalo semen from the local native **Neeli Ravi** breed and cattle semen from the **Sahiwal** breed, which is recognized as Pakistan's most important

Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

indigenous cattle breed. Commercially prepared frozen semen straws of both species were obtained and processed for analysis.

- The semen straws were thawed according to standard laboratory procedures, after which the samples were diluted at a 1:1 ratio using Triladyl semen extender to ensure optimal sperm dispersion and motility assessment.
- Following dilution, video recordings of each semen sample were captured using a smartphone Iphone 13, following standardized recording conditions to minimize variability. At the same time, aliquots of the same samples were transferred to a separate laboratory room for evaluation using Hamilton Thorn's CASA system, ensuring parallel assessment of identical samples.
- The entire process of thawing, dilution, video recording, and CASA evaluation was completed within a time frame of approximately two hours to minimize the effects of prolonged sample handling on semen quality parameters. After completion of the analysis, the generated data from both the smartphone-based assessments and the CASA system were collected, compiled, and stored.

Table 1MAKSA results comparison

Tests done	Key Semen Parameters	MAKSA	CASA	Difference %
Bull 1 HF 333	Concentration M/ml	**222**	20.1	**11x**
	Total Motility %	**75**	72	**−4.00%**
	Progressive Motility %	**18**	28	**55.56%**
Bull 2 HF 001	Concentration M/ml	**169**	18.6	**9x**
	Total Motility %	**75**	70	**−6.67%**
	Progressive Motility %	**43**	30	**−30.23%**
Bull 3 HF 555	Concentration M/ml	**122.9**	2.1	**60x**
	Total Motility %	**76**	65	**−14.47%**

Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

	Progressive Motility %	**49**	35	**−28.57%**

Discussion / Expert Insights

- **Total and progressive motility correlates well with CASA readings**, indicating the potential of the MAKSA system to become a world certification tool. Progressive motion sensitivity by the AI based algorithm works robustly which needs further calibrations.
- Concentration values vary and are dependent on the factors like dilution of the sample, relative concentration differences among the different fields of one sample.
- The system sometimes showed incompatibility with **phase contrast** microscopic recordings that is likely due to contrast differentiation between sperm and the background.
- Overall, MAKSA proved stable, reliable, and user-friendly under real-world field conditions

Recommendations

- Use bright-field microscopy only for CASA video recording; contrast/phase-contrast microscopy introduces halo artifacts and poor contrast, leading to inaccurate sperm detection and motility analysis.
- Standardize dilution, slide preparation, chamber depth, and temperature to minimize concentration variables and background fluid movement.
- To improve users' journey it's recommended to make videos inside the app under the supervision of the app and provide guided error prompts for focus, dilution, lighting, and recording mode.

Appendix

• Screenshot of the semen videos made and CASA results
• MAKSA PDF report samples

Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

Bull 1: CASA readings

minitube

QC Sample Report

Donor name HF 555
Donor ID 555, 28-02-2026
Life number
Species Bovine
Breed HF

Motility & Concentration of processed ejaculate

Analysis conditions

Sample analysis

Date of analysis		28/02/26
Date of production		28/02/26
Total no. of sperm analysed		298
Number of fields		4.00
Total motility	[%]	65.6
Progressive motility	[%]	35.67
Progressive fast motility	[%]	22.56
Progressive slow motility	[%]	27.88
Progressive circular motility	[%]	0.39
Local circular motility	[%]	--
Local motility	[%]	3.75
Immotile	[%]	43.05

Bull 1: MAKSA values

8:00

Analysis 28.01.2026



Total Objects:	48
Concentration (millions/ml^3):	122.9
Progressive Motility:	49%
Non-Progressive Motility:	24%
Immotility:	27%
Dilution Rate:	0
Volume of field :	0

BULL 2: CASA readings

minitube

QC Sample Report

Donor name HF 001
Donor ID 001, 28-02-2026
Life number
Species Bovine
Breed HF

Motility & Concentration of processed ejaculate

Analysis conditions

Sample analysis

Date of analysis		28/02/26
Date of production		28/02/26
Total no. of sperm analysed		343
Number of fields		4.00
Total motility	[%]	70.6
Progressive motility	[%]	30.65
Progressive fast motility	[%]	28.88
Progressive slow motility	[%]	21.97
Progressive circular motility	[%]	0.6
Local circular motility	[%]	--
Local motility	[%]	2.98

BULL 2: MAKSA values

8:00

Analysis 28.01.2026



Total Objects:	66
Concentration (millions/ml^3):	169
Progressive Motility:	43%
Non-Progressive Motility:	15%
Immotility:	42%

Docusign Envelope ID: 5FE48A0B-0F09-425B-BF24-6CC5B8D437B7

BULL 3: CASA readings	BULL 3: MAKSA values
	

Dept. Theriogenolog
Out fall Road
-54000 Lahore
Pakistan

minitube

QC Sample Report

Donor name	HF 333	Species	Bovine
Donor ID	333, 28-02-2026	Breed	HF
Life number			

Motility & Concentration of processed ejaculate

Analysis conditions

Sample analysis

Date of analysis		28/02/26
Date of production		28/02/26
Total no. of sperm analysed		223.00
Number of fields/chambers		4.00
Total motility	[%]	72.1
Progressive motility	[%]	28.9
Progressive fast motility	[%]	31.39
Progressive slow motility	[%]	21.97
Progressive circular motility	[%]	0.45
Local circular motility	[%]	--
Local motility	[%]	3.14
Immotile	[%]	43.05

Notes: Concentration/Dose=02-million
Not Valid for Court of Law

8:06

Analysis 28.01.2026

Total Objects:	87
Concentration (millions/ml^3):	222.7
Progressive Motility:	18%
Non-Progressive Motility:	40%
Immotility:	42%