

SHOP PACKET - HYDROCYN AQUA GEL

REF :	HW40F900	ODR. QTY :	900	RS 3 + 7 (QC TEST)	ISSUED DATE	14-Feb-22
ITEM CODE :	H0100016	PACK SIZE :	20/BOX		LOT NO	B21425
ITEM :	Hydrocyn Aqua Wound Gel (15g)	CUSTOMER	VMD		MFG. DATE	14-Feb-22
		ORD. NO	B086		EXP DATE:	2023-08


PROCESS	START QTY TUBE	OUTPUT TUBE	REJECT TUBE	OPERATOR	IPQC APPROVAL	RESULTS
HYDROCYN AQUA GEL PREPARATION						
TIME START : 12:46pm						
DATE : 08-02-22	910	910	NIL	Farah 1051	Roshni 12:54pm 26/02/22	2/2
TIME COMPLETE : 12:54pm						
DATE : 28-02-22						
CHECK : pH / VISCOSITY / FAC PR-HR001 QC-HR003						
LABELLING GEL TUBE						
TIME START : 8:55 AM						
DATE : 28-02-22	910	910	NIL	Vicki (1044)	Roshni 2:00pm 26/02/22	7/5
TIME COMPLETE : 2:00 PM						
DATE : 28-02-22						
GEL FILLING						
TIME START : 9:47am						
DATE : 28-02-22	910	910	910	Susi (1125)	Roshni 2:30pm 28/02/22	5/5
TIME COMPLETE : 2:30pm						
DATE : 28-02-22						
SEALING AND CUTTING GEL TUBE						
TIME START : 9:47am						
DATE : 28-02-22	910	903	7	FARAH (1051)	Roshni 9:40 am 01/03/2022	- 7 Tube Sealing reject replace with new Tube 28/02/22
TIME COMPLETE : 2:30pm						
DATE : 28-02-22						
LOT NUMBER AND EXPIRY DATE PRINTING						
TIME START : 8:40 am						
DATE : 01-03-22	903	903	NIL	ETA (932) Anniga (1129)	Roshni 9:40am 01/03/2022	5/5
TIME COMPLETE : 9:40 am						
DATE : 01-03-22						
BOX PACKING & FINAL CONTROL						
TIME START : 8:45 AM						
DATE : 01-03-22	903	903	NIL	Gangana (190) YUNHUA (1134) WIKMAH (1185)	Roshni 1:00pm 01/03/2022	Refer PCR
TIME COMPLETE : 11:19 am						
DATE : 01-03-22						
CARTON PACKING						
TIME START : 10:20 am						
DATE : 02-03-22	45 box 40tm	45 box 40tm	NIL	Herdit (1053)	Anniga 02/03/2022	Refer PCR
TIME COMPLETE : 10:30 am						
DATE : 02-03-22						

QUANTITY PRODUCED	910 pcs
QUANTITY FOR QC/RETAIN SAMPLE	3257 pcs
SHIPMENT QUANTITY PCS / CTNS :	900 pcs 45 box / 40tm
FINAL SHIPMENT DATE	
COMPLETE/PARTIAL :	COMPLETE

Checked by (PR), Signature : <i>[Signature]</i> Name : KATINI ROBINSON Date : 2/3/22	Approved by (QC), Signature : <i>[Signature]</i> Name : KATINI ROBINSON Date : 3/3/22
---	--



LABEL APPROVAL



✓
24/01/03/2020

✓
28/12/19

Hydrocyn aqua[®] gel

REF HW40F900
LOT B21425
Qty 20 pcs / box
Weight 15g / pc

VIGILENZ[®]
MALAYSIA

✓
24/01/03/2020

✓
28/12/19



Title: QC IN-PROCESS INSPECTION RECORD – HYDROCYN aqua GEL		 Bactiguard	
Document type: FORM	Doc Id: QC-IP004-01	Issue Date: 15/11/2021	Page: 1(3)


Reference / Lot No.	HW40F900 / B21425
Order Qty.	900

A. Before Filling

Batch No.	Container No.	pH (9.00 – 10.50)	FAC (100 – 200 ppm)	Viscosity (8,000 – 10,000 cps)	Gel is clear and free from contamination	Test disposition (PASS / FAIL)
61B26022201	C11	9.75	Comply	9898	✓	pass
61B26022202	C12	9.75	Comply	9898	✓	pass

B. Tube Labelling

Sampling / Frequency	Printed label is clear and legible	Reference no. corresponds to Shop Packet and approved artwork	Correct positioning of labelling	Test disposition (PASS / FAIL)
5/5	✓	✓	✓	pass
1/5	✓	✓	✓	pass
2/5	✓	✓	✓	pass
3/5	✓	✓	✓	pass
4/5	✓	✓	✓	pass
5/5	✓	✓	✓	pass


Title: QC IN-PROCESS INSPECTION RECORD – HYDROCYN aqua GEL			
Document type: FORM	Doc Id: * QC-IP004-01	Issue Date: 15/11/2021	Page: 2(3)

C. Filling Process

Sampling / Frequency	Weight of Sample						Leak Test	Test disposition (PASS / FAIL)
	Specification: 15g (16.5 – 17.8g) / 30g (31.6 – 33.3g) / 50g (51.9 – 53.4g) / 100g (102.8 – 104.7g)							
	1	2	3	4	5	Average		
5/5	16.97 g	17.12 g	17.16 g	17.09 g	17.04 g	17.07 g	✓	pass
1/5	17.20 g	17.01 g	17.11 g	17.08 g	17.14 g	17.11 g	✓	pass
2/5	17.41 g	17.32 g	17.19 g	17.16 g	17.21 g	17.26 g	✓	pass
3/5	17.31 g	17.09 g	17.02 g	17.17 g	17.43 g	17.20 g	✓	pass
4/5	17.53 g	17.72 g	17.32 g	17.07 g	17.62 g	17.45 g	✓	pass
5/5	17.63 g	17.72 g	17.45 g	17.53 g	17.09 g	17.48 g	✓	pass

D. Sealing and Cutting

Sampling / Frequency	Sealing is intact	No sharp edge and not slanted after sealing and cutting	Lot no. and expiry date on tube is correct as in Shop Packet and legible	Seal Width Specification: (15g : 4-6mm) (30g/ 50g/ 100g : 6-8mm)	Leak Test	Test disposition (PASS / FAIL)
5/5	✓	✓	✓	✓	✓	pass
1/5	✓	✓	✓	✓	✓	pass
2/5	✓	✓	✓	✓	✓	pass
3/5	✓	✓	✓	✓	✓	pass
4/5	✓	✓	✓	✓	✓	pass
5/5	✓	✓	✓	✓	✓	pass

Title: QC IN-PROCESS INSPECTION RECORD – HYDROCYN aqua GEL			
Document type: FORM	Doc Id: QC-IP004-01	Issue Date: 15/11/2021	Page: 3(3)

E. Printing on Primary Packaging

Sampling / Frequency	Print information must be clear / legible and conforms to specifications as stated in approved artwork	Lot no. and expiry date are corresponding to Shop Packet and approved artwork	Correct position of printing	Test disposition (PASS / FAIL)
5/5	✓	✓	✓	pass
1/5	✓	✓	✓	pass
2/5	✓	✓	✓	pass
3/5	✓	✓	✓	pass
4/5	✓	✓	✓	pass
5/5	✓	✓	✓	pass

F. Primary Packaging

Sampling / Frequency	Product is properly pack in primary packaging	Ref. no., lot no. and expiry date on primary box is correct and legible	Special requirement printing is clear and legible	Presence of secure stickers on both opening sides of the primary box	Presence of hologram sticker on the primary box	Test disposition (PASS / FAIL)
5/5	✓	✓	✓	✓	✓	pass
1/5	✓	✓	✓	✓	✓	pass
2/5	✓	✓	✓	✓	✓	pass
3/5	✓	✓	✓	✓	✓	pass
4/5	✓	✓	✓	✓	✓	pass
5/5	✓	✓	✓	✓	✓	pass

G. Conclusion

Overall in – process inspection: (PASS / FAIL)

Inspected By : Ray Poshan
 Name : Ray Poshan
 Date : 01/01/2022

VIGILENZ®

FINAL CONTROL TEST REPORT : HYDROCYN aqua® GEL

Doc. No. QC-FC006

Rev. 03

REF

HWYOF 900

QTY PRODUCED (pcs) :

900+3

LOT NO

B21435

RETAIN SAMPLE QTY (pcs) :

3

No	Inspection Criteria	Sampling Details	Sample Size	AQL		Results		Status (PASS/FAIL)
				Accept	Reject	Accept	Reject	
1	Legible and correct printing of label.	S3,0,15	80	0	1	80	0	pass
2	Availability of Hologram and Secure Sticker.	S3,0,15	80	0	1	80	0	pass
3	Free from contamination, deformed or damaged on packaging.	S3,0,15	80	0	1	80	0	pass
4	Proper sealing and no leakage.	S3,0,15	80	0	1	80	0	pass
5	Correct quantity in secondary box packaging.	S3,0,15	80	0	1	80	0	pass

Remarks :

N/P

Inspected By

Praj

Name

Prajne

Date

01/03/2022

Approved By

MARTIN

Name

MARTIN

Date

3/3/2022

RAW MATERIAL LOADING FORM (RMLF)

Doc. No.

PR-RM001

Rev.

06

REF

16/02/2022

LOT NO.

HW40F900

B21425

SHOP PACKET QTY (pcs)

900

CUSTOMER

VMD (MALAYSIA)

LO NO.

10/VMD/110222/01

RS

10

NO.	SYSTEM CODE	ITEM DESCRIPTION 1	DESCRIPTION 2	LOT NO.	UOM	QUANTITY (pcs)		RECEIVED BY
						REQUIRED	ISSUED	
1	P0100225	1.5ml Long Nozzle Tube with white color screw on cap	Diameter 22 x 80mm, White Colour Tube with Long Nozzle(5 layers), Open End, Gloss Finishing	H022210902	pcs	910	910	<i>Kanjana</i>
2	C0100026	RHEOLOGICAL ADDITIVE (SUPPLEMENT MP)	-	H080210801	gms	800	800	<i>Kanjana</i>
3	C0100048	Sodium Hypochloride, 10% Available Chlorine	0	C023210701	ml	200	200	<i>Kanjana</i>
4	C0100024	Pure Dry Vacuum Salt (25kg)	-	C021210601	gms	20	20	<i>Kanjana</i>
5	N/A	STOCK SOLUTION	N/A	<i>618-6023201</i> <i>618-6023202</i>	ltr	20	20	<i>Kanjana</i>

(2) PRODUCT PACKAGING

NO	SYSTEM CODE	ITEM DESCRIPTION 1	LOT NO.	QTY REQUIRED (pcs)	RECEIVED BY
1	P0100397	LBH-006-R4 - Hydrocyn Tube Label, Size: 65mm x 62mm, Material: Transparent PE	L220211202	911	<i>Kanjana</i>
2	P0100316	L63810 REV - ROUND HOLOGRAM LABEL, SIZE: 15.0mm X 15.0mm	L020211101	911	<i>Kanjana</i>
3	P0100849	Secure-Secure Round Sticker For Gel Tube (15mm) Size: 15mm (D), Material: Breakable PVC, Color: Black, Roll Form (15g only)	LB23220101	1821	<i>Kanjana</i>
4	P0100399	BX002, Inner Tube Box for Hydrocyn Gel - Size: 30mm(L) x 25mm(W) x 116mm(H), Material 310gsm 1 Side Coated Art Card (4 colors), Rev: 3	H076210801	911	<i>Kanjana</i>
5	P0100318	Outer Tube Box for Hydrocyn Gel - Size: 120 x 125 x 130mm, Material: 400gsm Grey Back	H077210801	45	<i>Kanjana</i>
6	P0100208	BLANK LABEL (Sharps Pad/ Gauze/ Gel) SIZE: 85mm X 65mm	L190210802	45	<i>Kanjana</i>
7		#N/A			

(3) OVER WRAPPING AND CARTON PACK

NO	SYSTEM CODE	ITEM DESCRIPTION 1	LOT NO.	QTY REQUIRED (pcs)	RECEIVED BY
1	P0100661	Gauze/Gel Carton-4LJ388mm X (W)293mm X (H)305mm LR150N/LR150 (BC FLUTE)	H114211101	4	<i>maulid</i>
2		#N/A			
3		#N/A			

(4) SPECIAL REQUIREMENT

NO	SYSTEM CODE	ITEM DESCRIPTION 1	LOT NO.	QTY REQUIRED (pcs)	RECEIVED BY
1	P0100324	"SAMPLE NOT FOR SALE" Sticker - Size: 35mm x 18mm, Material: Transparent PE, Colour: Red (2 layers)		NA	
2		#N/A			

ISSUED BY (MM) :

Signature :

LOVIN

Name :

LOVIN

Date :

16/02/2022

VERIFIED BY (PR) :

Signature :

RA RAJINI CANJANA

Name :

RA RAJINI CANJANA

Date :

21/2/22

VIGILENZ®	PRODUCT RELEASE TEST REPORT – HYDROCYN aqua® GEL		Doc No.	QC-PC008
			Rev.	02

Description	Hydrocyn Aqua wound Gel Tube (15g)	Reference No.	HW40F900	Date of Inspection	03/03/2022
		Lot No.	B21425	Sample Size	2 pcs x 15g

Release Criteria:

Packing Integrity				Visual Inspection	Chemical Testing		Test Disposition
Correct and legible printing of label	Proper sealing	Ability to dispense	Free from contamination/deformed/damaged	No leakage	Free from contamination / particles / foreign matters	pH (9.00 – 10.50)	FAC (100 – 200 ppm)
✓	✓	✓	✓	✓	✓	9.70	comply
							PASS / FAIL
							pass

✓ = Pass, X = Fail

Remarks: N/P

CONCLUSIONS: ☒ RELEASE ☐ ON-HOLD

Inspected By : Poy
 Name : ROSHNI
 Date : 03/03/2022

Approved By : [Signature]
 Name : AARTINI
 Date : 3/3/2022

