



INSPECT / SANITIZE PETG - Station Card

Inspect

- Check one frame for each bag to ensure it fits our shield; report if non-standard
- Check for any no-go printing errors such as layer separation; toss if issue



Sanitize

- 1 cup bleach to 3 gal water; set a timer to change in 2 hr
- Let parts rest submerged in bleach for 5 minutes
- DO NOT MOVE TO WATER (only next station can touch)

