



QC3 & QC4 Report on GE21-MODULE-M8-0012

QC3 Result



Test Date	2023-08-04
Temperature (°C) / Pressure (mBar)	21.32 / 1010.3
Pressure drop (mBar/hr)	1.09
Time constant (hr)	22.45

QC4 Result



Test Date	2023-08-04
The measured resistance R_m (MΩ)	4.907
Resistance Deviation (%)	1.86