


QC3 & QC4 Report on GE21-MODULE-M8-0009

QC3 Result




Test Date 2023-06-29

Temperature (°C) / Pressure (mBar) 21.63 / 1017.93

Pressure drop (mBar/hr) 5.91

Time constant (hr) 3.57

QC4 Result



Test Date 2023-06-28

The measured resistance R_m (MΩ) 4.955

Resistance Deviation (%) 0.91