


QC3 & QC4 Report on GE21-MODULE-M8-0008

QC3 Result




Test Date 2023-07-14

Temperature (°C) / Pressure (mBar) 21.81 / 1018.99

Pressure drop (mBar/hr) 4.21

Time constant (hr) 11.21

QC4 Result



Test Date 2023-07-14

The measured resistance R_m (MΩ) 4.943

Resistance Deviation (%) 1.13