

Children's Product Certificate

ISSUED VERSION 1 — October 7, 2026

Issued by Akoya Medical LLC

Certificate ID: AKY-CE-R11-CPC-20261003-001

Akoya Medical LLC certifies that the product identified below complies with the applicable children's product safety rule identified in Section 2, based on the testing and material determinations identified below.

1. Product identification

Akoya Pediatric Character Eyewear — Model AKY-CE, design revision R11 (historical identifier AKY-PED). Report name: "Kid's Animal Facemasks." Ages 3 and older. Non-toy, single-use visual shield for professional, continuously supervised clinical use while seated/reclined; not protective eyewear and not sent home.

Covered lot: AKY-CE-20261003-001. Artwork variants listed in report R261356: Lion, Tiger, Puppy, Kitty, Monkey and Bear, with identical tested construction. Current production selection: Lion, Tiger, Puppy and Kitty. Construction: CMYK-printed paperboard, adhesive-backed foam, white elastic cord and black adjuster, using the tested R11 assembly.

2. Applicable children's product safety rule

15 U.S.C. § 1278a (CPSIA Section 101) — Total lead content in accessible components of children's products: 100 ppm maximum. Supporting mechanical, flammability, phthalate and labeling results are recorded separately on page 2; those test headings are not all mandatory CPC citations for this 3+ non-toy product.

3. Domestic manufacturer certifying compliance

Akoya Medical LLC

333 Tamiami Trail South, Suite 317, Venice, FL 34285, USA

Telephone: 515-587-5863 | Email: info@akoyamedical.com

4. Individuals maintaining test records

Thomas Goetz Jr., c/o Akoya Medical LLC

333 Tamiami Trail South, Suite 317, Venice, FL 34285, USA

Email: tgoetz@akoyamedical.com | Telephone: 515-306-9753

Drew Zaun, c/o Akoya Medical LLC

333 Tamiami Trail South, Suite 317, Venice, FL 34285, USA

Email: dzaun@akoyamedical.com | Telephone: 515-587-5863 (company contact)

5. Date and place of manufacture / final assembly

October 2026; continuous first batch beginning October 3, 2026, under lot AKY-CE-20261003-001. Final assembly: Venice, Florida, USA. Production remains in progress as of October 6, 2026. Coverage is limited to conforming units made under this identified lot and unchanged tested configuration; actual production dates and quantities are recorded in the batch record.

6. Date and place of testing; basis for certification

AM Testing & Services, Inc., Punta Gorda, Florida, USA. Report R261356, dated October 5, 2026; samples received September 23, 2026. Report date is used in this certificate. Total lead results for foam, elastic cord and black adjuster: below detection limit (<10 ppm), against a 100 ppm limit. Paper and ordinary CMYK printing: 16 CFR § 1500.91(d)(5) and (6) material determinations; report lists paper with CMYK ink as exempt. See page 2 for the manufacturer's adhesive-coverage confirmation.

7. CPSC-accepted third-party testing laboratory

AM Testing & Services, Inc. — CPSC Laboratory ID 1342

24451 Sandhill Blvd., Unit A, Punta Gorda, FL 33983, USA

Telephone: 708-907-5252 | Report signatory: William M. Baumann, Laboratory Director

Supporting record and production scope

Companion to the issued CPC on page 1. This page records the test basis and manufacturer confirmations; it is not a separate certificate or an expansion of the laboratory report.

Completed report results

Report requirement / evaluation	Result and scope
16 CFR 1500 — use and abuse; 16 CFR 1501 — small parts	Pass. Sharp-point and small-part checks passed as received and after the reported impact, torque, tension and compression tests. Sharp-edge checks were N/A.
16 CFR 1500.44 — flammability	Pass; 0.058 inches/second. The report uses the 0.1 inches/second flammable-solid criterion. No 16 CFR 1610 testing is claimed.
15 U.S.C. § 1278a — total lead	Pass for foam, white elastic cord and black adjuster. Each <10 ppm. Paper with CMYK ink listed as exempt.
16 CFR 1307 / CPSIA Section 108 — phthalates	Pass. Black adjuster DEHP: 166 mg/kg, below the report's 1,000 mg/kg limit; other reported values ND. Retained as supplemental testing for the stated non-toy use.
Tracking label and country-of-origin marking	Pass. Report and Dan Bartels' October 6 email rely on the submitted labeling/packaging configuration.
Washington/Maine and California screening	Washington/Maine phthalate: Pass. California Prop. 65 entries: Complies, qualified by the report's similar-product basis; detailed tables are informational. These are not federal CPC citations.

Manufacturer confirmations recorded October 6, 2026

Thomas Goetz Jr. confirms that the PSA adhesive was tested as part of the foam, as previously communicated by the laboratory; it is not an outstanding request for separate adhesive testing. The report itself identifies the tested component as "Foam." This statement records Akoya's confirmation without altering the report's component description.

Thomas confirms that production materials and construction match the tested samples and that Akoya will use the submitted packaging. The flammability basis follows the final report's 16 CFR 1500.44 testing for this paperboard construction; 16 CFR 1610 is not claimed as tested or included as an unresolved testing request. No ASTM F963 toy certification or under-three age coverage is claimed.

Labeling, ongoing production and document control

Retain the reviewed R11/L06 mask warning and tracking information, the individually labeled bag, and the labeled outer box. Packaging is handled by the clinician; its tracking information still identifies the enclosed product and batch. Dan's October 6 email requires AM Testing review if that packaging changes.

Keep actual production dates, material/lot traceability, quantities and assembly checks against AKY-CE-20261003-001. A continuous batch may use its starting month/year on the CPC. Later lots or changed construction require an updated coverage record and assessment of any testing affected by changes; this certificate does not certify unidentified future production.

Version 1 issued by Akoya Medical LLC on October 7, 2026, as authorized by Thomas Goetz Jr., adopting the product scope, testing basis and manufacturer confirmations above. Akoya is responsible for this certificate's accuracy. Furnish the certificate to distributors and retailers and ensure it accompanies covered shipments physically or electronically. Retain supporting records and update coverage when required.

References retained with the certificate

AM Testing report R261356 (October 5, 2026); Dan Bartels email (October 6, 2026); Akoya_R11_Labeling_Package_Dan.pdf (October 5, 2026); manufacturer confirmations in this conversation (October 6, 2026). CPSC seven-element and continuous-batch guidance: <https://www.cpsc.gov/Business--Manufacturing/Testing-Certification/Childrens-Product-Certificate>