

Article

# Precision and Limits of Detection for Selected Commercially Available, Low-cost Carbon Dioxide and Methane Gas Sensors

Wesley T. Honeycutt <sup>1,†</sup>, M. Tyler Ley <sup>2</sup>, and Nicholas F. Materer <sup>3,\*</sup>

<sup>1</sup> Oklahoma State University, Department of Chemistry, 107 Physical Sciences, Stillwater, OK 74078, USA; wes.honeycutt@okstate.edu

<sup>2</sup> Oklahoma State University Department of Civil Engineering; 319C Engineering South, Stillwater, OK 74078, USA; tyler.ley@okstate.edu

<sup>3</sup> Oklahoma State University Department of Chemistry; 107 Physical Sciences, Stillwater, OK 74078, USA; materer@okstate.edu

\* Correspondence: materer@okstate.edu

† Current address: University of Oklahoma Biological Survey; 111 Chesapeake St.; Norman, OK 73019, USA

Version June 27, 2019 submitted to Sensors

**Abstract:** The performance of a sensor platform for environmental or industrial monitoring is sensitive to the cost and performance of the individual sensor elements. Thus, the detection limits, accuracy, and precision of commercially available, low-cost carbon dioxide and methane gas concentration sensors were evaluated by precise measurements at known gas concentrations. Sensors were selected based on market availability, cost, power consumption, detection range, and accuracy. A specially constructed gas mixing chamber, coupled to a precision bench-top analyzer, was used to characterize each sensor during a controlled exposure to known gas concentrations. For environmental monitoring, the selected carbon dioxide sensors were characterized around 400 ppm. For methane, the sensor response was first monitored at 0 ppm, close to the typical environmental background. The selected sensors were then evaluated at gas concentrations up to 1000 ppm. The determined detection limits accuracy, and precision provides a set of matrices that can be used to evaluate and select sensors for integration into a sensor platform for specific applications.

**Keywords:** Sensor systems and applications, Carbon Dioxide, Methane

## 1. Introduction

Environmental monitoring of local gas concentration is becoming increasingly important to ensure worker safety and for early identification of potential leaks. Sensor platforms for real time monitoring of carbon dioxide (CO<sub>2</sub>) and methane (CH<sub>4</sub>) gas concentrations are of interest as these gases can impact animal health and crop growth. Since these gases are also considered greenhouse gases, monitoring of industrial sites is important to understand their impact on the environment. Collection of multiple samples at several locations in a given area is critical to obtain reliable results, and this directly impacts the overall complexity and cost. Aspects of effective sampling, along with other considerations, are discussed by Keith et al. [1].

Potential sensor platforms must consider the limit of detection, precision, accuracy, reliability, and power consumption for each sensor integrated within the unit. Such devices include, but are not limited to, those used in HVAC air handlers [2,3], chemical processing units [4], oil well monitoring devices [5,6], and environmental monitoring [7–11]. Given the large number of commercially available sensors and papers discussing their use in various applications, there is limited literature that directly compares the precision, sensitivity and limits of detection of the sensor elements. This lack of information

hinders one's ability to select the optimal sensors for sensor platform designed for a given application. To address this issue, an array of commercially available, low-cost sensors for CO<sub>2</sub> and CH<sub>4</sub> were evaluated at concentrations typical to those required for environmental monitoring, and monitoring around carbon sequestration operations and oil fields.

Low-cost sensors currently available can be generally categorized by detection method. Principal among these methods are optical absorption, chemiresistive (based on the resistance changes of a material due to chemical reaction with an analyte [12]), and electrochemical. Since studies have cited concerns with electrochemical gas concentration sensors, such as a short lifetime and lack of robustness [13], only optical and chemiresistive sensors were selected for evaluation in this study. Optical methods of detecting CO<sub>2</sub> and CH<sub>4</sub> are based on measuring the absorption of light at 2352 cm<sup>-1</sup> and 3015 cm<sup>-1</sup>, respectively [14,15]. Since these sensors utilize the Beer-Lambert law to relate absorption to concentration, the calibration is only dependent on the geometry of the sensor and physical properties of the gas [16].

Lower cost optical sensors typically utilize nondispersive infrared (NDIR) sensing. This method utilizes a broad spectrum light source which is restricted by a narrow band pass filter across the absorbance maximum before reaching the detector. As a result of the simple construction requirements of NDIR, lower cost parts can be used and the design can be both more compact and robust. In general, NDIR detection is utilized for CO<sub>2</sub> due to its relatively large molar absorption coefficient, allowing for short path lengths to be used in devices. Use of NDIR detectors for CH<sub>4</sub> is limited in practical applications due to its lower absorption coefficient and overlapping symmetric C-H stretches. The overlapping stretches makes CH<sub>4</sub> difficult to distinguish from other common aliphatic gases such as ethane and propane [17]. Although selectivity of the NDIR sensors with respect to different analyte gases was not explored in this study, it should be considered when these sensors are integrated into a monitoring device.

Chemiresistive sensors for detection of CH<sub>4</sub> typically use a thin oxide film [13]. These sensors work by measuring resistance changes due to differences in the electron transport through the metal oxide film, in the presence of oxygen and reactive gases [18]. When the film is exposed to CH<sub>4</sub>, the molecule adsorbs and reacts with surface oxygen species resulting in a change in the electrical conductivity [19,20]. Since the sensor depends on chemical reactions involving both atmospheric oxygen and CH<sub>4</sub> for proper functioning, the measurement is dependent on the partial pressures of these gases. By the laws governing interactions of gases, the chemiresistive sensors are dependent on atmospheric properties such as relative humidity and temperature in addition to film preparation. Chemiresistive sensors are also known to respond to a range of hydrocarbon gases [21]. In many cases, the non-linear response curves for additional gases are included in the documentation provided by the manufacturer. However, similar to the optical sensors for CH<sub>4</sub>, the selectivity was not explored but should be considered when integrating these sensors into a sensor platform.

## 2. Methods

### 2.1. Sensor Selection

With two exceptions, the sensors selected were all commercially available in large volumes (at least 1000 units) at low-cost (defined here as less than \$100 per unit in bulk). Selected sensors were expected to have sensitivity at the environmental concentrations of CO<sub>2</sub> and CH<sub>4</sub>. For CO<sub>2</sub>, this level is around 400 ppm [22,23]. For CH<sub>4</sub>, the baseline atmospheric concentration is under 2 ppm [24–26]. In total, six CO<sub>2</sub> sensors were selected for testing, the K-30 SE-0018, COZIR AMB GC-020, Gascard CO<sub>2</sub>, MSH-P/CO<sub>2</sub>/NC/5/V/P/F, MSH-DP/HC/CO<sub>2</sub>/NC/P/F, and Telaire T6615. Given the prevalence of optical sensors for CO<sub>2</sub>, no chemiresistive sensor was selected for this analyte. For CH<sub>4</sub>, seven sensors were evaluated, the MQ-4, Gascard CH<sub>4</sub>, MSH-P/HC/NC/5/V/P/F, MSH-DP/HC/CO<sub>2</sub>/NC/P/F, TGS-2600, TGS-2610, and TGS-2611. The relevant properties, obtained from the manufacturer documentation, for the CO<sub>2</sub> (Table S1) and CH<sub>4</sub> sensors (Table S2) are

**Table 1.** Calibrated gas mixtures used to prepare various gas mixtures with a carrier gas.

Carbon Dioxide	Methane	Balance Gas
3000 ppm	3000 ppm	nitrogen
100 ppm	100 ppm	nitrogen
0 ppm	20 ppm	nitrogen

summarized in the supplementary materials. The selected sensors meet the requirements for use in portable low-power monitoring devices [27].


The K-30, COZIR, Dynament, and Telaire sensors are all NDIR CO<sub>2</sub> sensors. The Dynament hydrocarbon sensor (MSH-P/HC) was chosen as inexpensive candidates for CH<sub>4</sub> detection. Dynament also provides a dual gas NDIR sensor (MSH-DP/HC/CO<sub>2</sub>) designed to measure both CO<sub>2</sub> and CH<sub>4</sub> concentrations, which is attractive for applications requiring both low-cost and portability. Finally, CO<sub>2</sub> and CH<sub>4</sub> Gascard (sold by GHG Analytical) sensors were included, even though they were more expensive than the other chosen NDIR sensors. However, the Gascard sensors are still significantly less expensive than typical bench-top instruments. The inclusion of pressure and temperature sensors make them attractive enough to compensate for the expense.

The low-cost chemiresistive-based CH<sub>4</sub> sensors studied here are typically used in gas warning systems [28]. Selected sensors include the MQ-4 from Hanwei Electronics and TGS-2600, TGS-2610, and TGS-2611 manufactured by Figaro Engineering Inc. sensors. The TGS sensors are used in commercial CH<sub>4</sub> detectors, and the TGS-2600 sensor has been previously evaluated for atmospheric applications [29,30]. There are several different MQ versions optimized for hydrocarbon sensing. The MQ-4 sensor was chosen as this variant was specifically tuned for CH<sub>4</sub>. Chemiresistive sensors required a minimum conditioning period or “burn-in” time. This “burn-in” time is a period (often exceeding 24 hours) during which the heating element is allowed to constantly heat the sensing element of the sensor. The “burn-in” time was met or exceeded for all chemiresistive sensors.

## 2.2. Analyte Preparation and Sensor Testing

To test these sensors under controlled conditions, a gas mixing apparatus (Figure 1) was constructed which prepares gas mixtures of known concentration by controlling the positive pressure flow of gas streams into a mixing chamber. A high-quality bench-top analyzer (California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer), sensitive to both CO<sub>2</sub> and CH<sub>4</sub>, was used as a precision reference. This setup allows gas flows of a known concentration to be prepared from a calibrated gas cylinder by mixing with a carrier gas. The calibrated gas contained a mixture of gases at concentrations listed in Table 1 depending on the specific experiment being performed. Since the objective is to address the accuracy, precision, sensitivity, and limits of detection, only CO<sub>2</sub> and CH<sub>4</sub> were utilized. However, selectivity requirements must be considered when these sensors are integrated into a monitoring platform. For CO<sub>2</sub> experiments, the carrier gas was typically nitrogen. For CH<sub>4</sub>, medical grade air was utilized as a carrier gas since the chemiresistive sensors required oxygen to correctly measure the CH<sub>4</sub> concentration. All reported concentrations were determined using the bench-top analyzer. Thus, the concentration ranges and all related values are reported here not as whole numbers but direct observations.


The calibrated gas mixtures were provided by and certified within  $\pm 2\%$  by Airgas Inc. The calibrated gas was diluted using either air or nitrogen gas by a set of mass flow controllers to produce specific concentration of the analyte gases. The solenoid valves in the gas mixing apparatus were used to send either the undiluted calibrated gas or the carrier gas to analyzer for calibration purposes. This ability allows the analyzer to be periodically calibrated. The uncertainty in the gas concentration results in a systematic error. Around atmospheric levels of CH<sub>4</sub>, this error is less than 0.4 ppm. For atmospheric levels of CO<sub>2</sub>, the error is approximately 8 ppm. Since the fluctuations around the mean were independent of the exact concentration (see Section 3.3), this systematic error will not effect the results from the baseline noise or the International Union of Applied Chemistry (IUPAC) limit of



**Figure 1.** Component diagram of controlled gas exposure apparatus with chamber for diffusion-type sensors.

detection, discussed below. For the sensor calibrations, uncertainty in the absolute concentration of the carrier gas will introduce an error in the intercept but not the slope. While any systematic error will have small effect on the calibration-corrected limit of detection, the random uncertainty dominates any potential systematic bias of a few ppm.

Multiple sensors were operated concurrently as shown in Figure 2. The “Gas Mixer” in this diagram is the gas mixing apparatus shown in Figure 1, and the “Gas Analyzer” is the California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer. There are two types of sampling methods used by the selected sensors; diffusion and flow (see Table S1 and S2 in the supplemental materials). Flow sampling sensors are designed such that the analyte gases must physically pass through a sampling chamber within the sensor. In terms of applications, these sensors typically require additional filters to be added upstream to prevent inclusion of particulate matter, which may damage the sensor or introduce measurement error. In the experimental setup (Figure 2), the flow capable sensors were connected directly to gas flow from the mixing apparatus. In addition to the flow sensors, another common sampling technique relies on diffusion through a fine mesh screen or film. The diffusion method allows the sensing element to be exposed to the analyte gas while still being



**Figure 2.** The diagram shows the flow of gas from the mixing apparatus to Gascard sensors, small enclosures containing sensors, the large gas flow chamber, and finally to the California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer. The gas mixing apparatus is shown in Figure 1. The reported concentrations are based on the value detected by the California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer.

protected from particulate contaminants. This method also eliminates the extra pump and particulate filters required for sensors that require gas flow. The small “sensor flow box” enclosures depicted in Figure 2 contained the K-30 CO<sub>2</sub> sensor and a socketable chemiresistive CH<sub>4</sub> sensor (such as the MQ-4 or TGS series sensors). The larger “gas flow chamber” contained the COZIR, Telaire, Dynament, and K-30 sensors. Given the large total volume, data collected from the sensors in this chamber were analyzed only after the system reached a constant concentration following the introduction of a gas with known concentration of CO<sub>2</sub> or CH<sub>4</sub>. In some experiments, performed to accurately measure the response time for chemiresistive sensors, a smaller sensor enclosure (internal volume 2.54 cm<sup>3</sup>) was used to minimize the time to reach a stable concentration.

### 2.3. Sensor Output and Data Collection

$$\zeta = \frac{V_{measured}}{V_{cc} - V_{measured}} \times \frac{1}{R_{ref}} \quad (1)$$

The output of each sensor was logged as a function of time using software provided by the manufacturer as part of development kits when provided, or using an Arduino microcontroller with a prototyping board and microSD card. Due to the unique interface requirements of each sensor, development kits were purchased when possible. In all cases, the internal electronics within the selected optical sensors process the option data and provide a digital output. For example, the K-30 reports the concentration directly, while the Gascard reports a digital value between zero and one, with one being the maximum concentration (30,000 ppm CO<sub>2</sub> for the sensor used in this study). For the Gascard, scaling was applied to the output before analyzing. The chemiresistive sensors were energized using the recommended voltage for the heater element and the response was measured using

a 12-bit A/D converter, after buffering and filtering, across a reference resistor (a 10 k $\Omega$  resistor was used for  $R_{ref}$ ). The A/D converter was referenced using the 5 V supply, which was also connected to the sensing element of the chemiresistive sensor. Equation 1 describes the conversion of the measured voltage ( $V_{measured}$ ) to the conductivity ( $\varsigma$ ), measured in Siemens (S).

#### 2.4. Calibration and Fitting Procedures

Calibration curve of each sensor was determined by varying concentrations and recording the output with time. The CO<sub>2</sub> sensors were calibrated using gas concentrations from 34.5 to 1020 ppm. For the CH<sub>4</sub> sensors, calibration curves were generated from gas concentrations between 1.85 and 995 ppm. The procedure starts with a constant flow of the carrier gas until stable baseline is obtained. Next, the analyte gas at a known concentration is introduced. After 24 h, the carrier gas is reintroduced and the system is again allowed to stabilize before the next measurement. During this procedure, concentration data from each sensor was continually collected. From these data, the average baseline and response to each gas concentration was extracted. The average response at each concentration was determined from the data collected after the system stabilized (include any overshoot or ringing) after each new concentration.

$$f(x) = \frac{a \times b \times x}{1 + b \times x} \quad (2)$$

The fitting procedures for the optical and chemiresistive sensors differ. The raw output of the optical absorption-based sensors is expected to follow Beer-Lambert law. In all cases, the internal electronics within the selected optical sensors process the absorbance and provide a digital output. For the optical sensors, the accuracy and precision of the internal calibration are quantified by linear regression. The calibration curves for the chemiresistive are non-linear, and modeled using a Langmuir-like or Langmuirian form (Equation 2). Equation 2 converges to a linear function the ppm approaches zero, which simplifies the limit of detection calculations discussed below. The parameters  $a$  and  $b$ , along with the asymptotic standard errors, are determined by fitting the experimental data using the Levenberg-Marquardt algorithm in *gnuplot* [31]. Additional non-linear functions have been suggested by kinetic analysis [32,33]. Of these, a power law model, where the  $\log$  of the output is proportional to the  $\log$  of the concentration in ppm, is considered.


#### 2.5. Precision and Baseline Noise Tests

Precision of the sensors was determined by a 20 h to 30 h data collection period at a known concentration and uniform flow. These experiments were performed around the baseline atmospheric concentration for each analyte gas, which is approximately 400 ppm for CO<sub>2</sub> [22,23] and under 2 ppm for CH<sub>4</sub> [24–26].

#### 2.6. Limit of Detection Determination

The limit of detection is the minimum concentration that can be detected as significantly different from the background [34–36]. IUPAC defines the limit of detection as three times the standard deviation ( $\sigma$ ) from the background. For the sensors discussed here, the raw output from the sensors must be transformed into a concentration, and any uncertainties in the calibration will affect the limit of detection. For measurements requiring a calibration curve, Long and Winefordner provide a review of the various definitions as well as several examples [34]. For optical sensors, their procedure to include the uncertainties in the slope and intercept in the calculated the limit of detection was utilized. For the non-linear chemiresistive sensors, the asymptotic standard uncertainties in  $a$  and  $b$  were propagation to correct the IUPAC limit of detection.






**Figure 3.** This combined figure of the response curve of the K-30 (plotted with “+” and “x”) and Gascard (plotted with “\*” and “□”) shows consistent discrepancy between both sensors during gas concentration increases. Both sensors, in the presence of clean air, were subjected to injection of CO<sub>2</sub> of 577 and 1433 ppm (exact concentrations determined by the California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer) before the chamber was flushed with clean air. The response of the sensors to both the increase and decrease in concentration of CO<sub>2</sub> was quicker than the sampling time of the experiment.

### 3. Results and Discussion

#### 3.1. Sensor Response

Figure 3 shows the response of two selected CO<sub>2</sub> sensors, the K-30 and Gascard, to 600 ppm and 1400 ppm concentration steps of CO<sub>2</sub>. The K-30 output was used directly, while the Gascard data was multiplied by the provided scaling factor before plotting. This data demonstrates relative similarity in behavior of the selected sensors in the presence of concentration changes well above the limit of detection. It was observed that all the optical sensors responded quickly to an increase in CO<sub>2</sub> inside the environmental chamber, limited only by the internal sampling rate of the sensor.

Figure 4 depicts a plot of the response of an MQ-4 sensor upon exposure to five different concentrations of CH<sub>4</sub> between 500 ppm and 2200 ppm, which could represent a CH<sub>4</sub> leak. In our setup, the MQ-4 produces significant overshoot. Some of the overshoot is inherent to the sensor, but could also be a result of the small interruption in the gas flow during a concentration change. For all concentrations, a stable value within 2.5% of the mean was produced after  $78 \pm 10$  s, when averaged over all experiments. The settling time for the MQ-4 sensor did not appear to be concentration dependent. Shorter settling time and less overshoot was observed for the TGS sensors.




**Figure 4.** The MQ-4 response to difference  $\text{CH}_4$  concentrations determined by the California Analytical Instruments, Inc. ZRE Non-Dispersive Infrared Analyzer.

### 3.2. Calibration Result

Typical calibration plots for an optical and a chemiresistive sensors are shown in Figure 5. The data points are the measured values at each concentration, while the line is a fit to the data. The upper plot in Figure 5 is a calibration curve obtained from the Gascard sensor, along with the linear fit. The slope is  $1.02 \pm 0.02$  indicating an excellent calibration and precision. Above their respective detection limits, good linear fits were obtained for each optical sensor. Linear regression results show the slopes for the selected sensors are all close to one, indicating good precision. For the  $\text{CO}_2$  sensors, the slopes range from 1.11 to 0.86, with the MSH-P/ $\text{CO}_2$  and Telair as the high and low outliers, respectively. Excluding these sensors, the average slope is 0.99. For the  $\text{CH}_4$  sensors, the slopes are 0.98 (Gascard), 1.03 (MSH-DP/ $\text{HC}/\text{CO}_2$ ), and 1.20 (MSH-P/ $\text{HC}$ ). The fact that a majority of the slopes are close to one is expected as the optical path length and the physical properties of the analyte determine the response. In contrast to the slopes, the value of the vertical intercept or the response at zero concentration vary widely, exceeding several 100 ppm in many cases. Large variations are found even between sensors from the same manufacturer. Since the intercept depends on the intensity of the source and the detector, these variations are not unexpected. To address this issue, a majority of the optical sensors tested provide an easy way to adjust the intercept. For example, the K-30 sensors can auto-calibrate the zero based on known average background level of  $\text{CO}_2$ . In general, it was found the intercept must be adjusted before use.

The lower plot in Figure 5 shows the response of a chemiresistive  $\text{CH}_4$  sensor (MQ-4, response is expressed in conductivity) with non-linear fit to a Langmuirian form (Equation 2). For ppm concentration above 500 ppm, acceptable results were also obtained using a power law model. However, this model produced a fit that increased too rapidly with ppm at lower concentration.






**Figure 5.** The response of the Gascard and MQ-4 sensors with respect to the concentration reported from the California Analytical Instruments Inc. ZRE Non-Dispersive Infrared Analyzer for different concentrations of either CO<sub>2</sub> or CH<sub>4</sub>, respectively. This plot reports values directly from the instrument. The top part of the figure shows the response of an optical CO<sub>2</sub> sensor with the expected linear behavior. The bottom part of the figure shows the response of a chemiresistive CH<sub>4</sub> sensor (MQ-4, expressed in conductivity rather than concentration as this is the directly reported value of the sensor) with a fit to the non-linear response.

At high concentrations, this model did not accurately represent a real sensor where the response approaches a constant. A lower concentration reference gas was required to precisely prepare an analyte gas at 100 ppm and below. This action resulted in a gap between 100 and 500 ppm in the lower plot. It is also significant that the chemiresistive sensors shows a rapid increase in conductivity with concentration below 100 ppm when compared to the optical sensors.

Unlike optical sensors for CO<sub>2</sub>, which only needs reading at zero ppm to be experimentally established, chemiresistive sensors require the measurement of a complete non-linear calibration curve to ensure accuracy and precision. The output of the chemiresistive sensors are also dependent on temperature and humidity, further complicating the calibration [37]. Although the selected chemiresistive sensors have the requisite precision to provide a clear indication of a change (based on number of  $\sigma$  above the noise, see Table 2 along with the associated discussion below) at the ppm level, obtaining an accurate result is challenging.

### 3.3. Precision and Baseline Noise Tests

The measurement precision and baseline noise of each sensor was performed as described in Section 2.5. In most cases, the precision of the sensors was determined by a 20 h to 30 h data collection time at a known concentration (400 ppm for CO<sub>2</sub> and 0 ppm for CH<sub>4</sub> based on previously discussed atmospheric baselines) under uniform flow. In general, the optical sensors displayed flat baselines. Chemiresistive sensors typically display baseline changes due to humidity and temperature, which must be taken into account in monitoring applications. Figure 6 shows a typical data set for the




**Figure 6.** Long term baseline data were collected at the atmospheric baseline conditions (using a bottle of compressed medical grade breathing air (approximately 0 ppm  $\text{CH}_4$ ), in the gas mixing chamber. The reported values are in units of conductivity obtained directly from the instrument, and cannot be reported as concentration for inter-unit comparison. The baseline fluctuations can be observed in these plots.

TGS-2611 and MQ-4 sensor. TGS-2611 sensor displayed significantly less baseline drift than the MQ-4 sensor.

Since the initial Fourier analysis showed no significant periodic variations, distribution of the digitized sensor output around the mean response of the sensor was utilized to quantify the precision and baseline noise. The data stream (sensor output with time) from each sensor was subtracted from the mean response of the sensor, and a histogram of these differences was created. Since the digitized sensor outputs have a finite number of possible output values, no additional bins were created while producing the analysis. Although the data utilized for this analysis were obtained at typical environmental concentrations, additional experiments found that the measured deviation around the mean for each sensor was independent of the concentration of analyte gas. There was insignificant correlation between the concentration of the analyte gas and the  $\sigma$  obtained at each concentration, as quantified by the Pearson's Correlation Coefficient ( $\rho = -0.173$ ).

The resulting histograms, along with a best-fit Gaussian peak, are shown in Figures 7 and 8 for the  $\text{CO}_2$  and  $\text{CH}_4$  sensors, respectively. In Table 2,  $\sigma_{GAUSS}$  is the standard deviation determined from the Gaussian fit. The standard deviation ( $\sigma$ ) calculated from the background subtracted sensor data is almost identical to  $\sigma_{GAUSS}$ . In Figures 7 and 8, the abscissa (x-coordinate) was scaled by the standard deviation of the data set and the area normalized to one. The resolution of the dual gas Dynament (MSH-DP/HC/CO<sub>2</sub>) sensor was insufficient to properly determine the fluctuations around the mean, as this sensor oscillated between two values with time at the given concentration. The limited responses is consistent with the resolution stated in their documentation, which lists 0.01% or 100 ppm as the low end of the concentration range. The fluctuation around the mean for the single gas Dynament (MSH-P/CO<sub>2</sub>) sensor is slightly less than the stated resolution of 50 ppm.

To determine the quality of the Gaussian fits in Figures 7 and 8, the root-mean-squared error (RMSE) between the fit and the experimentally generated histogram was calculated for each sensor.




**Figure 7.** Frequency distribution of the digitized sensor output around the mean response of the sensor (points), along with Gaussian non-linear for each CO<sub>2</sub> sensors around the typical environmental baseline concentration of 400 ppm. Any deviation between the center of the Gaussian fit and the mean are due to significant asymmetry of the frequency distribution around the mean.

These values are also listed in Table 2. For RMSE, a lower value indicates that the Gaussian function fits closely to the data points, whereas a higher value indicates a poorer fit. The values listed in Table 2 range from roughly 0.150 to 0.300. As the probability density function to which the data were fit was normalized to 1, the RMSE values are unitless. Since RMSE is a measure of fit, these values can be used in conjunction with  $\sigma_{GAUSS}$  to characterize the sensors.

Of the CO<sub>2</sub> sensors (see Figure 7 and Table 2), the Gascard and K-30 sensors produced the smallest standard deviation around the mean,  $\sigma_{GAUSS}$ , or the highest precision. The  $\sigma_{GAUSS}$  value of the Telaire sensor was approximately 2 times than that of either the Gascard or the K-30 sensor. The COZIR and Dynament (MSH-P/CO<sub>2</sub>) sensors generated  $\sigma_{GAUSS}$  values which were significantly greater than that of the Gascard, Telaire and K-30 sensors. Finally, Table 2) shows that the dual gas Dynament (MSH-P/HC/CO<sub>2</sub>) sensor produced the largest  $\sigma_{GAUSS}$  or deviation around the mean. The fluctuations around the mean for Dynament sensors are still lower than the quoted resolutions in the manufacturer's documentation of 50 and 100 ppm, respectively, for these two sensors.

The Gascard sensor for CO<sub>2</sub> produced a normal response around the mean detected value with low RMSE. At times, there were some large fluctuations, several standard deviations around the mean response, in the output of the Gascard. This suggests that some minimal digital filtering may be required. For the K-30 sensor, a good Gaussian fit was produced (small RMSE). However, the distribution for the K-30 was skewed toward higher concentration values and appeared to contain two overlapping peaks. This caused the  $\sigma_{GAUSS}$  to be slightly larger and the RMSE to be artificially inflated. Careful analysis of this double peak showed each peak had a similar standard deviation



**Figure 8.** Frequency distribution of the digitized sensor output around the mean response of the sensor (points), along with Gaussian non-linear fit for each  $\text{CH}_4$  sensor around the typical environmental baseline concentration of 0 ppm. Any deviation between the center of the Gaussian fit and the mean are due to significant asymmetry of the frequency distribution around the mean.

( $\sigma$ ), suggesting a small shift in the mean measured value. Since the K-30 sensor periodically adjusts for changing backgrounds in the firmware to ensure a normal output concentration of 400 ppm, the observed change in the reported mean concentration in the middle of the run was likely caused by this period background adjustment. It is also possible the periodic temperature and pressure variations, which were corrected by the Gascard sensor, were responsible for the observed skew in the K-30. In general, precise work requires incorporation of pressure and temperature sensors into potential environmental sensing units to allow ppm corrections to be performed at time of measurement.

The probability distribution of responses around the mean by the GE Telaire sensor is a single peak with a  $\sigma_{GAUSS}$  that is approximately 2 to 3 times as large as that of the K-30 and Gascard sensors. Since the GE Telaire and K-30 sensor share comparable sensing mechanisms and path lengths, this larger  $\sigma_{GAUSS}$  was initially surprising. A direct comparison of the response for both the Telaire and K-30 sensors both with and without ambient light, showed that, unlike the K-30 sensor, the Telaire sensor was sensitive to ambient light level. Given that long data collection times are required to produce reliable histograms, the ambient light conditions changed over the course of data collection. It was suspected that the larger deviations around the mean for the Telaire sensor and resulting  $\sigma_{GAUSS}$  values are due to changes in the ambient light. Further experiments with the sensor shielded from ambient light substantiate this suspicion.

For the optical  $\text{CH}_4$  sensors (see Figure 8 and Table 2), the Gascard for  $\text{CH}_4$  produced a  $\sigma_{GAUSS}$  with a low RMSE. The single gas Dynament (MSH-P/HC/) hydrocarbon sensor produced a very low  $\sigma_{GAUSS}$  and performed well in terms of precision. The  $\sigma_{GAUSS}$  for the dual-gas Dynament

**Table 2.** The first two numbered are the computed standard deviation in ppm of Gaussian fitted probability distributions and root-mean-squared error in the fit. The next two numbers are the computed IUPAC ( $3\sigma_{GAUSS}$ ) and calibration uncertainty corrected Limits of Detections in ppm.

	Sensor	$\sigma_{GAUSS}$ (ppm)	RMSE	IUPAC	Corrected <sup>†</sup>
Carbon Dioxide	K-30 SE-0018	1.91	0.219	5.7	27.1
	COZIR AMB GC-020	14.1	0.304	42.3	80.6
	Gascard CO <sub>2</sub>	2.12	0.223	6.4	32.1
	MSH-DP/HC/CO <sub>2</sub>	86.4	0.197	260	254
	MSH-P/CO <sub>2</sub>	17.6	0.217	52.8	68.0
	Telaire T6615	4.42	0.185	13.3	31.1
CH <sub>4</sub> /Hydrocarbon	MQ-4	26.7 <sup>‡</sup>		80.0	82.0
	Gascard CH <sub>4</sub>	35.7	0.222	110	151
	MSH-P/HC	3.54	0.152	10.6	170
	TGS-2600	36.1	0.225	110	117
	TGS-2610	37.1	0.237	111	113
	TGS-2611	5.4	0.208	16.3	16.3

<sup>†</sup> A linear calibration was utilized for the optical sensor, while the chemiresistive utilized the non-linear Langmuirian fit. See text.

<sup>‡</sup> This value was determined directly from the experimental response. See text.

(MSH-DP/HC/CO<sub>2</sub>) sensor was not included in this table. As mentioned previously, the dual-gas Dynament (MSH-DP/HC/CO<sub>2</sub>) sensor only reported two values for CH<sub>4</sub> around the mean rather than a distribution of several values. This result was consistent with the 100 ppm resolution quoted in the manufacturer's documentation for concentrations of less than 10% CH<sub>4</sub>.

The result for the chemiresistive CH<sub>4</sub> sensors are also shown in Figure 8, with a  $\sigma_{GAUSS}$  and a RMSE results in Table 2. The distribution curve for the MQ-4 sensor was not included. Since the MQ-4 sensor displayed significant baseline drift when compared to the TGS-2611 sensor (Figure 6), the standard deviation was instead calculated directly from a relatively flat region of the baseline. This different treatment was not inconsistent with the use of the sensor in many applications where drift is a result of temperature and humidity. In general, the use of a dynamic background subtraction algorithm is required for chemiresistive sensors. The  $\sigma_{GAUSS}$  results of the TGS-2600 and TGS-2610 were similar, as expected due to their comparable sensing mechanisms. The baseline noise, as quantified by  $\sigma_{GAUSS}$ , for the CH<sub>4</sub> optimized TGS-2611 sensor was lower than the other TGS sensors.

### 3.4. Limits of Detection

The limits of detection were determined as discussed in Section 2.6 and are listed in Table 2. It should be noted these tests were carried out in a controlled environment and are best-case values. Of the tested CO<sub>2</sub> sensors, the Gascard, K-30, and Telaire sensors have comparable low limits of detection. The COZIR and Dynament single analyte sensor (MSH-P/CO<sub>2</sub>) have the next highest limits of detection, and the Dynament dual analyte sensor (MSH-DP/HC/CO<sub>2</sub>) has the highest. These results reflect the ordering of  $\sigma_{GAUSS}$  in Table 2. These results can also be correlated with the optical path lengths of the sensors, Larger path lengths resulting in greater sensitivity.

The limit of detection for the Dynament single analyte CH<sub>4</sub> sensor (MSH-P/HC) demonstrates the important influence of calibration uncertainties on the reported limit. For this sensor, the 3.54 ppm precision ( $\sigma_{GAUSS}$ ) or deviation around the mean was significantly lower than the 50 ppm resolution

quoted in the manufacturer's documentation at the low-end of the 1% concentration range of the sensor. The IUPAC limit of detection is 10.6 ppm. After correcting for the calibration error, the limit of detection is 170 ppm. This larger limit of detection is more consistent with expectations based the stated resolution of this sensor.

Of the tested chemiresistive CH<sub>4</sub> sensors, the TGS-2611 sensor had the lowest limit of detection under controlled conditions, primarily due to its stable baseline, which influences both the background noise and the quality of the calibration. Similar sensitivities have been reported for the TGS-2600 [29]. Both the non-methane optimized TGS-2600 and TGS-2610 sensors had similar limits of detection which were slightly larger than the CH<sub>4</sub> optimized TGS-2611 and the MQ-4. The MQ-4 sensor has precision between the TGS-2611 and both the TGS-2600 and TGS-2610 sensors. The 82 ppm limit of detection for the MQ-4 sensor is consistent with scatter in the calibration curve at concentration under 100 ppm (see Figure 4).

#### 4. Conclusion

The performance of a sensor platform for environmental or industrial monitoring is sensitive to the cost and performance of the individual sensor elements. The detection limits, accuracy, and precision for a range of potential sensors for measuring ppm concentrations of CO<sub>2</sub> and CH<sub>4</sub> for low-cost monitoring instrumentation quantified under consistent conditions and experimental methodologies. These parameters were quantified around the baseline atmospheric concentration for each analyte gas, which is approximately 400 ppm for CO<sub>2</sub> [22,23] and under 2 ppm for CH<sub>4</sub> [24–26]. The selected sensors were also evaluated at gas concentrations exceeding 1000 ppm, in order to mimic a leak from a CO<sub>2</sub> sequestration location or a gas well.

For CO<sub>2</sub>, the Gascard sensor had high-performance based on the detection limits and precision reported in Table 2, but is comparatively more expensive than the other sensors investigated. The Gascard sensor is also limited by reliance on active sampling, which requires a pump that introduces a source of additional power consumption, a mechanical failure point, and sampling complexity. Nevertheless, given that concentrations are typically reported in parts-per notation, the inclusion of temperature and pressure corrections on the Gascard sensor allow it to automatically correct the reported ppm value and eliminates the need for additional sensors. The lower cost K-30 and Telaire sensors also had good performance. The ability of the K-30 and Telaire sensors to operate by passive diffusion rather than mechanically pumped flow also reduces the cost and complexity of a potential monitoring system. The K-30 sensor, along with suggested corrections for improved accuracy, has also been investigated for ambient air monitoring. [38] One unique feature of the K-30 sensor is that it has an option to autozero and correct reported output by assuming the minimum concentration of a multi-day run is 400 ppm. If implemented, this feature has advantages for remote instrumentation.

For low-cost chemiresistive CH<sub>4</sub> sensors, the performance as expected at concentration greater than the measured limit of detection (Table 2) or, equivalently, greater than approximately 100 ppm. However, these sensors do not meet the ppm or even sub-ppm sensitivity required for environmental monitoring. The more expansive but still relatively low-cost optical NDIR CH<sub>4</sub> sensors (the GasCard CH<sub>4</sub> and MSH-P / HC) also perform well at higher concentrations, but lack the required sensitivity for monitoring environmental levels of CH<sub>4</sub> around the global average. Although progress has been made to produce viable sensors [15], larger, more costly devices are still required for precision measurements at the ppm level. With an increasing focus on remote sensing technology, a new generation of higher-sensitivity low-cost CH<sub>4</sub> sensors will be required to produce accurate data at expected atmospheric concentrations. The reported detection limits and precision for a range of available low-cost sensors, all measured using the same methodology and experimental equipment, provides a solid reference to which future improvements can be benchmarked against.

**Supplementary Materials:** The following are available online at <http://www.mdpi.com/1424-8220/xx/1/5/s1>: Table S1: Manufacturer listed properties of evaluated CO<sub>2</sub> sensors, Table S2: Manufacturer listed properties of evaluated CH<sub>4</sub> sensors.

**Author Contributions:** Conceptualization, X.X. and Y.Y.; methodology, X.X.; software, X.X.; validation, X.X., Y.Y. and Z.Z.; formal analysis, X.X.; investigation, X.X.; resources, X.X.; data curation, X.X.; writing—original draft preparation, X.X.; writing—review and editing, X.X.; visualization, X.X.; supervision, X.X.; project administration, X.X.; funding acquisition, Y.Y.

**Funding:** The authors acknowledge the financial support of this work by U.S. Department of Energy under contract DE-FOA-0000798. Additional support was provided by Oklahoma State University.

**Conflicts of Interest:** The authors declare no conflict of interest.

## Abbreviations

The following abbreviations are used in this manuscript:

CO <sub>2</sub>	Carbon dioxide
CH <sub>4</sub>	Methane
NDIR	Nondispersive Infrared
RMSE	Root-mean-squared error

- Keith, L.H.; Crummett, W.; Deegan, J.; Libby, R.A.; Taylor, J.K.; Wentler, G. Principles of Environmental Analysis. *Analytical Chemistry* **1983**, *55*, 2210–2218. doi:10.1021/ac00264a003.
- Yang, Z.; Li, N.; Becerik-Gerber, B.; Orosz, M. A Systematic Approach to Occupancy Modeling in Ambient Sensor-Rich Buildings. *SIMULATION* **2014**, *90*, 960–977. doi:10.1177/0037549713489918.
- Chung, W.Y.; Lee, S.C. A Selective AQS System with Artificial Neural Network in Automobile. *Proceedings of the Eleventh International Meeting on Chemical Sensors IMCS-11IMCS 2006IMCS 11* **2008**, *130*, 258–263. doi:10.1016/j.snb.2007.07.138.
- Won, W.; Lee, K.S. Nonlinear Observer with Adaptive Grid Allocation for a Fixed-Bed Adsorption Process. *Computers & Chemical Engineering* **2012**, *46*, 69–77. doi:10.1016/j.compchemeng.2012.07.001.
- Yi, P.; Xiao, L.; Zhang, Y. Remote Real-Time Monitoring System for Oil and Gas Well Based on Wireless Sensor Networks. *Mechanic Automation and Control Engineering (MACE), 2010 International Conference On*, 2010, pp. 2427–2429. doi:10.1109/MACE.2010.5535870.
- Somov, A.; Baranov, A.; Spirjakin, D.; Spirjakin, A.; Sleptsov, V.; Passerone, R. Deployment and Evaluation of a Wireless Sensor Network for Methane Leak Detection. *Selected Papers from the 26th European Conference on Solid-State Transducers Kraków, Poland, 9-12 September 2012* **2013**, *202*, 217–225. doi:10.1016/j.sna.2012.11.047.
- Pering, T.; Tamburello, G.; McGonigle, A.; Aiuppa, A.; Cannata, A.; Giudice, G.; Patanè, D. High Time Resolution Fluctuations in Volcanic Carbon Dioxide Degassing from Mount Etna. *Journal of Volcanology and Geothermal Research* **2014**, *270*, 115–121. doi:10.1016/j.jvolgeores.2013.11.014.
- Black, R.; (Mick) Meyer, C.; Yates, A.; Zweiten, L.V.; Mueller, J. Formation of Artefacts While Sampling Emissions of PCDD/PCDF from Open Burning of Biomass. *Chemosphere* **2012**, *88*, 352–357. doi:10.1016/j.chemosphere.2012.03.046.
- Guohua, H.; Lvy, W.; Yanhong, M.; Lingxia, Z. Study of Grass Carp (*Ctenopharyngodon Idellus*) Quality Predictive Model Based on Electronic Nose. *Sensors and Actuators B: Chemical* **2012**, *166–167*, 301–308. doi:10.1016/j.snb.2012.02.066.
- Karunanithi, S.; Din, N.M.; Hakimie, H.; Hua, C.K.; Omar, R.C.; Yee, T.C. Performance of Lab-scale Solar Powered Wireless Landfill Monitoring System. *Energy and Environment, 2009. ICEE 2009. 3rd International Conference On*, 2009, pp. 443–448. doi:10.1109/ICEENVIRON.2009.5398610.
- Shendell, D.G.; Therkorn, J.H.; Yamamoto, N.; Meng, Q.; Kelly, S.W.; Foster, C.A. Outdoor Near-Roadway, Community and Residential Pollen, Carbon Dioxide and Particulate Matter Measurements in the Urban Core of an Agricultural Region in Central CA. *Atmospheric Environment* **2012**, *50*, 103–111. doi:10.1016/j.atmosenv.2011.12.056.
- Wetchakun, K.; Samerjai, T.; Tamaekong, N.; Liawhiran, C.; Siri Wong, C.; Kruefu, V.; Wisitsoraat, A.; Tuantranont, A.; Phanichphant, S. Semiconducting Metal Oxides as Sensors for Environmentally Hazardous Gases. *Sensors and Actuators B: Chemical* **2011**, *160*, 580–591. doi:https://doi.org/10.1016/j.snb.2011.08.032.



13. Neri, G. First Fifty Years of Chemoresistive Gas Sensors. *Chemosensors* **2015**, *3*, 1–20. doi:10.3390/chemosensors3010001.
14. Frodl, R.; Tille, T. A High-Precision NDIR Gas Sensor for Automotive Applications. *IEEE Sensors Journal* **2006**, *6*, 1697–1705. doi:10.1109/JSEN.2006.884440.
15. Zhu, Z.; Xu, Y.; Jiang, B. A One Ppm NDIR Methane Gas Sensor with Single Frequency Filter Denoising Algorithm. *Sensors (Basel, Switzerland)* **2012**, *12*, 12729–12740. doi:10.3390/s120912729.
16. Bacsik, Z.; Mink, J.; Keresztury, G. FTIR Spectroscopy of the Atmosphere. I. Principles and Methods. *Applied Spectroscopy Reviews* **2004**, *39*, 295–363. doi:10.1081/ASR-200030192.
17. Coblenz Society, Inc.. Evaluated Infrared Reference Spectra. In *NIST Chemistry WebBook*; Lindstrom, P.J.; Mallard, W.G., Eds.; Number 69 in NIST Standard Reference Database, National Institute of Standards and Technology: Gaithersburg MD, 20899.
18. Albert, K.J.; Lewis, N.S.; Schauer, C.L.; Sotzing, G.A.; Stitzel, S.E.; Vaid, T.P.; Walt, D.R. Cross-Reactive Chemical Sensor Arrays. *Chemical Reviews* **2000**, *100*, 2595–2626. doi:10.1021/cr980102w.
19. Wang, C.; Yin, L.; Zhang, L.; Xiang, D.; Gao, R. Metal Oxide Gas Sensors: Sensitivity and Influencing Factors. *Sensors* **2010**, *10*, 2088–2106.
20. Prudenziati, M.; Morten, B. Thick-Film Sensors: An Overview. *Sensors and Actuators* **1986**, *10*, 65–82. doi:10.1016/0250-6874(86)80035-X.
21. Sekhar, P.K.; Kysar, J.; Brosha, E.L.; Kreller, C.R. Development and Testing of an Electrochemical Methane Sensor. *Sensors and Actuators B: Chemical* **2016**, *228*, 162 – 167. doi:https://doi.org/10.1016/j.snb.2015.12.100.
22. Blasing, T. Recent Greenhouse Gas Concentrations. Technical report, U. S. Department of Energy, CDIAC, 2016.
23. Dlugokencky, E. Trends in Atmospheric Methane. Technical report, NOAA/ESRL, 2016.
24. Turner, A.J.; Jacob, D.J.; Benmergui, J.; Wofsy, S.C.; Maasakkers, J.D.; Butz, A.; Hasekamp, O.; Biraud, S.C. A Large Increase in U.S. Methane Emissions over the Past Decade Inferred from Satellite Data and Surface Observations. *Geophysical Research Letters* **2016**, *43*, 2218–2224. doi:10.1002/2016GL067987.
25. Bamberger, I.; Stieger, J.; Buchmann, N.; Eugster, W. Spatial Variability of Methane: Attributing Atmospheric Concentrations to Emissions. *Environmental Pollution* **2014**, *190*, 65–74. doi:10.1016/j.envpol.2014.03.028.
26. Dlugokencky, E.; Tans, P. Trends in Atmospheric Carbon Dioxide. Technical report, NOAA/ESRL, 2016.
27. Karpov, E.c.; Karpov, c.F.; Suchkov, c.; Mironov, S.; Baranov, A.; Sleptsov, V.; Calliari, L. Energy Efficient Planar Catalytic Sensor for Methane Measurement. *Sensors and Actuators A: Physical* **2013**, *194*, 176 – 180. doi:https://doi.org/10.1016/j.sna.2013.01.057.
28. Chiu, S.W.; Tang, K.T. Towards a Chemiresistive Sensor-Integrated Electronic Nose: A Review. *Sensors* **2013**, *13*, 14214–14247. doi:10.3390/s131014214.
29. Eugster, W.; Kling, G.W. Performance of a Low-Cost Methane Sensor for Ambient Concentration Measurements in Preliminary Studies. *Atmospheric Measurement Techniques* **2012**, *5*, 1925–1934. doi:10.5194/amt-5-1925-2012.
30. van den Bossche, M.; Rose, N.T.; Wekker, S.F.J.D. Potential of a Low-Cost Gas Sensor for Atmospheric Methane Monitoring. *Sensors and Actuators B: Chemical* **2017**, *238*, 501 – 509. doi:https://doi.org/10.1016/j.snb.2016.07.092.
31. Williams, T.; Kelley, C. Gnuplot 5.0: An Interactive Plotting Program, 2016.
32. Barsan, N.; Schweizer-Berberich, M.; Göpel, W. Fundamental and Practical Aspects in the Design of Nanoscaled SnO<sub>2</sub> Gas Sensors: A Status Report. *Fresenius' Journal of Analytical Chemistry* **1999**, *365*, 287–304. doi:10.1007/s002160051490.
33. Ahlers, S.; Müller, G.; Doll, T. A Rate Equation Approach to the Gas Sensitivity of Thin Film Metal Oxide Materials. *Sensors and Actuators B: Chemical* **2005**, *107*, 587–599. doi:10.1016/j.snb.2004.11.020.
34. Long, G.L.; Winefordner, J.D. Limit of Detection A Closer Look at the IUPAC Definition. *Analytical Chemistry* **1983**, *55*, 712A–724A. doi:10.1021/ac00258a724.
35. Currie, L.A. Detection: International Update, and Some Emerging Di-Lemmas Involving Calibration, the Blank, and Multiple Detection Decisions. *Chemometrics and Intelligent Laboratory Systems* **1997**, *31*, 151–181. doi:https://doi.org/10.1016/S0169-7439(97)00009-9.

36. Mocak J.; Bond A. M.; Mitchell S.; Scollary G.. A Statistical Overview of Standard (IUPAC and ACS) and New Procedures for Determining the Limits of Detection and Quantification: Application to Voltammetric and Stripping Techniques (Technical Report). *Pure and Applied Chemistry* **2009**, *69*, 297. doi:10.1351/pac199769020297.
37. Benkstein, K.D.; Rogers, P.H.; Montgomery, C.B.; Jin, C.; Raman, B.; Semancik, S. Analytical Capabilities of Chemiresistive Microsensor Arrays in a Simulated Martian Atmosphere. *Sensors and Actuators B: Chemical* **2014**, *197*, 280 – 291. doi:https://doi.org/10.1016/j.snb.2014.02.088.
38. Martin, C.R.; Zeng, N.; Karion, A.; Dickerson, R.R.; Ren, X.; Turpie, B.N.; Weber, K.J. Evaluation and environmental correction of ambient CO<sub>2</sub> measurements from a low-cost NDIR sensor. *Atmospheric Measurement Techniques* **2017**, *10*, 2383–2395. doi:10.5194/amt-10-2383-2017.

© 2019 by the authors. Submitted to *Sensors* for possible open access publication under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).