

Tool Summary

Tool	Type	Least Count	Tolerance (ILE)	Rule Used (%)	Last Digits Allowed	Example
Ruler (Left Side)	Analog	0.5	0.5	100	0, 5	13.5 ± 0.5 mm
Ruler (Right Side)	Analog	1	0.5	50	0, 5	13.5 + 0.5 mm
Digital Caliper	Digital	0.01	0.01	100	0 thru 9	29.97 ± .01 mm
Scale	Analog	0.1	0.1	100	0 thru 9	19.3 ± .1 gm
Micrometer Caliper	Analog	0.01	0.005	50	0, 5	6.435 ± .005 mm
Graduated Cylinder	Analog	1	0.5	50	0, 5	22.5 + 0.5 ml

mm = millimeter, gm = gram, ml = milliliter