Серийный номер термотестера: 8196 Название измерения:

Период измерения: 15 мин

Выбран период с 09:00 30.06.2020 по 17:59 01.07.2020 Количество измерений: 132

Задан диапазон температур между: 2°C и 8°C Количество нарушений: 79

+ - нарушение температурного режима

Номер	Дата	Время	T, °C	Нар.
1	30.06.2020	09:12	26.0	+
2	30.06.2020	09:27	26.1	+
3	30.06.2020	09:42	26.2	+
4	30.06.2020	09:57	26.4	+
5	30.06.2020	10:12	26.5	+
6	30.06.2020	10:27	26.7	+
7	30.06.2020	10:42	26.8	+
8	30.06.2020	10:57	27.0	+
9	30.06.2020	11:12	27.1	+
10	30.06.2020	11:27	27.3	+
11	30.06.2020	11:42	27.4	+
12	30.06.2020	11:57	27.6	+
13	30.06.2020	12:12	27.7	+
14	30.06.2020	12:27	27.8	+
15	30.06.2020	12:42	27.9	+
16	30.06.2020	12:57	28.0	+
17	30.06.2020	13:12	28.1	+
18	30.06.2020	13:27	32.9	+
19	30.06.2020	13:42	34.4	+
20	30.06.2020	13:57	34.8	+
21	30.06.2020	14:12	35.3	+
22	30.06.2020	14:27	35.3	+
23	30.06.2020	14:42	32.7	+
24	30.06.2020	14:57	31.8	+
25	30.06.2020	15:12	31.3	+
26	30.06.2020	15:27	30.9	+
27	30.06.2020	15:42	30.5	+
28	30.06.2020	15:57	30.6	+
29	30.06.2020	16:12	19.3	+
30	30.06.2020	16:27	15.1	+

Номер	Дата	Время	T, °C	Нар.
31	30.06.2020	16:42	10.2	+
32	30.06.2020	16:57	8.1	+
33	30.06.2020	17:12	9.3	+
34	30.06.2020	17:27	8.0	+
35	30.06.2020	17:42	9.0	+
36	30.06.2020	17:57	8.1	+
37	30.06.2020	18:12	8.2	+
38	30.06.2020	18:27	8.3	+
39	30.06.2020	18:42	7.8	
40	30.06.2020	18:57	7.7	
41	30.06.2020	19:12	8.2	+
42	30.06.2020	19:27	8.4	+
43	30.06.2020	19:42	8.3	+
44	30.06.2020	19:57	7.9	
45	30.06.2020	20:12	7.6	
46	30.06.2020	20:27	7.5	
47	30.06.2020	20:42	8.1	+
48	30.06.2020	20:57	7.6	
49	30.06.2020	21:12	7.9	
50	30.06.2020	21:27	7.8	
51	30.06.2020	21:42	8.4	+
52	30.06.2020	21:57	7.7	
53	30.06.2020	22:12	7.4	
54	30.06.2020	22:27	7.8	
55	30.06.2020	22:42	8.2	+
56	30.06.2020	22:57	7.6	
57	30.06.2020	23:12	8.1	+
58	30.06.2020	23:27	7.5	
59	30.06.2020	23:42	8.0	+
60	30.06.2020	23:57	7.4	

Номер	Дата	Время	T, °C	Нар.
61	01.07.2020	00:12	7.9	
62	01.07.2020	00:27	7.4	
63	01.07.2020	00:42	7.8	
64	01.07.2020	00:57	8.0	+
65	01.07.2020	01:12	7.6	
66	01.07.2020	01:27	7.1	
67	01.07.2020	01:42	7.6	
68	01.07.2020	01:57	7.6	
69	01.07.2020	02:12	7.5	
70	01.07.2020	02:27	7.8	
71	01.07.2020	02:42	7.2	
72	01.07.2020	02:57	7.6	
73	01.07.2020	03:12	7.0	
74	01.07.2020	03:27	7.4	
75	01.07.2020	03:42	7.3	
76	01.07.2020	03:57	7.2	
77	01.07.2020	04:12	7.6	
78	01.07.2020	04:27	7.0	
79	01.07.2020	04:42	7.4	
80	01.07.2020	04:57	6.8	
81	01.07.2020	05:12	7.3	
82	01.07.2020	05:27	7.1	
83	01.07.2020	05:42	7.0	
84	01.07.2020	05:57	7.4	
85	01.07.2020	06:12	6.8	
86	01.07.2020	06:27	7.2	
87	01.07.2020	06:42	6.7	
88	01.07.2020	06:57	7.1	
89	01.07.2020	07:12	6.6	
90	01.07.2020	07:27	7.1	

Номер	Дата	Время	т, °С	Нар.
91	01.07.2020	07:42	6.7	
92	01.07.2020	07:57	7.2	
93	01.07.2020	08:12	6.8	
94	01.07.2020	08:27	6.9	
95	01.07.2020	08:42	7.0	
96	01.07.2020	08:57	6.7	
97	01.07.2020	09:12	6.6	
98	01.07.2020	09:27	7.0	
99	01.07.2020	09:42	7.0	
100	01.07.2020	09:57	7.4	
101	01.07.2020	10:12	9.8	+
102	01.07.2020	10:27	12.1	+
103	01.07.2020	10:42	18.3	+
104	01.07.2020	10:57	19.4	+
105	01.07.2020	11:12	19.6	+
106	01.07.2020	11:27	19.7	+
107	01.07.2020	11:42	19.7	+
108	01.07.2020	11:57	19.7	+
109	01.07.2020	12:12	25.4	+
110	01.07.2020	12:27	29.0	+
111	01.07.2020	12:42	28.1	+
112	01.07.2020	12:57	27.9	+
113	01.07.2020	13:12	27.9	+
114	01.07.2020	13:27	28.1	+
115	01.07.2020	13:42	28.4	+
116	01.07.2020	13:57	28.5	+
117	01.07.2020	14:12	28.8	+
118	01.07.2020	14:27	28.8	+
119	01.07.2020	14:42	28.5	+
120	01.07.2020	14:57	27.8	+

Серийный номер термотестера: 8196 Название измерения:

Период измерения: 15 мин

Выбран период с 09:00 30.06.2020 по 17:59 01.07.2020 Количество измерений: 132

Задан диапазон температур между: 2°C и 8°C Количество нарушений: 79

+ - нарушение температурного режима

Номер	Дата	Время	T, °C	Нар.
121	01.07.2020	15:12	27.8	+
122	01.07.2020	15:27	27.4	+
123	01.07.2020	15:42	26.7	+
124	01.07.2020	15:57	26.4	+
125	01.07.2020	16:12	26.1	+
126	01.07.2020	16:27	26.3	+
127	01.07.2020	16:42	26.0	+
128	01.07.2020	16:57	26.1	+
129	01.07.2020	17:12	26.2	+
130	01.07.2020	17:27	26.2	+
131	01.07.2020	17:42	26.4	+
132	01.07.2020	17:57	26.7	+

Номер	Дата	Время	T, °C	Нар.

Номер	Дата	Время	T, °C	Нар.

Номер	Дата	Время	T, °C	Нар.

Серийный номер термотестера: 8196 Название измерения:

Период измерения: 15 мин

Выбран период с 09:00 30.06.2020 по 17:59 01.07.2020 Количество измерений: 132

Задан диапазон температур между: 2°C и 8°C Количество нарушений: 79



