

Thermo Scientific Varioskan Flash Spectral Scanning Multimode Reader

Optimal performance for demanding research assays



Thermo Scientific Varioskan Flash together with Skanlt Software offers optimal performance for demanding research assays

Thermo Scientific Varioskan Flash is a spectral scanning multimode reader, including fluorescence intensity, time-resolved fluorescence (TRF), photometric, and luminometric detection technologies. It offers optimal performance for demanding research assays with unlimited wavelength selection, up to three onboard dispensers, utmost optical performance and the advanced Thermo Scientific Skanlt Software

Unlimited wavelength selection for assay optimization

The Thermo Scientific Varioskan Flash spectral scanning multimode reader combines fluorescence intensity, time-resolved fluorescence (TRF), photometric, and optional luminometric detection technologies.

It provides unlimited wavelength selection, and thereby allows both spectral analysis and measurement at any single wavelength. This gives ultimate flexibility for identifying the optimal measurement wavelength for any assay, now and in the future.

High-performance optical technology

High-quality performance is obtained with the optimized optics of Varioskan® Flash. In fluorometry, stray light transmission is minimized using quadruple monochromators. This ensures ultimate spectral quality that quarantees superior assay sensitivity. In addition, the quadruple monochromator optics offers excellent performance and flexibility in measuring either single or multilabel assays over a wide concentration range. Furthermore, the bottom reading possibility enables the user to select the optimal reading position for fluorometric measurement.

In photometry, double monochromator design produces a very low stray light level ensuring excellent linearity. Outstanding accuracy and precision are guaranteed with the perfect beam optics.

Varioskan Flash has two dedicated detection optics for luminometry: Scanning optics that is excellent for studying and optimizing luminometric assay parameters, and the high sensitivity Varioskan LumiSens optics, specially designed for measurement of any luminometric assay with excellent sensitivity, including multilabel assays requiring wavelength selection.


Onboard dispensers for exact follow-up of kinetic reactions


For easy and accurate reagent additions, Varioskan Flash can be equipped with up to three onboard dispensers.

The dispenser facilitates the work and allows easier optimization of assays. The instrument supports simultaneous dispensing and measurement, thereby enabling follow-up of kinetic reactions directly from reaction onset. This capability is essential for flash type luminescence reactions, Ca²⁺ flux studies and other rapid kinetic applications. The ability to add

reagents in any order or in any phase of the kinetic assay allows execution of sequential multistep assays, such as ATP and reporter gene applications.

In addition, automated dispensing guarantees reproducible dispensing from day to day and from person to person.





The utmost sensitivity and a wide dynamic range is achieved by Varioskan Flash's optical system with quadruple monochromators.

Varioskan Flash supports various detection technologies.

(BRET, FRET, TR-FRET)

Varioskan Flash capabilities provide versatility for assays

Assays requiring any measurement wavelength from low UV to near IR can be performed due to a wide spectral range of the Varioskan Flash. Photometric pathlength correction is ideal for direct measurement of DNA, RNA and proteins. Fluorometric UV measurement of fluorescent amino acids is an efficient tool in labelfree assays for protein structure changes.

Fluorescent cell based assays provide high-performance with Varioskan Flash due to selectable top or bottom reading, flexibility in plate formats and a high-precision incubator. With TRF technology, Varioskan Flash offers great flexibility for high-performance cellular assays without interference from biological background. Timeresolved fluorescence resonance energy transfer (TR-FRET) assays are easily optimized using the TRF spectral scanning function of the Varioskan Flash. Additional TRF delay time optimization using automatic τ value calculation makes it straightforward to define the best possible measurement parameters.

The luminometric monochromator mode of Varioskan Flash offers a unique possibility to optimize measurement wavelengths in luminometric multilabel assays. Thereafter, assays can be performed with superior sensitivity with the normal or filter mode of the dedicated Varioskan LumiSens optics.

High-performance incubator for controlled assay conditions

To maintain optimal and constant reaction conditions, Varioskan Flash has a high-performance onboard incubator that ensures controlled assay conditions. It is ideal for cellular assays, enzyme assays and other applications where temperature control is essential.

Specially designed for automation

Varioskan Flash has been designed for easy integration with automated systems. The robotic plate carrier of Varioskan Flash is specially designed for convenient robot arm access, allowing microplate gripping in both portrait and landscape configurations. Thermo Scientific Skanlt Software also has a special remote control interface for integration with automated systems and LIMS.


Logical assay setup using SkanIt Software

The powerful Skanlt® Software allows easy assay optimization, flexible data handling and convenient report formatting. The unique steplist feature makes assay setup highly visual and flexible, and the workflow logical and easy to follow.

There are two editions of Skanlt Software available: a Research Edition for scientists working in life science research, and a Drug Discovery Edition offering features needed for compliance with the FDA's 21 CFR Part 11, for the drug discovery industry.

Applications:

- Apoptosis assays
- Ca2+ flux assays
- Cell proliferation
- · Cellular assays
- Cytotoxicity and ADMETox
- Direct DNA, RNA and protein quantitation
- ELISA/FIA/TRF-ELISA assays
- Enzyme kinetic studies
- Europium assays
- FRET assays
- TR-FRET assays
- BRET assays
- GPCR assays
- Ion channel assays
- · Kinase assays
- Multilabel assays
- Reporter gene assays
- Signal transduction
- Tryptophan and tyrosine UV fluorescence



The Steplist in Skanlt Software makes assay setup visual and logical.



Varioskan Flash spectral scanning multimode reader



Onboard dispensers enable optimization of kinetic assays and facilitate assay development.

Plate types 5-1556-well plates Double excitation and double emission monochromators	Fluorescence Intensity/Time-Resolve	d Fluorescence
Wavelength selection Double excitation and double emission monochromators	•	
Existation wavelength range Emission wavelength range Excitation/emission bandwidth Sm and 12 mm/12 mm Light source Sensitivity/dynamic range Fluorescence intensity, top reading: < 0.4 fmol fluorescein/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading: < 4 fmol fluorescein/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading: < 4 fmol fluorescein/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading: < 20 amol Europium/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading: < 20 amol Europium/well, > 5 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 5 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 5 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Fluorescence, top reading: < 120 amol ATP/well plates Fluorescence, top reading: < 120 amol ATP/well plate Fluorescence, top reading: <		·
Existation/emission bandwidth Existation/emission bandwidth Light source Sensitivity/dynamic range Fluorescence intensity, top reading: - 0.4 fmol fluorescein/well, - 6 decades, 384-well plate fluorescence intensity, bottom reading: - 4 fmol fluorescein/well, - 5 decades, 384-well plate fluorescence intensity, bottom reading: - 4 fmol fluorescein/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, top reading: - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, september fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Europeum/well, - 5 decades, 384-well plate fluorescence, and - 120 amol Security fluorescence, 384-well plates Wavelength range Wavelength range Wavelength selection Double monochromators Wavelength range 200 – 1000 nm Accuracy 200 – 1000 Abs, whichever is greater, at 200 – 399 nm (0 – 2 Abs) 210 – 22% or 0000 Abs, whichever is greater, at 200 – 399 nm (0 – 2 Abs) 210 – 22% or 0000 Abs, whichever is greater, at 200 – 399 nm (0 – 2 Abs) 210 – 22% or 0000 Abs, whichever is greater, at 200 – 1000 nm (0 – 3 Abs) 210 – 22% or 0000 Abs, whichever is greater, at 200 – 1000 nm (0 – 3		
Excitation/emission bandwidth Light source Xenon flash lamp Fluorescence intensity, top reading < 0.4 fmol fluorescenin/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading; < 24 fmol fluorescenin/well, > 5 decades, 384-well plate fluorescence intensity, bottom reading; < 24 fmol fluorescenin/well, > 5 decades, 384-well plate fluorescence intensity, bottom reading; < 20 amol Europium/well, > 5 decades, 384-well plate fluorescence, top reading < 20 amol Europium/well, > 5 decades, 384-well plate fluorescence intensity, bottom reading; < 24 fmol fluorescenin/well, > 5 decades, 384-well plate fluorescence intensity, bottom reading; < 24 fmol fluorescenin/well, > 5 decades, 384-well plate fluorescence intensity, bottom reading; < 24 fmol fluorescenin/well, > 5 decades, 384-well plates Luminometry Plate types 8 - 158-well plates and double monochromators Wavelength range 8 - 160 fm m, spectral scanning 270 - 840 nm 9 Avelength selection Wavelength range 200 - 1000 nm 8 andwidth 5 mm Light source Xenon flash lamp 1 - 3 Abs; 384-well plate; at 450 nm, ± 2% 0 - 3 Abs; 384-well plate; at 450 nm, ± 2% 1 - 2 x or 0.003 Abs, whichever is greater, at 200 - 399 nm (0 - 2 Abs) 1 + 1% or 0.003 Abs, whichever is greater, at 400 - 1000 nm (0 - 3 Abs) Dispenser up to 3, automatic adapte outloaded on maximum well volume Accuracy 1 - 10 0.001 µ, with 1 µ increments (1 mi syringe) Automatic adapte outloaded on maximum well volume Accuracy 1 - 10 0.001 µ, with 1 µ increments (1 mi syringe, 0.40 nm tip) Dispensing speed 1 - 10 0.001 µ, with 1 µ increments (1 mi syringe, 0.40 nm tip) Dispensing speed 3 - 10 yellowell, 1 mil syringe, 0.40 nm tip) Incubator From ambient + 4**C to 45*C at ambient 25*C Shaker Orbital with adjustable speed and diameter General Features Measurement speed 96 well p		
Light source Xenon flash lamp Fluorescence intensity, top reading: < 0.4 fmol fluorescein/well, > 6 decades, 384-well plate Fluorescence intensity, bottom reading: < 4 fmol fluorescein/well, > 5 decades, 384-well plate Fluorescence intensity, bottom reading: < 4 fmol fluorescein/well, > 6 decades, 384-well plate Time-resorved fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Time-resorved fluorescence, top reading: < 120 amol Europium/well, > 6 decades, 384-well plate Plate types 6 = 1536-well plates, spectral scanning 270 = 940 rm Wavelength selection All wavelengths, filters and double monochromators Wavelength range 380 = 707 nm, spectral scanning 270 = 940 rm Sensitivity/dynamic range Photometry Plate types 6 = 384-well plates Wavelength selection Double monochromators Wavelength selection Double monochromators Wavelength range 200 = 1000 nm Bandwidth Snm Light source Xenon flash lamp Light source Linear measurement range 0 = 4 Abs (98-well plates) at 450 rm, ± 2% 0 = 3 Abs (384-well plate) at 450 rm, ± 2% 0 = 3 Abs (384-well plate) at 450 rm, ± 2% Accuracy ± 2% or 0003 Abs, whichever is greater, at 400 = 1000 nm (0 = 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 rm (0 = 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 rm (0 = 3 Abs) In (standand), Sml (on request) Dispensing volume 1 = 1000 upl, with 1 µl increments (1 ml syringe) Accuracy < 0.2 µl or 2%, whichever is greater, at 450 rm mitp) Precision 5 = 19 µl < 5%, 20 = 1000 upl <2% (1 ml syringe, 0.40 mm tip) Dispensing speed Measurement modes Measurement modes Measurement modes Measurement modes Measurement modes Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1538-well plate in in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed Overall dimensions		
Sensitivity/dynamic range		
Fluorescence intensity, bottom reading: <4 fmol fluorescenin/well, > 5.5 decades, 384-well plate Time-resolved fluorescence, top reading: <120 amol Europium/well, > 6 decades, 384-well plate Luminometry Plate types 6 – 1936-well plates, spectral scanning 6 – 394-well plates Wavelength selection All wavelengths, filters and double monochromators Wevelength range 360 – 670 nm, spectral scanning 270 – 840 nm Sensitivity/dynamic range <7 amol Alf/well, > 7 decades, flash Alf reaction, 384-well plate Photometry Plate types 6 – 384-well platas Wavelength selection Double monochromators Wavelength range 200 – 1000 nm Bandwidth 5 nm Light source Xenon flash lamp Linear measurement range 0 – 4 Abs (96-well plate) at 450 nm, ± 2% 0 – 3 Abs (384-well plate) at 450 nm, ± 2% 0 – 3 Abs (384-well plate) at 450 nm, ± 2% 0 – 3 Abs (384-well plate) at 450 nm, ± 2% 0 – 3 Abs (384-well plate) at 450 nm (0 – 2 Abs) 1 five or 1003 Abs, whichever is greater, at 200 – 390 nm (0 – 2 Abs) 2 five or 1003 Abs, whichever is greater, at 400 – 1000 nm (0 – 3 Abs) Precision SD <0.003 Abs, swhichever is greater, at 450 nm (0 – 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 – 384-well plates 3 yringe size 1 mil (standard), 5 mil (on request) Dispensing volume 1 – 10000 µ, with 1 µ increments (1 ml syringe) Automatic selety control based on maximum well volume Accuracy <0.2 µ or 2%, whichever is greater, 5 – 10 000 µ (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate 80 s, 384-well plate 60 s, 384-well plate 6		·
Plate types 6 – 1536-well plates, spectral scanning 6 – 384-well plates Wavelength selection All wavelengths, filters and double monochromators Wavelength range 300 – 670 nm, spectral scanning 270 – 840 nm Sensitivity/dynamic range < 7 amol ATP/well, > 7 decades, flash ATP reaction, 384-well plate Photometry Plate types 6 – 384-well plates Wavelength selection Double monochromators Wavelength range 200 – 1000 nm Bandwidth 5 nm Light source Xenon flash lamp Linear measurement range 0 – 4 Abs (96-well plate) at 450 nm, ± 2%, 0 – 3 Abs (384-well plate) at 450 nm, ± 2%, 0 – 3 Abs (384-well plate) at 450 nm, ± 2%, 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-well plate) at 450 nm, ± 2 – 3 abs (384-wel	Sensitivity/ dynamic range	Fluorescence intensity, bottom reading: < 4 fmol fluorescein/well, > 5.5 decades, 384-well plate
Wavelength selection All wavelengths, filters and double monochromators Wavelength range Sensitivity/dynamic range Photometry Plate types 6 − 384-well plates Wavelength selection Double monochromators Wavelength selection Double monochromators Wavelength range 200 − 1000 nm Bandwidth 5 nm Light source Xenon flash lamp 0 − 4 Abs [96-well plate] at 450 nm, ± 2% 0 − 3 Abs [384-well plate] at 450 nm, ± 2% 4 ccuracy 2 2 % or 0.003 Abs, whichever is greater, at 200 − 399 nm (0 − 2 Abs) 1 % or 0.003 Abs, whichever is greater, at 400 − 1000 nm (0 − 3 Abs) Precision Dispenser Up to 3, automatic dispensing position control Plate types 5 − 384-well plates Syringe size 1 m (standard), 5 ml (on request) Dispensing volume 1 − 10 000 µl, with 1 µl increments [1 ml syringe) Accuracy 4 c 2 µl or 2 m with 1 µl increments [1 ml syringe] Automatic safety control based on maximum well volume Accuracy 4 c 2 µl µl c 5%, 20 − 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 50 s 384-well plate 50 s 384-well plate 60 s 384-well plate 60 s 384-well plate 90 s 384-well plate 1 − 10 000 µl × with 1 µl increments [1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 50 s 384-well plate 60 s 384-well plate 60 s 384-well plate 90 s 384-well plate 1 ml (standard), 5 ml (on request) 1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 90 s 384-well plate 90 s 384-well plate 90 s 384-well plate 1 ml syringe, 0.40 mm tip) Dispensing speed 90 s 384-well plate 90 well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed Overall dimensions 5 40 mm (W) x 500 mm (H)	Luminometry	
Wavelength range 360 – 670 nm, spectral scanning 270 – 840 nm	Plate types	6 – 1536-well plates, spectral scanning 6 – 384-well plates
Sensitivity/dynamic range <7 amol ATP/well, >7 decades, flash ATP reaction, 384-well plate	Wavelength selection	All wavelengths, filters and double monochromators
Photometry Plate types 6 – 384-well plates Wavelength selection Double monochromators Wavelength range 200 – 1000 nm Bandwidth 5 nm Light source Xenon flash lamp 0 – 4 Abs (96-well plate) at 450 nm, ± 2% 0 – 3 Abs (384-well plate) at 450 nm, ± 2% 2 % or 0.003 Abs, whichever is greater, at 200 – 399 nm (0 – 2 Abs) ± 1% or 0.003 Abs, whichever is greater, at 400 – 1000 nm (0 – 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 – 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 – 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 – 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 – 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 µl < 5%, 20 – 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Wavelength range	360 – 670 nm, spectral scanning 270 – 840 nm
Plate types 6 − 384-well plates Wavelength selection Double monochromators Wavelength range 200 − 1000 nm Bandwidth 5 nm Light source Xenon flash lamp 0 − 4 Abs (96-well plate) at 450 nm, ± 2% 0 − 3 Abs (384-well plate) at 450 nm, ± 2% Accuracy ± 2% or 0.003 Abs, whichever is greater, at 200 − 399 nm (0 − 2 Abs) ± 1% or 0.003 Abs, whichever is greater, at 400 − 1000 nm (0 − 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 − 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 − 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 − 10.000 μl, with 1 μl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 μl or 2%, whichever is greater, 5 − 10.000 μl (1 ml syringe, 0.40 mm tip) Precision 5 − 19 μl < 5%, 20 − 10.000 μl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate 6 (5 μl/Well), 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 − 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Sensitivity/dynamic range	< 7 amol ATP/well, > 7 decades, flash ATP reaction, 384-well plate
Wavelength selection Double monochromators Wavelength range 200 − 1000 nm Bandwidth 5 nm Light source Xenon flash lamp Under measurement range 0 − 4 Abs (96-well plate) at 450 nm, ± 2% 0 − 3 Abs (384-well plate) at 450 nm, ± 2% Accuracy ± 2% or 0.003 Abs, whichever is greater, at 200 − 399 nm (0 − 2 Abs) ± 1% or 0.003 Abs, whichever is greater, at 450 − 1000 nm (0 − 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 − 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 − 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 − 10 000 µl, with 1 µl increments (1 ml syringe) Accuracy < 0.2 µl or 2%, whickever is greater, 5 − 10 000 µl (1 ml syringe). 0.40 mm tip) Precision 5 − 19 µl < 5%, 20 − 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate 6 yl /well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4*°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic,	Photometry	
Wavelength range 200 – 1000 nm	Plate types	6 – 384-well plates
Bandwidth 5 nm Light source Xenon flash lamp Linear measurement range 0 − 4 Abs (96-well plate) at 450 nm, ± 2% 0 − 3 Abs (384-well plate) at 450 nm, ± 2% Accuracy ± 2% or 0.003 Abs, whichever is greater, at 200 − 399 nm (0 − 2 Abs) ± 1% or 0.003 Abs, whichever is greater, at 400 − 1000 nm (0 − 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 − 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 − 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 − 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 − 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 − 19 µl < 5%, 20 − 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate 90 s, 3	Wavelength selection	Double monochromators
Light source Linear measurement range 0 - 4 Abs (96-well plate) at 450 nm, ± 2% 0 - 3 Abs (384-well plate) at 450 nm, ± 2% 4 2% or 0.003 Abs, whichever is greater, at 200 - 399 nm (0 - 2 Abs) ± 1% or 0.003 Abs, whichever is greater, at 400 - 1000 nm (0 - 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 - 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 - 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 - 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 - 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 - 19 µl < 5%, 20 - 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 334-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed 0 < 2 s/well, 400 - 500 mm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Wavelength range	200 – 1000 nm
Linear measurement range	Bandwidth	5 nm
Accuracy	Light source	Xenon flash lamp
# 1% or 0.003 Abs, whichever is greater, at 400 – 1000 nm (0 – 3 Abs) Precision SD < 0.001 Abs or CV < 0.5%, whichever is greater, at 450 nm (0 – 3 Abs) Dispenser up to 3, automatic dispensing position control Plate types 6 – 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 – 10 000 μl, with 1 μl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy <0.2 μl or 2%, whichever is greater, 5 – 10 000 μl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 μl < 5%, 20 – 10 000 μl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 μl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Linear measurement range	
Dispenser up to 3, automatic dispensing position control Plate types 6 − 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 − 10 000 μl, with 1 μl increments (1 ml syringe)	Accuracy	, , ,
Plate types 6 – 384-well plates Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 – 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 – 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 µl < 5%, 20 – 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 384-well plate 80 s, 384-well plate 65 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 – 500 mm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (U) x 500 mm (H)	Precision	SD $<$ 0.001 Abs or CV $<$ 0.5%, whichever is greater, at 450 nm (0 $-$ 3 Abs)
Syringe size 1 ml (standard), 5 ml (on request) Dispensing volume 1 – 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy <0.2 µl or 2%, whichever is greater, 5 – 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 µl < 5%, 20 – 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Dispenser	up to 3, automatic dispensing position control
Dispensing volume 1 – 10 000 µl, with 1 µl increments (1 ml syringe) Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 – 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 µl < 5%, 20 – 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Plate types	6 – 384-well plates
Automatic safety control based on maximum well volume Accuracy < 0.2 µl or 2%, whichever is greater, 5 – 10 000 µl (1 ml syringe, 0.40 mm tip) Precision 5 – 19 µl < 5%, 20 – 10 000 µl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Syringe size	1 ml (standard), 5 ml (on request)
Precision 5 – 19 μl < 5%, 20 – 10 000 μl < 2% (1 ml syringe, 0.40 mm tip) Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 μl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed < 2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Dispensing volume	
Dispensing speed 30 s, 96-well plate 80 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Accuracy	< 0.2 μl or 2%, whichever is greater, 5 – 10 000 μl (1 ml syringe, 0.40 mm tip)
80 s, 384-well plate (5 µl/well, 1 ml syringe, 0.40 mm tip) Incubator From ambient + 4°C to 45°C at ambient 25°C Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Precision	5 – 19 μl < 5%, 20 – 10 000 μl < 2% (1 ml syringe, 0.40 mm tip)
Shaker Orbital with adjustable speed and diameter General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Dispensing speed	80 s, 384-well plate
General Features Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Incubator	From ambient + 4°C to 45°C at ambient 25°C
Measurement modes Endpoint, kinetic, spectral scanning with all detection technologies Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Shaker	Orbital with adjustable speed and diameter
Measurement speed 96-well plate in 15 s, 384-well plate in 45 s, and 1536-well plate in 135 s (minimum kinetic interval time from A1 back to A1) Spectral scanning speed <2 s/well, 400 – 500 nm, 1 flash, 2 nm steps Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	General Features	
A1 back to A1) Spectral scanning speed	Measurement modes	Endpoint, kinetic, spectral scanning with all detection technologies
Overall dimensions 540 mm (W) x 580 mm (D) x 500 mm (H)	Measurement speed	
	Spectral scanning speed	< 2 s/well, 400 – 500 nm, 1 flash, 2 nm steps
	Overall dimensions	

Ordering Information	
Product Code	Description
5250030	Varioskan Flash
5250040	Varioskan Flash, including bottom reading
5250500	Varioskan LumiSens option (also enabling luminometric spectral scanning)
5250510	Dispenser option (1st, 2nd or 3rd)

North America: USA/Canada 800 522 7763

Europe: Austria +43 1 801 40 0, Belgium +32 2 482 30 30, France +33 2 2803 2180, Germany national toll free 08001-536 376, Germany international +49 6184 90 6940, Italy +39 02 02 95059 448, Netherlands +31 76 571 4440, Nordic countries +358 9 329 100, Russia/CIS +7 (495) 739 76 41, Spain/Portugal +34 93 223 09 18, Switzerland +41 44 454 12 12, UK/Ireland +44 870 609 9203 **Asia:** China +86 21 6865 4588 or +86 10 8419 3588, India toll free 1800 22 8374 India +91 22 6716 2200, Japan +81 45 453 9220, Other Asian countries +852 2885 4613 **Countries not listed:** +49 6184 90 6940 or +33 2 2803 2180

www.thermo.com/readingroom www.thermo.com/mpi

© 2009 Thermo Fisher Scientific Inc.
All rights reserved. All trademarks
are the property of Thermo Fisher
Scientific Inc., and its subsidiaries.
Specifications, terms and pricing are
subject to change. Not all products
are available in all countries. Please
consult your local sales representative for details.

