Compton Streuung

Elektronen außer Rand und Band

Friedrich Schüßler, Volker Karle

April 24, 2015

Assistent: Kilian Rosbach

Inhaltsverzeichnis

Foto des Aufbaus

Aufbau ohne Elektronik



Aufbau mit Elektronik



Zerfallsschemata von ^{137}Cs und ^{22}Na





Wie sieht ein Compton Peak aus? Klein-Nishina Formel!



NaI Szintillator, ^{137}Cs Probe, mit PVC



NaI Szintillator, ^{137}Cs Probe, mit PVC



Sichtbare Peaks und Kanten für beide Szintillisatoren

Probe	Peak/Kante	E / keV	NaI / Channel	PVC / Channel
^{137}Cs	Photo	662	8040.59 ± 0.03	
	Compton	477	5720 ± 4	178.9 ± 0.3
	Rückstreu	183	2510 ± 12	
^{22}Na	Photo	511	6347 ± 3	
	Compton	341	4000 ± 2000	108 ± 2
	Photo	1277	14180 ± 20	
	Compton	1064	12000 ± 4000	414 ± 4

Linearer fit für NaI Szintillator



^{22}Na sample (measurement time 16.5h)



^{137}Cs sample (measurement time 6h)



Linear fit


^{22}Na
^{137}Cs
Energy / keV
Channel
0
200
400
600
800
1000
1200
1400
0
100
200
300
400
500
600
700
Ch.(E) = $aE + b$
$a = [0.413 \pm 0.003]$ Ch./keV
$b = [-20 \pm 2]$Ch.

Peaks and fitting results of ^{137}Cs

Name	Energy	Channel
Photo peak	662 keV	8040.59 ± 0.03
Compton edge	477 keV	5720 ± 4
Escape peak	183 keV	2510 ± 12

^{137}Cs sample (measurement time 2.7h)



^{22}Na sample (measurement time about 1h)



Linear fit


Photo peak(Cs)
Compton(Cs)
Escape peak(Cs)
1.Photo peak(Na)
1.Compton edge(Na)
2.Photo peak(Na)
2.Compton edge(Na)
Ch.(E) = aE + b
a = [11.632 ± 0.012] Ch./keV
b = [340 ± 8]Ch.
Energy / keV
Channel
0
500
1000
1500
2000
0
2000
4000
6000
8000
10000
12000
14000
16000

Energieerhaltung



Differentieller Wirkungsquerschnitt



Energy Conservationation

- Comparison of peak energies for different angles

Background of the PS scintillator with coincidence and random coincidences (measurem. time 13.4h and 1h)

Energy of electrons: Rate of coincident events of PS scintillator at angle of $\theta = 90°$



Energy of electrons: Rate of coincident events of PS scintillator at angle of $\theta = 15°$



Background of NaI scintillator with coincidences (measurem. time 62h)



Energy of photons: Rate of coincident events of NaI scintillator at angle $\theta = 30°$



Energy of photons: Rate of coincident events of NaI scintillator at angle $\theta = 90°$



Now to the result: combining all those peaks...

NaI szintillator, ^{137}Cs Probe, ohne PVC