



World Leader in Testing Services
for the Medical Device Industry

2261 Tracy Road
Northwood, OH 43619
Phone 419-666-9459
Fax 419-666-2954

Master Bond, Inc.
154 Hobart Street
Hackensack, NJ 07601
Attn: DR. WALTER BRENNER

LAB NO. 91T-09580-00
P.O. NO. 06021091
ID NO. Not Supplied

CERTIFICATE OF COMPLIANCE

USP BIOLOGICAL TESTS
CLASSIFICATION VI

Test Article: Epoxy System EP21LV

ACUTE SYSTEMIC TOXICITY (USP): The saline, alcohol in saline, polyethylene glycol 400 and cottonseed oil extracts of the test article injected into mice did not produce a significantly greater systemic reaction than the blank extractant.

INTRACUTANEOUS TOXICITY (USP): The saline, alcohol in saline, polyethylene glycol 400 and cottonseed oil extracts of the test article injected intracutaneously in rabbits did not produce a significantly greater tissue reaction than the blank extractant.

IMPLANTATION TEST (USP): The macroscopic reaction of the test article implanted 7 days was not significant as compared to the USP Negative control plastic.

The sample of test article extracted at a ratio of 60sq.cm/20 ml and at a temperature of 50°C for 72 hour(s) met the requirements of a USP Class VI Plastic.