

# COLDPROOF

DIGITAL PHARMACOVIGILANCE REPORT . . .

## COLDPROOF EXCURSION REPORT

---

Fridge ID Domestic/Bar Fridge (Non-Compliant)  
Date Range Analysed Not specified

|                      |           |
|----------------------|-----------|
| GEYER Classification | GEYER RED |
| Outcome Status       | CRITICAL  |

# Clinical Summary

|                               |                                     |
|-------------------------------|-------------------------------------|
| Fridge Name                   | Domestic/Bar Fridge (Non-Compliant) |
| Model                         | Model not specified                 |
| Date Range                    | Not specified                       |
| Total Data Points             | 32717                               |
| Number of Excursions          | 29                                  |
| Longest Excursion             | 0.0 minutes                         |
| Time Above 8°C                | 0.0 minutes                         |
| Time Below 2°C                | 0.0 minutes                         |
| Classification Rule Triggered | REPORTABLE BREACH                   |

## GEYER Classification

|                         |
|-------------------------|
| <b>GEYER RED</b>        |
| Reason: Critical breach |

## Pharmacist Review

**Summary Interpretation:** The refrigerator demonstrated temperature behaviour that may be inconsistent with Strive-for-5 guidelines. A risk-based assessment is indicated.


**Action:**

- Quarantine stock where appropriate.
- Supervisor pharmacist review required.
- Document the outcome in the local clinical governance system.

# Fridge Signature

|                       |                  |
|-----------------------|------------------|
| Mean temperature      | N/A              |
| Min / max             | 5.4 °C / 12.7 °C |
| Drift                 | N/A              |
| Spike frequency       | N/A              |
| Cooling return slope  | N/A              |
| Power-cycle detection | N/A              |
| Sensor noise index    | N/A              |

## Fridge Signature Trace



## Interpretation Notes

- Wide range and/or high variability suggests domestic/non-compliant fridge.

# Excursions (Audit Record)

| Start Time       | End Time         | Duration (min) | Peak Temp (°C) | GEYER Rule     | Notes |
|------------------|------------------|----------------|----------------|----------------|-------|
| 09-09-2024 00:00 | 09-09-2024 00:00 | 0.0            | 12.7           | Reportable     |       |
| 09-09-2024 00:00 | 09-09-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-09-2024 00:00 | 09-09-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 10.9           | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.2            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 11.2           | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.3            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.2            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.3            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 9.9            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.4            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-10-2024 00:00 | 09-10-2024 00:00 | 0.0            | 8.2            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 9.1            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 8.1            | Non-reportable |       |
| 09-11-2024 00:00 | 09-11-2024 00:00 | 0.0            | 11.6           | Non-reportable |       |

## Clinical Interpretation

The recorded excursions indicate periods where product temperatures may have been outside the recommended 2–8°C range. The pattern and duration of excursions should be interpreted in the context of vaccine stability data, storage conditions, and local policy.

### Final Recommendation: Review Required

ColdProof recommends that stock stored during the flagged excursions be quarantined and assessed by a supervising pharmacist, with outcomes documented in accordance with Strive-for-5 and local clinical governance.